

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q2807
 Lab File ID: VW032055.D Date Analyzed: 08/11/2025
 Instrument ID: MSVOA_W Time Analyzed: 10:09
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	223324	7.96	406973	8.85	378090	11.63
UPPER LIMIT	446648	8.459	813946	9.349	756180	12.129
LOWER LIMIT	111662	7.459	203487	8.349	189045	11.129
EPA SAMPLE NO.						
COMP-4	176773	7.95	364442	8.85	350298	11.63
COMP-5	187413	7.97	390308	8.86	377426	11.64
COMP-6	171327	7.97	362068	8.85	358834	11.63
VW0811SBL01	176404	7.96	391811	8.85	359327	11.64
VW0811SBS01	224181	7.96	427747	8.85	394820	11.63
VW0811SBSD01	226773	7.96	441696	8.85	400240	11.64

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.