

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: POWE02  
 Lab Code: ACE SDG NO.: Q2807  
 Lab File ID: VW032055.D Date Analyzed: 08/11/2025  
 Instrument ID: MSVOA\_W Time Analyzed: 10:09  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 178995        | 13.556 |  |  |  |  |
| UPPER LIMIT    | 357990        | 14.056 |  |  |  |  |
| LOWER LIMIT    | 89497.5       | 13.056 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| COMP-4         | 165686        | 13.56  |  |  |  |  |
| COMP-5         | 181852        | 13.56  |  |  |  |  |
| COMP-6         | 176343        | 13.56  |  |  |  |  |
| VW0811SBL01    | 169733        | 13.56  |  |  |  |  |
| VW0811SBS01    | 191594        | 13.56  |  |  |  |  |
| VW0811SBSD01   | 195548        | 13.56  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.