

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q2810

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB136780

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	08/12/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	08/12/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	08/12/2025

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q2810

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB136825

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: CCV1 COD	mg/L	48.931	50	98	95-105	08/15/2025
Sample ID: CCV2 COD	mg/L	48.931	50	98	95-105	08/15/2025
Sample ID: CCV3 COD	mg/L	49.946	50	100	95-105	08/15/2025