

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2810
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136764

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136764BS							
BOD5	mg/L	198	208		105	1	84.6-115.4	08/08/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2810
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136765

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136765BS							
TSS	mg/L	550	531		96	1	90-110	08/11/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2810
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136766

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136766BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	08/11/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2810
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136767

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136767BS							
TPH	mg/L	20.0	17.1		86	1	78-114	08/11/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2810
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136767

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136767BSD							
TPH	mg/L	20.0	16.9		84	1	78-114	08/11/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q2810
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB136825

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136825BS							
COD	mg/L	50	47.9		96	1	90-110	08/15/2025