

Initial and Continuing Calibration Blank Summary

Client: Ardmore Chemical

SDG No.: Q2811

Project: PVSC Monthly 2025

RunNo.: LB136770

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	0.0014	0.0025	J	0.0012	0.005	08/11/2025
Sample ID: CCB1 Cyanide	mg/L	0.0015	0.0025	J	0.0012	0.005	08/11/2025
Sample ID: CCB2 Cyanide	mg/L	0.0016	0.0025	J	0.0012	0.005	08/11/2025