

Duplicate Sample Summary

Client:	Ardmore Chemical	SDG No.:	Q2811
Project:	PVSC Monthly 2025	Sample ID:	Q2789-02
Client ID:	CompDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	377		387		1	2.62		08/11/2025

Duplicate Sample Summary

Client:	Ardmore Chemical	SDG No.:	Q2811
Project:	PVSC Monthly 2025	Sample ID:	Q2811-01
Client ID:	EFF-WASTE WATERDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.0020	J	0.0019	J	1	5		08/11/2025

Duplicate Sample Summary

Client:	Ardmore Chemical	SDG No.:	Q2811
Project:	PVSC Monthly 2025	Sample ID:	Q2811-01
Client ID:	EFF-WASTE WATERMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.039		0.039		1	0		08/11/2025

Duplicate Sample Summary

Client:	Ardmore Chemical	SDG No.:	Q2811
Project:	PVSC Monthly 2025	Sample ID:	Q2811-02
Client ID:	EFF-WASTE WATERDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	605		617		1	2		08/08/2025