

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q2811

Project: PVSC Monthly 2025

Run No.: LB136764

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136764BS							
BOD5	mg/L	198	208		105	1	84.6-115.4	08/08/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q2811

Project: PVSC Monthly 2025

Run No.: LB136765

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136765BS							
TSS	mg/L	550	531		96	1	90-110	08/11/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q2811

Project: PVSC Monthly 2025

Run No.: LB136770

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169182BS							
Cyanide	mg/L	0.1	0.098		98	1	85-115	08/11/2025