



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Ardmore Chemical

SDG No.: Q2811

Contract: ARDM01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Lead	11.1	12.0	93	65 - 135	P	08/12/2025	11:56	LB136789
	Zinc	40.0	40.0	100	65 - 135	P	08/12/2025	11:56	LB136789