

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Ardmore Chemical</u>	level: <u>low</u>	sdg no.: <u>Q2811</u>
contract: <u>ARDM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2811-02</u>	client id: <u>EFF-WASTE WATERMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2811-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	430		1.76	J	500	86		P
Zinc	ug/L	75 - 125	303		203		100	99		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Ardmore Chemical **level:** low **sdg no.:** Q2811
contract: ARDM01 **lab code:** ACE
matrix: Water **sample id:** Q2811-02 **client id:** EFF-WASTE WATERMSD
Percent Solids for Sample: NA **Spiked ID:** Q2811-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	465		1.76	J	500	93		P
Zinc	ug/L	75 - 125	342		203		100	138	N	P