

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081525

Review By	Rahul	Review On	8/18/2025 8:39:29 AM
Supervise By	Jagrut	Supervise On	8/18/2025 8:43:38 AM
SubDirectory	BF081525	HP Acquire Method	BNA_F
		HP Processing Method	bf081525

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13170,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143415.D	15 Aug 2025 13:59		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF143416.D	15 Aug 2025 14:29		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF143417.D	15 Aug 2025 14:58		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF143418.D	15 Aug 2025 15:28		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF143419.D	15 Aug 2025 15:58	Compound#41,54 Kept on LR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF143420.D	15 Aug 2025 16:28	Benzaldehyde failed in the Calibration.	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF143421.D	15 Aug 2025 16:58		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF143422.D	15 Aug 2025 17:27		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF143423.D	15 Aug 2025 17:57	compound #77 removed.	RC/JU	Ok
10	SSTDICV040	ICVBF081525	BF143424.D	15 Aug 2025 18:57		RC/JU	Ok
11	PB169229BL	PB169229BL	BF143425.D	15 Aug 2025 19:27		RC/JU	Ok
12	PB169229BS	PB169229BS	BF143426.D	15 Aug 2025 19:56		RC/JU	Ok,M
13	SSTDCCC040	SSTDCCC040EC	BF143427.D	15 Aug 2025 20:26		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081825

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF081825	HP Acquire Method	BNA_F	HP Processing Method	bf081525
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13170,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143428.D	18 Aug 2025 08:37		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143429.D	18 Aug 2025 09:06		RC/JU	Ok
3	PB169212TB	PB169212TB	BF143430.D	18 Aug 2025 09:35		RC/JU	Ok
4	PB169256BS	PB169256BS	BF143431.D	18 Aug 2025 10:04		RC/JU	Ok,NR
5	PB169256BSD	PB169256BSD	BF143432.D	18 Aug 2025 10:33		RC/JU	Ok,NR
6	PB169256BL	PB169256BL	BF143433.D	18 Aug 2025 11:02		RC/JU	Ok
7	Q2832-10	TG-S05	BF143434.D	18 Aug 2025 11:41		RC/JU	Ok,NR
8	Q2844-01	PAD-081325	BF143435.D	18 Aug 2025 12:10		RC/JU	Ok
9	Q2837-01	TW-WTS-13	BF143436.D	18 Aug 2025 12:39		RC/JU	Ok
10	Q2811-02	EFF-WASTE WATER	BF143437.D	18 Aug 2025 13:08	Surrogate Fail	RC/JU	Ok
11	Q2844-06	MOO-25-0223	BF143438.D	18 Aug 2025 13:37	Internal Standard Failed.	RC/JU	Ok,NR
12	Q2844-11	MOO-25-0228	BF143439.D	18 Aug 2025 14:06		RC/JU	Ok
13	Q2844-09	MOO-25-0226	BF143440.D	18 Aug 2025 14:36		RC/JU	Ok
14	Q2844-04	MOO-25-0234-0235	BF143441.D	18 Aug 2025 15:05	Internal Standard Fail	RC/JU	ReRun
15	Q2844-02	MOO-25-0224-0225	BF143442.D	18 Aug 2025 15:34	Internal Standard Fail	RC/JU	ReRun
16	Q2874-03	ELI-HQ-CONCRETE	BF143443.D	18 Aug 2025 16:04		RC/JU	Ok

M : Manual Integration