

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ARDM01
 Lab Code: ACE SDG NO.: Q2811
 Lab File ID: VN087494.D Date Analyzed: 08/08/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:39
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	57936	7.80	323175	9.08	292007	11.85
UPPER LIMIT	115872	8.3	646350	9.582	584014	12.347
LOWER LIMIT	28968	7.3	161588	8.582	146004	11.347
EPA SAMPLE NO.						
EFF-WASTE WATER	57038	7.80	315848	9.08	279788	11.85
VN0808WBL01	56553	7.80	310808	9.08	276148	11.85
VN0808WBS01	58186	7.80	327589	9.08	296672	11.85
VN0808WBSD01	59024	7.79	330139	9.08	290582	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.