

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2812

Client ID: SSTD020678

Date Analyzed: 09/09/2025

Lab File ID: BP025692.D

Time Analyzed: 10:31

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	890621	17.178	904409	21.619	986597	24.983
UPPER LIMIT	1781240	17.678	1808820	22.119	1973190	25.483
LOWER LIMIT	445311	16.678	452205	21.119	493299	24.483
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	654774	17.18	749188	21.61	796486	24.98

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.