

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG No.: Q2812
 Instrument ID: MSVOA_X Calibration Date/Time: 08/12/2025 09:53
 Lab File ID: VX047284.D Init. Calib. Date(s): 07/21/2025 07/21/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 11:32
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Text butyl alcohol	250.000	259.15	-3.657995	20