

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX072125

Review By	John Carlone	Review On	7/22/2025 8:51:02 AM
Supervise By	Maresh Dadoda	Supervise On	7/22/2025 1:41:20 PM
SubDirectory	VX072125	HP Acquire Method	HP Processing Method 82X072125W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134843
Initial Calibration Stds	VP134844,VP134845,VP134846,VP134847,VP134848,VP134849
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP134850
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047054.D	21 Jul 2025 08:53		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX047055.D	21 Jul 2025 09:47		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX047056.D	21 Jul 2025 10:08	Comp. #11 is on Linear Regression	JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX047057.D	21 Jul 2025 10:29		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX047058.D	21 Jul 2025 10:50		JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX047059.D	21 Jul 2025 11:11		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX047060.D	21 Jul 2025 11:32		JC/MD	Ok,M
8	IBLK	IBLK	VX047061.D	21 Jul 2025 11:52		JC/MD	Ok
9	VSTDICV050	ICVVX072125	VX047062.D	21 Jul 2025 13:17		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX080825

Review By	John Carlone	Review On	8/12/2025 9:05:05 AM
Supervise By	Maresh Dadoda	Supervise On	8/12/2025 1:09:25 PM
SubDirectory	VX080825	HP Acquire Method	HP Processing Method 82X072125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135056
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135057,VP135058

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047262.D	08 Aug 2025 10:04		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047263.D	08 Aug 2025 11:04	pH#Lot#V12668	JC/MD	Ok,M
3	VX0808MBL01	VX0808MBL01	VX047264.D	08 Aug 2025 11:30		JC/MD	Ok
4	VX0808WBL01	VX0808WBL01	VX047265.D	08 Aug 2025 12:19		JC/MD	Ok
5	VX0808WBS01	VX0808WBS01	VX047266.D	08 Aug 2025 13:16		JC/MD	Ok,M
6	VX0808WBSD01	VX0808WBSD01	VX047267.D	08 Aug 2025 13:45		JC/MD	Ok,M
7	PB169150TB	PB169150TB	VX047268.D	08 Aug 2025 14:06		JC/MD	Ok
8	Q2779-04	TP-4	VX047269.D	08 Aug 2025 14:27	vial A pH#5.0	JC/MD	Ok,M
9	Q2779-08	TP-5	VX047270.D	08 Aug 2025 14:48	vial A pH#5.0	JC/MD	Ok
10	Q2783-03	RT5414	VX047271.D	08 Aug 2025 15:09	vial A pH#5.0	JC/MD	Ok
11	Q2798-04	TP-6	VX047272.D	08 Aug 2025 15:30	vial A pH#5.0	JC/MD	Ok
12	Q2800-02	VNJ-222	VX047273.D	08 Aug 2025 15:51	vial A pH#5.0	JC/MD	Ok
13	Q2791-01	SLUDGE	VX047274.D	08 Aug 2025 16:12	vial A pH#5.0 Surrogate fail	JC/MD	ReRun
14	IBLK	IBLK	VX047275.D	08 Aug 2025 16:33		JC/MD	Ok
15	Q2791-01	SLUDGE	VX047276.D	08 Aug 2025 16:54	vial B pH#5.0	JC/MD	Ok,M
16	IBLK	IBLK	VX047277.D	08 Aug 2025 17:15		JC/MD	Ok
17	Q2812-01	STORAGE-BLANK-SO	VX047278.D	08 Aug 2025 17:36	vial A pH<2	JC/MD	Ok
18	Q2812-02	STORAGE-BLANK-WA	VX047279.D	08 Aug 2025 17:57	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX080825

Review By	John Carlone	Review On	8/12/2025 9:05:05 AM
Supervise By	Maresh Dadoda	Supervise On	8/12/2025 1:09:25 PM
SubDirectory	VX080825	HP Acquire Method	HP Processing Method 82X072125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135056
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135057,VP135058

19	Q2812-03	STORAGE-BLANK-WA	VX047280.D	08 Aug 2025 18:18	vial A pH<2	JC/MD	Ok
20	Q2812-04	STORAGE-BLANK-SA	VX047281.D	08 Aug 2025 18:39	vial A pH<2 Surrogate fail	JC/MD	ReRun
21	VSTDCCC050	VSTDCCC050EC	VX047282.D	08 Aug 2025 19:00		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX081225

Review By	John Carlone	Review On	8/13/2025 8:53:36 AM
Supervise By		Supervise On	
SubDirectory	VX081225	HP Acquire Method	HP Processing Method 82X072125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135091		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135092,VP135093		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047283.D	12 Aug 2025 07:51		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047284.D	12 Aug 2025 09:53	pH#Lot#V12668	JC/MD	Ok,NS
3	VX0812MBL01	VX0812MBL01	VX047285.D	12 Aug 2025 10:29		JC/MD	Ok
4	VX0812WBL01	VX0812WBL01	VX047286.D	12 Aug 2025 10:50		JC/MD	Not Ok
5	VX0812WBL02	VX0812WBL02	VX047287.D	12 Aug 2025 11:17		JC/MD	Ok
6	VX0812WBS01	VX0812WBS01	VX047288.D	12 Aug 2025 11:52		JC/MD	Ok,NS
7	Q2812-04	STORAGE-BLANK-SAL	VX047289.D	12 Aug 2025 12:22	vial B pH<2	JC/MD	Ok
8	Q2814-01	TW-84SB-E	VX047290.D	12 Aug 2025 12:43	vial A pH<2	JC/MD	Ok
9	Q2814-02	TW-17M-N	VX047291.D	12 Aug 2025 13:05	vial A pH<2	JC/MD	Ok
10	VX0812WBSD01	VX0812WBSD01	VX047292.D	12 Aug 2025 13:26		JC/MD	Ok,NS
11	Q2814-03	TW-17M-W	VX047293.D	12 Aug 2025 13:51	vial A pH<2	JC/MD	Ok
12	Q2814-04	TW-82H-N	VX047294.D	12 Aug 2025 14:13	vial A pH<2	JC/MD	Ok
13	Q2814-05	TW-82H-E	VX047295.D	12 Aug 2025 14:34	vial A pH<2	JC/MD	Ok
14	Q2814-06	TW-82H-S	VX047296.D	12 Aug 2025 14:56	vial A pH<2 Surrogate Fail	JC/MD	ReRun
15	Q2814-07	TW-82H-W	VX047297.D	12 Aug 2025 15:18	vial A pH<2 Surrogate Fail	JC/MD	ReRun
16	Q2814-08	TW-38M-E	VX047298.D	12 Aug 2025 15:39	vial A pH<2	JC/MD	Ok,NS
17	Q2814-09	TW-38M-W	VX047299.D	12 Aug 2025 16:01	vial A pH<2	JC/MD	Ok
18	Q2814-10	TW-38M-N	VX047300.D	12 Aug 2025 16:23	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX081225

Review By	John Carlone	Review On	8/13/2025 8:53:36 AM
Supervise By		Supervise On	
SubDirectory	VX081225	HP Acquire Method	HP Processing Method 82X072125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135091		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135092,VP135093		

19	Q2814-11	TW-38M-S	VX047301.D	12 Aug 2025 16:44	vial A pH<2	JC/MD	Ok
20	Q2814-12	TW-BP-S	VX047302.D	12 Aug 2025 17:06	vial A pH<2 Surrogate Fail	JC/MD	ReRun
21	Q2814-13	TW-BP-N	VX047303.D	12 Aug 2025 17:28	vial A pH<2	JC/MD	Ok
22	Q2814-14	TW-BP-E	VX047304.D	12 Aug 2025 17:49	vial A pH<2	JC/MD	Ok
23	Q2814-15	TW-BP-W	VX047305.D	12 Aug 2025 18:11	vial A pH<2	JC/MD	Ok
24	Q2814-16	FB	VX047306.D	12 Aug 2025 18:32	vial A pH<2 FB;Hit of Com.#16	JC/MD	ReRun
25	VSTDCCC050	VSTDCCC050EC	VX047307.D	12 Aug 2025 18:54		JC/MD	Ok,NS

M : Manual Integration