

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2813

Project: Pre Treatment Plant 2025

RunNo.: LB136763

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.495	0.50	99	90-110	08/08/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.503	0.5	101	90-110	08/08/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.491	0.5	98	90-110	08/08/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2813

Project: Pre Treatment Plant 2025

RunNo.: LB136771

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.505	0.50	101	90-110	08/11/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.493	0.50	99	90-110	08/11/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.503	0.50	101	90-110	08/11/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2813

Project: Pre Treatment Plant 2025

RunNo.: LB136784

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.99	1	99	90-110	08/12/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.97	1	97	90-110	08/12/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.97	1	97	90-110	08/12/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	08/12/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.99	1	99	90-110	08/12/2025