

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2813

Project: Pre Treatment Plant 2025

Run No.: LB136763

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136763BS							
Orthophosphate as P	mg/L	0.5	0.51		102	1	90-110	08/08/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2813

Project: Pre Treatment Plant 2025

Run No.: LB136764

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136764BS							
BOD5	mg/L	198	208		105	1	84.6-115.4	08/08/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2813

Project: Pre Treatment Plant 2025

Run No.: LB136765

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136765BS							
TSS	mg/L	550	531		96	1	90-110	08/11/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2813

Project: Pre Treatment Plant 2025

Run No.: LB136766

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136766BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	08/11/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2813

Project: Pre Treatment Plant 2025

Run No.: LB136771

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169196BS							
Phosphorus, Total	mg/L	0.50	0.50		100	1	90-110	08/11/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2813

Project: Pre Treatment Plant 2025

Run No.: LB136784

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169199BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	08/12/2025