

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2814
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1254CCC500 **Data File :** PO112933.D **Time Analyzed:** 09:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.556	5.461	5.661	518.260	500.000	3.7
Aroclor-1254-2	5.705	5.611	5.811	534.480	500.000	6.9
Aroclor-1254-3	6.108	6.014	6.214	517.950	500.000	3.6
Aroclor-1254-4	6.338	6.244	6.444	513.790	500.000	2.8
Aroclor-1254-5	6.756	6.662	6.862	515.850	500.000	3.2
Decachlorobiphenyl	8.687	8.595	8.795	50.880	50.000	1.8
Tetrachloro-m-xylene	3.666	3.570	3.770	56.410	50.000	12.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2814
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1254CCC500 **Data File :** PO112933.D **Time Analyzed:** 09:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.529	5.435	5.635	480.390	500.000	-3.9
Aroclor-1254-2	5.677	5.583	5.783	480.730	500.000	-3.9
Aroclor-1254-3	6.077	5.984	6.184	489.620	500.000	-2.1
Aroclor-1254-4	6.306	6.211	6.411	485.870	500.000	-2.8
Aroclor-1254-5	6.720	6.627	6.827	509.000	500.000	1.8
Decachlorobiphenyl	8.632	8.541	8.741	53.380	50.000	6.8
Tetrachloro-m-xylene	3.658	3.563	3.763	53.310	50.000	6.6