

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2814
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112945.D **Time Analyzed:** 15:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.749	4.655	4.855	492.510	500.000	-1.5
Aroclor-1016-2	4.768	4.674	4.874	491.760	500.000	-1.6
Aroclor-1016-3	4.825	4.731	4.931	477.380	500.000	-4.5
Aroclor-1016-4	4.944	4.851	5.051	484.610	500.000	-3.1
Aroclor-1016-5	5.201	5.108	5.308	496.740	500.000	-0.7
Aroclor-1260-1	6.237	6.145	6.345	451.200	500.000	-9.8
Aroclor-1260-2	6.427	6.333	6.533	437.550	500.000	-12.5
Aroclor-1260-3	6.793	6.700	6.900	443.170	500.000	-11.4
Aroclor-1260-4	7.052	6.959	7.159	410.660	500.000	-17.9
Aroclor-1260-5	7.295	7.203	7.403	406.730	500.000	-18.7
Decachlorobiphenyl	8.685	8.593	8.793	39.400	50.000	-21.2
Tetrachloro-m-xylene	3.665	3.569	3.769	51.520	50.000	3.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2814
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112945.D **Time Analyzed:** 15:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.733	4.638	4.838	468.860	500.000	-6.2
Aroclor-1016-2	4.751	4.656	4.856	470.600	500.000	-5.9
Aroclor-1016-3	4.926	4.831	5.031	459.360	500.000	-8.1
Aroclor-1016-4	4.968	4.873	5.073	462.900	500.000	-7.4
Aroclor-1016-5	5.180	5.085	5.285	466.490	500.000	-6.7
Aroclor-1260-1	6.207	6.113	6.313	438.420	500.000	-12.3
Aroclor-1260-2	6.396	6.302	6.502	443.180	500.000	-11.4
Aroclor-1260-3	6.547	6.453	6.653	454.330	500.000	-9.1
Aroclor-1260-4	7.016	6.923	7.123	450.560	500.000	-9.9
Aroclor-1260-5	7.258	7.164	7.364	471.600	500.000	-5.7
Decachlorobiphenyl	8.633	8.540	8.740	46.800	50.000	-6.4
Tetrachloro-m-xylene	3.659	3.562	3.762	49.870	50.000	-0.3