

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2814
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL06 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1254CCC500 **Data File :** PO112948.D **Time Analyzed:** 16:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.554	5.461	5.661	512.350	500.000	2.5
Aroclor-1254-2	5.703	5.611	5.811	524.120	500.000	4.8
Aroclor-1254-3	6.106	6.014	6.214	483.810	500.000	-3.2
Aroclor-1254-4	6.336	6.244	6.444	482.870	500.000	-3.4
Aroclor-1254-5	6.754	6.662	6.862	490.050	500.000	-2.0
Decachlorobiphenyl	8.686	8.595	8.795	46.630	50.000	-6.7
Tetrachloro-m-xylene	3.665	3.570	3.770	55.050	50.000	10.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2814
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL06 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1254CCC500 **Data File :** PO112948.D **Time Analyzed:** 16:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.530	5.435	5.635	477.220	500.000	-4.6
Aroclor-1254-2	5.677	5.583	5.783	478.260	500.000	-4.3
Aroclor-1254-3	6.077	5.984	6.184	476.060	500.000	-4.8
Aroclor-1254-4	6.305	6.211	6.411	469.550	500.000	-6.1
Aroclor-1254-5	6.720	6.627	6.827	485.160	500.000	-3.0
Decachlorobiphenyl	8.632	8.541	8.741	51.230	50.000	2.5
Tetrachloro-m-xylene	3.659	3.563	3.763	52.780	50.000	5.6