

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074177.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 15:03
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 15:21:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 15:20:45 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.657	3.805	56653742	212.2E6	50.000	50.000
2)	SA Decachlor...	10.434	8.826	49220928	317.2E6	50.000	50.000
Target Compounds							
16)	L4 AR-1242-1	5.808	4.905	17959256	178.1E6	500.000	500.000
17)	L4 AR-1242-2	5.830	4.963	26531667	82516708	500.000	500.000
18)	L4 AR-1242-3	5.893	5.084	17362372	46044427	500.000	500.000
19)	L4 AR-1242-4	5.990	5.164	14345906	63188535	500.000	505.329m
20)	L4 AR-1242-5	6.719	5.687	17203244	74177320	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074177.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 15:03
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 15:21:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 15:20:45 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.657 min
Delta R.T.: 0.000 min
Response: 56653742
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025




#1 Tetrachloro-m-xylene

R.T.: 3.805 min
Delta R.T.: 0.000 min
Response: 212235944
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.434 min
Delta R.T.: 0.000 min
Response: 49220928
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.826 min
Delta R.T.: 0.000 min
Response: 317160143
Conc: 50.00 ng/ml




#16 AR-1242-1

R.T.: 5.808 min
Delta R.T.: 0.000 min
Response: 17959256
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025




#16 AR-1242-1

R.T.: 4.905 min
Delta R.T.: 0.000 min
Response: 178087793
Conc: 500.00 ng/ml




#17 AR-1242-2

R.T.: 5.830 min
Delta R.T.: 0.000 min
Response: 26531667
Conc: 500.00 ng/ml



#17 AR-1242-2

R.T.: 4.963 min
Delta R.T.: 0.000 min
Response: 82516708
Conc: 500.00 ng/ml




#18 AR-1242-3

R.T.: 5.893 min
Delta R.T.: 0.000 min
Response: 17362372
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025




#18 AR-1242-3

R.T.: 5.084 min
Delta R.T.: 0.000 min
Response: 46044427
Conc: 500.00 ng/ml




#19 AR-1242-4

R.T.: 5.990 min
Delta R.T.: 0.000 min
Response: 14345906
Conc: 500.00 ng/ml



#19 AR-1242-4

R.T.: 5.164 min
Delta R.T.: 0.000 min
Response: 63188535
Conc: 505.33 ng/ml m




#20 AR-1242-5

R.T.: 6.719 min
Delta R.T.: 0.000 min
Response: 17203244
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025



#20 AR-1242-5

R.T.: 5.687 min
Delta R.T.: 0.000 min
Response: 74177320
Conc: 500.00 ng/ml