

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: FIRS02

Lab Code: ACE
SDG NO.: Q2814

Instrument ID: ECD_L
Date(s) Analyzed: 07/28/2025 07/28/2025

GC Column: ZB-MR1
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.21	6.11	6.31	29153500
		2	6.61	6.51	6.71	24445000
		3	7.02	6.92	7.12	110273000
		4	7.11	7.01	7.21	81046400
		5	7.89	7.79	7.99	57873100

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: FIRS02

Lab Code: ACE SDG NO.: Q2814

Instrument ID: ECD_L Date(s) Analyzed: 07/28/2025 07/28/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.07	4.97	5.17	34047600
		2	5.76	5.66	5.86	44078700
		3	6.04	5.94	6.14	44632800
		4	6.67	6.57	6.77	147063000
		5	7.11	7.01	7.21	92006100