

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2814

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/28/2025 07/28/2025

Client Sample No. (PEM): PEM - PL096594.D Date Analyzed: 07/28/2025

Lab Sample No.(PEM): PEM Time Analyzed: 16:25

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.017	8.920	9.120	21.270	20.000	6.4
Tetrachloro-m-xylene	3.535	3.480	3.590	20.960	20.000	4.8
alpha-BHC	3.983	3.930	4.030	10.440	10.000	4.4
beta-BHC	4.498	4.450	4.550	10.520	10.000	5.2
gamma-BHC (Lindane)	4.311	4.260	4.360	10.550	10.000	5.5
Endrin	6.543	6.470	6.610	54.020	50.000	8.0
4,4'-DDT	6.990	6.920	7.060	109.500	100.000	9.5
Methoxychlor	7.463	7.390	7.530	267.200	250.000	6.9

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/28/2025 07/28/2025

Client Sample No. (PEM): PEM - PL096594.D Date Analyzed: 07/28/2025

Lab Sample No.(PEM): PEM Time Analyzed: 16:25

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.992	7.890	8.090	20.870	20.000	4.4
Tetrachloro-m-xylene	2.828	2.780	2.880	20.780	20.000	3.9
alpha-BHC	3.333	3.280	3.380	10.270	10.000	2.7
beta-BHC	3.962	3.910	4.010	10.840	10.000	8.4
gamma-BHC (Lindane)	3.666	3.620	3.720	10.440	10.000	4.4
Endrin	5.710	5.640	5.780	53.630	50.000	7.3
4,4'-DDT	6.109	6.040	6.180	111.210	100.000	11.2
Methoxychlor	6.682	6.610	6.750	253.290	250.000	1.3