

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2814

Client ID : SSTDCCC040

Date Analyzed: 08/15/2025

Lab File ID: BP025422.D

Time Analyzed: 10:36

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	163210	7.79	699113	10.55	458301	14.41
UPPER LIMIT	326420	8.29	1398230	11.054	916602	14.907
LOWER LIMIT	81605	7.29	349557	10.054	229151	13.907
EPA SAMPLE NO.						
01 TW-82H-ERE	137678	7.79	540790	10.55	370556	14.40
02 TW-518R-S	138753	7.78	565678	10.55	378580	14.40
03 TW-518R-N	144702	7.79	586287	10.55	406797	14.41
04 TW-518R-W	131204	7.79	536271	10.56	359414	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.