

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2814

Client ID : SSTDCCC040

Date Analyzed: 08/20/2025

Lab File ID: BP025512.D

Time Analyzed: 18:27

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	166180	7.79	619421	10.55	355189	14.40
UPPER LIMIT	332360	8.29	1238840	11.054	710378	14.895
LOWER LIMIT	83090	7.29	309711	10.054	177595	13.895
EPA SAMPLE NO.						
01 TW-518R-NRE	227667	7.79	906269	10.56	565733	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.