

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG No.: Q2814

Instrument ID: MSVOA_X

Calibration Date/Time: 08/12/2025 09:53

Lab File ID: VX047284.D

Init. Calib. Date(s): 07/21/2025 07/21/2025

Heated Purge: (Y/N) N

Init. Calib. Time(s): 09:47 11:32

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.664	0.700		5.42	20
Chloromethane	0.657	0.619	0.1	-5.78	20
Vinyl Chloride	0.706	0.710		0.57	20
Bromomethane	0.384	0.423		10.16	20
Chloroethane	0.464	0.423		-8.84	20
Trichlorofluoromethane	1.053	1.060		0.67	20
1,1,2-Trichlorotrifluoroethane	0.634	0.627		-1.1	20
1,1-Dichloroethene	0.628	0.616		-1.91	20
Acetone	0.250	0.262		4.8	20
Carbon Disulfide	1.818	1.741		-4.24	20
Methyl tert-butyl Ether	1.912	2.026		5.96	20
Methyl Acetate	0.663	0.767		15.69	20
Methylene Chloride	0.686	0.675		-1.6	20
trans-1,2-Dichloroethene	0.643	0.644		0.16	20
1,1-Dichloroethane	1.214	1.182	0.1	-2.64	20
Cyclohexane	1.164	1.087		-6.61	20
2-Butanone	0.363	0.348		-4.13	20
Carbon Tetrachloride	0.562	0.553		-1.6	20
cis-1,2-Dichloroethene	0.761	0.747		-1.84	20
Bromochloromethane	0.559	0.571		2.15	20
Chloroform	1.236	1.217		-1.54	20
1,1,1-Trichloroethane	1.074	1.058		-1.49	20
Methylcyclohexane	0.657	0.628		-4.41	20
Benzene	1.523	1.515		-0.52	20
1,2-Dichloroethane	0.536	0.553		3.17	20
Trichloroethene	0.391	0.385		-1.53	20
1,2-Dichloropropane	0.381	0.386		1.31	20
Bromodichloromethane	0.569	0.591		3.87	20
4-Methyl-2-Pentanone	0.449	0.476		6.01	20
Toluene	0.936	0.940		0.43	20
t-1,3-Dichloropropene	0.538	0.566		5.2	20
cis-1,3-Dichloropropene	0.588	0.609		3.57	20
1,1,2-Trichloroethane	0.351	0.368		4.84	20
2-Hexanone	0.297	0.332		11.78	20
Dibromochloromethane	0.412	0.442		7.28	20
1,2-Dibromoethane	0.367	0.380		3.54	20
Tetrachloroethene	0.360	0.352		-2.22	20
Chlorobenzene	1.171	1.165	0.3	-0.51	20

All other compounds must meet a minimum RRF of 0.010.

RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2814</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>08/12/2025</u> <u>09:53</u>
Lab File ID:	<u>VX047284.D</u>	Init. Calib. Date(s):	<u>07/21/2025</u> <u>07/21/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:47</u> <u>11:32</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	2.048	2.055		0.34	20
m/p-Xylenes	0.766	0.769		0.39	20
o-Xylene	0.729	0.741		1.65	20
Styrene	1.250	1.290		3.2	20
Bromoform	0.300	0.320	0.1	6.67	20
Isopropylbenzene	3.945	3.828		-2.97	20
1,1,2,2-Tetrachloroethane	1.127	1.141	0.3	1.24	20
1,3-Dichlorobenzene	1.798	1.747		-2.84	20
1,4-Dichlorobenzene	1.821	1.758		-3.46	20
1,2-Dichlorobenzene	1.718	1.683		-2.04	20
1,2-Dibromo-3-Chloropropane	0.216	0.231		6.94	20
1,2,4-Trichlorobenzene	1.135	1.121		-1.23	20
1,2,3-Trichlorobenzene	1.094	1.054		-3.66	20
1,2-Dichloroethane-d4	0.623	0.583		-6.42	20
Dibromofluoromethane	0.309	0.317		2.59	20
Toluene-d8	1.000	0.887		-11.3	20
4-Bromofluorobenzene	0.431	0.427		-0.93	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.