

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081325\
 Data File : VX047309.D
 Acq On : 13 Aug 2025 09:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 03:23:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	65	0.00
2 T	Dichlorodifluoromethane	0.664	0.684	-3.0	58	0.00
3 P	Chloromethane	0.657	0.610	7.2	57	0.00
4 C	Vinyl Chloride	0.706	0.722	-2.3#	60	0.00
5 T	Bromomethane	0.384	0.390	-1.6	57	0.00
6 T	Chloroethane	0.464	0.440	5.2	61	0.00
7 T	Trichlorofluoromethane	1.053	1.065	-1.1	60	0.00
8 T	Diethyl Ether	0.371	0.402	-8.4	67	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.634	0.626	1.3	61	0.00
10 T	Methyl Iodide	0.647	1.084	-67.5#	97	0.00
11 T	Tert butyl alcohol	0.080	0.077	3.8	69	0.00
12 CM	1,1-Dichloroethene	0.628	0.628	0.0	61	0.00
13 T	Acrolein	0.134	0.209	-56.0#	98	0.00
14 T	Allyl chloride	1.136	1.122	1.2	59	0.00
15 T	Acrylonitrile	0.297	0.333	-12.1	66	0.00
16 T	Acetone	0.250	0.291	-16.4	76	0.00
17 T	Carbon Disulfide	1.818	1.715	5.7	58	0.00
18 T	Methyl Acetate	0.663	0.813	-22.6	77	0.00
19 T	Methyl tert-butyl Ether	1.912	2.118	-10.8	67	0.00
20 T	Methylene Chloride	0.686	0.714	-4.1	65	0.00
21 T	trans-1,2-Dichloroethene	0.643	0.659	-2.5	62	0.00
22 T	Diisopropyl ether	2.184	2.343	-7.3	65	0.00
23 T	Vinyl Acetate	1.752	1.942	-10.8	65	0.00
24 P	1,1-Dichloroethane	1.214	1.233	-1.6	62	0.00
25 T	2-Butanone	0.363	0.375	-3.3	63	0.00
26 T	2,2-Dichloropropane	0.940	0.944	-0.4	63	0.00
27 T	cis-1,2-Dichloroethene	0.761	0.797	-4.7	64	0.00
28 T	Bromochloromethane	0.559	0.601	-7.5	64	0.00
29 T	Tetrahydrofuran	0.234	0.235	-0.4	60	0.00
30 C	Chloroform	1.236	1.275	-3.2#	65	0.00
31 T	Cyclohexane	1.164	1.079	7.3	58	0.00
32 T	1,1,1-Trichloroethane	1.074	1.081	-0.7	63	0.00
33 S	1,2-Dichloroethane-d4	0.623	0.607	2.6	69	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	67	0.00
35 S	Dibromofluoromethane	0.309	0.319	-3.2	72	0.00
36 T	1,1-Dichloropropene	0.526	0.508	3.4	62	0.00
37 T	Ethyl Acetate	0.485	0.562	-15.9	73	0.00
38 T	Carbon Tetrachloride	0.562	0.542	3.6	61	0.00
39 T	Methylcyclohexane	0.657	0.604	8.1	57	0.00
40 TM	Benzene	1.523	1.554	-2.0	63	0.00
41 T	Methacrylonitrile	0.254	0.262	-3.1	60	0.00
42 TM	1,2-Dichloroethane	0.536	0.577	-7.6	67	0.00
43 T	Isopropyl Acetate	0.756	0.838	-10.8	66	0.00
44 TM	Trichloroethene	0.391	0.389	0.5	62	0.00
45 C	1,2-Dichloropropane	0.381	0.396	-3.9#	64	0.00
46 T	Dibromomethane	0.267	0.295	-10.5	67	0.00
47 T	Bromodichloromethane	0.569	0.606	-6.5	65	0.00
48 T	Methyl methacrylate	0.412	0.435	-5.6	66	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081325\
 Data File : VX047309.D
 Acq On : 13 Aug 2025 09:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 03:23:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.005	0.0	62	0.02
50 S	Toluene-d8	1.000	0.880	12.0	63	0.00
51 T	4-Methyl-2-Pentanone	0.449	0.493	-9.8	66	0.00
52 CM	Toluene	0.936	0.966	-3.2#	63	0.00
53 T	t-1,3-Dichloropropene	0.538	0.562	-4.5	63	0.00
54 T	cis-1,3-Dichloropropene	0.588	0.617	-4.9	63	0.00
55 T	1,1,2-Trichloroethane	0.351	0.384	-9.4	68	0.00
56 T	Ethyl methacrylate	0.520	0.589	-13.3	66	0.00
57 T	1,3-Dichloropropane	0.610	0.658	-7.9	67	0.00
58 T	2-Chloroethyl Vinyl ether	0.289	0.296	-2.4	61	0.00
59 T	2-Hexanone	0.297	0.345	-16.2	67	0.00
60 T	Dibromochloromethane	0.412	0.448	-8.7	67	0.00
61 T	1,2-Dibromoethane	0.367	0.399	-8.7	67	0.00
62 S	4-Bromofluorobenzene	0.431	0.428	0.7	70	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	67	0.00
64 T	Tetrachloroethene	0.360	0.349	3.1	62	0.00
65 PM	Chlorobenzene	1.171	1.184	-1.1	65	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.417	-5.8	67	0.00
67 C	Ethyl Benzene	2.048	2.053	-0.2#	63	0.00
68 T	m/p-Xylenes	0.766	0.773	-0.9	63	0.00
69 T	o-Xylene	0.729	0.745	-2.2	64	0.00
70 T	Styrene	1.250	1.297	-3.8	64	0.00
71 P	Bromoform	0.300	0.324	-8.0	67	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	70	0.00
73 T	Isopropylbenzene	3.945	3.857	2.2	62	0.00
74 T	N-amyl acetate	1.592	1.625	-2.1	64	0.00
75 P	1,1,2,2-Tetrachloroethane	1.127	1.184	-5.1	68	0.00
76 T	1,2,3-Trichloropropane	0.925	0.954	-3.1	67	0.00
77 T	Bromobenzene	0.946	0.959	-1.4	66	0.00
78 T	n-propylbenzene	4.714	4.599	2.4	63	0.00
79 T	2-Chlorotoluene	2.818	2.750	2.4	63	0.00
80 T	1,3,5-Trimethylbenzene	3.259	3.221	1.2	64	0.00
81 T	trans-1,4-Dichloro-2-butene	0.359	0.193	46.2#	34#	0.00
82 T	4-Chlorotoluene	3.289	3.264	0.8	64	0.00
83 T	tert-Butylbenzene	3.271	3.199	2.2	62	0.00
84 T	1,2,4-Trimethylbenzene	3.251	3.228	0.7	63	0.00
85 T	sec-Butylbenzene	4.087	3.893	4.7	62	0.00
86 T	p-Isopropyltoluene	3.389	3.290	2.9	62	0.00
87 T	1,3-Dichlorobenzene	1.798	1.804	-0.3	66	0.00
88 T	1,4-Dichlorobenzene	1.821	1.773	2.6	65	0.00
89 T	n-Butylbenzene	3.178	2.998	5.7	60	0.00
90 T	Hexachloroethane	0.606	0.594	2.0	63	0.00
91 T	1,2-Dichlorobenzene	1.718	1.708	0.6	65	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.216	0.232	-7.4	68	0.00
93 T	1,2,4-Trichlorobenzene	1.135	1.133	0.2	65	0.00
94 T	Hexachlorobutadiene	0.386	0.361	6.5	62	0.00
95 T	Naphthalene	3.328	3.447	-3.6	64	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081325\
Data File : VX047309.D
Acq On : 13 Aug 2025 09:44
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 03:23:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.094	1.081	1.2	64	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5