

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	USACE018-44 DOD	Date Received:	
Client Sample ID:	VX0812WBSD01	SDG No.:	Q2814
Lab Sample ID:	VX0812WBSD01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047292.D	1	08/12/25 13:26	VX081225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	20.3		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	17.6		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	19.5		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	18.3		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	17.6		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.4		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.2		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.2		0.23	0.75	1.00	ug/L
67-64-1	Acetone	94.3		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	18.4		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.5		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	20.0		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.5		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.4		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.4		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	18.9		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	91.1		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.8		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.2		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	23.9		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	19.8		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.2		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.3		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.4		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.4		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.1		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	18.9		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.1		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	98.7		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.9		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	First Environment, Inc.		Date Collected:	
Project:	USACE018-44 DOD		Date Received:	
Client Sample ID:	VX0812WBSD01		SDG No.:	Q2814
Lab Sample ID:	VX0812WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047292.D	1	08/12/25 13:26	VX081225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.2		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.2		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	19.7		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	100		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	19.5		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.2		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.7		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.2		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.1		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	38.9		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	19.6		0.12	0.50	1.00	ug/L
100-42-5	Styrene	19.6		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	19.2		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	18.2		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	18.4		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	17.9		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.3		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	18.3		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	18.6		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	17.3		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.5		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	48.9		81 - 118		98%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	45.3		89 - 112		91%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.3		85 - 114		103%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	232000	5.568				
540-36-3	1,4-Difluorobenzene	397000	6.769				
3114-55-4	Chlorobenzene-d5	360000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	188000	12.018				

Report of Analysis

Client: First Environment, Inc.

Date Collected:

Project: USACE018-44 DOD

Date Received:

Client Sample ID: VX0812WBSD01

SDG No.: Q2814

Lab Sample ID: VX0812WBSD01

Matrix: Water

Analytical Method: 8260D

% Solid: 0

Sample Wt/Vol: 5 Units: mL

Final Vol: 5000 uL

Soil Aliquot Vol: _____ uL

Test: VOC-TCLVOA-10

GC Column: DB-624UI ID : 0.18

Level : LOW

Prep Method :

File ID/Qc Batch: Dilution:

Date Analyzed

Prep Batch ID

VX047292.D 1

08/12/25 13:26

VX081225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products