

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2814

Lab File ID: VX047284.D

Date Analyzed: 08/12/2025

Instrument ID: MSVOA X

Time Analyzed: 09:53

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	165820	12.018				
UPPER LIMIT	331640	12.518				
LOWER LIMIT	82910	11.518				
EPA SAMPLE NO.						
TW-84SB-E	200668	12.02				
TW-17M-N	322981	12.02				
TW-17M-W	200683	12.02				
TW-82H-N	196916	12.02				
TW-82H-E	196582	12.02				
TW-82H-S	228813	12.02				
TW-82H-W	291317	12.02				
TW-38M-E	208948	12.02				
TW-38M-W	180111	12.02				
TW-38M-N	205788	12.02				
TW-38M-S	176128	12.02				
TW-BP-S	106606	12.02				
TW-BP-N	179716	12.02				
TW-BP-E	181876	12.02				
TW-BP-W	180840	12.02				
VX0812WBL02	313492	12.02				
VX0812WBS01	207271	12.02				
VX0812WBSD01	187721	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: FIRS02
 Lab Code: ACE SDG NO.: Q2814
 Lab File ID: VX047284.D Date Analyzed: 08/12/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:53
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	165820	12.018				
UPPER LIMIT	331640	12.518				
LOWER LIMIT	82910	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.