

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2814</u>
Lab File ID:	<u>VX047309.D</u>	Date Analyzed:	<u>08/13/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:44</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	156059	12.018				
UPPER LIMIT	312118	12.518				
LOWER LIMIT	78029.5	11.518				
EPA SAMPLE NO.						
TW-82H-SRE	164920	12.02				
TW-82H-WRE	157483	12.02				
TW-BP-SRE	155373	12.02				
FB	170270	12.02				
TW-518R-N	158976	12.02				
TW-518R-E	156454	12.02				
TW-518R-W	169163	12.02				
VX0813WBL01	172532	12.02				
VX0813WBS01	166290	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.