

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: FIRS02
 Lab Code: ACE SDG NO.: Q2814
 Lab File ID: VX047336.D Date Analyzed: 08/14/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:37
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	163540	12.018				
UPPER LIMIT	327080	12.518				
LOWER LIMIT	81770	11.518				
EPA SAMPLE NO.						
TW-518R-S	217218	12.02				
TW-518R-NRE	229471	12.02				
TW-518R-ERE	231965	12.02				
TB	212249	12.02				
VX0814WBL01	203879	12.02				
VX0814WBS02	182232	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.