

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2814

Lab File ID: VX047054.D

BFB Injection Date: 07/21/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:53

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.4
75	30.0 - 60.0% of mass 95	52.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	72.5
175	5.0 - 9.0% of mass 174	5.3 (7.3) 1
176	95.0 - 101.0% of mass 174	69.5 (95.9) 1
177	5.0 - 9.0% of mass 176	5 (7.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX047055.D	07/21/2025	09:47
VSTDIC005	VSTDIC005	VX047056.D	07/21/2025	10:08
VSTDIC020	VSTDIC020	VX047057.D	07/21/2025	10:29
VSTDIC050	VSTDIC050	VX047058.D	07/21/2025	10:50
VSTDIC100	VSTDIC100	VX047059.D	07/21/2025	11:11
VSTDIC150	VSTDIC150	VX047060.D	07/21/2025	11:32