

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2814

Lab File ID: VX047308.D

BFB Injection Date: 08/13/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:30

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.3
75	30.0 - 60.0% of mass 95	51.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.3 (0.4) 1
174	50.0 - 100.0% of mass 95	72.3
175	5.0 - 9.0% of mass 174	5.5 (7.6) 1
176	95.0 - 101.0% of mass 174	71.1 (98.4) 1
177	5.0 - 9.0% of mass 176	4.6 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX047309.D	08/13/2025	09:44
VX0813WBL01	VX0813WBL01	VX047311.D	08/13/2025	11:02
VX0813WBS01	VX0813WBS01	VX047312.D	08/13/2025	11:24
TW-BP-SRE	Q2814-12RE	VX047314.D	08/13/2025	12:13
FB	Q2814-16	VX047315.D	08/13/2025	12:35
TW-82H-SRE	Q2814-06RE	VX047317.D	08/13/2025	13:18
TW-82H-WRE	Q2814-07RE	VX047318.D	08/13/2025	13:40
TW-518R-N	Q2814-18	VX047320.D	08/13/2025	14:24
TW-518R-E	Q2814-19	VX047321.D	08/13/2025	14:45
TW-518R-W	Q2814-20	VX047322.D	08/13/2025	15:07
VSTDCCC050EC	VSTDCCC050	VX047334.D	08/13/2025	19:50