

DATA PACKAGE

GC SEMI-VOLATILES

PROJECT NAME : USACE018-44 DOD

FIRST ENVIRONMENT, INC.
10 Park Place, Bldg 1A, Suite 504

Butler, NJ - 07405
Phone No: 973-334-0003

ORDER ID : Q2815
ATTENTION : Al Smith



Laboratory Certification ID # 20012



1) PESTICIDE-TCL Data	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	12
9) QC Data Summary For Pesticide-TCL	13
9.1) Deuterated Monitoring Compound Summary	14
9.2) LCS/LCSD Summary	16
9.3) Method Blank Summary	18
10) Sample Data	19
10.1) TW-705R-S	20
10.2) TW-705R-SDL	26
10.3) TW-22M-W	32
11) Calibration Data Summary	39
11.1) Initial Calibration Data	40
11.1.1) PL072825	40
11.2) Continued Calibration Data	143
11.2.1) PL096819.D	143
11.2.2) PL096828.D	161
11.2.3) PL096837.D	179
11.2.4) PL096841.D	197
11.2.5) PL096850.D	215
11.2.6) PL096858.D	233
11.2.7) PEM Files	251
11.3) RESCHK Data	291
11.4) Analytical Seq	299
12) Compound Detection Summary	301
13) QC Sample Data	308
13.1) Method Blank Data	309
13.2) PIBLK Data	314
13.3) LCS Data	349
13.4) LCSD Data	365

Table Of Contents for Q2815

14) Manual Integration	381
15) Analytical Runlogs	396
16) Extraction Logs	409
16.1) PB169225.pdf	409
16.2) PB169225IC.pdf	411
17) Standard Prep Logs	412
18) Shipping Document	480
18.1) Chain Of Custody	481
18.2) ROC	483
18.3) Lab Certificate	486
18.4) Internal COC	487
19) Not Reviewed Data	491

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Cover Page

Order ID : Q2815

Project ID : USACE018-44 DOD

Client : First Environment, Inc.

Lab Sample Number

Q2815-01
Q2815-02
Q2815-03
Q2815-04
Q2815-05
Q2815-06
Q2815-07
Q2815-08
Q2815-09
Q2815-10
Q2815-11
Q2815-12
Q2815-13
Q2815-14
Q2815-15
Q2815-16
Q2815-17
Q2815-18
Q2815-19
Q2815-20
Q2815-21
Q2815-22
Q2815-23
Q2815-24
Q2815-25
Q2815-26

Client Sample Number

TW-705R-S
TW-10PC-W
TW-10P-E
TW-10P-S
TW-10P-W
TW-10P-N
TW-88H-E
TW-88H-N
TW-88H-W
TW-88H-S
TW-22M-W
TW-22M-S
TW-22M-E
TW-22M-N
TW-17M-E
TW-17M-S
TW-84SB-S
TW-84SB-W
DUP
TW-11M-W
TW-11M-E
TW-11M-S
TW-11M-N
TB
TW-11M-W
FB

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:03 pm, Aug 22, 2025

Date: 8/21/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

First Environment, Inc.

Project Name: USACE018-44 DOD

Project # N/A

Order ID # Q2815

Test Name: Pesticide-TCL

A. Number of Samples and Date of Receipt:

2 Water samples were received on 08/08/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Pesticide-TCL. This data package contains results for Pesticide-TCL.

C. Analytical Techniques:

The analysis was performed on instrument ECD_L. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df.; Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of Pesticide-TCLs was based on method 8081B and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 20% in the Initial Calibration (Method PL072825.M) for Endrin aldehyde in 2nd column, this Compound is passing on Linear regression.

The Continuous Calibration met the requirements.

Sample TW-705R-S was diluted due to high concentration.

E. Additional Comments:

The not QT review data is reported in the Miscellaneous.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:03 pm, Aug 22, 2025

Signature _____

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2815

MATRIX: Water

METHOD: 8081B/3510

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified.			✓
2. Standard Summary Submitted.			✓
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD. The %RSD is greater than 20% in the Initial Calibration (Method PL072825.M) for Endrin aldehyde in 2nd column, this Compound is passing on Linear regression. The Continuous Calibration met the requirements.			✓
4. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
6. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all compounds. The Blank Spike Duplicate met requirements for all compounds. The RPD were met for all analysis.			✓
7. Retention Time Shift Meet Criteria (if applicable) Comments:			✓
8. Extraction Holding Time Met If not met, list number of days exceeded for each sample:			✓



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

	NA	NO	YES
9. Analysis Holding Time Met			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			

ADDITIONAL COMMENTS:

Sample TW-705R-S was diluted due to high concentration.

The not QT review data is reported in the Miscellaneous.

QA REVIEW

REVIEWED

Sohil Jodhani, QA/QC Director , 8/22/2025, 11:53:18 AM

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2815

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 08/21/2025

LAB CHRONICLE

OrderID:	Q2815	OrderDate:	8/11/2025 10:33:09 AM
Client:	First Environment, Inc.	Project:	USACE018-44 DOD
Contact:	Al Smith	Location:	D41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2815-01	TW-705R-S	WATER	PCB	8082A	08/06/25	08/12/25	08/13/25	08/08/25
			Pesticide-TCL	8081B		08/12/25	08/18/25	
Q2815-01DL	TW-705R-SDL	WATER	Pesticide-TCL	8081B	08/06/25	08/12/25	08/18/25	08/08/25
Q2815-11	TW-22M-W	WATER	PCB	8082A	08/08/25	08/12/25	08/13/25	08/08/25
			Pesticide-TCL	8081B		08/12/25	08/15/25	
Q2815-21	TW-11M-E	WATER	PCB	8082A	08/08/25	08/12/25	08/13/25	08/08/25
Q2815-23	TW-11M-N	WATER	PCB	8082A	08/08/25	08/12/25	08/13/25	08/08/25
Q2815-26	FB	WATER	PCB	8082A	08/08/25	08/12/25	08/13/25	08/08/25

Hit Summary Sheet
SW-846

SDG No.: Q2815

Order ID: Q2815

Client: First Environment, Inc.

Project ID: USACE018-44 DOD

Sample ID	Client ID		Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : TW-705R-S									
Q2815-01	TW-705R-S	WATER	4,4-DDE	0.54		0.0037	0.010	0.050	ug/L
Q2815-01	TW-705R-S	WATER	4,4-DDD	1.60	E	0.0071	0.025	0.050	ug/L
Total Concentration:				2.140					
Client ID : TW-705R-SDL									
Q2815-01DL	TW-705R-SDL	WATER	4,4-DDE	0.54	D	0.0074	0.020	0.10	ug/L
Q2815-01DL	TW-705R-SDL	WATER	4,4-DDD	1.50	D	0.014	0.050	0.10	ug/L
Total Concentration:				2.040					
Client ID : TW-22M-W									
Q2815-11	TW-22M-W	WATER	Dieldrin	0.0093	J	0.0036	0.010	0.051	ug/L
Q2815-11	TW-22M-W	WATER	4,4-DDT	0.065	P	0.0035	0.010	0.051	ug/L
Q2815-11	TW-22M-W	WATER	alpha-Chlordane	0.016	J	0.0035	0.010	0.051	ug/L
Total Concentration:				0.090					



QC SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Surrogate Summary

SDG No.: **Q2815**

Client: **First Environment, Inc.**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PL096593.D	PIBLK-PL096593.D	Decachlorobiphen	1	20	19.2	96		30	135
		Tetrachloro-m-xyl	1	20	18.6	93		44	124
		Decachlorobiphen	2	20	19.1	95		30	135
		Tetrachloro-m-xyl	2	20	18.3	92		44	124
I.BLK-PL096817.D	PIBLK-PL096817.D	Decachlorobiphen	1	20	22.6	113		30	135
		Tetrachloro-m-xyl	1	20	21.1	105		44	124
		Decachlorobiphen	2	20	19.3	97		30	135
		Tetrachloro-m-xyl	2	20	19.9	100		44	124
PB169225BL	PB169225BL	Decachlorobiphen	1	20	20.5	102		30	135
		Tetrachloro-m-xyl	1	20	19.6	98		44	124
		Decachlorobiphen	2	20	20.6	103		30	135
		Tetrachloro-m-xyl	2	20	19.7	99		44	124
PB169225BS	PB169225BS	Decachlorobiphen	1	20	20.5	102		30	135
		Tetrachloro-m-xyl	1	20	19.1	96		44	124
		Decachlorobiphen	2	20	20.9	105		30	135
		Tetrachloro-m-xyl	2	20	19.7	99		44	124
PB169225BSD	PB169225BSD	Decachlorobiphen	1	20	20.2	101		30	135
		Tetrachloro-m-xyl	1	20	18.7	94		44	124
		Decachlorobiphen	2	20	20.8	104		30	135
		Tetrachloro-m-xyl	2	20	19.3	96		44	124
I.BLK-PL096827.D	PIBLK-PL096827.D	Decachlorobiphen	1	20	21.9	109		30	135
		Tetrachloro-m-xyl	1	20	20.2	101		44	124
		Decachlorobiphen	2	20	22.1	110		30	135
		Tetrachloro-m-xyl	2	20	20.6	103		44	124
I.BLK-PL096835.D	PIBLK-PL096835.D	Decachlorobiphen	1	20	21.5	107		30	135
		Tetrachloro-m-xyl	1	20	20.4	102		44	124
		Decachlorobiphen	2	20	20.8	104		30	135
		Tetrachloro-m-xyl	2	20	20.5	102		44	124
Q2815-11	TW-22M-W	Decachlorobiphen	1	20	12.7	64		30	135
		Tetrachloro-m-xyl	1	20	16.8	84		44	124
		Decachlorobiphen	2	20	12.5	63		30	135
		Tetrachloro-m-xyl	2	20	18.0	90		44	124
I.BLK-PL096840.D	PIBLK-PL096840.D	Decachlorobiphen	1	20	22.8	114		30	135
		Tetrachloro-m-xyl	1	20	20.7	103		44	124
		Decachlorobiphen	2	20	22.0	110		30	135
		Tetrachloro-m-xyl	2	20	20.6	103		44	124
I.BLK-PL096849.D	PIBLK-PL096849.D	Decachlorobiphen	1	20	22.3	111		30	135
		Tetrachloro-m-xyl	1	20	21.1	106		44	124
		Decachlorobiphen	2	20	21.0	105		30	135
		Tetrachloro-m-xyl	2	20	21.3	107		44	124
Q2815-01	TW-705R-S	Decachlorobiphen	1	20	10.5	52		30	135

Surrogate Summary

SDG No.: Q2815

Client: First Environment, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2815-01	TW-705R-S	Tetrachloro-m-xyl	1	20	12.9	65		44	124
		Decachlorobiphen	2	20	9.27	46		30	135
Q2815-01DL	TW-705R-SDL	Tetrachloro-m-xyl	2	20	14.2	71		44	124
		Decachlorobiphen	1	20	10.7	53		30	135
		Tetrachloro-m-xyl	1	20	13.5	67		44	124
		Decachlorobiphen	2	20	10.0	50		30	135
		Tetrachloro-m-xyl	2	20	14.7	74		44	124
I.BLK-PL096857.D	PIBLK-PL096857.D	Decachlorobiphen	1	20	22.5	112		30	135
		Tetrachloro-m-xyl	1	20	21.1	105		44	124
		Decachlorobiphen	2	20	22.0	110		30	135
		Tetrachloro-m-xyl	2	20	21.0	105		44	124

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2815</u>	Analytical Method:	<u>8081B</u>
Client:	<u>First Environment, Inc.</u>	Datafile :	<u>PL096823.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB169225BS (Column 1)	alpha-BHC	0.5	0.50	ug/L	100				54	138		
	beta-BHC	0.5	0.50	ug/L	100				56	136		
	delta-BHC	0.5	0.50	ug/L	101				52	142		
	gamma-BHC (Lindane)	0.5	0.50	ug/L	99				59	134		
	Heptachlor	0.5	0.53	ug/L	106				54	130		
	Aldrin	0.5	0.49	ug/L	99				45	134		
	Heptachlor epoxide	0.5	0.51	ug/L	101				61	133		
	Endosulfan I	0.5	0.49	ug/L	98				62	126		
	Dieldrin	0.5	0.49	ug/L	99				60	136		
	4,4'-DDE	0.5	0.49	ug/L	98				57	135		
	Endrin	0.5	0.51	ug/L	102				60	138		
	Endosulfan II	0.5	0.51	ug/L	101				52	135		
	4,4'-DDD	0.5	0.50	ug/L	101				56	143		
	Endosulfan sulfate	0.5	0.48	ug/L	96				62	133		
	4,4'-DDT	0.5	0.50	ug/L	99				51	143		
	Methoxychlor	0.5	0.49	ug/L	97				54	145		
	Endrin ketone	0.5	0.49	ug/L	98				58	134		
	Endrin aldehyde	0.5	0.53	ug/L	106				51	132		
	alpha-Chlordane	0.5	0.50	ug/L	100				60	129		
	gamma-Chlordane	0.5	0.50	ug/L	101				56	136		
	alpha-BHC	0.5	0.50	ug/L	100				54	138		
PB169225BS (Column 2)	beta-BHC	0.5	0.50	ug/L	99				56	136		
	delta-BHC	0.5	0.50	ug/L	100				52	142		
	gamma-BHC (Lindane)	0.5	0.50	ug/L	100				59	134		
	Heptachlor	0.5	0.49	ug/L	98				54	130		
	Aldrin	0.5	0.50	ug/L	99				45	134		
	Heptachlor epoxide	0.5	0.49	ug/L	98				61	133		
	Endosulfan I	0.5	0.48	ug/L	95				62	126		
	Dieldrin	0.5	0.49	ug/L	98				60	136		
	4,4'-DDE	0.5	0.49	ug/L	99				57	135		
	Endrin	0.5	0.48	ug/L	97				60	138		
	Endosulfan II	0.5	0.49	ug/L	97				52	135		
	4,4'-DDD	0.5	0.49	ug/L	99				56	143		
	Endosulfan sulfate	0.5	0.48	ug/L	97				62	133		
	4,4'-DDT	0.5	0.49	ug/L	97				51	143		
	Methoxychlor	0.5	0.48	ug/L	97				54	145		
	Endrin ketone	0.5	0.52	ug/L	104				58	134		
	Endrin aldehyde	0.5	0.55	ug/L	110				51	132		
	alpha-Chlordane	0.5	0.49	ug/L	97				60	129		
	gamma-Chlordane	0.5	0.50	ug/L	99				56	136		

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2815</u>	Analytical Method:	<u>8081B</u>
Client:	<u>First Environment, Inc.</u>	Datafile :	<u>PL096824.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169225BSD (Column 1)	alpha-BHC	0.5	0.48	ug/L	96	4			54	138	20
	beta-BHC	0.5	0.49	ug/L	97	3			56	136	20
	delta-BHC	0.5	0.49	ug/L	98	3			52	142	20
	gamma-BHC (Lindane)	0.5	0.48	ug/L	95	4			59	134	20
	Heptachlor	0.5	0.52	ug/L	103	3			54	130	20
	Aldrin	0.5	0.48	ug/L	96	3			45	134	20
	Heptachlor epoxide	0.5	0.49	ug/L	99	2			61	133	20
	Endosulfan I	0.5	0.48	ug/L	97	1			62	126	20
	Dieldrin	0.5	0.49	ug/L	97	2			60	136	20
	4,4'-DDE	0.5	0.50	ug/L	99	1			57	135	20
	Endrin	0.5	0.49	ug/L	98	4			60	138	20
	Endosulfan II	0.5	0.51	ug/L	102	1			52	135	20
	4,4'-DDD	0.5	0.49	ug/L	99	2			56	143	20
	Endosulfan sulfate	0.5	0.48	ug/L	95	1			62	133	20
	4,4'-DDT	0.5	0.48	ug/L	96	3			51	143	20
	Methoxychlor	0.5	0.47	ug/L	93	4			54	145	20
	Endrin ketone	0.5	0.48	ug/L	96	2			58	134	20
	Endrin aldehyde	0.5	0.52	ug/L	104	2			51	132	20
	alpha-Chlordane	0.5	0.49	ug/L	98	2			60	129	20
	gamma-Chlordane	0.5	0.49	ug/L	98	3			56	136	20
	alpha-BHC	0.5	0.48	ug/L	97	3			54	138	20
	beta-BHC	0.5	0.48	ug/L	96	3			56	136	20
	delta-BHC	0.5	0.48	ug/L	97	3			52	142	20
	gamma-BHC (Lindane)	0.5	0.48	ug/L	97	3			59	134	20
	Heptachlor	0.5	0.48	ug/L	95	3			54	130	20
	Aldrin	0.5	0.48	ug/L	96	3			45	134	20
PB169225BSD (Column 2)	Heptachlor epoxide	0.5	0.48	ug/L	96	2			61	133	20
	Endosulfan I	0.5	0.46	ug/L	92	3			62	126	20
	Dieldrin	0.5	0.48	ug/L	96	2			60	136	20
	4,4'-DDE	0.5	0.48	ug/L	96	3			57	135	20
	Endrin	0.5	0.47	ug/L	94	3			60	138	20
	Endosulfan II	0.5	0.48	ug/L	95	2			52	135	20
	4,4'-DDD	0.5	0.48	ug/L	97	2			56	143	20
	Endosulfan sulfate	0.5	0.47	ug/L	95	2			62	133	20
	4,4'-DDT	0.5	0.47	ug/L	94	3			51	143	20
	Methoxychlor	0.5	0.47	ug/L	94	3			54	145	20
	Endrin ketone	0.5	0.49	ug/L	98	6			58	134	20
	Endrin aldehyde	0.5	0.54	ug/L	108	2			51	132	20
	alpha-Chlordane	0.5	0.47	ug/L	95	2			60	129	20
	gamma-Chlordane	0.5	0.48	ug/L	97	2			56	136	20

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB169225BL

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Lab Sample ID: PB169225BL

Lab File ID: PL096822.D

Matrix: (soil/water) WATER

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 08/12/2025

Date Analyzed (1): 08/15/2025

Date Analyzed (2): 08/15/2025

Time Analyzed (1): 16:07

Time Analyzed (2): 16:07

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB169225BS	PB169225BS	PL096823.D	08/15/2025	08/15/2025
PB169225BSD	PB169225BSD	PL096824.D	08/15/2025	08/15/2025
TW-22M-W	Q2815-11	PL096839.D	08/15/2025	08/15/2025
TW-705R-S	Q2815-01	PL096855.D	08/18/2025	08/18/2025

COMMENTS: _____



SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/06/25
Project:	USACE018-44 DOD	Date Received:	08/08/25
Client Sample ID:	TW-705R-S	SDG No.:	Q2815
Lab Sample ID:	Q2815-01	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096855.D	1	08/12/25 10:20	08/18/25 16:30	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.010	U	0.0039	0.010	0.050	ug/L
319-85-7	beta-BHC	0.010	U	0.0049	0.010	0.050	ug/L
319-86-8	delta-BHC	0.025	U	0.011	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.010	U	0.0037	0.010	0.050	ug/L
76-44-8	Heptachlor	0.010	U	0.0027	0.010	0.050	ug/L
309-00-2	Aldrin	0.010	U	0.0036	0.010	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0096	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.010	U	0.0031	0.010	0.050	ug/L
60-57-1	Dieldrin	0.010	U	0.0036	0.010	0.050	ug/L
72-55-9	4,4-DDE	0.54		0.0037	0.010	0.050	ug/L
72-20-8	Endrin	0.010	U	0.0032	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0079	0.025	0.050	ug/L
72-54-8	4,4-DDD	1.60	E	0.0071	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.010	U	0.0037	0.010	0.050	ug/L
50-29-3	4,4-DDT	0.010	U	0.0035	0.010	0.050	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0093	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.011	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.010	U	0.0035	0.010	0.050	ug/L
5103-74-2	gamma-Chlordane	0.010	U	0.0039	0.010	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.17	0.50	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	10.5		30 - 135		52%	SPK: 20
877-09-8	Tetrachloro-m-xylene	14.2		44 - 124		71%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/06/25
Project:	USACE018-44 DOD	Date Received:	08/08/25
Client Sample ID:	TW-705R-S	SDG No.:	Q2815
Lab Sample ID:	Q2815-01	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096855.D	1	08/12/25 10:20	08/18/25 16:30	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096855.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 16:30
Operator : AR\AJ
Sample : Q2815-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TW-705R-S

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.529	2.823	41094470	68001961	12.920m	14.233
28) SA Decachlor...	9.005	7.987	25022927	40225957	10.493m	9.272
Target Compounds						
12) B 4,4'-DDE	6.161	5.299	173.6E6	251.1E6	54.062	45.534
16) A 4,4'-DDD	6.669	5.852	405.9E6	635.7E6	160.537	135.109

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096855.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 16:30
Operator : AR\AJ
Sample : Q2815-01
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
TW-705R-S

Manual Integrations
APPROVED


Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Response_ Signal: PL096855.D\ECD1A.ch



#1 Tetrachloro-m-xylene

R.T.: 3.529 min
Delta R.T.: -0.006 min
Response: 41094470
Conc: 12.92 ng/ml

Instrument :

ECD_L

ClientSampleId :

TW-705R-S


Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/19/2025

Supervised By :mohammad ahmed 08/20/2025


Time Response_ Signal: PL096855.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.823 min
Delta R.T.: -0.005 min
Response: 68001961
Conc: 14.23 ng/ml


Time Response_ Signal: PL096855.D\ECD1A.ch



#12 4,4'-DDE

R.T.: 6.161 min
Delta R.T.: -0.006 min
Response: 173606744
Conc: 54.06 ng/ml


Time Response_ Signal: PL096855.D\ECD2B.ch



#12 4,4'-DDE

R.T.: 5.299 min
Delta R.T.: -0.005 min
Response: 251094235
Conc: 45.53 ng/ml

Signal: PL096855.D\ECD1A.ch



#16 4,4'-DDD


R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 405896514
Conc: 160.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-705R-S

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025


Signal: PL096855.D\ECD2B.ch



#16 4,4'-DDD

R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 635727139
Conc: 135.11 ng/ml


Signal: PL096855.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 9.005 min
Delta R.T.: -0.011 min
Response: 25022927
Conc: 10.49 ng/ml m

Signal: PL096855.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 40225957
Conc: 9.27 ng/ml

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/06/25
Project:	USACE018-44 DOD	Date Received:	08/08/25
Client Sample ID:	TW-705R-SDL	SDG No.:	Q2815
Lab Sample ID:	Q2815-01DL	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096856.D	2	08/12/25 10:20	08/18/25 16:44	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.020	UD	0.0078	0.020	0.10	ug/L
319-85-7	beta-BHC	0.020	UD	0.0098	0.020	0.10	ug/L
319-86-8	delta-BHC	0.050	UD	0.022	0.050	0.10	ug/L
58-89-9	gamma-BHC (Lindane)	0.020	UD	0.0074	0.020	0.10	ug/L
76-44-8	Heptachlor	0.020	UD	0.0054	0.020	0.10	ug/L
309-00-2	Aldrin	0.020	UD	0.0072	0.020	0.10	ug/L
1024-57-3	Heptachlor epoxide	0.050	UD	0.019	0.050	0.10	ug/L
959-98-8	Endosulfan I	0.020	UD	0.0062	0.020	0.10	ug/L
60-57-1	Dieldrin	0.020	UD	0.0072	0.020	0.10	ug/L
72-55-9	4,4-DDE	0.54	D	0.0074	0.020	0.10	ug/L
72-20-8	Endrin	0.020	UD	0.0064	0.020	0.10	ug/L
33213-65-9	Endosulfan II	0.050	UD	0.016	0.050	0.10	ug/L
72-54-8	4,4-DDD	1.50	D	0.014	0.050	0.10	ug/L
1031-07-8	Endosulfan Sulfate	0.020	UD	0.0074	0.020	0.10	ug/L
50-29-3	4,4-DDT	0.020	UD	0.0070	0.020	0.10	ug/L
72-43-5	Methoxychlor	0.050	UD	0.022	0.050	0.10	ug/L
53494-70-5	Endrin ketone	0.050	UD	0.019	0.050	0.10	ug/L
7421-93-4	Endrin aldehyde	0.050	UD	0.022	0.050	0.10	ug/L
5103-71-9	alpha-Chlordane	0.020	UD	0.0070	0.020	0.10	ug/L
5103-74-2	gamma-Chlordane	0.020	UD	0.0078	0.020	0.10	ug/L
8001-35-2	Toxaphene	1.00	UD	0.34	1.00	2.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	10.7		30 - 135		53%	SPK: 20
877-09-8	Tetrachloro-m-xylene	14.7		44 - 124		74%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/06/25
Project:	USACE018-44 DOD	Date Received:	08/08/25
Client Sample ID:	TW-705R-SDL	SDG No.:	Q2815
Lab Sample ID:	Q2815-01DL	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096856.D	2	08/12/25 10:20	08/18/25 16:44	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096856.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 16:44
Operator : AR\AJ
Sample : Q2815-01DL 2X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TW-705R-SDL

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.529	2.825	21430701	35190031	6.738m	7.365
2) SA Decachlor...	9.005	7.986	12736470	21675838	5.341	4.996
Target Compounds						
12) B 4,4'-DDE	6.159	5.299	87221532	123.0E6	27.161	22.307
16) A 4,4'-DDD	6.668	5.851	190.3E6	320.5E6	75.265	68.109

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096856.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 16:44
Operator : AR\AJ
Sample : Q2815-01DL 2X
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
TW-705R-SDL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.529 min
Delta R.T.: -0.006 min
Response: 21430701
Conc: 6.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-705R-SDL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 35190031
Conc: 7.37 ng/ml



#12 4,4'-DDE

R.T.: 6.159 min
Delta R.T.: -0.007 min
Response: 87221532
Conc: 27.16 ng/ml




#12 4,4'-DDE

R.T.: 5.299 min
Delta R.T.: -0.005 min
Response: 123007575
Conc: 22.31 ng/ml

Signal: PL096856.D\ECD1A.ch

#16 4,4'-DDD



R.T.: 6.668 min
Delta R.T.: -0.007 min
Response: 190297206
Conc: 75.26 ng/ml


Instrument :
ECD_L
ClientSampleId :
TW-705R-SDL

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Signal: PL096856.D\ECD2B.ch


#16 4,4'-DDD



R.T.: 5.851 min
Delta R.T.: -0.005 min
Response: 320473038
Conc: 68.11 ng/ml

Signal: PL096856.D\ECD1A.ch


#28 Decachlorobiphenyl



R.T.: 9.005 min
Delta R.T.: -0.011 min
Response: 12736470
Conc: 5.34 ng/ml

Signal: PL096856.D\ECD2B.ch

#28 Decachlorobiphenyl



R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 21675838
Conc: 5.00 ng/ml

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/08/25
Project:	USACE018-44 DOD	Date Received:	08/08/25
Client Sample ID:	TW-22M-W	SDG No.:	Q2815
Lab Sample ID:	Q2815-11	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	990	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096839.D	1	08/12/25 10:20	08/15/25 21:15	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.010	U	0.0039	0.010	0.051	ug/L
319-85-7	beta-BHC	0.010	U	0.0049	0.010	0.051	ug/L
319-86-8	delta-BHC	0.025	U	0.011	0.025	0.051	ug/L
58-89-9	gamma-BHC (Lindane)	0.010	U	0.0037	0.010	0.051	ug/L
76-44-8	Heptachlor	0.010	U	0.0027	0.010	0.051	ug/L
309-00-2	Aldrin	0.010	U	0.0036	0.010	0.051	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0097	0.025	0.051	ug/L
959-98-8	Endosulfan I	0.010	U	0.0031	0.010	0.051	ug/L
60-57-1	Dieldrin	0.0093	J	0.0036	0.010	0.051	ug/L
72-55-9	4,4-DDE	0.010	U	0.0037	0.010	0.051	ug/L
72-20-8	Endrin	0.010	U	0.0032	0.010	0.051	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0080	0.025	0.051	ug/L
72-54-8	4,4-DDD	0.025	U	0.0072	0.025	0.051	ug/L
1031-07-8	Endosulfan Sulfate	0.010	U	0.0037	0.010	0.051	ug/L
50-29-3	4,4-DDT	0.065	P	0.0035	0.010	0.051	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.051	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0094	0.025	0.051	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.011	0.025	0.051	ug/L
5103-71-9	alpha-Chlordane	0.016	J	0.0035	0.010	0.051	ug/L
5103-74-2	gamma-Chlordane	0.010	U	0.0039	0.010	0.051	ug/L
8001-35-2	Toxaphene	0.51	U	0.17	0.51	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	12.7		30 - 135		64%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.0		44 - 124		90%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/08/25
Project:	USACE018-44 DOD	Date Received:	08/08/25
Client Sample ID:	TW-22M-W	SDG No.:	Q2815
Lab Sample ID:	Q2815-11	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	990	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
Prep Method :	3510C	Test:	Pesticide-TCL
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096839.D	1	08/12/25 10:20	08/15/25 21:15	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096839.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 21:15
 Operator : AR\AJ
 Sample : Q2815-11
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TW-22M-W

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 08/18/2025
 Supervised By : mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:57:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.530	2.826	53404978	85851950	16.790m	17.969
28) SA Decachlor...	9.003	7.985	30309997	54215259	12.710m	12.497m
Target Compounds						
11) B alpha-Chl...	5.995	5.102	6222106	9254060	1.614m	1.569m
13) MA Dieldrin	6.309	5.429	3076047	5442748	0.829m	0.921m
17) MA 4,4'-DDT	6.982	6.120	7635862	32723031	2.663	6.469 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096839.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 21:15
Operator : AR\AJ
Sample : Q2815-11
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
TW-22M-W


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:57:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 53404978
Conc: 16.79 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-22M-W


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 85851950
Conc: 17.97 ng/ml




#11 alpha-Chlordane

R.T.: 5.995 min
Delta R.T.: -0.001 min
Response: 6222106
Conc: 1.61 ng/ml m



#11 alpha-Chlordane

R.T.: 5.102 min
Delta R.T.: -0.013 min
Response: 9254060
Conc: 1.57 ng/ml m




#13 Dieldrin

R.T.: 6.309 min
Delta R.T.: -0.007 min
Response: 3076047
Conc: 0.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-22M-W


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#13 Dieldrin

R.T.: 5.429 min
Delta R.T.: -0.005 min
Response: 5442748
Conc: 0.92 ng/ml m




#17 4,4'-DDT

R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 7635862
Conc: 2.66 ng/ml



#17 4,4'-DDT

R.T.: 6.120 min
Delta R.T.: 0.011 min
Response: 32723031
Conc: 6.47 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.003 min
Delta R.T.: -0.013 min
Response: 30309997
Conc: 12.71 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-22M-W

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025



#28 Decachlorobiphenyl

R.T.: 7.985 min
Delta R.T.: -0.007 min
Response: 54215259
Conc: 12.50 ng/ml m



CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2815</u>
Instrument ID:	<u>ECD_L</u>	Calibration Date(s):	<u>07/28/2025</u> <u>07/28/2025</u>
		Calibration Times:	<u>16:52</u> <u>17:47</u>

 GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PL096596.D</u>	RT 075 = <u>PL096597.D</u>
RT 050 = <u>PL096598.D</u>	RT 025 = <u>PL096599.D</u>	RT 005 = <u>PL096600.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
4,4'-DDD	6.68	6.68	6.68	6.68	6.67	6.68	6.58	6.78
4,4'-DDE	6.17	6.17	6.17	6.17	6.17	6.17	6.07	6.27
4,4'-DDT	6.99	6.99	6.99	6.99	6.99	6.99	6.89	7.09
Aldrin	5.24	5.24	5.24	5.24	5.24	5.24	5.14	5.34
alpha-BHC	3.98	3.98	3.98	3.98	3.98	3.98	3.88	4.08
alpha-Chlordane	6.00	6.00	6.00	6.00	6.00	6.00	5.90	6.10
beta-BHC	4.50	4.50	4.50	4.50	4.50	4.50	4.40	4.60
Decachlorobiphenyl	9.02	9.02	9.02	9.02	9.02	9.02	8.92	9.12
delta-BHC	4.74	4.74	4.74	4.74	4.74	4.74	4.64	4.84
Dieldrin	6.32	6.32	6.32	6.32	6.32	6.32	6.22	6.42
Endosulfan I	6.04	6.04	6.04	6.04	6.04	6.04	5.94	6.14
Endosulfan II	6.76	6.76	6.76	6.76	6.75	6.75	6.65	6.85
Endosulfan sulfate	7.12	7.12	7.12	7.12	7.12	7.12	7.02	7.22
Endrin	6.54	6.54	6.54	6.54	6.54	6.54	6.44	6.64
Endrin aldehyde	6.88	6.88	6.88	6.88	6.88	6.88	6.78	6.98
Endrin ketone	7.60	7.60	7.60	7.60	7.60	7.60	7.50	7.70
gamma-BHC (Lindane)	4.31	4.31	4.31	4.31	4.31	4.31	4.21	4.41
gamma-Chlordane	5.92	5.92	5.92	5.92	5.92	5.92	5.82	6.02
Heptachlor	4.90	4.90	4.90	4.90	4.90	4.90	4.80	5.00
Heptachlor epoxide	5.66	5.66	5.66	5.66	5.66	5.66	5.56	5.76
Methoxychlor	7.46	7.46	7.46	7.46	7.46	7.46	7.36	7.56
Tetrachloro-m-xylene	3.54	3.54	3.54	3.54	3.53	3.53	3.43	3.63

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2815</u>
Instrument ID:	<u>ECD_L</u>	Calibration Date(s):	<u>07/28/2025</u> <u>07/28/2025</u>
		Calibration Times:	<u>16:52</u> <u>17:47</u>

 GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PL096596.D</u>	RT 075 = <u>PL096597.D</u>
RT 050 = <u>PL096598.D</u>	RT 025 = <u>PL096599.D</u>	RT 005 = <u>PL096600.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
4,4'-DDD	5.86	5.86	5.86	5.86	5.86	5.86	5.76	5.96
4,4'-DDE	5.30	5.30	5.30	5.30	5.30	5.30	5.20	5.40
4,4'-DDT	6.11	6.11	6.11	6.11	6.11	6.11	6.01	6.21
Aldrin	4.30	4.30	4.30	4.30	4.30	4.30	4.20	4.40
alpha-BHC	3.33	3.33	3.33	3.33	3.33	3.33	3.23	3.43
alpha-Chlordane	5.12	5.12	5.12	5.12	5.12	5.11	5.01	5.21
beta-BHC	3.96	3.96	3.96	3.96	3.96	3.96	3.86	4.06
Decachlorobiphenyl	7.99	7.99	7.99	7.99	7.99	7.99	7.89	8.09
delta-BHC	4.20	4.20	4.20	4.20	4.19	4.19	4.09	4.29
Dieldrin	5.44	5.44	5.43	5.43	5.43	5.43	5.33	5.53
Endosulfan I	5.17	5.17	5.17	5.17	5.17	5.17	5.07	5.27
Endosulfan II	6.00	6.00	6.00	6.00	6.00	6.00	5.90	6.10
Endosulfan sulfate	6.40	6.40	6.40	6.40	6.40	6.40	6.30	6.50
Endrin	5.71	5.71	5.71	5.71	5.71	5.71	5.61	5.81
Endrin aldehyde	6.18	6.18	6.18	6.18	6.18	6.18	6.08	6.28
Endrin ketone	6.91	6.91	6.91	6.91	6.91	6.91	6.81	7.01
gamma-BHC (Lindane)	3.67	3.67	3.67	3.67	3.66	3.66	3.56	3.76
gamma-Chlordane	5.05	5.05	5.05	5.05	5.05	5.05	4.95	5.15
Heptachlor	4.01	4.01	4.01	4.01	4.01	4.01	3.91	4.11
Heptachlor epoxide	4.80	4.80	4.80	4.80	4.80	4.80	4.70	4.90
Methoxychlor	6.68	6.68	6.68	6.68	6.68	6.68	6.58	6.78
Tetrachloro-m-xylene	2.83	2.83	2.83	2.83	2.83	2.83	2.73	2.93

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_L

Contract: FIRS02
SDG NO.: Q2815
Calibration Date(s): 07/28/2025 07/28/2025
Calibration Times: 16:52 17:47

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 =	PL096596.D	CF 075 =	PL096597.D		
CF 050 =	PL096598.D	CF 025 =	PL096599.D	CF 005 =	PL096600.D		
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	2615150000	2542280000	2523890000	2500980000	2459540000	2528370000	2
4,4'-DDE	3382020000	3179740000	3147240000	3113410000	3233830000	3211250000	3
4,4'-DDT	2980050000	2901890000	2884740000	2837350000	2734400000	2867690000	3
Aldrin	4418080000	4273520000	4268380000	4243090000	4301950000	4301000000	2
alpha-BHC	4873630000	4685670000	4645170000	4549100000	4378210000	4626350000	4
alpha-Chlordane	3883510000	3780330000	3801880000	3877780000	3929720000	3854650000	2
beta-BHC	1772460000	1750490000	1781040000	1856990000	1863870000	1804970000	3
Decachlorobiphenyl	2289370000	2250980000	2316430000	2450150000	2616360000	2384660000	6
delta-BHC	4161630000	4012240000	3978460000	3955030000	3841950000	3989860000	3
Dieldrin	3849470000	3737590000	3712880000	3678490000	3577310000	3711150000	3
Endosulfan I	3668500000	3524950000	3536580000	3570970000	3680850000	3596370000	2
Endosulfan II	3069420000	3008630000	3042800000	3175650000	3716720000	3202640000	9
Endosulfan sulfate	2875180000	2814890000	2851640000	2870280000	2956480000	2873690000	2
Endrin	3129990000	2990370000	2986700000	3066080000	2960790000	3026790000	2
Endrin aldehyde	2129160000	2102960000	2124970000	2164420000	2208060000	2145920000	2
Endrin ketone	3023290000	2975820000	3003920000	3029220000	3001370000	3006720000	1
gamma-BHC (Lindane)	4576650000	4410990000	4390520000	4378210000	4360850000	4423440000	2
gamma-Chlordane	3912350000	3800880000	3800750000	3806650000	3705560000	3805240000	2
Heptachlor	4215340000	4096230000	4129500000	4154230000	4183150000	4155690000	1
Heptachlor epoxide	3888140000	3791890000	3847180000	3911860000	3843640000	3856540000	1
Methoxychlor	1449590000	1444460000	1478010000	1507550000	1462160000	1468350000	2
Tetrachloro-m-xylene	3176500000	3106160000	3136600000	3226680000	3257980000	3180780000	2

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_L

Contract: FIRS02
SDG NO.: Q2815
Calibration Date(s): 07/28/2025 07/28/2025
Calibration Times: 16:52 17:47

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 =	PL096596.D	CF 075 =	PL096597.D		
CF 050 =	PL096598.D	CF 025 =	PL096599.D	CF 005 =	PL096600.D		
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	4739770000	4648800000	4693360000	4702740000	4741750000	4705280000	1
4,4'-DDE	5653430000	5561070000	5540610000	5437860000	5379080000	5514410000	2
4,4'-DDT	5212890000	5114220000	5118490000	5045850000	4800030000	5058300000	3
Aldrin	6310530000	6202530000	6230090000	6182780000	6109380000	6207060000	1
alpha-BHC	7308810000	7159610000	7137170000	7045380000	6734230000	7077040000	3
alpha-Chlordane	5806170000	5731520000	5800200000	5896310000	6260360000	5898910000	4
beta-BHC	2740170000	2730810000	2782730000	2873420000	2977360000	2820900000	4
Decachlorobiphenyl	4108230000	4098750000	4223500000	4429960000	4831160000	4338320000	7
delta-BHC	6621180000	6509520000	6495090000	6469880000	6272460000	6473630000	2
Dieldrin	5973840000	5884270000	5897970000	5923090000	5853560000	5906550000	1
Endosulfan I	5357100000	5308830000	5362730000	5541910000	6086480000	5531410000	6
Endosulfan II	5116760000	5062950000	5122660000	5175210000	5203600000	5136240000	1
Endosulfan sulfate	5007540000	4977340000	5062730000	5135950000	5247510000	5086210000	2
Endrin	5415430000	5344790000	5376200000	5344400000	5549010000	5405970000	2
Endrin aldehyde	3589550000	3619430000	3751370000	3952130000	5791870000	4140870000	23
Endrin ketone	5489410000	5439780000	5529300000	5876580000	5513690000	5569750000	3
gamma-BHC (Lindane)	6730030000	6610730000	6634310000	6589530000	6452200000	6603360000	2
gamma-Chlordane	5958660000	5857990000	5898300000	5915340000	5770280000	5880110000	1
Heptachlor	6694450000	6605950000	6658580000	6695280000	6655490000	6661950000	1
Heptachlor epoxide	5706360000	5635230000	5698940000	5730960000	5785230000	5711340000	1
Methoxychlor	2677460000	2673930000	2731560000	2788840000	2830690000	2740500000	3
Tetrachloro-m-xylene	4740810000	4674260000	4731000000	4815490000	4927290000	4777770000	2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: FIRS02

Lab Code: ACE SDG NO.: Q2815

Instrument ID: ECD_L Date(s) Analyzed: 07/28/2025 07/28/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.21	6.11	6.31	29153500
		2	6.61	6.51	6.71	24445000
		3	7.02	6.92	7.12	110273000
		4	7.11	7.01	7.21	81046400
		5	7.89	7.79	7.99	57873100



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: FIRS02


Lab Code: ACE SDG NO.: Q2815

Instrument ID: ECD_L Date(s) Analyzed: 07/28/2025 07/28/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.07	4.97	5.17	34047600
		2	5.76	5.66	5.86	44078700
		3	6.04	5.94	6.14	44632800
		4	6.67	6.57	6.77	147063000
		5	7.11	7.01	7.21	92006100

Endrin aldehyde #2



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096596.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 16:52
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 07:44:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 07:36:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.828	317.6E6	474.1E6	101.272	100.207
28)	SA Decachlor...	9.015	7.992	228.9E6	410.8E6	98.832	97.271
Target Compounds							
2)	A alpha-BHC	3.983	3.333	487.4E6	730.9E6	104.918	102.405
3)	MA gamma-BHC...	4.311	3.665	457.7E6	673.0E6	104.239	101.443
4)	MA Heptachlor	4.903	4.014	421.5E6	669.4E6	102.079	100.539
5)	MB Aldrin	5.243	4.297	441.8E6	631.1E6	103.507	101.291
6)	B beta-BHC	4.497	3.962	177.2E6	274.0E6	99.519	98.471
7)	B delta-BHC	4.743	4.195	416.2E6	662.1E6	104.604	101.941
8)	B Heptachlo...	5.663	4.799	388.8E6	570.6E6	101.065	100.130
9)	A Endosulfan I	6.044	5.170	366.8E6	535.7E6	103.730	99.895
10)	B gamma-Chl...	5.916	5.051	391.2E6	595.9E6	102.936	101.023
11)	B alpha-Chl...	5.997	5.115	388.4E6	580.6E6	102.147	100.103
12)	B 4,4'-DDE	6.167	5.304	338.2E6	565.3E6	107.460	102.036
13)	MA Dieldrin	6.316	5.435	384.9E6	597.4E6	103.679	101.286
14)	MA Endrin	6.543	5.709	313.0E6	541.5E6	104.797	100.730
15)	B Endosulfa...	6.755	6.001	306.9E6	511.7E6	100.875	99.885
16)	A 4,4'-DDD	6.676	5.857	261.5E6	474.0E6	103.616	100.989
17)	MA 4,4'-DDT	6.989	6.109	298.0E6	521.3E6	103.304	101.844
18)	B Endrin al...	6.883	6.179	212.9E6	359.0E6	100.197	95.686
19)	B Endosulfa...	7.118	6.403	287.5E6	500.8E6	100.826	98.910
20)	A Methoxychlor	7.462	6.681	145.0E6	267.7E6	98.077	98.019
21)	B Endrin ke...	7.597	6.907	302.3E6	548.9E6	100.645	99.278
22)	Mirex	8.076	7.098	237.9E6	412.0E6	97.547	96.837



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096596.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 16:52
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 07:44:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 07:36:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Response_ Signal: PL096596.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 317649612
Conc: 101.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100


Time Response_ Signal: PL096596.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 474080616
Conc: 100.21 ng/ml


Time Response_ Signal: PL096596.D\ECD1A.ch



#2 alpha-BHC

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 487363323
Conc: 104.92 ng/ml


Time Response_ Signal: PL096596.D\ECD2B.ch



#2 alpha-BHC

R.T.: 3.333 min
Delta R.T.: 0.000 min
Response: 730880517
Conc: 102.40 ng/ml

Response_ Signal: PL096596.D\ECD1A.ch




#3 gamma-BHC (Lindane)

R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 457665420
Conc: 104.24 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100


Time Response_ Signal: PL096596.D\ECD2B.ch



#3 gamma-BHC (Lindane)

R.T.: 3.665 min
Delta R.T.: 0.000 min
Response: 673002560
Conc: 101.44 ng/ml


Time Response_ Signal: PL096596.D\ECD1A.ch



#4 Heptachlor

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 421533951
Conc: 102.08 ng/ml


Time Response_ Signal: PL096596.D\ECD2B.ch



#4 Heptachlor

R.T.: 4.014 min
Delta R.T.: 0.000 min
Response: 669445389
Conc: 100.54 ng/ml

Response_ Signal: PL096596.D\ECD1A.ch




#5 Aldrin

R.T.: 5.243 min
Delta R.T.: 0.000 min
Response: 441807820
Conc: 103.51 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100


Time Response_ Signal: PL096596.D\ECD2B.ch



#5 Aldrin

R.T.: 4.297 min
Delta R.T.: 0.000 min
Response: 631053302
Conc: 101.29 ng/ml


Time Response_ Signal: PL096596.D\ECD1A.ch



#6 beta-BHC

R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 177246384
Conc: 99.52 ng/ml

Time Response_ Signal: PL096596.D\ECD2B.ch




#6 beta-BHC

R.T.: 3.962 min
Delta R.T.: 0.000 min
Response: 274017290
Conc: 98.47 ng/ml

Response_ Signal: PL096596.D\ECD1A.ch

#7 delta-BHC




R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 416162516
Conc: 104.60 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Time Response_ Signal: PL096596.D\ECD2B.ch


#7 delta-BHC



R.T.: 4.195 min
Delta R.T.: 0.000 min
Response: 662118416
Conc: 101.94 ng/ml

Time Response_ Signal: PL096596.D\ECD1A.ch


#8 Heptachlor epoxide



R.T.: 5.663 min
Delta R.T.: 0.000 min
Response: 388813747
Conc: 101.06 ng/ml

Time Response_ Signal: PL096596.D\ECD2B.ch

#8 Heptachlor epoxide




R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 570635747
Conc: 100.13 ng/ml

Response_ Signal: PL096596.D\ECD1A.ch

#9 Endosulfan I

R.T.: 6.044 min
Delta R.T.: 0.000 min
Response: 366849911
Conc: 103.73 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDICC100



Time Response_ Signal: PL096596.D\ECD2B.ch

#9 Endosulfan I


R.T.: 5.170 min
Delta R.T.: 0.000 min
Response: 535710057
Conc: 99.90 ng/ml



Time Response_ Signal: PL096596.D\ECD1A.ch

#10 gamma-Chlordane


R.T.: 5.916 min
Delta R.T.: 0.000 min
Response: 391234530
Conc: 102.94 ng/ml




Time Response_ Signal: PL096596.D\ECD2B.ch

#10 gamma-Chlordane

R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 595865751
Conc: 101.02 ng/ml



Response_ Signal: PL096596.D\ECD1A.ch




#11 alpha-Chlordane

R.T.: 5.997 min
Delta R.T.: 0.000 min
Response: 388351267
Conc: 102.15 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100


Time Response_ Signal: PL096596.D\ECD2B.ch



#11 alpha-Chlordane

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 580616706
Conc: 100.10 ng/ml


Time Response_ Signal: PL096596.D\ECD1A.ch



#12 4,4'-DDE


R.T.: 6.167 min
Delta R.T.: 0.000 min
Response: 338202440
Conc: 107.46 ng/ml

Time Response_ Signal: PL096596.D\ECD2B.ch



#12 4,4'-DDE


R.T.: 5.304 min
Delta R.T.: 0.000 min
Response: 565343423
Conc: 102.04 ng/ml



#13 Dieldrin


R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 384947334
Conc: 103.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100




#13 Dieldrin

R.T.: 5.435 min
Delta R.T.: 0.000 min
Response: 597384135
Conc: 101.29 ng/ml



#14 Endrin


R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 312998820
Conc: 104.80 ng/ml



#14 Endrin

R.T.: 5.709 min
Delta R.T.: 0.000 min
Response: 541542527
Conc: 100.73 ng/ml

Response_ Signal: PL096596.D\ECD1A.ch




#15 Endosulfan II

R.T.: 6.755 min
Delta R.T.: 0.000 min
Response: 306942268
Conc: 100.88 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100


Time Response_ Signal: PL096596.D\ECD2B.ch



#15 Endosulfan II

R.T.: 6.001 min
Delta R.T.: 0.000 min
Response: 511675779
Conc: 99.88 ng/ml


Time Response_ Signal: PL096596.D\ECD1A.ch



#16 4,4'-DDD

R.T.: 6.676 min
Delta R.T.: 0.000 min
Response: 261514558
Conc: 103.62 ng/ml

Time Response_ Signal: PL096596.D\ECD2B.ch



#16 4,4'-DDD

R.T.: 5.857 min
Delta R.T.: 0.000 min
Response: 473976803
Conc: 100.99 ng/ml

Response_ Signal: PL096596.D\ECD1A.ch

#17 4,4'-DDT

R.T.: 6.989 min
Delta R.T.: 0.000 min
Response: 298004870
Conc: 103.30 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Time 6.60 6.70 6.80 6.90 7.00 7.10 7.20 7.30 7.40

Response_ Signal: PL096596.D\ECD2B.ch

#17 4,4'-DDT

R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 521289312
Conc: 101.84 ng/ml

Time 5.90 5.95 6.00 6.05 6.10 6.15 6.20 6.25

Response_ Signal: PL096596.D\ECD1A.ch

#18 Endrin aldehyde

R.T.: 6.883 min
Delta R.T.: 0.000 min
Response: 212916204
Conc: 100.20 ng/ml


Time 6.70 6.80 6.90 7.00 7.10

Response_ Signal: PL096596.D\ECD2B.ch

#18 Endrin aldehyde


R.T.: 6.179 min
Delta R.T.: 0.000 min
Response: 358954906
Conc: 95.69 ng/ml

Time 5.90 6.00 6.10 6.20 6.30 6.40




R.T.: 7.118 min
Delta R.T.: 0.000 min
Response: 287518219
Conc: 100.83 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDICC100




R.T.: 6.403 min
Delta R.T.: 0.000 min
Response: 500753542
Conc: 98.91 ng/ml



R.T.: 7.462 min
Delta R.T.: 0.000 min
Response: 144959027
Conc: 98.08 ng/ml



R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 267745710
Conc: 98.02 ng/ml



R.T.: 7.597 min
Delta R.T.: 0.000 min
Response: 302328647
Conc: 100.64 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Time

Response_

Signal: PL096596.D\ECD2B.ch

#21 Endrin ketone




R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 548940584
Conc: 99.28 ng/ml

Time

Response_

Signal: PL096596.D\ECD1A.ch

#22 Mirex




R.T.: 8.076 min
Delta R.T.: 0.000 min
Response: 237884655
Conc: 97.55 ng/ml

Time

Response_

Signal: PL096596.D\ECD2B.ch

#22 Mirex



R.T.: 7.098 min
Delta R.T.: 0.001 min
Response: 411955568
Conc: 96.84 ng/ml


Time

Response_

Signal: PL096596.D\ECD1A.ch

#28 Decachlorobiphenyl

ICAL Form



R.T.: 9.015 min
Delta R.T.: 0.000 min
Response: 228937183
Conc: 98.83 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Time

Response

Signal: PL096596.D\ECD2B.ch

#28 Decachlorobiphenyl



R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 410823200
Conc: 97.27 ng/ml

Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096597.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 17:06
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 07:45:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 07:36:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.828	233.0E6	350.6E6	74.272	74.101
28)	SA Decachlor...	9.016	7.992	168.8E6	307.4E6	72.881	72.785
Target Compounds							
2)	A alpha-BHC	3.983	3.334	351.4E6	537.0E6	75.654	75.236
3)	MA gamma-BHC...	4.311	3.665	330.8E6	495.8E6	75.350	74.733
4)	MA Heptachlor	4.903	4.014	307.2E6	495.4E6	74.396	74.407
5)	MB Aldrin	5.243	4.297	320.5E6	465.2E6	75.090	74.668
6)	B beta-BHC	4.497	3.961	131.3E6	204.8E6	73.714	73.601
7)	B delta-BHC	4.743	4.195	300.9E6	488.2E6	75.637	75.167
8)	B Heptachlo...	5.662	4.799	284.4E6	422.6E6	73.922	74.162
9)	A Endosulfan I	6.044	5.170	264.4E6	398.2E6	74.753	74.246
10)	B gamma-Chl...	5.916	5.051	285.1E6	439.3E6	75.003	74.487
11)	B alpha-Chl...	5.996	5.115	283.5E6	429.9E6	74.575	74.112
12)	B 4,4'-DDE	6.167	5.304	238.5E6	417.1E6	75.774	75.277
13)	MA Dieldrin	6.316	5.435	280.3E6	441.3E6	75.499	74.826
14)	MA Endrin	6.543	5.709	224.3E6	400.9E6	75.092	74.562
15)	B Endosulfa...	6.755	6.001	225.6E6	379.7E6	74.158	74.126
16)	A 4,4'-DDD	6.675	5.857	190.7E6	348.7E6	75.547	74.288
17)	MA 4,4'-DDT	6.990	6.109	217.6E6	383.6E6	75.446	74.937
18)	B Endrin al...	6.884	6.179	157.7E6	271.5E6	74.223	72.362
19)	B Endosulfa...	7.118	6.403	211.1E6	373.3E6	74.033	73.735
20)	A Methoxychlor	7.462	6.681	108.3E6	200.5E6	73.298	73.418
21)	B Endrin ke...	7.597	6.907	223.2E6	408.0E6	74.299	73.786
22)	Mirex	8.076	7.098	177.1E6	309.7E6	72.625	72.793



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096597.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 17:06
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 07:45:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 07:36:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 232961822
Conc: 74.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 350569391
Conc: 74.10 ng/ml




#2 alpha-BHC

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 351424905
Conc: 75.65 ng/ml



#2 alpha-BHC

R.T.: 3.334 min
Delta R.T.: 0.000 min
Response: 536970875
Conc: 75.24 ng/ml



R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 330824316
Conc: 75.35 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Time

Response_

Signal: PL096597.D\ECD2B.ch

#3 gamma-BHC (Lindane)




R.T.: 3.665 min
Delta R.T.: 0.000 min
Response: 495804891
Conc: 74.73 ng/ml

Time

Response_

Signal: PL096597.D\ECD1A.ch

#4 Heptachlor




R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 307217245
Conc: 74.40 ng/ml

Time

Response_


Signal: PL096597.D\ECD2B.ch

#4 Heptachlor



R.T.: 4.014 min
Delta R.T.: 0.000 min
Response: 495446498
Conc: 74.41 ng/ml

Time



R.T.: 5.243 min
Delta R.T.: 0.000 min
Response: 320513876
Conc: 75.09 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Time

Response_

Signal: PL096597.D\ECD2B.ch

#5 Aldrin




R.T.: 4.297 min
Delta R.T.: 0.000 min
Response: 465189508
Conc: 74.67 ng/ml

Time

Response_

Signal: PL096597.D\ECD1A.ch

#6 beta-BHC




R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 131286587
Conc: 73.71 ng/ml

Time

Response_

Signal: PL096597.D\ECD2B.ch


#6 beta-BHC



R.T.: 3.961 min
Delta R.T.: 0.000 min
Response: 204810930
Conc: 73.60 ng/ml

Time

Response_ Signal: PL096597.D\ECD1A.ch




#7 delta-BHC

R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 300917635
Conc: 75.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075


Time Response_ Signal: PL096597.D\ECD2B.ch



#7 delta-BHC

R.T.: 4.195 min
Delta R.T.: 0.000 min
Response: 488213801
Conc: 75.17 ng/ml


Time Response_ Signal: PL096597.D\ECD1A.ch



#8 Heptachlor epoxide


R.T.: 5.662 min
Delta R.T.: 0.000 min
Response: 284391514
Conc: 73.92 ng/ml

Time Response_ Signal: PL096597.D\ECD2B.ch



#8 Heptachlor epoxide

R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 422642097
Conc: 74.16 ng/ml



R.T.: 6.044 min
Delta R.T.: 0.000 min
Response: 264371314
Conc: 74.75 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Time

Response_

Signal: PL096597.D\ECD2B.ch

#9 Endosulfan I




R.T.: 5.170 min
Delta R.T.: 0.000 min
Response: 398162363
Conc: 74.25 ng/ml

Time

Response_

Signal: PL096597.D\ECD1A.ch

#10 gamma-Chlordane




R.T.: 5.916 min
Delta R.T.: 0.000 min
Response: 285066035
Conc: 75.00 ng/ml

Time

Response_

Signal: PL096597.D\ECD2B.ch

#10 gamma-Chlordane



R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 439349234
Conc: 74.49 ng/ml


Time

Response_

Signal: PL096597.D\ECD1A.ch

#11 alpha-Chlordane

ICAL Form



R.T.: 5.996 min
Delta R.T.: 0.000 min
Response: 283525040
Conc: 74.57 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Time

Response_

Signal: PL096597.D\ECD2B.ch

#11 alpha-Chlordane




R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 429864118
Conc: 74.11 ng/ml

Time

Response_

Signal: PL096597.D\ECD1A.ch

#12 4,4'-DDE




R.T.: 6.167 min
Delta R.T.: 0.000 min
Response: 238480667
Conc: 75.77 ng/ml

Time

Response_


Signal: PL096597.D\ECD2B.ch

#12 4,4'-DDE



R.T.: 5.304 min
Delta R.T.: 0.000 min
Response: 417080341
Conc: 75.28 ng/ml


Time



#13 Dieldrin


R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 280319278
Conc: 75.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#13 Dieldrin

R.T.: 5.435 min
Delta R.T.: 0.000 min
Response: 441320179
Conc: 74.83 ng/ml



#14 Endrin


R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 224278091
Conc: 75.09 ng/ml



#14 Endrin

R.T.: 5.709 min
Delta R.T.: 0.000 min
Response: 400859370
Conc: 74.56 ng/ml

Response_ Signal: PL096597.D\ECD1A.ch




#15 Endosulfan II

R.T.: 6.755 min
Delta R.T.: 0.000 min
Response: 225647134
Conc: 74.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075


Time Response_ Signal: PL096597.D\ECD2B.ch



#15 Endosulfan II

R.T.: 6.001 min
Delta R.T.: 0.000 min
Response: 379721248
Conc: 74.13 ng/ml


Time Response_ Signal: PL096597.D\ECD1A.ch



#16 4,4'-DDD


R.T.: 6.675 min
Delta R.T.: 0.000 min
Response: 190671313
Conc: 75.55 ng/ml

Time Response_ Signal: PL096597.D\ECD2B.ch



#16 4,4'-DDD


R.T.: 5.857 min
Delta R.T.: 0.000 min
Response: 348660112
Conc: 74.29 ng/ml



#17 4,4'-DDT


R.T.: 6.990 min
Delta R.T.: 0.000 min
Response: 217641704
Conc: 75.45 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#17 4,4'-DDT

R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 383566750
Conc: 74.94 ng/ml




#18 Endrin aldehyde

R.T.: 6.884 min
Delta R.T.: 0.000 min
Response: 157721830
Conc: 74.22 ng/ml



#18 Endrin aldehyde


R.T.: 6.179 min
Delta R.T.: 0.000 min
Response: 271456969
Conc: 72.36 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.118 min
Delta R.T.: 0.000 min
Response: 211116580
Conc: 74.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075




#19 Endosulfan Sulfate

R.T.: 6.403 min
Delta R.T.: 0.000 min
Response: 373300857
Conc: 73.74 ng/ml




#20 Methoxychlor

R.T.: 7.462 min
Delta R.T.: 0.000 min
Response: 108334576
Conc: 73.30 ng/ml



#20 Methoxychlor

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 200544716
Conc: 73.42 ng/ml




R.T.: 7.597 min
Delta R.T.: 0.000 min
Response: 223186820
Conc: 74.30 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Time Response_ Signal: PL096597.D\ECD2B.ch


#21 Endrin ketone



R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 407983144
Conc: 73.79 ng/ml

Time Response_ Signal: PL096597.D\ECD1A.ch


#22 Mirex




R.T.: 8.076 min
Delta R.T.: 0.000 min
Response: 177107922
Conc: 72.63 ng/ml

Time Response_ Signal: PL096597.D\ECD2B.ch

#22 Mirex




R.T.: 7.098 min
Delta R.T.: 0.000 min
Response: 309671656
Conc: 72.79 ng/ml



R.T.: 9.016 min
Delta R.T.: 0.000 min
Response: 168823757
Conc: 72.88 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC075



#28 Decachlorobiphenyl

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 307406585
Conc: 72.78 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096598.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 17:19
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 07:45:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 07:36:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.535	2.828	156.8E6	236.5E6	50.000	50.000
28) SA Decachlor...	9.016	7.992	115.8E6	211.2E6	50.000	50.000
Target Compounds						
2) A alpha-BHC	3.982	3.333	232.3E6	356.9E6	50.000	50.000
3) MA gamma-BHC...	4.310	3.665	219.5E6	331.7E6	50.000	50.000
4) MA Heptachlor	4.903	4.014	206.5E6	332.9E6	50.000	50.000
5) MB Aldrin	5.243	4.297	213.4E6	311.5E6	50.000	50.000
6) B beta-BHC	4.497	3.961	89051761	139.1E6	50.000	50.000
7) B delta-BHC	4.743	4.195	198.9E6	324.8E6	50.000	50.000
8) B Heptachlo...	5.662	4.799	192.4E6	284.9E6	50.000	50.000
9) A Endosulfan I	6.044	5.169	176.8E6	268.1E6	50.000	50.000
10) B gamma-Chl...	5.916	5.051	190.0E6	294.9E6	50.000	50.000
11) B alpha-Chl...	5.997	5.115	190.1E6	290.0E6	50.000	50.000
12) B 4,4'-DDE	6.166	5.304	157.4E6	277.0E6	50.000	50.000
13) MA Dieldrin	6.316	5.434	185.6E6	294.9E6	50.000	50.000
14) MA Endrin	6.542	5.709	149.3E6	268.8E6	50.000	50.000
15) B Endosulfa...	6.755	6.001	152.1E6	256.1E6	50.000	50.000
16) A 4,4'-DDD	6.675	5.856	126.2E6	234.7E6	50.000	50.000
17) MA 4,4'-DDT	6.989	6.109	144.2E6	255.9E6	50.000	50.000
18) B Endrin al...	6.884	6.179	106.2E6	187.6E6	50.000	50.000
19) B Endosulfa...	7.118	6.402	142.6E6	253.1E6	50.000	50.000
20) A Methoxychlor	7.462	6.681	73900571	136.6E6	50.000	50.000
21) B Endrin ke...	7.597	6.907	150.2E6	276.5E6	50.000	50.000
22) Mirex	8.076	7.097	121.9E6	212.7E6	50.000	50.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096598.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 17:19
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 07:45:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 07:36:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 156830016
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 236549853
Conc: 50.00 ng/ml



#2 alpha-BHC


R.T.: 3.982 min
Delta R.T.: 0.000 min
Response: 232258676
Conc: 50.00 ng/ml



#2 alpha-BHC

R.T.: 3.333 min
Delta R.T.: 0.000 min
Response: 356858706
Conc: 50.00 ng/ml

Response_ Signal: PL096598.D\ECD1A.ch




#3 gamma-BHC (Lindane)

R.T.: 4.310 min
Delta R.T.: 0.000 min
Response: 219526089
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050


Time Response_ Signal: PL096598.D\ECD2B.ch



#3 gamma-BHC (Lindane)

R.T.: 3.665 min
Delta R.T.: 0.000 min
Response: 331715507
Conc: 50.00 ng/ml


Time Response_ Signal: PL096598.D\ECD1A.ch



#4 Heptachlor


R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 206474907
Conc: 50.00 ng/ml

Time Response_ Signal: PL096598.D\ECD2B.ch



#4 Heptachlor


R.T.: 4.014 min
Delta R.T.: 0.000 min
Response: 332929030
Conc: 50.00 ng/ml



#5 Aldrin


R.T.: 5.243 min
Delta R.T.: 0.000 min
Response: 213418935
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#5 Aldrin

R.T.: 4.297 min
Delta R.T.: 0.000 min
Response: 311504423
Conc: 50.00 ng/ml




#6 beta-BHC

R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 89051761
Conc: 50.00 ng/ml



#6 beta-BHC


R.T.: 3.961 min
Delta R.T.: 0.000 min
Response: 139136358
Conc: 50.00 ng/ml



#7 delta-BHC


R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 198923133
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#7 delta-BHC

R.T.: 4.195 min
Delta R.T.: 0.000 min
Response: 324754328
Conc: 50.00 ng/ml




#8 Heptachlor epoxide

R.T.: 5.662 min
Delta R.T.: 0.000 min
Response: 192358965
Conc: 50.00 ng/ml



#8 Heptachlor epoxide


R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 284946767
Conc: 50.00 ng/ml



#9 Endosulfan I


R.T.: 6.044 min
Delta R.T.: 0.000 min
Response: 176829163
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#9 Endosulfan I

R.T.: 5.169 min
Delta R.T.: 0.000 min
Response: 268136431
Conc: 50.00 ng/ml




#10 gamma-Chlordane

R.T.: 5.916 min
Delta R.T.: 0.000 min
Response: 190037398
Conc: 50.00 ng/ml



#10 gamma-Chlordane


R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 294914893
Conc: 50.00 ng/ml



#11 alpha-Chlordane


R.T.: 5.997 min
Delta R.T.: 0.000 min
Response: 190094073
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#11 alpha-Chlordane

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 290010102
Conc: 50.00 ng/ml




#12 4,4'-DDE

R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 157362162
Conc: 50.00 ng/ml



#12 4,4'-DDE


R.T.: 5.304 min
Delta R.T.: 0.000 min
Response: 277030353
Conc: 50.00 ng/ml



#13 Dieldrin


R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 185644019
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#13 Dieldrin

R.T.: 5.434 min
Delta R.T.: 0.000 min
Response: 294898417
Conc: 50.00 ng/ml



#14 Endrin

R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 149335231
Conc: 50.00 ng/ml



#14 Endrin


R.T.: 5.709 min
Delta R.T.: 0.000 min
Response: 268809924
Conc: 50.00 ng/ml

Response_ Signal: PL096598.D\ECD1A.ch

#15 Endosulfan II

R.T.: 6.755 min
Delta R.T.: 0.000 min
Response: 152139869
Conc: 50.00 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDICC050



Time Response_ Signal: PL096598.D\ECD2B.ch

#15 Endosulfan II


R.T.: 6.001 min
Delta R.T.: 0.000 min
Response: 256133247
Conc: 50.00 ng/ml



Time Response_ Signal: PL096598.D\ECD1A.ch

#16 4,4'-DDD


R.T.: 6.675 min
Delta R.T.: 0.000 min
Response: 126194489
Conc: 50.00 ng/ml



Time Response_ Signal: PL096598.D\ECD2B.ch

#16 4,4'-DDD

R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 234668190
Conc: 50.00 ng/ml




Response_ Signal: PL096598.D\ECD1A.ch

#17 4,4'-DDT

R.T.: 6.989 min
Delta R.T.: 0.000 min
Response: 144237122
Conc: 50.00 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDICC050



Time Response_ Signal: PL096598.D\ECD2B.ch

#17 4,4'-DDT


R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 255924551
Conc: 50.00 ng/ml



Time Response_ Signal: PL096598.D\ECD1A.ch

#18 Endrin aldehyde


R.T.: 6.884 min
Delta R.T.: 0.000 min
Response: 106248721
Conc: 50.00 ng/ml




Time Response_ Signal: PL096598.D\ECD2B.ch

#18 Endrin aldehyde

R.T.: 6.179 min
Delta R.T.: 0.000 min
Response: 187568317
Conc: 50.00 ng/ml






#19 Endosulfan Sulfate


R.T.: 7.118 min
Delta R.T.: 0.000 min
Response: 142581938
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#19 Endosulfan Sulfate

R.T.: 6.402 min
Delta R.T.: 0.000 min
Response: 253136738
Conc: 50.00 ng/ml




#20 Methoxychlor

R.T.: 7.462 min
Delta R.T.: 0.000 min
Response: 73900571
Conc: 50.00 ng/ml



#20 Methoxychlor


R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 136578007
Conc: 50.00 ng/ml



#21 Endrin ketone


R.T.: 7.597 min
Delta R.T.: 0.000 min
Response: 150195843
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050




#21 Endrin ketone

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 276465208
Conc: 50.00 ng/ml




#22 Mirex

R.T.: 8.076 min
Delta R.T.: 0.000 min
Response: 121932948
Conc: 50.00 ng/ml



#22 Mirex


R.T.: 7.097 min
Delta R.T.: 0.000 min
Response: 212706537
Conc: 50.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.016 min
Delta R.T.: 0.000 min
Response: 115821567
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC050



#28 Decachlorobiphenyl

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 211174905
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096599.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 17:33
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDICC025

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 07:45:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 07:36:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.828	80667083	120.4E6	25.718	25.447
28)	SA Decachlor...	9.016	7.992	61253778	110.7E6	26.443	26.222
Target Compounds							
2)	A alpha-BHC	3.983	3.333	113.7E6	176.1E6	24.483	24.678
3)	MA gamma-BHC...	4.311	3.665	109.5E6	164.7E6	24.930	24.831
4)	MA Heptachlor	4.903	4.014	103.9E6	167.4E6	25.150	25.138
5)	MB Aldrin	5.242	4.297	106.1E6	154.6E6	24.852	24.810
6)	B beta-BHC	4.497	3.961	46424655	71835591	26.066	25.815
7)	B delta-BHC	4.743	4.195	98875722	161.7E6	24.853	24.903
8)	B Heptachlo...	5.662	4.799	97796537	143.3E6	25.420	25.140
9)	A Endosulfan I	6.044	5.170	89274310	138.5E6	25.243	25.835
10)	B gamma-Chl...	5.916	5.051	95166219	147.9E6	25.039	25.072
11)	B alpha-Chl...	5.996	5.115	96944619	147.4E6	25.499	25.414
12)	B 4,4'-DDE	6.166	5.303	77835138	135.9E6	24.731	24.536
13)	MA Dieldrin	6.316	5.434	91962286	148.1E6	24.768	25.106
14)	MA Endrin	6.542	5.709	76652120	133.6E6	25.664	24.852
15)	B Endosulfa...	6.755	6.001	79391270	129.4E6	26.092	25.256
16)	A 4,4'-DDD	6.675	5.856	62524397	117.6E6	24.773	25.050
17)	MA 4,4'-DDT	6.989	6.109	70933756	126.1E6	24.589	24.645
18)	B Endrin al...	6.884	6.177	54110594	98803191	25.464	26.338m
19)	B Endosulfa...	7.117	6.403	71756969	128.4E6	25.163	25.362
20)	A Methoxychlor	7.462	6.681	37688830	69721101	25.500	25.524
21)	B Endrin ke...	7.597	6.905	75730573	146.9E6	25.211	26.570m
22)	Mirex	8.076	7.097	63693911	111.6E6	26.118	26.231

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096599.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 17:33
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDICC025

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 07:45:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 07:36:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 80667083
Conc: 25.72 ng/ml

Instrument :

ECD_L


ClientSampleId :

PSTDICC025

Manual Integrations
APPROVED


Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 120387362
Conc: 25.45 ng/ml



#2 alpha-BHC


R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 113727400
Conc: 24.48 ng/ml



#2 alpha-BHC

R.T.: 3.333 min
Delta R.T.: 0.000 min
Response: 176134463
Conc: 24.68 ng/ml

Response_ Signal: PL096599.D\ECD1A.ch



#3 gamma-BHC (Lindane)


R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 109455262
Conc: 24.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025


Time Response_ Signal: PL096599.D\ECD2B.ch



#3 gamma-BHC (Lindane)

R.T.: 3.665 min
Delta R.T.: 0.000 min
Response: 164738265
Conc: 24.83 ng/ml


Time Response_ Signal: PL096599.D\ECD1A.ch



#4 Heptachlor


R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 103855678
Conc: 25.15 ng/ml

Time Response_ Signal: PL096599.D\ECD2B.ch



#4 Heptachlor

R.T.: 4.014 min
Delta R.T.: 0.000 min
Response: 167381908
Conc: 25.14 ng/ml




#5 Aldrin

R.T.: 5.242 min
Delta R.T.: 0.000 min
Response: 106077325
Conc: 24.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#5 Aldrin

R.T.: 4.297 min
Delta R.T.: 0.000 min
Response: 154569484
Conc: 24.81 ng/ml



#6 beta-BHC

R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 46424655
Conc: 26.07 ng/ml




#6 beta-BHC

R.T.: 3.961 min
Delta R.T.: 0.000 min
Response: 71835591
Conc: 25.81 ng/ml

Response_ Signal: PL096599.D\ECD1A.ch

#7 delta-BHC



R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 98875722
Conc: 24.85 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDICC025

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025

Time Response_ Signal: PL096599.D\ECD2B.ch


#7 delta-BHC



R.T.: 4.195 min
Delta R.T.: 0.000 min
Response: 161747114
Conc: 24.90 ng/ml

Time Response_ Signal: PL096599.D\ECD1A.ch


#8 Heptachlor epoxide




R.T.: 5.662 min
Delta R.T.: 0.000 min
Response: 97796537
Conc: 25.42 ng/ml

Time Response_ Signal: PL096599.D\ECD2B.ch

#8 Heptachlor epoxide



R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 143274077
Conc: 25.14 ng/ml




#9 Endosulfan I

R.T.: 6.044 min
Delta R.T.: 0.000 min
Response: 89274310
Conc: 25.24 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#9 Endosulfan I

R.T.: 5.170 min
Delta R.T.: 0.000 min
Response: 138547710
Conc: 25.84 ng/ml




#10 gamma-Chlordane

R.T.: 5.916 min
Delta R.T.: 0.000 min
Response: 95166219
Conc: 25.04 ng/ml



#10 gamma-Chlordane

R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 147883477
Conc: 25.07 ng/ml




#11 alpha-Chlordane

R.T.: 5.996 min
Delta R.T.: 0.000 min
Response: 96944619
Conc: 25.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#11 alpha-Chlordane

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 147407858
Conc: 25.41 ng/ml




#12 4,4'-DDE

R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 77835138
Conc: 24.73 ng/ml



#12 4,4'-DDE

R.T.: 5.303 min
Delta R.T.: 0.000 min
Response: 135946511
Conc: 24.54 ng/ml




#13 Dieldrin

R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 91962286
Conc: 24.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#13 Dieldrin

R.T.: 5.434 min
Delta R.T.: 0.000 min
Response: 148077198
Conc: 25.11 ng/ml




#14 Endrin

R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 76652120
Conc: 25.66 ng/ml



#14 Endrin

R.T.: 5.709 min
Delta R.T.: 0.000 min
Response: 133610110
Conc: 24.85 ng/ml




#15 Endosulfan II

R.T.: 6.755 min
Delta R.T.: 0.000 min
Response: 79391270
Conc: 26.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#15 Endosulfan II

R.T.: 6.001 min
Delta R.T.: 0.000 min
Response: 129380147
Conc: 25.26 ng/ml




#16 4,4'-DDD

R.T.: 6.675 min
Delta R.T.: 0.000 min
Response: 62524397
Conc: 24.77 ng/ml



#16 4,4'-DDD

R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 117568590
Conc: 25.05 ng/ml




#17 4,4'-DDT

R.T.: 6.989 min
Delta R.T.: 0.000 min
Response: 70933756
Conc: 24.59 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#17 4,4'-DDT

R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 126146252
Conc: 24.65 ng/ml




#18 Endrin aldehyde

R.T.: 6.884 min
Delta R.T.: 0.000 min
Response: 54110594
Conc: 25.46 ng/ml



#18 Endrin aldehyde

R.T.: 6.177 min
Delta R.T.: -0.001 min
Response: 98803191
Conc: 26.34 ng/ml m




#19 Endosulfan Sulfate

R.T.: 7.117 min
Delta R.T.: 0.000 min
Response: 71756969
Conc: 25.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#19 Endosulfan Sulfate

R.T.: 6.403 min
Delta R.T.: 0.000 min
Response: 128398736
Conc: 25.36 ng/ml




#20 Methoxychlor

R.T.: 7.462 min
Delta R.T.: 0.000 min
Response: 37688830
Conc: 25.50 ng/ml



#20 Methoxychlor

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 69721101
Conc: 25.52 ng/ml




#21 Endrin ketone

R.T.: 7.597 min
Delta R.T.: 0.000 min
Response: 75730573
Conc: 25.21 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#21 Endrin ketone

R.T.: 6.905 min
Delta R.T.: -0.001 min
Response: 146914534
Conc: 26.57 ng/ml m




#22 Mirex

R.T.: 8.076 min
Delta R.T.: 0.000 min
Response: 63693911
Conc: 26.12 ng/ml



#22 Mirex

R.T.: 7.097 min
Delta R.T.: 0.000 min
Response: 111590978
Conc: 26.23 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.016 min
Delta R.T.: 0.000 min
Response: 61253778
Conc: 26.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025



#28 Decachlorobiphenyl

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 110748988
Conc: 26.22 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096600.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 17:47
 Operator : AR\AJ
 Sample : PSTDICC005
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDICC005

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 07:45:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 07:36:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.534	2.827	16289913	24636469	5.193	5.207
28) SA Decachlor...	9.016	7.992	13081800	24155779	5.647	5.719
Target Compounds						
2) A alpha-BHC	3.981	3.333	21891026	33671129	4.713	4.718
3) MA gamma-BHC...	4.310	3.664	21804234	32260997	4.966	4.863
4) MA Heptachlor	4.901	4.013	20915741	33277461	5.065m	4.998
5) MB Aldrin	5.240	4.296	21509762	30546909	5.039m	4.903
6) B beta-BHC	4.497	3.961	9319336	14886777	5.233	5.350
7) B delta-BHC	4.742	4.194	19209773	31362311	4.828	4.829
8) B Heptachlo...	5.660	4.798	19218202	28926139	4.995m	5.076
9) A Endosulfan I	6.043	5.169	18404240	30432392	5.204	5.675
10) B gamma-Chl...	5.915	5.050	18527805	28851412	4.875	4.891
11) B alpha-Chl...	5.996	5.115	19648624	31301796	5.168	5.397
12) B 4,4'-DDE	6.165	5.303	16169148	26895387	5.138	4.854
13) MA Dieldrin	6.315	5.434	17886564	29267822	4.817	4.962
14) MA Endrin	6.541	5.709	14803966	27745055	4.957	5.161
15) B Endosulfa...	6.753	6.000	18583621	26018008	6.107m	5.079
16) A 4,4'-DDD	6.674	5.856	12297723	23708739	4.873	5.052
17) MA 4,4'-DDT	6.989	6.108	13672011	24000169	4.739	4.689
18) B Endrin al...	6.884	6.177	11040304	28959356	5.195	7.720m#
19) B Endosulfa...	7.117	6.402	14782385	26237528	5.184	5.182
20) A Methoxychlor	7.462	6.681	7310793	14153462	4.946	5.181
21) B Endrin ke...	7.597	6.907	15006830	27568439	4.996	4.986
22) Mirex	8.076	7.097	13350221	24190284	5.474	5.686

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096600.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 17:47
Operator : AR\AJ
Sample : PSTDICC005
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDICC005

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 07/29/2025


Supervised By :mohammad ahmed 07/30/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 07:45:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 07:36:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Response_ Signal: PL096600.D\ECD1A.ch



#1 Tetrachloro-m-xylene

R.T.: 3.534 min
Delta R.T.: 0.000 min
Response: 16289913
Conc: 5.19 ng/ml

Instrument :

ECD_L

ClientSampleId :

PSTDICC005


Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025


Time Response_ Signal: PL096600.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.827 min
Delta R.T.: 0.000 min
Response: 24636469
Conc: 5.21 ng/ml


Time Response_ Signal: PL096600.D\ECD1A.ch



#2 alpha-BHC


R.T.: 3.981 min
Delta R.T.: 0.000 min
Response: 21891026
Conc: 4.71 ng/ml

Time Response_ Signal: PL096600.D\ECD2B.ch



#2 alpha-BHC

R.T.: 3.333 min
Delta R.T.: 0.000 min
Response: 33671129
Conc: 4.72 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.310 min
Delta R.T.: 0.000 min
Response: 21804234
Conc: 4.97 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#3 gamma-BHC (Lindane)

R.T.: 3.664 min
Delta R.T.: 0.000 min
Response: 32260997
Conc: 4.86 ng/ml




#4 Heptachlor

R.T.: 4.901 min
Delta R.T.: -0.002 min
Response: 20915741
Conc: 5.06 ng/ml m



#4 Heptachlor

R.T.: 4.013 min
Delta R.T.: 0.000 min
Response: 33277461
Conc: 5.00 ng/ml




#5 Aldrin

R.T.: 5.240 min
Delta R.T.: -0.002 min
Response: 21509762
Conc: 5.04 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#5 Aldrin

R.T.: 4.296 min
Delta R.T.: 0.000 min
Response: 30546909
Conc: 4.90 ng/ml




#6 beta-BHC

R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 9319336
Conc: 5.23 ng/ml



#6 beta-BHC

R.T.: 3.961 min
Delta R.T.: 0.000 min
Response: 14886777
Conc: 5.35 ng/ml




#7 delta-BHC

R.T.: 4.742 min
Delta R.T.: 0.000 min
Response: 19209773
Conc: 4.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#7 delta-BHC

R.T.: 4.194 min
Delta R.T.: 0.000 min
Response: 31362311
Conc: 4.83 ng/ml




#8 Heptachlor epoxide

R.T.: 5.660 min
Delta R.T.: -0.002 min
Response: 19218202
Conc: 5.00 ng/ml m



#8 Heptachlor epoxide

R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 28926139
Conc: 5.08 ng/ml




#9 Endosulfan I

R.T.: 6.043 min
Delta R.T.: 0.000 min
Response: 18404240
Conc: 5.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#9 Endosulfan I

R.T.: 5.169 min
Delta R.T.: 0.000 min
Response: 30432392
Conc: 5.67 ng/ml




#10 gamma-Chlordane

R.T.: 5.915 min
Delta R.T.: 0.000 min
Response: 18527805
Conc: 4.87 ng/ml



#10 gamma-Chlordane

R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 28851412
Conc: 4.89 ng/ml




#11 alpha-Chlordane

R.T.: 5.996 min
Delta R.T.: 0.000 min
Response: 19648624
Conc: 5.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#11 alpha-Chlordane

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 31301796
Conc: 5.40 ng/ml




#12 4,4'-DDE

R.T.: 6.165 min
Delta R.T.: 0.000 min
Response: 16169148
Conc: 5.14 ng/ml



#12 4,4'-DDE

R.T.: 5.303 min
Delta R.T.: 0.000 min
Response: 26895387
Conc: 4.85 ng/ml




#13 Dieldrin

R.T.: 6.315 min
Delta R.T.: 0.000 min
Response: 17886564
Conc: 4.82 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#13 Dieldrin

R.T.: 5.434 min
Delta R.T.: 0.000 min
Response: 29267822
Conc: 4.96 ng/ml




#14 Endrin

R.T.: 6.541 min
Delta R.T.: -0.001 min
Response: 14803966
Conc: 4.96 ng/ml



#14 Endrin

R.T.: 5.709 min
Delta R.T.: 0.000 min
Response: 27745055
Conc: 5.16 ng/ml




#15 Endosulfan II

R.T.: 6.753 min
Delta R.T.: -0.002 min
Response: 18583621
Conc: 6.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#15 Endosulfan II

R.T.: 6.000 min
Delta R.T.: 0.000 min
Response: 26018008
Conc: 5.08 ng/ml




#16 4,4'-DDD

R.T.: 6.674 min
Delta R.T.: 0.000 min
Response: 12297723
Conc: 4.87 ng/ml



#16 4,4'-DDD

R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 23708739
Conc: 5.05 ng/ml




#17 4,4'-DDT

R.T.: 6.989 min
Delta R.T.: 0.000 min
Response: 13672011
Conc: 4.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#17 4,4'-DDT

R.T.: 6.108 min
Delta R.T.: 0.000 min
Response: 24000169
Conc: 4.69 ng/ml




#18 Endrin aldehyde

R.T.: 6.884 min
Delta R.T.: 0.000 min
Response: 11040304
Conc: 5.20 ng/ml



#18 Endrin aldehyde

R.T.: 6.177 min
Delta R.T.: -0.002 min
Response: 28959356
Conc: 7.72 ng/ml m



#19 Endosulfan Sulfate

R.T.: 7.117 min
Delta R.T.: 0.000 min
Response: 14782385
Conc: 5.18 ng/ml

Instrument :

ECD_L

ClientSampleId :


PSTDICC005

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025




#19 Endosulfan Sulfate

R.T.: 6.402 min
Delta R.T.: 0.000 min
Response: 26237528
Conc: 5.18 ng/ml




#20 Methoxychlor

R.T.: 7.462 min
Delta R.T.: 0.000 min
Response: 7310793
Conc: 4.95 ng/ml



#20 Methoxychlor

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 14153462
Conc: 5.18 ng/ml



#21 Endrin ketone

R.T.: 7.597 min
Delta R.T.: 0.000 min
Response: 15006830
Conc: 5.00 ng/ml

Instrument :

ECD_L

ClientSampleId :


PSTDICC005

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025




#21 Endrin ketone

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 27568439
Conc: 4.99 ng/ml




#22 Mirex

R.T.: 8.076 min
Delta R.T.: 0.000 min
Response: 13350221
Conc: 5.47 ng/ml



#22 Mirex

R.T.: 7.097 min
Delta R.T.: 0.000 min
Response: 24190284
Conc: 5.69 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.016 min
Delta R.T.: 0.000 min
Response: 13081800
Conc: 5.65 ng/ml

Instrument :

ECD_L

ClientSampleId :


PSTDICC005

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025



#28 Decachlorobiphenyl

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 24155779
Conc: 5.72 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096603.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 18:27
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 08:11:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 08:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.827	154.7E6	305.4E6	50.000	50.000
28)	SA Decachlor...	9.015	7.991	113.6E6	210.4E6	50.000	50.000
Target Compounds							
23)	Chlordane-1	4.689	3.838	80999277	100.1E6	500.000	500.000
24)	Chlordane-2	5.213	4.418	83409909	115.0E6	500.000	500.000
25)	Chlordane-3	5.917	5.052	326.5E6	336.1E6	500.000	500.000
26)	Chlordane-4	6.001	5.115	403.5E6	300.6E6	500.000	500.000
27)	Chlordane-5	6.838	6.009	63996555	122.4E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096603.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 18:27
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 08:11:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 08:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Response_ Signal: PL096603.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 154680305
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500


Time Response_ Signal: PL096603.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.827 min
Delta R.T.: 0.000 min
Response: 305382843
Conc: 50.00 ng/ml


Time Response_ Signal: PL096603.D\ECD1A.ch



#23 Chlordane-1


R.T.: 4.689 min
Delta R.T.: 0.000 min
Response: 80999277
Conc: 500.00 ng/ml

Time Response_ Signal: PL096603.D\ECD2B.ch



#23 Chlordane-1

R.T.: 3.838 min
Delta R.T.: 0.000 min
Response: 100066760
Conc: 500.00 ng/ml



R.T.: 5.213 min
Delta R.T.: 0.000 min
Response: 83409909
Conc: 500.00 ng/ml


Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Time

Response_

Signal: PL096603.D\ECD2B.ch

#24 Chlordane-2




R.T.: 4.418 min
Delta R.T.: 0.000 min
Response: 115003204
Conc: 500.00 ng/ml

Time

Response_

Signal: PL096603.D\ECD1A.ch

#25 Chlordane-3




R.T.: 5.917 min
Delta R.T.: 0.000 min
Response: 326513783
Conc: 500.00 ng/ml

Time

Response_


Signal: PL096603.D\ECD2B.ch

#25 Chlordane-3



R.T.: 5.052 min
Delta R.T.: 0.000 min
Response: 336093734
Conc: 500.00 ng/ml

Response_ Signal: PL096603.D\ECD1A.ch




#26 Chlordane-4

R.T.: 6.001 min
Delta R.T.: 0.000 min
Response: 403531433
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500


Time Response_ Signal: PL096603.D\ECD2B.ch



#26 Chlordane-4

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 300621016
Conc: 500.00 ng/ml


Time Response_ Signal: PL096603.D\ECD1A.ch



#27 Chlordane-5


R.T.: 6.838 min
Delta R.T.: 0.000 min
Response: 63996555
Conc: 500.00 ng/ml

Time Response_ Signal: PL096603.D\ECD2B.ch



#27 Chlordane-5


R.T.: 6.009 min
Delta R.T.: 0.000 min
Response: 122398851
Conc: 500.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.015 min
Delta R.T.: 0.000 min
Response: 113631254
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500



#28 Decachlorobiphenyl

R.T.: 7.991 min
Delta R.T.: 0.000 min
Response: 210412831
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096608.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 19:36
 Operator : AR\AJ
 Sample : PTOXICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PTOXICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 08:35:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 08:34:10 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.535	2.828	165.3E6	248.4E6	50.000	50.000
7) SA Decachlor...	9.015	7.992	122.2E6	228.1E6	50.000	50.000
Target Compounds						
2) Toxaphene-1	6.211	5.073	14576752	17023794	500.000	500.000
3) Toxaphene-2	6.610	5.758	12222511	22039332	500.000	500.000
4) Toxaphene-3	7.024	6.039	55136340	22316397	500.000	500.000
5) Toxaphene-4	7.114	6.673	40523177	73531411	500.000	500.000
6) Toxaphene-5	7.892	7.114	28936559	46003063	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096608.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 19:36
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PTOXICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 08:35:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 08:34:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 165314687
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PTOXICC500




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 248396246
Conc: 50.00 ng/ml




#2 Toxaphene-1

R.T.: 6.211 min
Delta R.T.: 0.000 min
Response: 14576752
Conc: 500.00 ng/ml



#2 Toxaphene-1


R.T.: 5.073 min
Delta R.T.: 0.000 min
Response: 17023794
Conc: 500.00 ng/ml



#3 Toxaphene-2


R.T.: 6.610 min
Delta R.T.: 0.000 min
Response: 12222511
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PTOXICC500




#3 Toxaphene-2

R.T.: 5.758 min
Delta R.T.: 0.000 min
Response: 22039332
Conc: 500.00 ng/ml




#4 Toxaphene-3

R.T.: 7.024 min
Delta R.T.: 0.000 min
Response: 55136340
Conc: 500.00 ng/ml



#4 Toxaphene-3


R.T.: 6.039 min
Delta R.T.: 0.000 min
Response: 22316397
Conc: 500.00 ng/ml



#5 Toxaphene-4


R.T.: 7.114 min
Delta R.T.: 0.000 min
Response: 40523177
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PTOXICC500




#5 Toxaphene-4

R.T.: 6.673 min
Delta R.T.: 0.000 min
Response: 73531411
Conc: 500.00 ng/ml




#6 Toxaphene-5

R.T.: 7.892 min
Delta R.T.: 0.000 min
Response: 28936559
Conc: 500.00 ng/ml



#6 Toxaphene-5


R.T.: 7.114 min
Delta R.T.: 0.000 min
Response: 46003063
Conc: 500.00 ng/ml



#7 Decachlorobiphenyl

R.T.: 9.015 min
Delta R.T.: 0.000 min
Response: 122190524
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PTOXICC500



#7 Decachlorobiphenyl

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 228076815
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096611.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 20:17
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :

ECD_L

Client SampleId :

ICVPL072825

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 08:47:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 08:20:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.828	161.0E6	247.3E6	50.630	51.756
28)	SA Decachlor...	9.015	7.992	117.1E6	221.9E6	49.100	51.159
Target Compounds							
2)	A alpha-BHC	3.983	3.334	237.2E6	373.0E6	51.282	52.705
3)	MA gamma-BHC...	4.311	3.666	223.6E6	346.0E6	50.545	52.400
4)	MA Heptachlor	4.903	4.014	211.3E6	346.4E6	50.843	51.992
5)	MB Aldrin	5.243	4.297	215.6E6	324.1E6	50.122	52.220
6)	B beta-BHC	4.497	3.961	91859349	145.4E6	50.893	51.555
7)	B delta-BHC	4.742	4.195	202.5E6	338.7E6	50.758m	52.325
8)	B Heptachlo...	5.662	4.799	194.2E6	295.6E6	50.346	51.754
9)	A Endosulfan I	6.044	5.170	178.4E6	277.2E6	49.613	50.116
10)	B gamma-Chl...	5.916	5.051	193.4E6	305.7E6	50.837	51.997
11)	B alpha-Chl...	5.996	5.115	191.7E6	300.1E6	49.735	50.869
12)	B 4,4'-DDE	6.166	5.304	158.1E6	286.3E6	49.223	51.924
13)	MA Dieldrin	6.316	5.435	186.8E6	306.3E6	50.341	51.858
14)	MA Endrin	6.543	5.709	147.3E6	272.8E6	48.674	50.462
15)	B Endosulfa...	6.755	6.001	158.8E6	265.2E6	49.582	51.632
16)	A 4,4'-DDD	6.675	5.856	126.9E6	242.0E6	50.209	51.423
17)	MA 4,4'-DDT	6.989	6.109	144.5E6	261.7E6	50.400	51.730
18)	B Endrin al...	6.883	6.179	108.5E6	197.4E6	50.570	53.386
19)	B Endosulfa...	7.117	6.402	143.9E6	262.8E6	50.068	51.670
20)	A Methoxychlor	7.462	6.681	74330342	140.5E6	50.622	51.276
21)	B Endrin ke...	7.597	6.907	153.3E6	289.7E6	50.977	52.006
22)	Mirex	8.076	7.097	123.5E6	221.2E6	49.793	50.730

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096611.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 20:17
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

ICVPL072825

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 08:47:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 08:20:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 161043183
Conc: 50.63 ng/ml


Instrument :

ECD_L

ClientSampleId :
ICVPL072825


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 247278851
Conc: 51.76 ng/ml




#2 alpha-BHC

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 237249143
Conc: 51.28 ng/ml



#2 alpha-BHC

R.T.: 3.334 min
Delta R.T.: 0.000 min
Response: 372995453
Conc: 52.71 ng/ml



R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 223584100
Conc: 50.55 ng/ml

Instrument :

ECD_L

ClientSampleId :


ICVPL072825

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025

Time Response_ Signal: PL096611.D\ECD2B.ch


#3 gamma-BHC (Lindane)



R.T.: 3.666 min
Delta R.T.: 0.000 min
Response: 346014885
Conc: 52.40 ng/ml

Time Response_ Signal: PL096611.D\ECD1A.ch


#4 Heptachlor




R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 211285693
Conc: 50.84 ng/ml

Time Response_ Signal: PL096611.D\ECD2B.ch

#4 Heptachlor



R.T.: 4.014 min
Delta R.T.: 0.000 min
Response: 346369326
Conc: 51.99 ng/ml




#5 Aldrin

R.T.: 5.243 min
Delta R.T.: 0.000 min
Response: 215573068
Conc: 50.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#5 Aldrin

R.T.: 4.297 min
Delta R.T.: 0.000 min
Response: 324132499
Conc: 52.22 ng/ml




#6 beta-BHC

R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 91859349
Conc: 50.89 ng/ml



#6 beta-BHC

R.T.: 3.961 min
Delta R.T.: 0.000 min
Response: 145431130
Conc: 51.55 ng/ml




#7 delta-BHC

R.T.: 4.742 min
Delta R.T.: 0.000 min
Response: 202516027
Conc: 50.76 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#7 delta-BHC

R.T.: 4.195 min
Delta R.T.: 0.000 min
Response: 338731728
Conc: 52.32 ng/ml




#8 Heptachlor epoxide

R.T.: 5.662 min
Delta R.T.: 0.000 min
Response: 194162959
Conc: 50.35 ng/ml



#8 Heptachlor epoxide

R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 295584103
Conc: 51.75 ng/ml




#9 Endosulfan I

R.T.: 6.044 min
Delta R.T.: 0.000 min
Response: 178427332
Conc: 49.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#9 Endosulfan I

R.T.: 5.170 min
Delta R.T.: 0.000 min
Response: 277213283
Conc: 50.12 ng/ml




#10 gamma-Chlordane

R.T.: 5.916 min
Delta R.T.: 0.000 min
Response: 193445906
Conc: 50.84 ng/ml



#10 gamma-Chlordane

R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 305748780
Conc: 52.00 ng/ml



#11 alpha-Chlordane

R.T.: 5.996 min
Delta R.T.: 0.000 min
Response: 191709507
Conc: 49.73 ng/ml

Instrument :


ECD_L

ClientSampleId :

ICVPL072825


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#11 alpha-Chlordane

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 300071767
Conc: 50.87 ng/ml




#12 4,4'-DDE

R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 158068181
Conc: 49.22 ng/ml



#12 4,4'-DDE

R.T.: 5.304 min
Delta R.T.: 0.000 min
Response: 286332408
Conc: 51.92 ng/ml




#13 Dieldrin

R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 186821697
Conc: 50.34 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#13 Dieldrin

R.T.: 5.435 min
Delta R.T.: 0.000 min
Response: 306303510
Conc: 51.86 ng/ml




#14 Endrin

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 147327440
Conc: 48.67 ng/ml



#14 Endrin

R.T.: 5.709 min
Delta R.T.: 0.000 min
Response: 272795312
Conc: 50.46 ng/ml



#15 Endosulfan II

R.T.: 6.755 min
Delta R.T.: 0.000 min
Response: 158792104
Conc: 49.58 ng/ml

Instrument :

ECD_L

ClientSampleId :


ICVPL072825

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025




#15 Endosulfan II

R.T.: 6.001 min
Delta R.T.: 0.000 min
Response: 265195055
Conc: 51.63 ng/ml




#16 4,4'-DDD

R.T.: 6.675 min
Delta R.T.: 0.000 min
Response: 126947170
Conc: 50.21 ng/ml



#16 4,4'-DDD

R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 241958902
Conc: 51.42 ng/ml




#17 4,4'-DDT

R.T.: 6.989 min
Delta R.T.: 0.000 min
Response: 144531756
Conc: 50.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#17 4,4'-DDT

R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 261665047
Conc: 51.73 ng/ml




#18 Endrin aldehyde

R.T.: 6.883 min
Delta R.T.: 0.000 min
Response: 108518783
Conc: 50.57 ng/ml



#18 Endrin aldehyde

R.T.: 6.179 min
Delta R.T.: 0.000 min
Response: 197425966
Conc: 53.39 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.117 min
Delta R.T.: 0.000 min
Response: 143881166
Conc: 50.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#19 Endosulfan Sulfate

R.T.: 6.402 min
Delta R.T.: 0.000 min
Response: 262804486
Conc: 51.67 ng/ml



#20 Methoxychlor


R.T.: 7.462 min
Delta R.T.: 0.000 min
Response: 74330342
Conc: 50.62 ng/ml



#20 Methoxychlor

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 140523051
Conc: 51.28 ng/ml

Response_ Signal: PL096611.D\ECD1A.ch



#21 Endrin ketone


R.T.: 7.597 min
Delta R.T.: 0.000 min
Response: 153274359
Conc: 50.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025


Time Response_ Signal: PL096611.D\ECD2B.ch



#21 Endrin ketone

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 289661773
Conc: 52.01 ng/ml


Response_ Signal: PL096611.D\ECD1A.ch



#22 Mirex


R.T.: 8.076 min
Delta R.T.: 0.000 min
Response: 123454153
Conc: 49.79 ng/ml

Time Response_ Signal: PL096611.D\ECD2B.ch



#22 Mirex

R.T.: 7.097 min
Delta R.T.: 0.000 min
Response: 221225472
Conc: 50.73 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.015 min
Delta R.T.: -0.001 min
Response: 117087022
Conc: 49.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025



#28 Decachlorobiphenyl

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 221944303
Conc: 51.16 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/15/2025

Initial Calibration Date(s): 07/28/2025

07/28/2025

Continuing Calib Time: 10:58

Initial Calibration Time(s): 16:52

17:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.01	9.02	8.92	9.12	0.01
Tetrachloro-m-xylene	3.53	3.54	3.44	3.64	0.01
alpha-BHC	3.98	3.98	3.88	4.08	0.00
beta-BHC	4.49	4.50	4.40	4.60	0.01
delta-BHC	4.74	4.74	4.64	4.84	0.00
gamma-BHC (Lindane)	4.30	4.31	4.21	4.41	0.01
Heptachlor	4.90	4.90	4.80	5.00	0.00
Aldrin	5.24	5.24	5.14	5.34	0.00
Heptachlor epoxide	5.66	5.66	5.56	5.76	0.00
Endosulfan I	6.04	6.04	5.94	6.14	0.00
Dieldrin	6.31	6.32	6.22	6.42	0.01
4,4'-DDE	6.16	6.17	6.07	6.27	0.01
Endrin	6.54	6.54	6.44	6.64	0.00
Endosulfan II	6.75	6.76	6.66	6.86	0.01
4,4'-DDD	6.67	6.68	6.58	6.78	0.01
Endosulfan sulfate	7.11	7.12	7.02	7.22	0.01
4,4'-DDT	6.99	6.99	6.89	7.09	0.00
Methoxychlor	7.46	7.46	7.36	7.56	0.00
Endrin ketone	7.59	7.60	7.50	7.70	0.01
Endrin aldehyde	6.88	6.88	6.78	6.98	0.00
alpha-Chlordane	5.99	6.00	5.90	6.10	0.01
gamma-Chlordane	5.91	5.92	5.82	6.02	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/15/2025

Initial Calibration Date(s): 07/28/2025

07/28/2025

Continuing Calib Time: 10:58

Initial Calibration Time(s): 16:52

17:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.99	7.99	7.89	8.09	0.00
Tetrachloro-m-xylene	2.82	2.83	2.73	2.93	0.01
alpha-BHC	3.33	3.33	3.23	3.43	0.00
beta-BHC	3.96	3.96	3.86	4.06	0.00
delta-BHC	4.19	4.20	4.10	4.30	0.01
gamma-BHC (Lindane)	3.66	3.67	3.57	3.77	0.01
Heptachlor	4.01	4.01	3.91	4.11	0.00
Aldrin	4.29	4.30	4.20	4.40	0.01
Heptachlor epoxide	4.79	4.80	4.70	4.90	0.01
Endosulfan I	5.17	5.17	5.07	5.27	0.01
Dieldrin	5.43	5.43	5.33	5.53	0.00
4,4'-DDE	5.30	5.30	5.20	5.40	0.00
Endrin	5.70	5.71	5.61	5.81	0.01
Endosulfan II	6.00	6.00	5.90	6.10	0.00
4,4'-DDD	5.85	5.86	5.76	5.96	0.01
Endosulfan sulfate	6.40	6.40	6.30	6.50	0.00
4,4'-DDT	6.10	6.11	6.01	6.21	0.01
Methoxychlor	6.68	6.68	6.58	6.78	0.00
Endrin ketone	6.90	6.91	6.81	7.01	0.01
Endrin aldehyde	6.17	6.18	6.08	6.28	0.01
alpha-Chlordane	5.11	5.12	5.02	5.22	0.01
gamma-Chlordane	5.05	5.05	4.95	5.15	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096819.D **Time Analyzed:** 10:58

COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
4,4'-DDD	6.669	6.575	6.775	53.850	50.000	7.7
4,4'-DDE	6.162	6.066	6.266	57.100	50.000	14.2
4,4'-DDT	6.985	6.889	7.089	55.260	50.000	10.5
Aldrin	5.237	5.143	5.343	55.460	50.000	10.9
alpha-BHC	3.976	3.882	4.082	56.620	50.000	13.2
alpha-Chlordane	5.992	5.897	6.097	55.430	50.000	10.9
beta-BHC	4.491	4.397	4.597	54.870	50.000	9.7
Decachlorobiphenyl	9.009	8.916	9.116	53.580	50.000	7.2
delta-BHC	4.738	4.643	4.843	56.320	50.000	12.6
Dieldrin	6.311	6.216	6.416	54.300	50.000	8.6
Endosulfan I	6.039	5.944	6.144	55.090	50.000	10.2
Endosulfan II	6.749	6.655	6.855	52.840	50.000	5.7
Endosulfan sulfate	7.113	7.018	7.218	52.570	50.000	5.1
Endrin	6.536	6.442	6.642	55.480	50.000	11.0
Endrin aldehyde	6.878	6.784	6.984	53.660	50.000	7.3
Endrin ketone	7.592	7.497	7.697	53.120	50.000	6.2
gamma-BHC (Lindane)	4.304	4.210	4.410	53.000	50.000	6.0
gamma-Chlordane	5.911	5.816	6.016	56.310	50.000	12.6
Heptachlor	4.897	4.803	5.003	59.840	50.000	19.7
Heptachlor epoxide	5.657	5.562	5.762	53.470	50.000	6.9
Methoxychlor	7.457	7.362	7.562	54.900	50.000	9.8
Tetrachloro-m-xylene	3.530	3.435	3.635	54.570	50.000	9.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096819.D **Time Analyzed:** 10:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.852	5.756	5.956	51.050	50.000	2.1
4,4'-DDE	5.299	5.204	5.404	51.950	50.000	3.9
4,4'-DDT	6.103	6.009	6.209	52.520	50.000	5.0
Aldrin	4.291	4.197	4.397	51.080	50.000	2.2
alpha-BHC	3.328	3.233	3.433	53.290	50.000	6.6
alpha-Chlordane	5.110	5.015	5.215	51.520	50.000	3.0
beta-BHC	3.956	3.861	4.061	52.520	50.000	5.0
Decachlorobiphenyl	7.987	7.892	8.092	48.190	50.000	-3.6
delta-BHC	4.189	4.095	4.295	52.180	50.000	4.4
Dieldrin	5.430	5.334	5.534	51.280	50.000	2.6
Endosulfan I	5.165	5.069	5.269	50.070	50.000	0.1
Endosulfan II	5.996	5.901	6.101	56.750	50.000	13.5
Endosulfan sulfate	6.398	6.302	6.502	49.690	50.000	-0.6
Endrin	5.704	5.609	5.809	50.950	50.000	1.9
Endrin aldehyde	6.174	6.079	6.279	51.830	50.000	3.7
Endrin ketone	6.903	6.807	7.007	50.320	50.000	0.6
gamma-BHC (Lindane)	3.660	3.565	3.765	52.940	50.000	5.9
gamma-Chlordane	5.046	4.951	5.151	52.790	50.000	5.6
Heptachlor	4.009	3.914	4.114	51.850	50.000	3.7
Heptachlor epoxide	4.794	4.699	4.899	51.080	50.000	2.2
Methoxychlor	6.677	6.581	6.781	50.460	50.000	0.9
Tetrachloro-m-xylene	2.822	2.728	2.928	54.140	50.000	8.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096819.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 10:58
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:54:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.822	173.6E6	258.7E6	54.574	54.143
28)	SA Decachlor...	9.009	7.987	127.8E6	209.1E6	53.584	48.192
Target Compounds							
2)	A alpha-BHC	3.976	3.328	262.0E6	377.1E6	56.623m	53.290
3)	MA gamma-BHC...	4.304	3.660	234.5E6	349.6E6	53.005m	52.938
4)	MA Heptachlor	4.897	4.009	248.7E6	345.4E6	59.841	51.850
5)	MB Aldrin	5.237	4.291	238.5E6	317.1E6	55.458	51.080
6)	B beta-BHC	4.491	3.956	99032385	148.2E6	54.867m	52.522
7)	B delta-BHC	4.738	4.189	224.7E6	337.8E6	56.322	52.180
8)	B Heptachlo...	5.657	4.794	206.2E6	291.8E6	53.473	51.083
9)	A Endosulfan I	6.039	5.165	198.1E6	277.0E6	55.090	50.075
10)	B gamma-Chl...	5.911	5.046	214.3E6	310.4E6	56.305	52.789
11)	B alpha-Chl...	5.992	5.110	213.7E6	303.9E6	55.429	51.520
12)	B 4,4'-DDE	6.162	5.299	183.4E6	286.5E6	57.098	51.952
13)	MA Dieldrin	6.311	5.430	201.5E6	302.9E6	54.299	51.283
14)	MA Endrin	6.536	5.704	167.9E6	275.4E6	55.479m	50.951
15)	B Endosulfa...	6.749	5.996	169.2E6	291.5E6	52.839m	56.746
16)	A 4,4'-DDD	6.669	5.852	136.2E6	240.2E6	53.850m	51.048
17)	MA 4,4'-DDT	6.985	6.103	158.5E6	265.7E6	55.256	52.524m
18)	B Endrin al...	6.878	6.174	115.1E6	188.4E6	53.660m	51.827
19)	B Endosulfa...	7.113	6.398	151.1E6	252.7E6	52.569	49.688
20)	A Methoxychlor	7.457	6.677	80618417	138.3E6	54.904	50.460
21)	B Endrin ke...	7.592	6.903	159.7E6	280.3E6	53.123	50.322
22)	Mirex	8.071	7.092	128.9E6	216.7E6	51.989	49.693

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096819.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 10:58
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:54:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 173588834
Conc: 54.57 ng/ml

Instrument :

ECD_L

ClientSampleId :


PSTDCCC050

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025




#1 Tetrachloro-m-xylene

R.T.: 2.822 min
Delta R.T.: -0.006 min
Response: 258683720
Conc: 54.14 ng/ml



#2 alpha-BHC

R.T.: 3.976 min
Delta R.T.: -0.006 min
Response: 261958305
Conc: 56.62 ng/ml m



#2 alpha-BHC

R.T.: 3.328 min
Delta R.T.: -0.005 min
Response: 377136964
Conc: 53.29 ng/ml

Response_ Signal: PL096819.D\ECD1A.ch

#3 gamma-BHC (Lindane)

R.T.: 4.304 min
Delta R.T.: -0.006 min
Response: 234462579
Conc: 53.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Time

Response_ Signal: PL096819.D\ECD2B.ch

#3 gamma-BHC (Lindane)

R.T.: 3.660 min
Delta R.T.: -0.005 min
Response: 349570640
Conc: 52.94 ng/ml

Time

Response_ Signal: PL096819.D\ECD1A.ch

#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.005 min
Response: 248682326
Conc: 59.84 ng/ml

Time


Response_ Signal: PL096819.D\ECD2B.ch

#4 Heptachlor

R.T.: 4.009 min
Delta R.T.: -0.005 min
Response: 345419725
Conc: 51.85 ng/ml

Time

Response_ Signal: PL096819.D\ECD1A.ch



#5 Aldrin

R.T.: 5.237 min
Delta R.T.: -0.006 min
Response: 238527024
Conc: 55.46 ng/ml

Instrument :

ECD_L


ClientSampleId :

PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025


Time Response_ Signal: PL096819.D\ECD2B.ch



#5 Aldrin

R.T.: 4.291 min
Delta R.T.: -0.006 min
Response: 317057424
Conc: 51.08 ng/ml


Time Response_ Signal: PL096819.D\ECD1A.ch



#6 beta-BHC

R.T.: 4.491 min
Delta R.T.: -0.006 min
Response: 99032385
Conc: 54.87 ng/ml m

Time Response_ Signal: PL096819.D\ECD2B.ch




#6 beta-BHC

R.T.: 3.956 min
Delta R.T.: -0.005 min
Response: 148157872
Conc: 52.52 ng/ml

Response_ Signal: PL096819.D\ECD1A.ch

#7 delta-BHC



R.T.: 4.738 min
Delta R.T.: -0.004 min
Response: 224715371
Conc: 56.32 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Response_ Signal: PL096819.D\ECD2B.ch


#7 delta-BHC



R.T.: 4.189 min
Delta R.T.: -0.006 min
Response: 337794712
Conc: 52.18 ng/ml

Response_ Signal: PL096819.D\ECD1A.ch


#8 Heptachlor epoxide




R.T.: 5.657 min
Delta R.T.: -0.005 min
Response: 206221416
Conc: 53.47 ng/ml

Response_ Signal: PL096819.D\ECD2B.ch

#8 Heptachlor epoxide



R.T.: 4.794 min
Delta R.T.: -0.005 min
Response: 291750031
Conc: 51.08 ng/ml




#9 Endosulfan I

R.T.: 6.039 min
Delta R.T.: -0.004 min
Response: 198123782
Conc: 55.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#9 Endosulfan I

R.T.: 5.165 min
Delta R.T.: -0.004 min
Response: 276983051
Conc: 50.07 ng/ml




#10 gamma-Chlordane

R.T.: 5.911 min
Delta R.T.: -0.005 min
Response: 214254477
Conc: 56.31 ng/ml



#10 gamma-Chlordane

R.T.: 5.046 min
Delta R.T.: -0.005 min
Response: 310405952
Conc: 52.79 ng/ml



#11 alpha-Chlordane

R.T.: 5.992 min
Delta R.T.: -0.005 min
Response: 213658778
Conc: 55.43 ng/ml

Instrument :


ECD_L

ClientSampleId :

PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#11 alpha-Chlordane

R.T.: 5.110 min
Delta R.T.: -0.005 min
Response: 303910082
Conc: 51.52 ng/ml




#12 4,4'-DDE

R.T.: 6.162 min
Delta R.T.: -0.005 min
Response: 183355642
Conc: 57.10 ng/ml



#12 4,4'-DDE

R.T.: 5.299 min
Delta R.T.: -0.005 min
Response: 286484870
Conc: 51.95 ng/ml




#13 Dieldrin

R.T.: 6.311 min
Delta R.T.: -0.005 min
Response: 201510129
Conc: 54.30 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#13 Dieldrin

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 302906139
Conc: 51.28 ng/ml




#14 Endrin

R.T.: 6.536 min
Delta R.T.: -0.007 min
Response: 167924672
Conc: 55.48 ng/ml m



#14 Endrin

R.T.: 5.704 min
Delta R.T.: -0.005 min
Response: 275437287
Conc: 50.95 ng/ml




#15 Endosulfan II

R.T.: 6.749 min
Delta R.T.: -0.006 min
Response: 169225743
Conc: 52.84 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#15 Endosulfan II

R.T.: 5.996 min
Delta R.T.: -0.005 min
Response: 291461407
Conc: 56.75 ng/ml




#16 4,4'-DDD

R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 136152898
Conc: 53.85 ng/ml m



#16 4,4'-DDD

R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 240194386
Conc: 51.05 ng/ml




#17 4,4'-DDT

R.T.: 6.985 min
Delta R.T.: -0.004 min
Response: 158456518
Conc: 55.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#17 4,4'-DDT

R.T.: 6.103 min
Delta R.T.: -0.006 min
Response: 265682609
Conc: 52.52 ng/ml m




#18 Endrin aldehyde

R.T.: 6.878 min
Delta R.T.: -0.006 min
Response: 115149393
Conc: 53.66 ng/ml m



#18 Endrin aldehyde

R.T.: 6.174 min
Delta R.T.: -0.005 min
Response: 188393647
Conc: 51.83 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.113 min
Delta R.T.: -0.005 min
Response: 151067060
Conc: 52.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#19 Endosulfan Sulfate

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 252725524
Conc: 49.69 ng/ml




#20 Methoxychlor

R.T.: 7.457 min
Delta R.T.: -0.005 min
Response: 80618417
Conc: 54.90 ng/ml



#20 Methoxychlor

R.T.: 6.677 min
Delta R.T.: -0.004 min
Response: 138285984
Conc: 50.46 ng/ml




#21 Endrin ketone

R.T.: 7.592 min
Delta R.T.: -0.005 min
Response: 159725679
Conc: 53.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#21 Endrin ketone

R.T.: 6.903 min
Delta R.T.: -0.004 min
Response: 280283510
Conc: 50.32 ng/ml




#22 Mirex

R.T.: 8.071 min
Delta R.T.: -0.005 min
Response: 128899173
Conc: 51.99 ng/ml



#22 Mirex

R.T.: 7.092 min
Delta R.T.: -0.005 min
Response: 216703290
Conc: 49.69 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.009 min
Delta R.T.: -0.006 min
Response: 127780741
Conc: 53.58 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 209072671
Conc: 48.19 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/15/2025

Initial Calibration Date(s): 07/28/2025

07/28/2025

Continuing Calib Time: 18:05

Initial Calibration Time(s): 16:52

17:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.00	9.02	8.92	9.12	0.02
Tetrachloro-m-xylene	3.53	3.54	3.44	3.64	0.01
alpha-BHC	3.98	3.98	3.88	4.08	0.00
beta-BHC	4.49	4.50	4.40	4.60	0.01
delta-BHC	4.74	4.74	4.64	4.84	0.00
gamma-BHC (Lindane)	4.30	4.31	4.21	4.41	0.01
Heptachlor	4.90	4.90	4.80	5.00	0.00
Aldrin	5.24	5.24	5.14	5.34	0.00
Heptachlor epoxide	5.66	5.66	5.56	5.76	0.00
Endosulfan I	6.04	6.04	5.94	6.14	0.00
Dieldrin	6.31	6.32	6.22	6.42	0.01
4,4'-DDE	6.16	6.17	6.07	6.27	0.01
Endrin	6.54	6.54	6.44	6.64	0.01
Endosulfan II	6.75	6.76	6.66	6.86	0.01
4,4'-DDD	6.67	6.68	6.58	6.78	0.01
Endosulfan sulfate	7.11	7.12	7.02	7.22	0.01
4,4'-DDT	6.98	6.99	6.89	7.09	0.01
Methoxychlor	7.46	7.46	7.36	7.56	0.00
Endrin ketone	7.59	7.60	7.50	7.70	0.01
Endrin aldehyde	6.88	6.88	6.78	6.98	0.00
alpha-Chlordane	5.99	6.00	5.90	6.10	0.01
gamma-Chlordane	5.91	5.92	5.82	6.02	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/15/2025

Initial Calibration Date(s): 07/28/2025

07/28/2025

Continuing Calib Time: 18:05

Initial Calibration Time(s): 16:52

17:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.99	7.99	7.89	8.09	0.00
Tetrachloro-m-xylene	2.83	2.83	2.73	2.93	0.00
alpha-BHC	3.33	3.33	3.23	3.43	0.00
beta-BHC	3.96	3.96	3.86	4.06	0.00
delta-BHC	4.19	4.20	4.10	4.30	0.01
gamma-BHC (Lindane)	3.66	3.67	3.57	3.77	0.01
Heptachlor	4.01	4.01	3.91	4.11	0.00
Aldrin	4.29	4.30	4.20	4.40	0.01
Heptachlor epoxide	4.80	4.80	4.70	4.90	0.00
Endosulfan I	5.17	5.17	5.07	5.27	0.00
Dieldrin	5.43	5.43	5.33	5.53	0.00
4,4'-DDE	5.30	5.30	5.20	5.40	0.00
Endrin	5.71	5.71	5.61	5.81	0.01
Endosulfan II	6.00	6.00	5.90	6.10	0.00
4,4'-DDD	5.85	5.86	5.76	5.96	0.01
Endosulfan sulfate	6.40	6.40	6.30	6.50	0.00
4,4'-DDT	6.11	6.11	6.01	6.21	0.00
Methoxychlor	6.68	6.68	6.58	6.78	0.00
Endrin ketone	6.90	6.91	6.81	7.01	0.01
Endrin aldehyde	6.18	6.18	6.08	6.28	0.00
alpha-Chlordane	5.11	5.12	5.02	5.22	0.01
gamma-Chlordane	5.05	5.05	4.95	5.15	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096828.D **Time Analyzed:** 18:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.669	6.575	6.775	58.870	50.000	17.7
4,4'-DDE	6.160	6.066	6.266	55.770	50.000	11.5
4,4'-DDT	6.983	6.889	7.089	55.390	50.000	10.8
Aldrin	5.237	5.143	5.343	56.740	50.000	13.5
alpha-BHC	3.978	3.882	4.082	57.410	50.000	14.8
alpha-Chlordane	5.990	5.897	6.097	56.440	50.000	12.9
beta-BHC	4.491	4.397	4.597	57.600	50.000	15.2
Decachlorobiphenyl	9.003	8.916	9.116	54.120	50.000	8.2
delta-BHC	4.738	4.643	4.843	58.440	50.000	16.9
Dieldrin	6.310	6.216	6.416	56.090	50.000	12.2
Endosulfan I	6.038	5.944	6.144	55.200	50.000	10.4
Endosulfan II	6.747	6.655	6.855	55.660	50.000	11.3
Endosulfan sulfate	7.111	7.018	7.218	53.610	50.000	7.2
Endrin	6.535	6.442	6.642	56.360	50.000	12.7
Endrin aldehyde	6.878	6.784	6.984	58.580	50.000	17.2
Endrin ketone	7.590	7.497	7.697	54.130	50.000	8.3
gamma-BHC (Lindane)	4.304	4.210	4.410	57.370	50.000	14.7
gamma-Chlordane	5.910	5.816	6.016	57.270	50.000	14.5
Heptachlor	4.896	4.803	5.003	59.580	50.000	19.2
Heptachlor epoxide	5.656	5.562	5.762	58.260	50.000	16.5
Methoxychlor	7.456	7.362	7.562	53.330	50.000	6.7
Tetrachloro-m-xylene	3.531	3.435	3.635	56.840	50.000	13.7



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096828.D **Time Analyzed:** 18:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.853	5.756	5.956	53.650	50.000	7.3
4,4'-DDE	5.300	5.204	5.404	53.590	50.000	7.2
4,4'-DDT	6.105	6.009	6.209	53.130	50.000	6.3
Aldrin	4.294	4.197	4.397	54.340	50.000	8.7
alpha-BHC	3.331	3.233	3.433	55.370	50.000	10.7
alpha-Chlordane	5.112	5.015	5.215	53.060	50.000	6.1
beta-BHC	3.959	3.861	4.061	55.140	50.000	10.3
Decachlorobiphenyl	7.986	7.892	8.092	53.020	50.000	6.0
delta-BHC	4.192	4.095	4.295	55.180	50.000	10.4
Dieldrin	5.430	5.334	5.534	53.900	50.000	7.8
Endosulfan I	5.166	5.069	5.269	50.900	50.000	1.8
Endosulfan II	5.997	5.901	6.101	53.150	50.000	6.3
Endosulfan sulfate	6.398	6.302	6.502	52.730	50.000	5.5
Endrin	5.705	5.609	5.809	52.740	50.000	5.5
Endrin aldehyde	6.175	6.079	6.279	57.110	50.000	14.2
Endrin ketone	6.902	6.807	7.007	56.500	50.000	13.0
gamma-BHC (Lindane)	3.663	3.565	3.765	55.150	50.000	10.3
gamma-Chlordane	5.047	4.951	5.151	54.270	50.000	8.5
Heptachlor	4.011	3.914	4.114	53.780	50.000	7.6
Heptachlor epoxide	4.796	4.699	4.899	53.570	50.000	7.1
Methoxychlor	6.677	6.581	6.781	52.380	50.000	4.8
Tetrachloro-m-xylene	2.826	2.728	2.928	54.950	50.000	9.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096828.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 18:05
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

Client SampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:55:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.826	180.8E6	262.6E6	56.837	54.955
2)	SA Decachlor...	9.003	7.986	129.1E6	230.0E6	54.118m	53.015
Target Compounds							
2)	A alpha-BHC	3.978	3.331	265.6E6	391.8E6	57.410	55.367
3)	MA gamma-BHC...	4.304	3.663	253.8E6	364.2E6	57.374m	55.153
4)	MA Heptachlor	4.896	4.011	247.6E6	358.3E6	59.585m	53.779
5)	MB Aldrin	5.237	4.294	244.0E6	337.3E6	56.736	54.343
6)	B beta-BHC	4.491	3.959	104.0E6	155.5E6	57.599m	55.137
7)	B delta-BHC	4.738	4.192	233.2E6	357.2E6	58.436	55.177
8)	B Heptachlo...	5.656	4.796	224.7E6	306.0E6	58.256	53.572
9)	A Endosulfan I	6.038	5.166	198.5E6	281.6E6	55.197	50.904
10)	B gamma-Chl...	5.910	5.047	217.9E6	319.1E6	57.272	54.270
11)	B alpha-Chl...	5.990	5.112	217.6E6	313.0E6	56.444	53.056
12)	B 4,4'-DDE	6.160	5.300	179.1E6	295.5E6	55.767	53.587
13)	MA Dieldrin	6.310	5.430	208.1E6	318.4E6	56.087	53.901m
14)	MA Endrin	6.535	5.705	170.6E6	285.1E6	56.360m	52.735
15)	B Endosulfa...	6.747	5.997	178.3E6	273.0E6	55.661m	53.148
16)	A 4,4'-DDD	6.669	5.853	148.8E6	252.4E6	58.869	53.649
17)	MA 4,4'-DDT	6.983	6.105	158.8E6	268.7E6	55.393	53.128
18)	B Endrin al...	6.878	6.175	125.7E6	207.6E6	58.576	57.110
19)	B Endosulfa...	7.111	6.398	154.1E6	268.2E6	53.613	52.730
20)	A Methoxychlor	7.456	6.677	78302387	143.6E6	53.327	52.384
21)	B Endrin ke...	7.590	6.902	162.8E6	314.7E6	54.131	56.503
22)	Mirex	8.067	7.092	132.7E6	232.2E6	53.538m	53.257

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096828.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 18:05
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 180784769
Conc: 56.84 ng/ml

Instrument :

ECD_L

ClientSampleId :


PSTDCCC050

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025




#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 262561642
Conc: 54.95 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 265600436
Conc: 57.41 ng/ml



#2 alpha-BHC

R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 391837029
Conc: 55.37 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.304 min
Delta R.T.: -0.006 min
Response: 253789840
Conc: 57.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#3 gamma-BHC (Lindane)

R.T.: 3.663 min
Delta R.T.: -0.002 min
Response: 364194496
Conc: 55.15 ng/ml



#4 Heptachlor


R.T.: 4.896 min
Delta R.T.: -0.007 min
Response: 247615227
Conc: 59.58 ng/ml m



#4 Heptachlor

R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 358273285
Conc: 53.78 ng/ml

Response_ Signal: PL096828.D\ECD1A.ch



#5 Aldrin

R.T.: 5.237 min
Delta R.T.: -0.006 min
Response: 244022336
Conc: 56.74 ng/ml

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050


Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025


Response_ Signal: PL096828.D\ECD2B.ch



#5 Aldrin

R.T.: 4.294 min
Delta R.T.: -0.003 min
Response: 337308377
Conc: 54.34 ng/ml


Response_ Signal: PL096828.D\ECD1A.ch



#6 beta-BHC

R.T.: 4.491 min
Delta R.T.: -0.006 min
Response: 103963562
Conc: 57.60 ng/ml m

Response_ Signal: PL096828.D\ECD2B.ch




#6 beta-BHC

R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 155535841
Conc: 55.14 ng/ml

Response_ Signal: PL096828.D\ECD1A.ch

#7 delta-BHC



R.T.: 4.738 min
Delta R.T.: -0.004 min
Response: 233150376
Conc: 58.44 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Time Response_ Signal: PL096828.D\ECD2B.ch


#7 delta-BHC



R.T.: 4.192 min
Delta R.T.: -0.003 min
Response: 357192657
Conc: 55.18 ng/ml

Time Response_ Signal: PL096828.D\ECD1A.ch


#8 Heptachlor epoxide




R.T.: 5.656 min
Delta R.T.: -0.006 min
Response: 224665911
Conc: 58.26 ng/ml

Time Response_ Signal: PL096828.D\ECD2B.ch

#8 Heptachlor epoxide



R.T.: 4.796 min
Delta R.T.: -0.003 min
Response: 305967969
Conc: 53.57 ng/ml




#9 Endosulfan I

R.T.: 6.038 min
Delta R.T.: -0.006 min
Response: 198508279
Conc: 55.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#9 Endosulfan I

R.T.: 5.166 min
Delta R.T.: -0.003 min
Response: 281573435
Conc: 50.90 ng/ml




#10 gamma-Chlordane

R.T.: 5.910 min
Delta R.T.: -0.006 min
Response: 217933444
Conc: 57.27 ng/ml



#10 gamma-Chlordane

R.T.: 5.047 min
Delta R.T.: -0.004 min
Response: 319111731
Conc: 54.27 ng/ml




#11 alpha-Chlordane

R.T.: 5.990 min
Delta R.T.: -0.006 min
Response: 217573409
Conc: 56.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#11 alpha-Chlordane

R.T.: 5.112 min
Delta R.T.: -0.003 min
Response: 312970390
Conc: 53.06 ng/ml




#12 4,4'-DDE

R.T.: 6.160 min
Delta R.T.: -0.006 min
Response: 179083022
Conc: 55.77 ng/ml



#12 4,4'-DDE

R.T.: 5.300 min
Delta R.T.: -0.004 min
Response: 295499974
Conc: 53.59 ng/ml




#13 Dieldrin

R.T.: 6.310 min
Delta R.T.: -0.006 min
Response: 208146865
Conc: 56.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#13 Dieldrin

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 318371693
Conc: 53.90 ng/ml m




#14 Endrin

R.T.: 6.535 min
Delta R.T.: -0.007 min
Response: 170590132
Conc: 56.36 ng/ml m



#14 Endrin

R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 285083802
Conc: 52.74 ng/ml




#15 Endosulfan II

R.T.: 6.747 min
Delta R.T.: -0.008 min
Response: 178261308
Conc: 55.66 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#15 Endosulfan II

R.T.: 5.997 min
Delta R.T.: -0.004 min
Response: 272979738
Conc: 53.15 ng/ml



#16 4,4'-DDD

R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 148842821
Conc: 58.87 ng/ml



#16 4,4'-DDD

R.T.: 5.853 min
Delta R.T.: -0.003 min
Response: 252433717
Conc: 53.65 ng/ml

Response_ Signal: PL096828.D\ECD1A.ch

#17 4,4'-DDT

R.T.: 6.983 min
Delta R.T.: -0.006 min
Response: 158848972
Conc: 55.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Time 6.60 6.70 6.80 6.90 7.00 7.10 7.20 7.30

Response_ Signal: PL096828.D\ECD2B.ch

#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 268739257
Conc: 53.13 ng/ml

Time 5.90 5.95 6.00 6.05 6.10 6.15 6.20 6.25 6.30

Response_ Signal: PL096828.D\ECD1A.ch

#18 Endrin aldehyde

R.T.: 6.878 min
Delta R.T.: -0.006 min
Response: 125699655
Conc: 58.58 ng/ml

Time 6.70 6.80 6.90 7.00 7.10


Response_ Signal: PL096828.D\ECD2B.ch

#18 Endrin aldehyde

R.T.: 6.175 min
Delta R.T.: -0.004 min
Response: 207597724
Conc: 57.11 ng/ml

Time 5.90 6.00 6.10 6.20 6.30 6.40

Response_ Signal: PL096828.D\ECD1A.ch



#19 Endosulfan Sulfate

R.T.: 7.111 min
Delta R.T.: -0.007 min
Response: 154065949
Conc: 53.61 ng/ml

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050


Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025


Time Response_ Signal: PL096828.D\ECD2B.ch



#19 Endosulfan Sulfate

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 268193619
Conc: 52.73 ng/ml


Time Response_ Signal: PL096828.D\ECD1A.ch



#20 Methoxychlor


R.T.: 7.456 min
Delta R.T.: -0.007 min
Response: 78302387
Conc: 53.33 ng/ml

Time Response_ Signal: PL096828.D\ECD2B.ch



#20 Methoxychlor

R.T.: 6.677 min
Delta R.T.: -0.004 min
Response: 143559352
Conc: 52.38 ng/ml




#21 Endrin ketone

R.T.: 7.590 min
Delta R.T.: -0.007 min
Response: 162756760
Conc: 54.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 314706868
Conc: 56.50 ng/ml




#22 Mirex

R.T.: 8.067 min
Delta R.T.: -0.009 min
Response: 132739312
Conc: 53.54 ng/ml m



#22 Mirex

R.T.: 7.092 min
Delta R.T.: -0.005 min
Response: 232248246
Conc: 53.26 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.003 min
Delta R.T.: -0.012 min
Response: 129052895
Conc: 54.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 229997908
Conc: 53.02 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/15/2025

Initial Calibration Date(s): 07/28/2025

07/28/2025

Continuing Calib Time: 20:48

Initial Calibration Time(s): 16:52

17:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.00	9.02	8.92	9.12	0.02
Tetrachloro-m-xylene	3.53	3.54	3.44	3.64	0.01
alpha-BHC	3.98	3.98	3.88	4.08	0.00
beta-BHC	4.49	4.50	4.40	4.60	0.01
delta-BHC	4.74	4.74	4.64	4.84	0.00
gamma-BHC (Lindane)	4.31	4.31	4.21	4.41	0.00
Heptachlor	4.90	4.90	4.80	5.00	0.00
Aldrin	5.24	5.24	5.14	5.34	0.00
Heptachlor epoxide	5.66	5.66	5.56	5.76	0.00
Endosulfan I	6.04	6.04	5.94	6.14	0.00
Dieldrin	6.31	6.32	6.22	6.42	0.01
4,4'-DDE	6.16	6.17	6.07	6.27	0.01
Endrin	6.53	6.54	6.44	6.64	0.01
Endosulfan II	6.75	6.76	6.66	6.86	0.01
4,4'-DDD	6.67	6.68	6.58	6.78	0.01
Endosulfan sulfate	7.11	7.12	7.02	7.22	0.01
4,4'-DDT	6.98	6.99	6.89	7.09	0.01
Methoxychlor	7.46	7.46	7.36	7.56	0.01
Endrin ketone	7.59	7.60	7.50	7.70	0.01
Endrin aldehyde	6.88	6.88	6.78	6.98	0.00
alpha-Chlordane	5.99	6.00	5.90	6.10	0.01
gamma-Chlordane	5.91	5.92	5.82	6.02	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2815</u>
Continuing Calib Date:	<u>08/15/2025</u>	Initial Calibration Date(s):	<u>07/28/2025</u> <u>07/28/2025</u>
Continuing Calib Time:	<u>20:48</u>	Initial Calibration Time(s):	<u>16:52</u> <u>17:47</u>

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.99	7.99	7.89	8.09	0.00
Tetrachloro-m-xylene	2.83	2.83	2.73	2.93	0.00
alpha-BHC	3.33	3.33	3.23	3.43	0.00
beta-BHC	3.96	3.96	3.86	4.06	0.00
delta-BHC	4.19	4.20	4.10	4.30	0.01
gamma-BHC (Lindane)	3.66	3.67	3.57	3.77	0.01
Heptachlor	4.01	4.01	3.91	4.11	0.00
Aldrin	4.29	4.30	4.20	4.40	0.01
Heptachlor epoxide	4.80	4.80	4.70	4.90	0.00
Endosulfan I	5.17	5.17	5.07	5.27	0.01
Dieldrin	5.43	5.43	5.33	5.53	0.00
4,4'-DDE	5.30	5.30	5.20	5.40	0.00
Endrin	5.71	5.71	5.61	5.81	0.01
Endosulfan II	6.00	6.00	5.90	6.10	0.00
4,4'-DDD	5.85	5.86	5.76	5.96	0.01
Endosulfan sulfate	6.40	6.40	6.30	6.50	0.00
4,4'-DDT	6.10	6.11	6.01	6.21	0.01
Methoxychlor	6.68	6.68	6.58	6.78	0.00
Endrin ketone	6.90	6.91	6.81	7.01	0.01
Endrin aldehyde	6.17	6.18	6.08	6.28	0.01
alpha-Chlordane	5.11	5.12	5.02	5.22	0.01
gamma-Chlordane	5.05	5.05	4.95	5.15	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096837.D **Time Analyzed:** 20:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.666	6.575	6.775	57.930	50.000	15.9
4,4'-DDE	6.159	6.066	6.266	57.590	50.000	15.2
4,4'-DDT	6.982	6.889	7.089	52.850	50.000	5.7
Aldrin	5.236	5.143	5.343	56.540	50.000	13.1
alpha-BHC	3.978	3.882	4.082	57.800	50.000	15.6
alpha-Chlordane	5.989	5.897	6.097	57.290	50.000	14.6
beta-BHC	4.492	4.397	4.597	56.550	50.000	13.1
Decachlorobiphenyl	9.003	8.916	9.116	55.610	50.000	11.2
delta-BHC	4.738	4.643	4.843	58.560	50.000	17.1
Dieldrin	6.309	6.216	6.416	53.720	50.000	7.4
Endosulfan I	6.037	5.944	6.144	54.910	50.000	9.8
Endosulfan II	6.746	6.655	6.855	54.430	50.000	8.9
Endosulfan sulfate	7.110	7.018	7.218	52.570	50.000	5.1
Endrin	6.534	6.442	6.642	55.170	50.000	10.3
Endrin aldehyde	6.877	6.784	6.984	57.720	50.000	15.4
Endrin ketone	7.589	7.497	7.697	54.290	50.000	8.6
gamma-BHC (Lindane)	4.306	4.210	4.410	56.580	50.000	13.2
gamma-Chlordane	5.909	5.816	6.016	56.630	50.000	13.3
Heptachlor	4.897	4.803	5.003	59.810	50.000	19.6
Heptachlor epoxide	5.655	5.562	5.762	57.880	50.000	15.8
Methoxychlor	7.455	7.362	7.562	50.240	50.000	0.5
Tetrachloro-m-xylene	3.531	3.435	3.635	56.820	50.000	13.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096837.D **Time Analyzed:** 20:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.852	5.756	5.956	53.330	50.000	6.7
4,4'-DDE	5.300	5.204	5.404	52.450	50.000	4.9
4,4'-DDT	6.104	6.009	6.209	50.060	50.000	0.1
Aldrin	4.293	4.197	4.397	53.730	50.000	7.5
alpha-BHC	3.331	3.233	3.433	54.720	50.000	9.4
alpha-Chlordane	5.111	5.015	5.215	52.090	50.000	4.2
beta-BHC	3.959	3.861	4.061	54.180	50.000	8.4
Decachlorobiphenyl	7.986	7.892	8.092	53.590	50.000	7.2
delta-BHC	4.191	4.095	4.295	54.240	50.000	8.5
Dieldrin	5.429	5.334	5.534	53.320	50.000	6.6
Endosulfan I	5.165	5.069	5.269	50.950	50.000	1.9
Endosulfan II	5.996	5.901	6.101	51.630	50.000	3.3
Endosulfan sulfate	6.397	6.302	6.502	51.600	50.000	3.2
Endrin	5.705	5.609	5.809	50.690	50.000	1.4
Endrin aldehyde	6.174	6.079	6.279	55.330	50.000	10.7
Endrin ketone	6.902	6.807	7.007	55.370	50.000	10.7
gamma-BHC (Lindane)	3.662	3.565	3.765	54.430	50.000	8.9
gamma-Chlordane	5.047	4.951	5.151	53.120	50.000	6.2
Heptachlor	4.011	3.914	4.114	52.430	50.000	4.9
Heptachlor epoxide	4.795	4.699	4.899	52.480	50.000	5.0
Methoxychlor	6.675	6.581	6.781	49.390	50.000	-1.2
Tetrachloro-m-xylene	2.826	2.728	2.928	54.380	50.000	8.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096837.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 20:48
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:57:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.826	180.7E6	259.8E6	56.821	54.377
28)	SA Decachlor...	9.003	7.986	132.6E6	232.5E6	55.611m	53.593
Target Compounds							
2)	A alpha-BHC	3.978	3.331	267.4E6	387.3E6	57.804	54.723
3)	MA gamma-BHC...	4.306	3.662	250.3E6	359.4E6	56.575	54.433
4)	MA Heptachlor	4.897	4.011	248.6E6	349.3E6	59.815	52.431
5)	MB Aldrin	5.236	4.293	243.2E6	333.5E6	56.545	53.727
6)	B beta-BHC	4.492	3.959	102.1E6	152.8E6	56.545	54.179
7)	B delta-BHC	4.738	4.191	233.6E6	351.1E6	58.557	54.240
8)	B Heptachlo...	5.655	4.795	223.2E6	299.7E6	57.882	52.481
9)	A Endosulfan I	6.037	5.165	197.5E6	281.8E6	54.908	50.952
10)	B gamma-Chl...	5.909	5.047	215.5E6	312.4E6	56.629	53.124
11)	B alpha-Chl...	5.989	5.111	220.8E6	307.3E6	57.290	52.088
12)	B 4,4'-DDE	6.159	5.300	184.9E6	289.2E6	57.588	52.452
13)	MA Dieldrin	6.309	5.429	199.4E6	314.9E6	53.723	53.317m
14)	MA Endrin	6.534	5.705	167.0E6	274.0E6	55.174m	50.691
15)	B Endosulfa...	6.746	5.996	174.3E6	265.2E6	54.433m	51.634
16)	A 4,4'-DDD	6.666	5.852	146.5E6	250.9E6	57.934m	53.334
17)	MA 4,4'-DDT	6.982	6.104	151.5E6	253.2E6	52.846	50.063
18)	B Endrin al...	6.877	6.174	123.9E6	201.1E6	57.722	55.332
19)	B Endosulfa...	7.110	6.397	151.1E6	262.5E6	52.573	51.602
20)	A Methoxychlor	7.455	6.675	73772766	135.4E6	50.242	49.391m
21)	B Endrin ke...	7.589	6.902	163.2E6	308.4E6	54.287	55.366
22)	Mirex	8.067	7.092	131.0E6	227.3E6	52.849m	52.122

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096837.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 20:48
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:57:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 180734970
Conc: 56.82 ng/ml

Instrument :

ECD_L

ClientSampleId :


PSTDCCC050

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025




#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 259800574
Conc: 54.38 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 267422416
Conc: 57.80 ng/ml



#2 alpha-BHC

R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 387274350
Conc: 54.72 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.306 min
Delta R.T.: -0.005 min
Response: 250257374
Conc: 56.58 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 359440039
Conc: 54.43 ng/ml



#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.006 min
Response: 248571375
Conc: 59.81 ng/ml




#4 Heptachlor

R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 349295899
Conc: 52.43 ng/ml

Response_ Signal: PL096837.D\ECD1A.ch

#5 Aldrin



R.T.: 5.236 min
Delta R.T.: -0.007 min
Response: 243198185
Conc: 56.54 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Time Response_ Signal: PL096837.D\ECD2B.ch


#5 Aldrin



R.T.: 4.293 min
Delta R.T.: -0.004 min
Response: 333486511
Conc: 53.73 ng/ml

Response_ Signal: PL096837.D\ECD1A.ch


#6 beta-BHC



R.T.: 4.492 min
Delta R.T.: -0.005 min
Response: 102062353
Conc: 56.55 ng/ml

Time Response_ Signal: PL096837.D\ECD2B.ch


#6 beta-BHC



R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 152833961
Conc: 54.18 ng/ml

Response_ Signal: PL096837.D\ECD1A.ch

#7 delta-BHC



R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 233635443
Conc: 58.56 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Response_ Signal: PL096837.D\ECD2B.ch


#7 delta-BHC



R.T.: 4.191 min
Delta R.T.: -0.004 min
Response: 351128336
Conc: 54.24 ng/ml

Response_ Signal: PL096837.D\ECD1A.ch


#8 Heptachlor epoxide



R.T.: 5.655 min
Delta R.T.: -0.007 min
Response: 223225892
Conc: 57.88 ng/ml

Response_ Signal: PL096837.D\ECD2B.ch


#8 Heptachlor epoxide



R.T.: 4.795 min
Delta R.T.: -0.004 min
Response: 299736064
Conc: 52.48 ng/ml

Signal: PL096837.D\ECD1A.ch

#9 Endosulfan I



R.T.: 6.037 min
Delta R.T.: -0.007 min
Response: 197468119
Conc: 54.91 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Signal: PL096837.D\ECD2B.ch


#9 Endosulfan I



R.T.: 5.165 min
Delta R.T.: -0.004 min
Response: 281834782
Conc: 50.95 ng/ml

Signal: PL096837.D\ECD1A.ch


#10 gamma-Chlordane




R.T.: 5.909 min
Delta R.T.: -0.007 min
Response: 215486294
Conc: 56.63 ng/ml

Signal: PL096837.D\ECD2B.ch

#10 gamma-Chlordane



R.T.: 5.047 min
Delta R.T.: -0.004 min
Response: 312373777
Conc: 53.12 ng/ml




#11 alpha-Chlordane

R.T.: 5.989 min
Delta R.T.: -0.007 min
Response: 220831221
Conc: 57.29 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#11 alpha-Chlordane

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 307264640
Conc: 52.09 ng/ml




#12 4,4'-DDE

R.T.: 6.159 min
Delta R.T.: -0.007 min
Response: 184929287
Conc: 57.59 ng/ml



#12 4,4'-DDE

R.T.: 5.300 min
Delta R.T.: -0.004 min
Response: 289241221
Conc: 52.45 ng/ml




#13 Dieldrin

R.T.: 6.309 min
Delta R.T.: -0.007 min
Response: 199375068
Conc: 53.72 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#13 Dieldrin

R.T.: 5.429 min
Delta R.T.: -0.005 min
Response: 314921208
Conc: 53.32 ng/ml m




#14 Endrin

R.T.: 6.534 min
Delta R.T.: -0.009 min
Response: 166998774
Conc: 55.17 ng/ml m



#14 Endrin

R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 274035427
Conc: 50.69 ng/ml




#15 Endosulfan II

R.T.: 6.746 min
Delta R.T.: -0.009 min
Response: 174328952
Conc: 54.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#15 Endosulfan II

R.T.: 5.996 min
Delta R.T.: -0.005 min
Response: 265206255
Conc: 51.63 ng/ml




#16 4,4'-DDD

R.T.: 6.666 min
Delta R.T.: -0.009 min
Response: 146477846
Conc: 57.93 ng/ml m



#16 4,4'-DDD

R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 250949782
Conc: 53.33 ng/ml




#17 4,4'-DDT

R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 151544815
Conc: 52.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#17 4,4'-DDT

R.T.: 6.104 min
Delta R.T.: -0.005 min
Response: 253232066
Conc: 50.06 ng/ml




#18 Endrin aldehyde

R.T.: 6.877 min
Delta R.T.: -0.007 min
Response: 123866292
Conc: 57.72 ng/ml



#18 Endrin aldehyde

R.T.: 6.174 min
Delta R.T.: -0.005 min
Response: 201134186
Conc: 55.33 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.110 min
Delta R.T.: -0.008 min
Response: 151078229
Conc: 52.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#19 Endosulfan Sulfate

R.T.: 6.397 min
Delta R.T.: -0.005 min
Response: 262456598
Conc: 51.60 ng/ml




#20 Methoxychlor

R.T.: 7.455 min
Delta R.T.: -0.008 min
Response: 73772766
Conc: 50.24 ng/ml



#20 Methoxychlor

R.T.: 6.675 min
Delta R.T.: -0.006 min
Response: 135355889
Conc: 49.39 ng/ml m




#21 Endrin ketone

R.T.: 7.589 min
Delta R.T.: -0.007 min
Response: 163226737
Conc: 54.29 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 308375287
Conc: 55.37 ng/ml




#22 Mirex

R.T.: 8.067 min
Delta R.T.: -0.008 min
Response: 131031867
Conc: 52.85 ng/ml m



#22 Mirex

R.T.: 7.092 min
Delta R.T.: -0.005 min
Response: 227296194
Conc: 52.12 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.003 min
Delta R.T.: -0.013 min
Response: 132614135
Conc: 55.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 232505562
Conc: 53.59 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/15/2025

Initial Calibration Date(s): 07/28/2025

07/28/2025

Continuing Calib Time: 22:10

Initial Calibration Time(s): 16:52

17:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.00	9.02	8.92	9.12	0.02
Tetrachloro-m-xylene	3.53	3.54	3.44	3.64	0.01
alpha-BHC	3.98	3.98	3.88	4.08	0.00
beta-BHC	4.49	4.50	4.40	4.60	0.01
delta-BHC	4.74	4.74	4.64	4.84	0.00
gamma-BHC (Lindane)	4.31	4.31	4.21	4.41	0.00
Heptachlor	4.90	4.90	4.80	5.00	0.00
Aldrin	5.24	5.24	5.14	5.34	0.00
Heptachlor epoxide	5.66	5.66	5.56	5.76	0.00
Endosulfan I	6.04	6.04	5.94	6.14	0.00
Dieldrin	6.31	6.32	6.22	6.42	0.01
4,4'-DDE	6.16	6.17	6.07	6.27	0.01
Endrin	6.54	6.54	6.44	6.64	0.01
Endosulfan II	6.75	6.76	6.66	6.86	0.01
4,4'-DDD	6.67	6.68	6.58	6.78	0.01
Endosulfan sulfate	7.11	7.12	7.02	7.22	0.01
4,4'-DDT	6.98	6.99	6.89	7.09	0.01
Methoxychlor	7.46	7.46	7.36	7.56	0.01
Endrin ketone	7.59	7.60	7.50	7.70	0.01
Endrin aldehyde	6.88	6.88	6.78	6.98	0.00
alpha-Chlordane	5.99	6.00	5.90	6.10	0.01
gamma-Chlordane	5.91	5.92	5.82	6.02	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/15/2025

Initial Calibration Date(s): 07/28/2025

07/28/2025

Continuing Calib Time: 22:10

Initial Calibration Time(s): 16:52

17:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.99	7.99	7.89	8.09	0.00
Tetrachloro-m-xylene	2.83	2.83	2.73	2.93	0.00
alpha-BHC	3.33	3.33	3.23	3.43	0.00
beta-BHC	3.96	3.96	3.86	4.06	0.00
delta-BHC	4.19	4.20	4.10	4.30	0.01
gamma-BHC (Lindane)	3.66	3.67	3.57	3.77	0.01
Heptachlor	4.01	4.01	3.91	4.11	0.00
Aldrin	4.29	4.30	4.20	4.40	0.01
Heptachlor epoxide	4.80	4.80	4.70	4.90	0.00
Endosulfan I	5.17	5.17	5.07	5.27	0.00
Dieldrin	5.43	5.43	5.33	5.53	0.00
4,4'-DDE	5.30	5.30	5.20	5.40	0.00
Endrin	5.71	5.71	5.61	5.81	0.01
Endosulfan II	6.00	6.00	5.90	6.10	0.00
4,4'-DDD	5.85	5.86	5.76	5.96	0.01
Endosulfan sulfate	6.40	6.40	6.30	6.50	0.00
4,4'-DDT	6.10	6.11	6.01	6.21	0.01
Methoxychlor	6.68	6.68	6.58	6.78	0.00
Endrin ketone	6.90	6.91	6.81	7.01	0.01
Endrin aldehyde	6.18	6.18	6.08	6.28	0.00
alpha-Chlordane	5.11	5.12	5.02	5.22	0.01
gamma-Chlordane	5.05	5.05	4.95	5.15	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096841.D **Time Analyzed:** 22:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.667	6.575	6.775	58.210	50.000	16.4
4,4'-DDE	6.160	6.066	6.266	56.440	50.000	12.9
4,4'-DDT	6.982	6.889	7.089	51.530	50.000	3.1
Aldrin	5.236	5.143	5.343	56.430	50.000	12.9
alpha-BHC	3.978	3.882	4.082	57.750	50.000	15.5
alpha-Chlordane	5.990	5.897	6.097	56.050	50.000	12.1
beta-BHC	4.491	4.397	4.597	58.240	50.000	16.5
Decachlorobiphenyl	9.003	8.916	9.116	55.680	50.000	11.4
delta-BHC	4.738	4.643	4.843	58.350	50.000	16.7
Dieldrin	6.308	6.216	6.416	55.140	50.000	10.3
Endosulfan I	6.037	5.944	6.144	55.030	50.000	10.1
Endosulfan II	6.747	6.655	6.855	54.040	50.000	8.1
Endosulfan sulfate	7.109	7.018	7.218	53.510	50.000	7.0
Endrin	6.535	6.442	6.642	53.310	50.000	6.6
Endrin aldehyde	6.878	6.784	6.984	57.740	50.000	15.5
Endrin ketone	7.590	7.497	7.697	54.970	50.000	9.9
gamma-BHC (Lindane)	4.306	4.210	4.410	56.530	50.000	13.1
gamma-Chlordane	5.909	5.816	6.016	57.020	50.000	14.0
Heptachlor	4.897	4.803	5.003	59.100	50.000	18.2
Heptachlor epoxide	5.656	5.562	5.762	57.030	50.000	14.1
Methoxychlor	7.455	7.362	7.562	48.450	50.000	-3.1
Tetrachloro-m-xylene	3.530	3.435	3.635	56.770	50.000	13.5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025
Client Sample No.: CCAL04 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096841.D **Time Analyzed:** 22:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.852	5.756	5.956	54.710	50.000	9.4
4,4'-DDE	5.300	5.204	5.404	53.400	50.000	6.8
4,4'-DDT	6.104	6.009	6.209	48.100	50.000	-3.8
Aldrin	4.294	4.197	4.397	54.520	50.000	9.0
alpha-BHC	3.331	3.233	3.433	55.470	50.000	10.9
alpha-Chlordane	5.111	5.015	5.215	53.020	50.000	6.0
beta-BHC	3.959	3.861	4.061	55.080	50.000	10.2
Decachlorobiphenyl	7.986	7.892	8.092	53.560	50.000	7.1
delta-BHC	4.191	4.095	4.295	55.010	50.000	10.0
Dieldrin	5.429	5.334	5.534	53.610	50.000	7.2
Endosulfan I	5.166	5.069	5.269	50.530	50.000	1.1
Endosulfan II	5.997	5.901	6.101	52.570	50.000	5.1
Endosulfan sulfate	6.398	6.302	6.502	51.980	50.000	4.0
Endrin	5.705	5.609	5.809	51.830	50.000	3.7
Endrin aldehyde	6.175	6.079	6.279	55.480	50.000	11.0
Endrin ketone	6.902	6.807	7.007	55.180	50.000	10.4
gamma-BHC (Lindane)	3.663	3.565	3.765	55.180	50.000	10.4
gamma-Chlordane	5.047	4.951	5.151	54.230	50.000	8.5
Heptachlor	4.011	3.914	4.114	52.370	50.000	4.7
Heptachlor epoxide	4.795	4.699	4.899	53.390	50.000	6.8
Methoxychlor	6.675	6.581	6.781	48.230	50.000	-3.5
Tetrachloro-m-xylene	2.825	2.728	2.928	55.360	50.000	10.7

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096841.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 22:10
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/20/2025

Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:57:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.825	180.6E6	264.5E6	56.769	55.355
28)	SA Decachlor...	9.003	7.986	132.8E6	232.4E6	55.677m	53.559
Target Compounds							
2)	A alpha-BHC	3.978	3.331	267.2E6	392.6E6	57.750	55.472
3)	MA gamma-BHC...	4.306	3.663	250.1E6	364.4E6	56.533	55.182
4)	MA Heptachlor	4.897	4.011	245.6E6	348.9E6	59.102	52.366
5)	MB Aldrin	5.236	4.294	242.7E6	338.4E6	56.432	54.523
6)	B beta-BHC	4.491	3.959	105.1E6	155.4E6	58.235m	55.083
7)	B delta-BHC	4.738	4.191	232.8E6	356.1E6	58.353	55.011
8)	B Heptachlo...	5.656	4.795	220.0E6	304.9E6	57.034	53.391
9)	A Endosulfan I	6.037	5.166	197.9E6	279.5E6	55.031	50.530
10)	B gamma-Chl...	5.909	5.047	217.0E6	318.9E6	57.015	54.231
11)	B alpha-Chl...	5.990	5.111	216.1E6	312.8E6	56.054	53.024
12)	B 4,4'-DDE	6.160	5.300	181.3E6	294.5E6	56.444	53.398
13)	MA Dieldrin	6.308	5.429	204.6E6	316.7E6	55.140m	53.612m
14)	MA Endrin	6.535	5.705	161.4E6	280.2E6	53.308m	51.832
15)	B Endosulfa...	6.747	5.997	173.1E6	270.0E6	54.038m	52.570
16)	A 4,4'-DDD	6.667	5.852	147.2E6	257.4E6	58.210m	54.710
17)	MA 4,4'-DDT	6.982	6.104	147.8E6	243.3E6	51.533	48.098
18)	B Endrin al...	6.878	6.175	123.9E6	201.7E6	57.736	55.476
19)	B Endosulfa...	7.109	6.398	153.8E6	264.4E6	53.513m	51.978
20)	A Methoxychlor	7.455	6.675	71143490	132.2E6	48.451	48.229m
21)	B Endrin ke...	7.590	6.902	165.3E6	307.3E6	54.974	55.181
22)	Mirex	8.067	7.091	130.8E6	230.2E6	52.760m	52.777

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096841.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 22:10
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 08/20/2025

Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:57:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 180569649
Conc: 56.77 ng/ml

Instrument :

ECD_L

ClientSampleId :


PSTDCCC050

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 08/20/2025

Supervised By :mohammad ahmed 08/21/2025




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 264474564
Conc: 55.36 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 267171388
Conc: 57.75 ng/ml



#2 alpha-BHC

R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 392574947
Conc: 55.47 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.306 min
Delta R.T.: -0.005 min
Response: 250071176
Conc: 56.53 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#3 gamma-BHC (Lindane)

R.T.: 3.663 min
Delta R.T.: -0.002 min
Response: 364386933
Conc: 55.18 ng/ml




#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.005 min
Response: 245610481
Conc: 59.10 ng/ml



#4 Heptachlor

R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 348861586
Conc: 52.37 ng/ml




#5 Aldrin

R.T.: 5.236 min
Delta R.T.: -0.007 min
Response: 242712510
Conc: 56.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#5 Aldrin

R.T.: 4.294 min
Delta R.T.: -0.003 min
Response: 338425766
Conc: 54.52 ng/ml



#6 beta-BHC

R.T.: 4.491 min
Delta R.T.: -0.006 min
Response: 105112723
Conc: 58.24 ng/ml m




#6 beta-BHC

R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 155383025
Conc: 55.08 ng/ml

Response_ Signal: PL096841.D\ECD1A.ch

#7 delta-BHC



R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 232818505
Conc: 58.35 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025

Response_ Signal: PL096841.D\ECD2B.ch


#7 delta-BHC



R.T.: 4.191 min
Delta R.T.: -0.004 min
Response: 356121645
Conc: 55.01 ng/ml

Response_ Signal: PL096841.D\ECD1A.ch


#8 Heptachlor epoxide




R.T.: 5.656 min
Delta R.T.: -0.007 min
Response: 219954792
Conc: 57.03 ng/ml

Response_ Signal: PL096841.D\ECD2B.ch

#8 Heptachlor epoxide



R.T.: 4.795 min
Delta R.T.: -0.004 min
Response: 304932521
Conc: 53.39 ng/ml




#9 Endosulfan I

R.T.: 6.037 min
Delta R.T.: -0.006 min
Response: 197913152
Conc: 55.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#9 Endosulfan I

R.T.: 5.166 min
Delta R.T.: -0.003 min
Response: 279504144
Conc: 50.53 ng/ml




#10 gamma-Chlordane

R.T.: 5.909 min
Delta R.T.: -0.007 min
Response: 216956458
Conc: 57.02 ng/ml



#10 gamma-Chlordane

R.T.: 5.047 min
Delta R.T.: -0.004 min
Response: 318882402
Conc: 54.23 ng/ml




#11 alpha-Chlordane

R.T.: 5.990 min
Delta R.T.: -0.006 min
Response: 216069565
Conc: 56.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#11 alpha-Chlordane

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 312786250
Conc: 53.02 ng/ml




#12 4,4'-DDE

R.T.: 6.160 min
Delta R.T.: -0.006 min
Response: 181255368
Conc: 56.44 ng/ml



#12 4,4'-DDE

R.T.: 5.300 min
Delta R.T.: -0.004 min
Response: 294458512
Conc: 53.40 ng/ml




#13 Dieldrin

R.T.: 6.308 min
Delta R.T.: -0.008 min
Response: 204632048
Conc: 55.14 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#13 Dieldrin

R.T.: 5.429 min
Delta R.T.: -0.005 min
Response: 316662631
Conc: 53.61 ng/ml m



#14 Endrin

R.T.: 6.535 min
Delta R.T.: -0.008 min
Response: 161351892
Conc: 53.31 ng/ml m




#14 Endrin

R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 280200055
Conc: 51.83 ng/ml

Response_ Signal: PL096841.D\ECD1A.ch

#15 Endosulfan II



R.T.: 6.747 min
Delta R.T.: -0.008 min
Response: 173065387
Conc: 54.04 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025

Time Response_ Signal: PL096841.D\ECD2B.ch


#15 Endosulfan II



R.T.: 5.997 min
Delta R.T.: -0.004 min
Response: 270010863
Conc: 52.57 ng/ml

Time Response_ Signal: PL096841.D\ECD1A.ch


#16 4,4'-DDD



R.T.: 6.667 min
Delta R.T.: -0.008 min
Response: 147177033
Conc: 58.21 ng/ml m

Time Response_ Signal: PL096841.D\ECD2B.ch


#16 4,4'-DDD



R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 257426326
Conc: 54.71 ng/ml

Response_ Signal: PL096841.D\ECD1A.ch

#17 4,4'-DDT



R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 147780944
Conc: 51.53 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025

Time Response_ Signal: PL096841.D\ECD2B.ch


#17 4,4'-DDT



R.T.: 6.104 min
Delta R.T.: -0.005 min
Response: 243296006
Conc: 48.10 ng/ml

Time Response_ Signal: PL096841.D\ECD1A.ch


#18 Endrin aldehyde



R.T.: 6.878 min
Delta R.T.: -0.006 min
Response: 123895984
Conc: 57.74 ng/ml

Time Response_ Signal: PL096841.D\ECD2B.ch


#18 Endrin aldehyde



R.T.: 6.175 min
Delta R.T.: -0.004 min
Response: 201657062
Conc: 55.48 ng/ml

Response_ Signal: PL096841.D\ECD1A.ch

#19 Endosulfan Sulfate



R.T.: 7.109 min
Delta R.T.: -0.008 min
Response: 153779924
Conc: 53.51 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025

Time Response_ Signal: PL096841.D\ECD2B.ch


#19 Endosulfan Sulfate



R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 264373350
Conc: 51.98 ng/ml

Response_ Signal: PL096841.D\ECD1A.ch


#20 Methoxychlor




R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 71143490
Conc: 48.45 ng/ml

Time Response_ Signal: PL096841.D\ECD2B.ch

#20 Methoxychlor



R.T.: 6.675 min
Delta R.T.: -0.006 min
Response: 132170521
Conc: 48.23 ng/ml m




#21 Endrin ketone

R.T.: 7.590 min
Delta R.T.: -0.007 min
Response: 165290236
Conc: 54.97 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 307346587
Conc: 55.18 ng/ml




#22 Mirex

R.T.: 8.067 min
Delta R.T.: -0.009 min
Response: 130809589
Conc: 52.76 ng/ml m



#22 Mirex

R.T.: 7.091 min
Delta R.T.: -0.006 min
Response: 230155234
Conc: 52.78 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.003 min
Delta R.T.: -0.013 min
Response: 132771027
Conc: 55.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 232358139
Conc: 53.56 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/18/2025

Initial Calibration Date(s): 07/28/2025 07/28/2025

Continuing Calib Time: 12:53

Initial Calibration Time(s): 16:52 17:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.00	9.02	8.92	9.12	0.02
Tetrachloro-m-xylene	3.53	3.54	3.44	3.64	0.01
alpha-BHC	3.98	3.98	3.88	4.08	0.00
beta-BHC	4.49	4.50	4.40	4.60	0.01
delta-BHC	4.74	4.74	4.64	4.84	0.00
gamma-BHC (Lindane)	4.30	4.31	4.21	4.41	0.01
Heptachlor	4.90	4.90	4.80	5.00	0.00
Aldrin	5.24	5.24	5.14	5.34	0.00
Heptachlor epoxide	5.66	5.66	5.56	5.76	0.00
Endosulfan I	6.04	6.04	5.94	6.14	0.00
Dieldrin	6.31	6.32	6.22	6.42	0.01
4,4'-DDE	6.16	6.17	6.07	6.27	0.01
Endrin	6.53	6.54	6.44	6.64	0.01
Endosulfan II	6.75	6.76	6.66	6.86	0.01
4,4'-DDD	6.67	6.68	6.58	6.78	0.01
Endosulfan sulfate	7.11	7.12	7.02	7.22	0.01
4,4'-DDT	6.98	6.99	6.89	7.09	0.01
Methoxychlor	7.46	7.46	7.36	7.56	0.01
Endrin ketone	7.59	7.60	7.50	7.70	0.01
Endrin aldehyde	6.88	6.88	6.78	6.98	0.00
alpha-Chlordane	5.99	6.00	5.90	6.10	0.01
gamma-Chlordane	5.91	5.92	5.82	6.02	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/18/2025

Initial Calibration Date(s): 07/28/2025

07/28/2025

Continuing Calib Time: 12:53

Initial Calibration Time(s): 16:52

17:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.99	7.99	7.89	8.09	0.00
Tetrachloro-m-xylene	2.83	2.83	2.73	2.93	0.00
alpha-BHC	3.33	3.33	3.23	3.43	0.00
beta-BHC	3.96	3.96	3.86	4.06	0.00
delta-BHC	4.19	4.20	4.10	4.30	0.01
gamma-BHC (Lindane)	3.66	3.67	3.57	3.77	0.01
Heptachlor	4.01	4.01	3.91	4.11	0.00
Aldrin	4.29	4.30	4.20	4.40	0.01
Heptachlor epoxide	4.80	4.80	4.70	4.90	0.00
Endosulfan I	5.17	5.17	5.07	5.27	0.00
Dieldrin	5.43	5.43	5.33	5.53	0.00
4,4'-DDE	5.30	5.30	5.20	5.40	0.00
Endrin	5.71	5.71	5.61	5.81	0.01
Endosulfan II	6.00	6.00	5.90	6.10	0.00
4,4'-DDD	5.85	5.86	5.76	5.96	0.01
Endosulfan sulfate	6.40	6.40	6.30	6.50	0.00
4,4'-DDT	6.11	6.11	6.01	6.21	0.00
Methoxychlor	6.68	6.68	6.58	6.78	0.00
Endrin ketone	6.90	6.91	6.81	7.01	0.01
Endrin aldehyde	6.18	6.18	6.08	6.28	0.00
alpha-Chlordane	5.11	5.12	5.02	5.22	0.01
gamma-Chlordane	5.05	5.05	4.95	5.15	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096850.D **Time Analyzed:** 12:53

COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
4,4'-DDD	6.669	6.575	6.775	57.210	50.000	14.4
4,4'-DDE	6.160	6.066	6.266	55.120	50.000	10.2
4,4'-DDT	6.983	6.889	7.089	51.140	50.000	2.3
Aldrin	5.236	5.143	5.343	55.700	50.000	11.4
alpha-BHC	3.976	3.882	4.082	57.600	50.000	15.2
alpha-Chlordane	5.990	5.897	6.097	55.490	50.000	11.0
beta-BHC	4.491	4.397	4.597	57.540	50.000	15.1
Decachlorobiphenyl	9.003	8.916	9.116	53.820	50.000	7.6
delta-BHC	4.738	4.643	4.843	57.850	50.000	15.7
Dieldrin	6.309	6.216	6.416	54.640	50.000	9.3
Endosulfan I	6.037	5.944	6.144	54.130	50.000	8.3
Endosulfan II	6.748	6.655	6.855	56.140	50.000	12.3
Endosulfan sulfate	7.109	7.018	7.218	52.640	50.000	5.3
Endrin	6.534	6.442	6.642	52.070	50.000	4.1
Endrin aldehyde	6.877	6.784	6.984	55.540	50.000	11.1
Endrin ketone	7.590	7.497	7.697	54.010	50.000	8.0
gamma-BHC (Lindane)	4.304	4.210	4.410	56.550	50.000	13.1
gamma-Chlordane	5.909	5.816	6.016	56.430	50.000	12.9
Heptachlor	4.897	4.803	5.003	58.560	50.000	17.1
Heptachlor epoxide	5.656	5.562	5.762	56.230	50.000	12.5
Methoxychlor	7.455	7.362	7.562	49.070	50.000	-1.9
Tetrachloro-m-xylene	3.531	3.435	3.635	56.740	50.000	13.5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096850.D **Time Analyzed:** 12:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.852	5.756	5.956	52.090	50.000	4.2
4,4'-DDE	5.300	5.204	5.404	51.640	50.000	3.3
4,4'-DDT	6.105	6.009	6.209	47.630	50.000	-4.7
Aldrin	4.293	4.197	4.397	52.560	50.000	5.1
alpha-BHC	3.331	3.233	3.433	54.010	50.000	8.0
alpha-Chlordane	5.111	5.015	5.215	51.140	50.000	2.3
beta-BHC	3.959	3.861	4.061	53.850	50.000	7.7
Decachlorobiphenyl	7.987	7.892	8.092	49.740	50.000	-0.5
delta-BHC	4.191	4.095	4.295	53.360	50.000	6.7
Dieldrin	5.431	5.334	5.534	51.330	50.000	2.7
Endosulfan I	5.166	5.069	5.269	49.310	50.000	-1.4
Endosulfan II	5.997	5.901	6.101	50.630	50.000	1.3
Endosulfan sulfate	6.398	6.302	6.502	49.500	50.000	-1.0
Endrin	5.705	5.609	5.809	47.970	50.000	-4.1
Endrin aldehyde	6.175	6.079	6.279	52.470	50.000	4.9
Endrin ketone	6.902	6.807	7.007	53.400	50.000	6.8
gamma-BHC (Lindane)	3.663	3.565	3.765	53.650	50.000	7.3
gamma-Chlordane	5.047	4.951	5.151	52.310	50.000	4.6
Heptachlor	4.011	3.914	4.114	51.450	50.000	2.9
Heptachlor epoxide	4.795	4.699	4.899	51.660	50.000	3.3
Methoxychlor	6.676	6.581	6.781	46.580	50.000	-6.8
Tetrachloro-m-xylene	2.825	2.728	2.928	54.010	50.000	8.0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
 Data File : PL096850.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2025 12:53
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/19/2025

Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 16:14:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.825	180.5E6	258.1E6	56.743	54.013
28)	SA Decachlor...	9.003	7.987	128.3E6	215.8E6	53.820m	49.743
Target Compounds							
2)	A alpha-BHC	3.976	3.331	266.5E6	382.2E6	57.604m	54.010
3)	MA gamma-BHC...	4.304	3.663	250.2E6	354.3E6	56.552m	53.650
4)	MA Heptachlor	4.897	4.011	243.4E6	342.8E6	58.563	51.453
5)	MB Aldrin	5.236	4.293	239.5E6	326.2E6	55.696	52.556
6)	B beta-BHC	4.491	3.959	103.9E6	151.9E6	57.544m	53.854
7)	B delta-BHC	4.738	4.191	230.8E6	345.4E6	57.846	53.355
8)	B Heptachlo...	5.656	4.795	216.8E6	295.1E6	56.227	51.661
9)	A Endosulfan I	6.037	5.166	194.7E6	272.8E6	54.131	49.314
10)	B gamma-Chl...	5.909	5.047	214.7E6	307.6E6	56.428	52.309
11)	B alpha-Chl...	5.990	5.111	213.9E6	301.7E6	55.491	51.139
12)	B 4,4'-DDE	6.160	5.300	177.0E6	284.8E6	55.124	51.643
13)	MA Dieldrin	6.309	5.431	202.8E6	303.2E6	54.636	51.330
14)	MA Endrin	6.534	5.705	157.6E6	259.3E6	52.066m	47.971
15)	B Endosulfa...	6.748	5.997	179.8E6	260.0E6	56.140	50.628
16)	A 4,4'-DDD	6.669	5.852	144.7E6	245.1E6	57.213	52.092
17)	MA 4,4'-DDT	6.983	6.105	146.7E6	240.9E6	51.141	47.630
18)	B Endrin al...	6.877	6.175	119.2E6	190.7E6	55.544	52.471
19)	B Endosulfa...	7.109	6.398	151.3E6	251.8E6	52.640m	49.499
20)	A Methoxychlor	7.455	6.676	72047040	127.7E6	49.067	46.583
21)	B Endrin ke...	7.590	6.902	162.4E6	297.4E6	54.012	53.401
22)	Mirex	8.067	7.091	129.4E6	219.8E6	52.182m	50.393m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096850.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 12:53
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 08/19/2025

Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 16:14:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 180486809
Conc: 56.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 258062587
Conc: 54.01 ng/ml



#2 alpha-BHC

R.T.: 3.976 min
Delta R.T.: -0.006 min
Response: 266497544
Conc: 57.60 ng/ml m



#2 alpha-BHC

R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 382230166
Conc: 54.01 ng/ml

Response_ Signal: PL096850.D\ECD1A.ch

#3 gamma-BHC (Lindane)

R.T.: 4.304 min
Delta R.T.: -0.006 min
Response: 250153591
Conc: 56.55 ng/ml

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025

Supervised By :mohammad ahmed 08/20/2025

Time

Response_ Signal: PL096850.D\ECD2B.ch

#3 gamma-BHC (Lindane)

R.T.: 3.663 min
Delta R.T.: -0.002 min
Response: 354267631
Conc: 53.65 ng/ml

Time

Response_ Signal: PL096850.D\ECD1A.ch

#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.006 min
Response: 243370604
Conc: 58.56 ng/ml

Time

Response_ Signal: PL096850.D\ECD2B.ch


#4 Heptachlor

R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 342778742
Conc: 51.45 ng/ml

Time

Response_ Signal: PL096850.D\ECD1A.ch

#5 Aldrin



R.T.: 5.236 min
Delta R.T.: -0.006 min
Response: 239548150
Conc: 55.70 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Response_ Signal: PL096850.D\ECD2B.ch


#5 Aldrin



R.T.: 4.293 min
Delta R.T.: -0.004 min
Response: 326218108
Conc: 52.56 ng/ml

Response_ Signal: PL096850.D\ECD1A.ch


#6 beta-BHC



R.T.: 4.491 min
Delta R.T.: -0.006 min
Response: 103865692
Conc: 57.54 ng/ml m

Response_ Signal: PL096850.D\ECD2B.ch


#6 beta-BHC



R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 151916219
Conc: 53.85 ng/ml

Response_ Signal: PL096850.D\ECD1A.ch

#7 delta-BHC



R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 230798856
Conc: 57.85 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Time Response_ Signal: PL096850.D\ECD2B.ch


#7 delta-BHC



R.T.: 4.191 min
Delta R.T.: -0.004 min
Response: 345402486
Conc: 53.36 ng/ml

Time Response_ Signal: PL096850.D\ECD1A.ch


#8 Heptachlor epoxide




R.T.: 5.656 min
Delta R.T.: -0.006 min
Response: 216841548
Conc: 56.23 ng/ml

Time Response_ Signal: PL096850.D\ECD2B.ch

#8 Heptachlor epoxide



R.T.: 4.795 min
Delta R.T.: -0.004 min
Response: 295053706
Conc: 51.66 ng/ml




#9 Endosulfan I

R.T.: 6.037 min
Delta R.T.: -0.006 min
Response: 194673996
Conc: 54.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#9 Endosulfan I

R.T.: 5.166 min
Delta R.T.: -0.003 min
Response: 272775922
Conc: 49.31 ng/ml




#10 gamma-Chlordane

R.T.: 5.909 min
Delta R.T.: -0.006 min
Response: 214722800
Conc: 56.43 ng/ml



#10 gamma-Chlordane

R.T.: 5.047 min
Delta R.T.: -0.004 min
Response: 307585622
Conc: 52.31 ng/ml




#11 alpha-Chlordane

R.T.: 5.990 min
Delta R.T.: -0.007 min
Response: 21390002
Conc: 55.49 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#11 alpha-Chlordane

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 301663826
Conc: 51.14 ng/ml




#12 4,4'-DDE

R.T.: 6.160 min
Delta R.T.: -0.006 min
Response: 177016085
Conc: 55.12 ng/ml



#12 4,4'-DDE

R.T.: 5.300 min
Delta R.T.: -0.004 min
Response: 284781762
Conc: 51.64 ng/ml




#13 Dieldrin

R.T.: 6.309 min
Delta R.T.: -0.006 min
Response: 202760836
Conc: 54.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#13 Dieldrin

R.T.: 5.431 min
Delta R.T.: -0.003 min
Response: 303181110
Conc: 51.33 ng/ml




#14 Endrin

R.T.: 6.534 min
Delta R.T.: -0.008 min
Response: 157593192
Conc: 52.07 ng/ml m



#14 Endrin

R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 259330218
Conc: 47.97 ng/ml




#15 Endosulfan II

R.T.: 6.748 min
Delta R.T.: -0.007 min
Response: 179795338
Conc: 56.14 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#15 Endosulfan II

R.T.: 5.997 min
Delta R.T.: -0.004 min
Response: 260037643
Conc: 50.63 ng/ml




#16 4,4'-DDD

R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 144655392
Conc: 57.21 ng/ml



#16 4,4'-DDD

R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 245109783
Conc: 52.09 ng/ml




#17 4,4'-DDT

R.T.: 6.983 min
Delta R.T.: -0.006 min
Response: 146657638
Conc: 51.14 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 240928658
Conc: 47.63 ng/ml



#18 Endrin aldehyde

R.T.: 6.877 min
Delta R.T.: -0.006 min
Response: 119191848
Conc: 55.54 ng/ml



#18 Endrin aldehyde

R.T.: 6.175 min
Delta R.T.: -0.004 min
Response: 190732922
Conc: 52.47 ng/ml

Response_ Signal: PL096850.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.109 min
Delta R.T.: -0.008 min
Response: 151272628
Conc: 52.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Time 6.90 6.95 7.00 7.05 7.10 7.15 7.20 7.25 7.30

Response_ Signal: PL096850.D\ECD2B.ch

#19 Endosulfan Sulfate

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 251764095
Conc: 49.50 ng/ml

Time 6.20 6.30 6.40 6.50 6.60

Response_ Signal: PL096850.D\ECD1A.ch

#20 Methoxychlor

R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 72047040
Conc: 49.07 ng/ml


Time 7.25 7.30 7.35 7.40 7.45 7.50 7.55 7.60 7.65

Response_ Signal: PL096850.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.676 min
Delta R.T.: -0.005 min
Response: 127661189
Conc: 46.58 ng/ml

Time 6.50 6.55 6.60 6.65 6.70 6.75 6.80 6.85




#21 Endrin ketone

R.T.: 7.590 min
Delta R.T.: -0.007 min
Response: 162398327
Conc: 54.01 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 297432233
Conc: 53.40 ng/ml




#22 Mirex

R.T.: 8.067 min
Delta R.T.: -0.008 min
Response: 129378159
Conc: 52.18 ng/ml m



#22 Mirex

R.T.: 7.091 min
Delta R.T.: -0.006 min
Response: 219758637
Conc: 50.39 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.003 min
Delta R.T.: -0.012 min
Response: 128342711
Conc: 53.82 ng/ml

Instrument :

ECD_L

ClientSampleId :


PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/19/2025

Supervised By :mohammad ahmed 08/20/2025



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 215802739
Conc: 49.74 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/18/2025

Initial Calibration Date(s): 07/28/2025 07/28/2025

Continuing Calib Time: 17:11

Initial Calibration Time(s): 16:52 17:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.00	9.02	8.92	9.12	0.02
Tetrachloro-m-xylene	3.53	3.54	3.44	3.64	0.01
alpha-BHC	3.98	3.98	3.88	4.08	0.00
beta-BHC	4.49	4.50	4.40	4.60	0.01
delta-BHC	4.74	4.74	4.64	4.84	0.00
gamma-BHC (Lindane)	4.31	4.31	4.21	4.41	0.00
Heptachlor	4.90	4.90	4.80	5.00	0.00
Aldrin	5.24	5.24	5.14	5.34	0.00
Heptachlor epoxide	5.66	5.66	5.56	5.76	0.00
Endosulfan I	6.04	6.04	5.94	6.14	0.00
Dieldrin	6.31	6.32	6.22	6.42	0.01
4,4'-DDE	6.16	6.17	6.07	6.27	0.01
Endrin	6.53	6.54	6.44	6.64	0.01
Endosulfan II	6.75	6.76	6.66	6.86	0.01
4,4'-DDD	6.67	6.68	6.58	6.78	0.01
Endosulfan sulfate	7.11	7.12	7.02	7.22	0.01
4,4'-DDT	6.98	6.99	6.89	7.09	0.01
Methoxychlor	7.45	7.46	7.36	7.56	0.01
Endrin ketone	7.59	7.60	7.50	7.70	0.01
Endrin aldehyde	6.88	6.88	6.78	6.98	0.00
alpha-Chlordane	5.99	6.00	5.90	6.10	0.01
gamma-Chlordane	5.91	5.92	5.82	6.02	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/18/2025

Initial Calibration Date(s): 07/28/2025

07/28/2025

Continuing Calib Time: 17:11

Initial Calibration Time(s): 16:52

17:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.98	7.99	7.89	8.09	0.01
Tetrachloro-m-xylene	2.83	2.83	2.73	2.93	0.00
alpha-BHC	3.33	3.33	3.23	3.43	0.00
beta-BHC	3.96	3.96	3.86	4.06	0.00
delta-BHC	4.19	4.20	4.10	4.30	0.01
gamma-BHC (Lindane)	3.66	3.67	3.57	3.77	0.01
Heptachlor	4.01	4.01	3.91	4.11	0.00
Aldrin	4.29	4.30	4.20	4.40	0.01
Heptachlor epoxide	4.79	4.80	4.70	4.90	0.01
Endosulfan I	5.17	5.17	5.07	5.27	0.01
Dieldrin	5.43	5.43	5.33	5.53	0.00
4,4'-DDE	5.30	5.30	5.20	5.40	0.00
Endrin	5.70	5.71	5.61	5.81	0.01
Endosulfan II	6.00	6.00	5.90	6.10	0.00
4,4'-DDD	5.85	5.86	5.76	5.96	0.01
Endosulfan sulfate	6.40	6.40	6.30	6.50	0.00
4,4'-DDT	6.10	6.11	6.01	6.21	0.01
Methoxychlor	6.67	6.68	6.58	6.78	0.01
Endrin ketone	6.90	6.91	6.81	7.01	0.01
Endrin aldehyde	6.17	6.18	6.08	6.28	0.01
alpha-Chlordane	5.11	5.12	5.02	5.22	0.01
gamma-Chlordane	5.05	5.05	4.95	5.15	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL06 **Date Analyzed:** 08/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096858.D **Time Analyzed:** 17:11

COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
4,4'-DDD	6.667	6.575	6.775	57.290	50.000	14.6
4,4'-DDE	6.159	6.066	6.266	55.290	50.000	10.6
4,4'-DDT	6.982	6.889	7.089	49.120	50.000	-1.8
Aldrin	5.236	5.143	5.343	55.400	50.000	10.8
alpha-BHC	3.978	3.882	4.082	57.140	50.000	14.3
alpha-Chlordane	5.989	5.897	6.097	55.080	50.000	10.2
beta-BHC	4.491	4.397	4.597	57.580	50.000	15.2
Decachlorobiphenyl	9.003	8.916	9.116	53.130	50.000	6.3
delta-BHC	4.738	4.643	4.843	57.850	50.000	15.7
Dieldrin	6.307	6.216	6.416	54.620	50.000	9.2
Endosulfan I	6.037	5.944	6.144	54.090	50.000	8.2
Endosulfan II	6.746	6.655	6.855	52.410	50.000	4.8
Endosulfan sulfate	7.108	7.018	7.218	52.270	50.000	4.5
Endrin	6.533	6.442	6.642	50.940	50.000	1.9
Endrin aldehyde	6.876	6.784	6.984	54.910	50.000	9.8
Endrin ketone	7.589	7.497	7.697	54.340	50.000	8.7
gamma-BHC (Lindane)	4.306	4.210	4.410	56.040	50.000	12.1
gamma-Chlordane	5.909	5.816	6.016	56.400	50.000	12.8
Heptachlor	4.897	4.803	5.003	57.590	50.000	15.2
Heptachlor epoxide	5.655	5.562	5.762	58.000	50.000	16.0
Methoxychlor	7.454	7.362	7.562	47.530	50.000	-4.9
Tetrachloro-m-xylene	3.531	3.435	3.635	56.050	50.000	12.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL06 **Date Analyzed:** 08/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096858.D **Time Analyzed:** 17:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.851	5.756	5.956	53.230	50.000	6.5
4,4'-DDE	5.299	5.204	5.404	52.000	50.000	4.0
4,4'-DDT	6.104	6.009	6.209	46.430	50.000	-7.1
Aldrin	4.293	4.197	4.397	52.560	50.000	5.1
alpha-BHC	3.330	3.233	3.433	53.570	50.000	7.1
alpha-Chlordane	5.110	5.015	5.215	51.430	50.000	2.9
beta-BHC	3.958	3.861	4.061	53.580	50.000	7.2
Decachlorobiphenyl	7.984	7.892	8.092	51.580	50.000	3.2
delta-BHC	4.191	4.095	4.295	53.250	50.000	6.5
Dieldrin	5.430	5.334	5.534	51.480	50.000	3.0
Endosulfan I	5.165	5.069	5.269	49.760	50.000	-0.5
Endosulfan II	5.996	5.901	6.101	51.220	50.000	2.4
Endosulfan sulfate	6.397	6.302	6.502	50.320	50.000	0.6
Endrin	5.704	5.609	5.809	49.140	50.000	-1.7
Endrin aldehyde	6.174	6.079	6.279	53.000	50.000	6.0
Endrin ketone	6.901	6.807	7.007	52.880	50.000	5.8
gamma-BHC (Lindane)	3.662	3.565	3.765	53.210	50.000	6.4
gamma-Chlordane	5.046	4.951	5.151	52.580	50.000	5.2
Heptachlor	4.010	3.914	4.114	50.580	50.000	1.2
Heptachlor epoxide	4.794	4.699	4.899	51.630	50.000	3.3
Methoxychlor	6.674	6.581	6.781	46.500	50.000	-7.0
Tetrachloro-m-xylene	2.825	2.728	2.928	53.510	50.000	7.0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
 Data File : PL096858.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2025 17:11
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/19/2025

Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 19 06:00:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.825	178.3E6	255.7E6	56.053	53.513
28)	SA Decachlor...	9.003	7.984	126.7E6	223.8E6	53.127m	51.577m
Target Compounds							
2)	A alpha-BHC	3.978	3.330	264.4E6	379.1E6	57.144	53.569
3)	MA gamma-BHC...	4.306	3.662	247.9E6	351.4E6	56.045	53.208
4)	MA Heptachlor	4.897	4.010	239.3E6	336.9E6	57.586	50.577
5)	MB Aldrin	5.236	4.293	238.3E6	326.2E6	55.402	52.559
6)	B beta-BHC	4.491	3.958	103.9E6	151.1E6	57.575m	53.581
7)	B delta-BHC	4.738	4.191	230.8E6	344.7E6	57.852	53.254
8)	B Heptachlo...	5.655	4.794	223.7E6	294.9E6	58.001	51.630
9)	A Endosulfan I	6.037	5.165	194.5E6	275.2E6	54.088	49.757
10)	B gamma-Chl...	5.909	5.046	214.6E6	309.2E6	56.404	52.576
11)	B alpha-Chl...	5.989	5.110	212.3E6	303.4E6	55.083	51.430
12)	B 4,4'-DDE	6.159	5.299	177.6E6	286.7E6	55.293	51.999
13)	MA Dieldrin	6.307	5.430	202.7E6	304.1E6	54.620m	51.481
14)	MA Endrin	6.533	5.704	154.2E6	265.6E6	50.941m	49.140
15)	B Endosulfa...	6.746	5.996	167.9E6	263.1E6	52.411m	51.215
16)	A 4,4'-DDD	6.667	5.851	144.8E6	250.5E6	57.289	53.228
17)	MA 4,4'-DDT	6.982	6.104	140.9E6	234.9E6	49.125	46.430
18)	B Endrin al...	6.876	6.174	117.8E6	192.7E6	54.905	52.999
19)	B Endosulfa...	7.108	6.397	150.2E6	255.9E6	52.274m	50.320
20)	A Methoxychlor	7.454	6.674	69792447	127.4E6	47.531	46.496m
21)	B Endrin ke...	7.589	6.901	163.4E6	294.6E6	54.341	52.884
22)	Mirex	8.066	7.091	129.7E6	222.9E6	52.299m	51.123

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096858.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 17:11
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 08/19/2025

Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 178292660
Conc: 56.05 ng/ml

Instrument :

ECD_L

ClientSampleId :


PSTDCCC050

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 08/19/2025

Supervised By :mohammad ahmed 08/20/2025




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 255673747
Conc: 53.51 ng/ml



#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 264370535
Conc: 57.14 ng/ml



#2 alpha-BHC

R.T.: 3.330 min
Delta R.T.: -0.003 min
Response: 379109531
Conc: 53.57 ng/ml

Response_ Signal: PL096858.D\ECD1A.ch

#3 gamma-BHC (Lindane)

R.T.: 4.306 min
Delta R.T.: -0.005 min
Response: 247910221
Conc: 56.04 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Time Response_ Signal: PL096858.D\ECD2B.ch

#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 351354786
Conc: 53.21 ng/ml

Time Response_ Signal: PL096858.D\ECD1A.ch

#4 Heptachlor


R.T.: 4.897 min
Delta R.T.: -0.006 min
Response: 239310105
Conc: 57.59 ng/ml

Time Response_ Signal: PL096858.D\ECD2B.ch

#4 Heptachlor

R.T.: 4.010 min
Delta R.T.: -0.004 min
Response: 336938392
Conc: 50.58 ng/ml

Response_ Signal: PL096858.D\ECD1A.ch



#5 Aldrin


R.T.: 5.236 min
Delta R.T.: -0.007 min
Response: 238282661
Conc: 55.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025


Response_ Signal: PL096858.D\ECD2B.ch



#5 Aldrin

R.T.: 4.293 min
Delta R.T.: -0.004 min
Response: 326237655
Conc: 52.56 ng/ml


Response_ Signal: PL096858.D\ECD1A.ch



#6 beta-BHC

R.T.: 4.491 min
Delta R.T.: -0.006 min
Response: 103921472
Conc: 57.58 ng/ml m

Response_ Signal: PL096858.D\ECD2B.ch




#6 beta-BHC

R.T.: 3.958 min
Delta R.T.: -0.003 min
Response: 151147750
Conc: 53.58 ng/ml

Response_ Signal: PL096858.D\ECD1A.ch

#7 delta-BHC



R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 230820906
Conc: 57.85 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Time Response_ Signal: PL096858.D\ECD2B.ch


#7 delta-BHC



R.T.: 4.191 min
Delta R.T.: -0.004 min
Response: 344747683
Conc: 53.25 ng/ml

Time Response_ Signal: PL096858.D\ECD1A.ch


#8 Heptachlor epoxide




R.T.: 5.655 min
Delta R.T.: -0.007 min
Response: 223683999
Conc: 58.00 ng/ml

Time Response_ Signal: PL096858.D\ECD2B.ch

#8 Heptachlor epoxide



R.T.: 4.794 min
Delta R.T.: -0.005 min
Response: 294878831
Conc: 51.63 ng/ml




#9 Endosulfan I

R.T.: 6.037 min
Delta R.T.: -0.007 min
Response: 194520543
Conc: 54.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#9 Endosulfan I

R.T.: 5.165 min
Delta R.T.: -0.004 min
Response: 275227850
Conc: 49.76 ng/ml




#10 gamma-Chlordane

R.T.: 5.909 min
Delta R.T.: -0.007 min
Response: 214629910
Conc: 56.40 ng/ml



#10 gamma-Chlordane

R.T.: 5.046 min
Delta R.T.: -0.005 min
Response: 309151010
Conc: 52.58 ng/ml




#11 alpha-Chlordane

R.T.: 5.989 min
Delta R.T.: -0.008 min
Response: 212327187
Conc: 55.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#11 alpha-Chlordane

R.T.: 5.110 min
Delta R.T.: -0.005 min
Response: 303382700
Conc: 51.43 ng/ml




#12 4,4'-DDE

R.T.: 6.159 min
Delta R.T.: -0.007 min
Response: 177559416
Conc: 55.29 ng/ml



#12 4,4'-DDE

R.T.: 5.299 min
Delta R.T.: -0.005 min
Response: 286745809
Conc: 52.00 ng/ml




#13 Dieldrin

R.T.: 6.307 min
Delta R.T.: -0.009 min
Response: 202703077
Conc: 54.62 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#13 Dieldrin

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 304075067
Conc: 51.48 ng/ml




#14 Endrin

R.T.: 6.533 min
Delta R.T.: -0.010 min
Response: 154187177
Conc: 50.94 ng/ml m



#14 Endrin

R.T.: 5.704 min
Delta R.T.: -0.005 min
Response: 265648440
Conc: 49.14 ng/ml




#15 Endosulfan II

R.T.: 6.746 min
Delta R.T.: -0.009 min
Response: 167853238
Conc: 52.41 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#15 Endosulfan II

R.T.: 5.996 min
Delta R.T.: -0.005 min
Response: 263053855
Conc: 51.22 ng/ml



#16 4,4'-DDD

R.T.: 6.667 min
Delta R.T.: -0.008 min
Response: 144848017
Conc: 57.29 ng/ml




#16 4,4'-DDD

R.T.: 5.851 min
Delta R.T.: -0.005 min
Response: 250453109
Conc: 53.23 ng/ml

Signal: PL096858.D\ECD1A.ch

#17 4,4'-DDT



R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 140873930
Conc: 49.12 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Signal: PL096858.D\ECD2B.ch


#17 4,4'-DDT



R.T.: 6.104 min
Delta R.T.: -0.005 min
Response: 234856582
Conc: 46.43 ng/ml

Signal: PL096858.D\ECD1A.ch


#18 Endrin aldehyde



R.T.: 6.876 min
Delta R.T.: -0.007 min
Response: 117821788
Conc: 54.91 ng/ml


Signal: PL096858.D\ECD2B.ch

#18 Endrin aldehyde



R.T.: 6.174 min
Delta R.T.: -0.005 min
Response: 192653257
Conc: 53.00 ng/ml

Response_ Signal: PL096858.D\ECD1A.ch



#19 Endosulfan Sulfate


R.T.: 7.108 min
Delta R.T.: -0.009 min
Response: 150218945
Conc: 52.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025


Time Response_ Signal: PL096858.D\ECD2B.ch



#19 Endosulfan Sulfate

R.T.: 6.397 min
Delta R.T.: -0.005 min
Response: 255940668
Conc: 50.32 ng/ml


Time Response_ Signal: PL096858.D\ECD1A.ch



#20 Methoxychlor


R.T.: 7.454 min
Delta R.T.: -0.008 min
Response: 69792447
Conc: 47.53 ng/ml

Time Response_ Signal: PL096858.D\ECD2B.ch



#20 Methoxychlor

R.T.: 6.674 min
Delta R.T.: -0.007 min
Response: 127423332
Conc: 46.50 ng/ml m




#21 Endrin ketone

R.T.: 7.589 min
Delta R.T.: -0.008 min
Response: 163387029
Conc: 54.34 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#21 Endrin ketone

R.T.: 6.901 min
Delta R.T.: -0.006 min
Response: 294551207
Conc: 52.88 ng/ml



#22 Mirex

R.T.: 8.066 min
Delta R.T.: -0.010 min
Response: 129666396
Conc: 52.30 ng/ml m




#22 Mirex

R.T.: 7.091 min
Delta R.T.: -0.006 min
Response: 222941586
Conc: 51.12 ng/ml

Signal: PL096858.D\ECD1A.ch

#28 Decachlorobiphenyl



R.T.: 9.003 min
Delta R.T.: -0.013 min
Response: 126690919
Conc: 53.13 ng/ml


Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Signal: PL096858.D\ECD2B.ch

#28 Decachlorobiphenyl



R.T.: 7.984 min
Delta R.T.: -0.008 min
Response: 223758596
Conc: 51.58 ng/ml m

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/28/2025 07/28/2025

Client Sample No. (PEM): PEM - PL096594.D Date Analyzed: 07/28/2025

Lab Sample No.(PEM): PEM Time Analyzed: 16:25

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.017	8.920	9.120	21.270	20.000	6.4
Tetrachloro-m-xylene	3.535	3.480	3.590	20.960	20.000	4.8
alpha-BHC	3.983	3.930	4.030	10.440	10.000	4.4
beta-BHC	4.498	4.450	4.550	10.520	10.000	5.2
gamma-BHC (Lindane)	4.311	4.260	4.360	10.550	10.000	5.5
Endrin	6.543	6.470	6.610	54.020	50.000	8.0
4,4'-DDT	6.990	6.920	7.060	109.500	100.000	9.5
Methoxychlor	7.463	7.390	7.530	267.200	250.000	6.9

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/28/2025 07/28/2025

Client Sample No. (PEM): PEM - PL096594.D Date Analyzed: 07/28/2025

Lab Sample No.(PEM): PEM Time Analyzed: 16:25

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.992	7.890	8.090	20.870	20.000	4.4
Tetrachloro-m-xylene	2.828	2.780	2.880	20.780	20.000	3.9
alpha-BHC	3.333	3.280	3.380	10.270	10.000	2.7
beta-BHC	3.962	3.910	4.010	10.840	10.000	8.4
gamma-BHC (Lindane)	3.666	3.620	3.720	10.440	10.000	4.4
Endrin	5.710	5.640	5.780	53.630	50.000	7.3
4,4'-DDT	6.109	6.040	6.180	111.210	100.000	11.2
Methoxychlor	6.682	6.610	6.750	253.290	250.000	1.3

Data File: PEM
 PL096594.D **Date Acquired** 7/28/2025 16:25
Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.54	163494961.3	167390237.3	3895275.98	2.33
Endrin aldehyde	6.89	1570507.515			
Endrin ketone	7.60	2324768.467			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.71	289935231	306295121.7	16359890.7	5.34
Endrin aldehyde #2	6.18	12998852.66			
Endrin ketone #2	6.91	3361038.042			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	6.99	314003595.7	315320131.5	1316535.88	0.42
4,4'-DDE	0.00	0			
4,4'-DDD	6.68	1316535.881			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.11	562525716.3	564144685.1	1618968.83	0.29
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.86	1618968.831			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096594.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 16:25
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 07/29/2025
 Supervised By :mohammad ahmed 07/30/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 08:57:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 08:58:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.535	2.828	66676938	99261210	20.962	20.776
28) SA Decachlor...	9.017	7.992	50725219	90556842	21.271	20.874
Target Compounds						
2) A alpha-BHC	3.983	3.333	48281852	72680702	10.436	10.270
3) MA gamma-BHC...	4.311	3.666	46672698	68921884	10.551	10.437
6) B beta-BHC	4.498	3.962	18987747	30564619	10.520	10.835
14) MA Endrin	6.543	5.710	163.5E6	289.9E6	54.016	53.632
16) A 4,4'-DDD	6.677	5.858	1316536	1618969	0.521m	0.344m#
17) MA 4,4'-DDT	6.990	6.109	314.0E6	562.5E6	109.497	111.208
18) B Endrin al...	6.888	6.180	1570508	12998853	0.732m	0.219 #
20) A Methoxychlor	7.463	6.682	392.3E6	694.1E6	267.202	253.290
21) B Endrin ke...	7.598	6.905	2324768	3361038	0.773m	0.603m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096594.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 16:25
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/29/2025
Supervised By : mohammad ahmed 07/30/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 08:57:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 08:58:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 66676938
Conc: 20.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 99261210
Conc: 20.78 ng/ml




#2 alpha-BHC

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 48281852
Conc: 10.44 ng/ml



#2 alpha-BHC

R.T.: 3.333 min
Delta R.T.: 0.000 min
Response: 72680702
Conc: 10.27 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 46672698
Conc: 10.55 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#3 gamma-BHC (Lindane)

R.T.: 3.666 min
Delta R.T.: 0.000 min
Response: 68921884
Conc: 10.44 ng/ml




#6 beta-BHC

R.T.: 4.498 min
Delta R.T.: 0.000 min
Response: 18987747
Conc: 10.52 ng/ml



#6 beta-BHC

R.T.: 3.962 min
Delta R.T.: 0.000 min
Response: 30564619
Conc: 10.84 ng/ml




#14 Endrin

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 163494961
Conc: 54.02 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#14 Endrin

R.T.: 5.710 min
Delta R.T.: 0.000 min
Response: 289935231
Conc: 53.63 ng/ml




#16 4,4'-DDD

R.T.: 6.677 min
Delta R.T.: 0.002 min
Response: 1316536
Conc: 0.52 ng/ml m



#16 4,4'-DDD

R.T.: 5.858 min
Delta R.T.: 0.001 min
Response: 1618969
Conc: 0.34 ng/ml m




#17 4,4'-DDT

R.T.: 6.990 min
Delta R.T.: 0.000 min
Response: 314003596
Conc: 109.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#17 4,4'-DDT

R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 562525716
Conc: 111.21 ng/ml




#18 Endrin aldehyde

R.T.: 6.888 min
Delta R.T.: 0.004 min
Response: 1570508
Conc: 0.73 ng/ml m



#18 Endrin aldehyde

R.T.: 6.180 min
Delta R.T.: 0.001 min
Response: 12998853
Conc: 0.22 ng/ml




#20 Methoxychlor

R.T.: 7.463 min
Delta R.T.: 0.000 min
Response: 392347777
Conc: 267.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#20 Methoxychlor

R.T.: 6.682 min
Delta R.T.: 0.000 min
Response: 694139453
Conc: 253.29 ng/ml




#21 Endrin ketone

R.T.: 7.598 min
Delta R.T.: 0.000 min
Response: 2324768
Conc: 0.77 ng/ml m



#21 Endrin ketone

R.T.: 6.905 min
Delta R.T.: -0.001 min
Response: 3361038
Conc: 0.60 ng/ml m




#28 Decachlorobiphenyl

R.T.: 9.017 min
Delta R.T.: 0.000 min
Response: 50725219
Conc: 21.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025



#28 Decachlorobiphenyl

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 90556842
Conc: 20.87 ng/ml

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/28/2025 07/28/2025

Client Sample No. (PEM): PEM - PL096818.D Date Analyzed: 08/15/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:13

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.004	8.900	9.100	20.950	20.000	4.8
Tetrachloro-m-xylene	3.531	3.480	3.580	20.800	20.000	4.0
alpha-BHC	3.978	3.930	4.030	10.310	10.000	3.1
beta-BHC	4.493	4.440	4.540	10.430	10.000	4.3
gamma-BHC (Lindane)	4.306	4.260	4.360	10.580	10.000	5.8
Endrin	6.534	6.460	6.600	54.950	50.000	9.9
4,4'-DDT	6.983	6.910	7.050	107.280	100.000	7.3
Methoxychlor	7.455	7.380	7.530	264.500	250.000	5.8

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/28/2025 07/28/2025

Client Sample No. (PEM): PEM - PL096818.D Date Analyzed: 08/15/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:13

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.987	7.890	8.090	19.750	20.000	-1.3
Tetrachloro-m-xylene	2.826	2.780	2.880	20.270	20.000	1.4
alpha-BHC	3.330	3.280	3.380	9.730	10.000	-2.7
beta-BHC	3.959	3.910	4.010	10.680	10.000	6.8
gamma-BHC (Lindane)	3.662	3.610	3.710	9.790	10.000	-2.1
Endrin	5.706	5.640	5.780	49.720	50.000	-0.6
4,4'-DDT	6.105	6.030	6.180	103.400	100.000	3.4
Methoxychlor	6.676	6.610	6.750	224.750	250.000	-10.1

Data File: PEM
 PL096818.D **Date Acquired** 8/15/2025 10:13
Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.53	166315119.2	172620555	6305435.76	3.65
Endrin aldehyde	6.89	4383906.726			
Endrin ketone	7.59	1921529.035			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.71	268809359.4	284617085.2	15807725.8	5.55
Endrin aldehyde #2	6.17	14217820.72			
Endrin ketone #2	6.90	1589905.098			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	6.98	307659279.4	308585833.3	926553.893	0.30
4,4'-DDE	0.00	0			
4,4'-DDD	6.67	926553.893			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.11	523036482.6	525352443.4	2315960.8	0.44
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.86	2315960.8			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096818.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 10:13
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 08/20/2025
 Supervised By : mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:54:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.826	66146774	96842073	20.796	20.269
28)	SA Decachlor...	9.004	7.987	49966203	85675813	20.953	19.749
Target Compounds							
2)	A alpha-BHC	3.978	3.330	47707827	68862700	10.312	9.730
3)	MA gamma-BHC...	4.306	3.662	46816462	64625617	10.584	9.787
6)	B beta-BHC	4.493	3.959	18826992	30137711	10.431	10.684
14)	MA Endrin	6.534	5.706	166.3E6	268.8E6	54.948m	49.725
16)	A 4,4'-DDD	6.672	5.857	926554	2315961	0.366m	0.492m#
17)	MA 4,4'-DDT	6.983	6.105	307.7E6	523.0E6	107.285	103.402
18)	B Endrin al...	6.885	6.171	4383907	14217821	2.043m	3.911m#
20)	A Methoxychlor	7.455	6.676	388.4E6	615.9E6	264.496	224.748m
21)	B Endrin ke...	7.589	6.902	1921529	1589905	0.639m	0.285m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096818.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 10:13
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:54:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 66146774
Conc: 20.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 96842073
Conc: 20.27 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 47707827
Conc: 10.31 ng/ml



#2 alpha-BHC

R.T.: 3.330 min
Delta R.T.: -0.003 min
Response: 68862700
Conc: 9.73 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.306 min
Delta R.T.: -0.005 min
Response: 46816462
Conc: 10.58 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 64625617
Conc: 9.79 ng/ml




#6 beta-BHC

R.T.: 4.493 min
Delta R.T.: -0.004 min
Response: 18826992
Conc: 10.43 ng/ml



#6 beta-BHC

R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 30137711
Conc: 10.68 ng/ml




#14 Endrin

R.T.: 6.534 min
Delta R.T.: -0.009 min
Response: 166315119
Conc: 54.95 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#14 Endrin

R.T.: 5.706 min
Delta R.T.: -0.003 min
Response: 268809359
Conc: 49.72 ng/ml




#16 4,4'-DDD

R.T.: 6.672 min
Delta R.T.: -0.003 min
Response: 926554
Conc: 0.37 ng/ml m



#16 4,4'-DDD

R.T.: 5.857 min
Delta R.T.: 0.001 min
Response: 2315961
Conc: 0.49 ng/ml m




#17 4,4'-DDT

R.T.: 6.983 min
Delta R.T.: -0.006 min
Response: 307659279
Conc: 107.28 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 523036483
Conc: 103.40 ng/ml




#18 Endrin aldehyde

R.T.: 6.885 min
Delta R.T.: 0.001 min
Response: 4383907
Conc: 2.04 ng/ml m



#18 Endrin aldehyde

R.T.: 6.171 min
Delta R.T.: -0.008 min
Response: 14217821
Conc: 3.91 ng/ml m




#20 Methoxychlor

R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 388373800
Conc: 264.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#20 Methoxychlor

R.T.: 6.676 min
Delta R.T.: -0.005 min
Response: 615921683
Conc: 224.75 ng/ml m




#21 Endrin ketone

R.T.: 7.589 min
Delta R.T.: -0.008 min
Response: 1921529
Conc: 0.64 ng/ml m



#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 1589905
Conc: 0.29 ng/ml m




#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.011 min
Response: 49966203
Conc: 20.95 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 85675813
Conc: 19.75 ng/ml

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/28/2025 07/28/2025

Client Sample No. (PEM): PEM - PL096836.D Date Analyzed: 08/15/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.004	8.900	9.100	21.430	20.000	7.2
Tetrachloro-m-xylene	3.530	3.480	3.580	21.810	20.000	9.1
alpha-BHC	3.977	3.930	4.030	10.480	10.000	4.8
beta-BHC	4.491	4.440	4.540	11.400	10.000	14.0
gamma-BHC (Lindane)	4.305	4.250	4.360	10.790	10.000	7.9
Endrin	6.534	6.460	6.600	55.930	50.000	11.9
4,4'-DDT	6.982	6.910	7.050	103.770	100.000	3.8
Methoxychlor	7.455	7.380	7.530	252.800	250.000	1.1

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/28/2025 07/28/2025

Client Sample No. (PEM): PEM - PL096836.D Date Analyzed: 08/15/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.986	7.890	8.090	21.270	20.000	6.4
Tetrachloro-m-xylene	2.825	2.770	2.880	21.120	20.000	5.6
alpha-BHC	3.331	3.280	3.380	10.120	10.000	1.2
beta-BHC	3.959	3.910	4.010	11.180	10.000	11.8
gamma-BHC (Lindane)	3.662	3.610	3.710	10.300	10.000	3.0
Endrin	5.705	5.630	5.780	51.560	50.000	3.1
4,4'-DDT	6.105	6.030	6.180	98.010	100.000	-2.0
Methoxychlor	6.675	6.600	6.750	224.400	250.000	-10.2

Data File: PEM
 PL096836.D **Date Acquired** 8/15/2025 20:07
Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.53	169273585.4	178359859.4	9086274.03	Down 5.09
Endrin aldehyde	6.88	3224394.625			
Endrin ketone	7.59	5861879.407			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.71	278722025.6	298422781.7	19700756	6.60
Endrin aldehyde #2	6.17	12380464.72			
Endrin ketone #2	6.90	7320291.325			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	6.98	297579199	306434573.7	8855374.64	2.89
4,4'-DDE	0.00	0			
4,4'-DDD	6.67	8855374.638			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.10	495753031.5	507492446.8	11739415.3	2.31
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.85	11739415.3			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096836.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 20:07
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 08/20/2025
 Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:56:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.825	69382655	100.9E6	21.813	21.115
28)	SA Decachlor...	9.004	7.986	51106982	92284478	21.432	21.272
Target Compounds							
2)	A alpha-BHC	3.977	3.331	48480683	71648677	10.479	10.124
3)	MA gamma-BHC...	4.305	3.662	47716628	67995426	10.787	10.297
6)	B beta-BHC	4.491	3.959	20578100	31540061	11.401m	11.181
14)	MA Endrin	6.534	5.705	169.3E6	278.7E6	55.925m	51.558
16)	A 4,4'-DDD	6.666	5.852	8855375	11739415	3.502m	2.495 #
17)	MA 4,4'-DDT	6.982	6.105	297.6E6	495.8E6	103.770	98.008
18)	B Endrin al...	6.880	6.172	3224395	12380465	1.503m	3.406m#
20)	A Methoxychlor	7.455	6.675	371.2E6	615.0E6	252.799	224.401m
21)	B Endrin ke...	7.590	6.901	5861879	7320291	1.950	1.314m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096836.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 20:07
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:56:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 69382655
Conc: 21.81 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 100884757
Conc: 21.12 ng/ml




#2 alpha-BHC

R.T.: 3.977 min
Delta R.T.: -0.005 min
Response: 48480683
Conc: 10.48 ng/ml



#2 alpha-BHC

R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 71648677
Conc: 10.12 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.305 min
Delta R.T.: -0.005 min
Response: 47716628
Conc: 10.79 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 67995426
Conc: 10.30 ng/ml




#6 beta-BHC

R.T.: 4.491 min
Delta R.T.: -0.006 min
Response: 20578100
Conc: 11.40 ng/ml m



#6 beta-BHC

R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 31540061
Conc: 11.18 ng/ml




#14 Endrin

R.T.: 6.534 min
Delta R.T.: -0.009 min
Response: 169273585
Conc: 55.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#14 Endrin

R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 278722026
Conc: 51.56 ng/ml




#16 4,4'-DDD

R.T.: 6.666 min
Delta R.T.: -0.009 min
Response: 8855375
Conc: 3.50 ng/ml m



#16 4,4'-DDD

R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 11739415
Conc: 2.49 ng/ml




#17 4,4'-DDT

R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 297579199
Conc: 103.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 495753032
Conc: 98.01 ng/ml




#18 Endrin aldehyde

R.T.: 6.880 min
Delta R.T.: -0.004 min
Response: 3224395
Conc: 1.50 ng/ml m



#18 Endrin aldehyde

R.T.: 6.172 min
Delta R.T.: -0.007 min
Response: 12380465
Conc: 3.41 ng/ml m




#20 Methoxychlor

R.T.: 7.455 min
Delta R.T.: -0.008 min
Response: 371198217
Conc: 252.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#20 Methoxychlor

R.T.: 6.675 min
Delta R.T.: -0.006 min
Response: 614969762
Conc: 224.40 ng/ml m




#21 Endrin ketone

R.T.: 7.590 min
Delta R.T.: -0.007 min
Response: 5861879
Conc: 1.95 ng/ml



#21 Endrin ketone

R.T.: 6.901 min
Delta R.T.: -0.006 min
Response: 7320291
Conc: 1.31 ng/ml m




#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.012 min
Response: 51106982
Conc: 21.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 92284478
Conc: 21.27 ng/ml

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/28/2025 07/28/2025

Client Sample No. (PEM): PEM - PL096844.D Date Analyzed: 08/18/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:26

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.006	8.910	9.110	22.690	20.000	13.5
Tetrachloro-m-xylene	3.531	3.480	3.580	23.150	20.000	15.8
alpha-BHC	3.976	3.930	4.030	11.470	10.000	14.7
beta-BHC	4.491	4.440	4.540	12.270	10.000	22.7
gamma-BHC (Lindane)	4.304	4.250	4.350	11.570	10.000	15.7
Endrin	6.535	6.460	6.610	56.720	50.000	13.4
4,4'-DDT	6.983	6.910	7.050	107.520	100.000	7.5
Methoxychlor	7.455	7.380	7.530	255.160	250.000	2.1

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/28/2025 07/28/2025

Client Sample No. (PEM): PEM - PL096844.D Date Analyzed: 08/18/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:26

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.987	7.890	8.090	21.760	20.000	8.8
Tetrachloro-m-xylene	2.825	2.770	2.880	21.850	20.000	9.3
alpha-BHC	3.330	3.280	3.380	10.540	10.000	5.4
beta-BHC	3.958	3.910	4.010	11.590	10.000	15.9
gamma-BHC (Lindane)	3.662	3.610	3.710	10.710	10.000	7.1
Endrin	5.705	5.630	5.780	50.230	50.000	0.5
4,4'-DDT	6.105	6.030	6.180	101.060	100.000	1.1
Methoxychlor	6.676	6.610	6.750	222.840	250.000	-10.9

Data File: PEM
 PL096844.D **Date Acquired** 8/18/2025 10:26
Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.53	171673272.2	184182065.5	12508793.4	Down 6.79
Endrin aldehyde	6.88	3600201.624			
Endrin ketone	7.59	8908591.736			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.71	271542500.1	307658534.2	36116034.2	11.74
Endrin aldehyde #2	6.18	14799157.79			
Endrin ketone #2	6.90	21316876.39			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	6.98	308321781.4	317657653.8	9335872.4	2.94
4,4'-DDE	0.00	0			
4,4'-DDD	6.67	9335872.397			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.10	511212305.6	528753740.4	17541434.8	3.32
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.85	17541434.78			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
 Data File : PL096844.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2025 10:26
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 08/19/2025
 Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 12:38:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.531	2.825	73626080	104.4E6	23.147	21.848
28) SA Decachlor...	9.006	7.987	54106762	94409704	22.690	21.762
Target Compounds						
2) A alpha-BHC	3.976	3.330	53077849	74616765	11.473m	10.544
3) MA gamma-BHC...	4.304	3.662	51157442	70745882	11.565m	10.714
6) B beta-BHC	4.491	3.958	22153696	32707689	12.274m	11.595
14) MA Endrin	6.535	5.705	171.7E6	271.5E6	56.718m	50.230
16) A 4,4'-DDD	6.669	5.852	9335872	17541435	3.692	3.728
17) MA 4,4'-DDT	6.983	6.105	308.3E6	511.2E6	107.516	101.064
18) B Endrin al...	6.878	6.175	3600202	14799158	1.678m	4.071 #
20) A Methoxychlor	7.455	6.676	374.7E6	610.7E6	255.163	222.842m
21) B Endrin ke...	7.590	6.902	8908592	21316876	2.963	3.827 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096844.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 10:26
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 12:38:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 73626080
Conc: 23.15 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 104384511
Conc: 21.85 ng/ml




#2 alpha-BHC

R.T.: 3.976 min
Delta R.T.: -0.006 min
Response: 53077849
Conc: 11.47 ng/ml m



#2 alpha-BHC

R.T.: 3.330 min
Delta R.T.: -0.003 min
Response: 74616765
Conc: 10.54 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.304 min
Delta R.T.: -0.006 min
Response: 51157442
Conc: 11.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 70745882
Conc: 10.71 ng/ml




#6 beta-BHC

R.T.: 4.491 min
Delta R.T.: -0.006 min
Response: 22153696
Conc: 12.27 ng/ml m



#6 beta-BHC

R.T.: 3.958 min
Delta R.T.: -0.003 min
Response: 32707689
Conc: 11.59 ng/ml




#14 Endrin

R.T.: 6.535 min
Delta R.T.: -0.008 min
Response: 171673272
Conc: 56.72 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#14 Endrin

R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 271542500
Conc: 50.23 ng/ml




#16 4,4'-DDD

R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 9335872
Conc: 3.69 ng/ml



#16 4,4'-DDD

R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 17541435
Conc: 3.73 ng/ml




#17 4,4'-DDT

R.T.: 6.983 min
Delta R.T.: -0.006 min
Response: 308321781
Conc: 107.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 511212306
Conc: 101.06 ng/ml




#18 Endrin aldehyde

R.T.: 6.878 min
Delta R.T.: -0.006 min
Response: 3600202
Conc: 1.68 ng/ml m



#18 Endrin aldehyde

R.T.: 6.175 min
Delta R.T.: -0.004 min
Response: 14799158
Conc: 4.07 ng/ml




#20 Methoxychlor

R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 374669518
Conc: 255.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#20 Methoxychlor

R.T.: 6.676 min
Delta R.T.: -0.005 min
Response: 610697036
Conc: 222.84 ng/ml m




#21 Endrin ketone

R.T.: 7.590 min
Delta R.T.: -0.006 min
Response: 8908592
Conc: 2.96 ng/ml



#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 21316876
Conc: 3.83 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.006 min
Delta R.T.: -0.009 min
Response: 54106762
Conc: 22.69 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 94409704
Conc: 21.76 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096595.D
Acq On : 28 Jul 2025 16:38
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e

Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Title : GC Extractables
Last Update : Tue Jul 29 08:58:13 2025
Integrator: ChemStation

RT#1	RT#2	Resolution
3.535	5.916	100.00%
5.916	6.045	100.00%
6.045	6.166	100.00%
6.166	6.317	100.00%
6.317	7.118	100.00%
7.118	7.463	100.00%
7.463	7.598	100.00%
7.598	9.016	100.00%

Signal #2

2.828	5.052	100.00%
5.052	5.170	100.00%
5.170	5.304	100.00%
5.304	5.435	100.00%
5.435	6.403	100.00%
6.403	6.682	100.00%
6.682	6.907	100.00%
6.907	7.993	100.00%

PL072825.M Tue Jul 29 09:06:52 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 16:38
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 RESCHK

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/29/2025
 Supervised By :mohammad ahmed 07/30/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 08:57:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 08:58:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.535	2.828	53837096	81066871	16.926	16.968
28) SA Decachlor...	9.016	7.993	41967810	75251539	17.599	17.346
Target Compounds						
9) A Endosulfan I	6.045	5.169	28022040	44379475	7.792	8.023m
10) B gamma-Chl...	5.916	5.052	31638592	48929496	8.314	8.321
12) B 4,4'-DDE	6.166	5.304	54660848	96001382	17.022	17.409
13) MA Dieldrin	6.316	5.435	63095073	99918926	17.001	16.917
19) B Endosulfa...	7.118	6.403	50325612	88476396	17.513	17.395
20) A Methoxychlor	7.463	6.682	128.9E6	238.4E6	87.784	87.000
21) B Endrin ke...	7.598	6.907	51693529	95186865	17.193	17.090

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096595.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 16:38
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 08:57:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 08:58:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 53837096
Conc: 16.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 81066871
Conc: 16.97 ng/ml




#9 Endosulfan I

R.T.: 6.045 min
Delta R.T.: 0.001 min
Response: 28022040
Conc: 7.79 ng/ml



#9 Endosulfan I

R.T.: 5.169 min
Delta R.T.: 0.000 min
Response: 44379475
Conc: 8.02 ng/ml m




#10 gamma-Chlordane

R.T.: 5.916 min
Delta R.T.: 0.000 min
Response: 31638592
Conc: 8.31 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#10 gamma-Chlordane

R.T.: 5.052 min
Delta R.T.: 0.000 min
Response: 48929496
Conc: 8.32 ng/ml




#12 4,4'-DDE

R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 54660848
Conc: 17.02 ng/ml



#12 4,4'-DDE

R.T.: 5.304 min
Delta R.T.: 0.000 min
Response: 96001382
Conc: 17.41 ng/ml




#13 Dieldrin

R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 63095073
Conc: 17.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#13 Dieldrin

R.T.: 5.435 min
Delta R.T.: 0.000 min
Response: 99918926
Conc: 16.92 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.118 min
Delta R.T.: 0.000 min
Response: 50325612
Conc: 17.51 ng/ml



#19 Endosulfan Sulfate

R.T.: 6.403 min
Delta R.T.: 0.001 min
Response: 88476396
Conc: 17.40 ng/ml



#20 Methoxychlor

R.T.: 7.463 min
Delta R.T.: 0.000 min
Response: 128897891
Conc: 87.78 ng/ml


Instrument :

ECD_L

ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/29/2025
Supervised By :mohammad ahmed 07/30/2025




#20 Methoxychlor

R.T.: 6.682 min
Delta R.T.: 0.000 min
Response: 238424041
Conc: 87.00 ng/ml




#21 Endrin ketone

R.T.: 7.598 min
Delta R.T.: 0.000 min
Response: 51693529
Conc: 17.19 ng/ml



#21 Endrin ketone

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 95186865
Conc: 17.09 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.016 min
Delta R.T.: 0.000 min
Response: 41967810
Conc: 17.60 ng/ml

Instrument :

ECD_L

ClientSampleId :


RESCHK

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 07/29/2025

Supervised By :mohammad ahmed 07/30/2025



#28 Decachlorobiphenyl

R.T.: 7.993 min
Delta R.T.: 0.000 min
Response: 75251539
Conc: 17.35 ng/ml

Analytical Sequence

Client: First Environment, Inc.

SDG No.: Q2815

Project: USACE018-44 DOD

Instrument ID: ECD_L

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 07/28/2025

07/28/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	07/28/2025	16:11	PL096593.D	9.02	3.54
PEM	PEM	07/28/2025	16:25	PL096594.D	9.02	3.54
RESCHK	RESCHK	07/28/2025	16:38	PL096595.D	9.02	3.54
PSTDICC100	PSTDICC100	07/28/2025	16:52	PL096596.D	9.02	3.54
PSTDICC075	PSTDICC075	07/28/2025	17:06	PL096597.D	9.02	3.54
PSTDICC050	PSTDICC050	07/28/2025	17:19	PL096598.D	9.02	3.54
PSTDICC025	PSTDICC025	07/28/2025	17:33	PL096599.D	9.02	3.54
PSTDICC005	PSTDICC005	07/28/2025	17:47	PL096600.D	9.02	3.53
PCHLORICC500	PCHLORICC500	07/28/2025	18:27	PL096603.D	9.02	3.54
PTOXICC500	PTOXICC500	07/28/2025	19:36	PL096608.D	9.02	3.54
IBLK	IBLK	08/15/2025	09:11	PL096817.D	9.01	3.53
PEM	PEM	08/15/2025	10:13	PL096818.D	9.00	3.53
PSTDCCC050	PSTDCCC050	08/15/2025	10:58	PL096819.D	9.01	3.53
PB169225BL	PB169225BL	08/15/2025	16:07	PL096822.D	9.00	3.53
PB169225BS	PB169225BS	08/15/2025	16:21	PL096823.D	9.00	3.53
PB169225BSD	PB169225BSD	08/15/2025	16:34	PL096824.D	9.00	3.53
IBLK	IBLK	08/15/2025	17:51	PL096827.D	9.00	3.53
PSTDCCC050	PSTDCCC050	08/15/2025	18:05	PL096828.D	9.00	3.53
IBLK	IBLK	08/15/2025	19:54	PL096835.D	9.00	3.53
PEM	PEM	08/15/2025	20:07	PL096836.D	9.00	3.53
PSTDCCC050	PSTDCCC050	08/15/2025	20:48	PL096837.D	9.00	3.53
TW-22M-W	Q2815-11	08/15/2025	21:15	PL096839.D	9.00	3.53
IBLK	IBLK	08/15/2025	21:56	PL096840.D	9.00	3.53
PSTDCCC050	PSTDCCC050	08/15/2025	22:10	PL096841.D	9.00	3.53
PEM	PEM	08/18/2025	10:26	PL096844.D	9.01	3.53
IBLK	IBLK	08/18/2025	12:39	PL096849.D	9.01	3.53
PSTDCCC050	PSTDCCC050	08/18/2025	12:53	PL096850.D	9.00	3.53
TW-705R-S	Q2815-01	08/18/2025	16:30	PL096855.D	9.01	3.53
TW-705R-SDL	Q2815-01DL	08/18/2025	16:44	PL096856.D	9.01	3.53
IBLK	IBLK	08/18/2025	16:57	PL096857.D	9.00	3.53
PSTDCCC050	PSTDCCC050	08/18/2025	17:11	PL096858.D	9.00	3.53

Analytical Sequence

Client: First Environment, Inc.

SDG No.: Q2815

Project: USACE018-44 DOD

Instrument ID: ECD_L

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 07/28/2025

07/28/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	07/28/2025	16:11	PL096593.D	7.99	2.83
PEM	PEM	07/28/2025	16:25	PL096594.D	7.99	2.83
RESCHK	RESCHK	07/28/2025	16:38	PL096595.D	7.99	2.83
PSTDICC100	PSTDICC100	07/28/2025	16:52	PL096596.D	7.99	2.83
PSTDICC075	PSTDICC075	07/28/2025	17:06	PL096597.D	7.99	2.83
PSTDICC050	PSTDICC050	07/28/2025	17:19	PL096598.D	7.99	2.83
PSTDICC025	PSTDICC025	07/28/2025	17:33	PL096599.D	7.99	2.83
PSTDICC005	PSTDICC005	07/28/2025	17:47	PL096600.D	7.99	2.83
PCHLORICC500	PCHLORICC500	07/28/2025	18:27	PL096603.D	7.99	2.83
PTOXICC500	PTOXICC500	07/28/2025	19:36	PL096608.D	7.99	2.83
IBLK	IBLK	08/15/2025	09:11	PL096817.D	7.99	2.82
PEM	PEM	08/15/2025	10:13	PL096818.D	7.99	2.83
PSTDCCC050	PSTDCCC050	08/15/2025	10:58	PL096819.D	7.99	2.82
PB169225BL	PB169225BL	08/15/2025	16:07	PL096822.D	7.99	2.82
PB169225BS	PB169225BS	08/15/2025	16:21	PL096823.D	7.99	2.83
PB169225BSD	PB169225BSD	08/15/2025	16:34	PL096824.D	7.99	2.83
IBLK	IBLK	08/15/2025	17:51	PL096827.D	7.99	2.82
PSTDCCC050	PSTDCCC050	08/15/2025	18:05	PL096828.D	7.99	2.83
IBLK	IBLK	08/15/2025	19:54	PL096835.D	7.99	2.83
PEM	PEM	08/15/2025	20:07	PL096836.D	7.99	2.83
PSTDCCC050	PSTDCCC050	08/15/2025	20:48	PL096837.D	7.99	2.83
TW-22M-W	Q2815-11	08/15/2025	21:15	PL096839.D	7.99	2.83
IBLK	IBLK	08/15/2025	21:56	PL096840.D	7.99	2.83
PSTDCCC050	PSTDCCC050	08/15/2025	22:10	PL096841.D	7.99	2.83
PEM	PEM	08/18/2025	10:26	PL096844.D	7.99	2.83
IBLK	IBLK	08/18/2025	12:39	PL096849.D	7.99	2.82
PSTDCCC050	PSTDCCC050	08/18/2025	12:53	PL096850.D	7.99	2.83
TW-705R-S	Q2815-01	08/18/2025	16:30	PL096855.D	7.99	2.82
TW-705R-SDL	Q2815-01DL	08/18/2025	16:44	PL096856.D	7.99	2.83
IBLK	IBLK	08/18/2025	16:57	PL096857.D	7.98	2.83
PSTDCCC050	PSTDCCC050	08/18/2025	17:11	PL096858.D	7.98	2.83

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB169225BS

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Lab Sample ID: PB169225BS

Date(s) Analyzed: 08/15/2025 08/15/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.67	6.62	6.72	0.50	1.7
	2	5.85	5.80	5.90	0.49	
4,4'-DDE	1	6.16	6.11	6.21	0.49	0.9
	2	5.30	5.25	5.35	0.49	
4,4'-DDT	1	6.98	6.93	7.03	0.50	2.2
	2	6.10	6.05	6.15	0.49	
alpha-BHC	1	3.98	3.93	4.03	0.50	0.1
	2	3.33	3.28	3.38	0.50	
Aldrin	1	5.24	5.19	5.29	0.49	0.6
	2	4.29	4.24	4.34	0.50	
alpha-Chlordane	1	5.99	5.94	6.04	0.50	2.6
	2	5.11	5.06	5.16	0.49	
Endosulfan II	1	6.75	6.70	6.80	0.51	4.1
	2	6.00	5.95	6.05	0.49	
Endosulfan sulfate	1	7.11	7.06	7.16	0.48	0.8
	2	6.40	6.35	6.45	0.48	
beta-BHC	1	4.49	4.44	4.54	0.50	0.8
	2	3.96	3.91	4.01	0.50	
delta-BHC	1	4.74	4.69	4.79	0.50	0.5
	2	4.19	4.14	4.24	0.50	
Endosulfan I	1	6.04	5.99	6.09	0.49	2.9
	2	5.17	5.12	5.22	0.48	
Dieldrin	1	6.31	6.26	6.36	0.49	0.3
	2	5.43	5.38	5.48	0.49	
Endrin aldehyde	1	6.88	6.83	6.93	0.53	3.3
	2	6.17	6.12	6.22	0.55	
Methoxychlor	1	7.45	7.40	7.50	0.49	0.6
	2	6.68	6.63	6.73	0.48	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB169225BS

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Lab Sample ID: PB169225BS

Date(s) Analyzed: 08/15/2025 08/15/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin ketone	1	7.59	7.54	7.64	0.49	5.6
	2	6.90	6.85	6.95	0.52	
gamma-BHC (Lindane)	1	4.30	4.25	4.35	0.50	1
	2	3.66	3.61	3.71	0.50	
Heptachlor	1	4.90	4.85	4.95	0.53	7.6
	2	4.01	3.96	4.06	0.49	
Heptachlor epoxide	1	5.66	5.61	5.71	0.51	3.1
	2	4.80	4.75	4.85	0.49	
gamma-Chlordane	1	5.91	5.86	5.96	0.50	1.4
	2	5.05	5.00	5.10	0.50	
Endrin	1	6.53	6.48	6.58	0.51	4.8
	2	5.71	5.66	5.76	0.48	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB169225BSD

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Lab Sample ID: PB169225BSD

Date(s) Analyzed: 08/15/2025

08/15/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan sulfate	1	7.11	7.06	7.16	0.48	0.1
	2	6.40	6.35	6.45	0.47	
Endrin ketone	1	7.59	7.54	7.64	0.48	2.5
	2	6.90	6.85	6.95	0.49	
alpha-BHC	1	3.98	3.93	4.03	0.48	0.4
	2	3.33	3.28	3.38	0.48	
Aldrin	1	5.24	5.19	5.29	0.48	0
	2	4.29	4.24	4.34	0.48	
beta-BHC	1	4.49	4.44	4.54	0.49	0.8
	2	3.96	3.91	4.01	0.48	
delta-BHC	1	4.74	4.69	4.79	0.49	1.1
	2	4.19	4.14	4.24	0.48	
Endosulfan I	1	6.04	5.99	6.09	0.48	4.5
	2	5.17	5.12	5.22	0.46	
alpha-Chlordane	1	5.99	5.94	6.04	0.49	2.8
	2	5.11	5.06	5.16	0.47	
4,4'-DDE	1	6.16	6.11	6.21	0.50	3
	2	5.30	5.25	5.35	0.48	
Dieldrin	1	6.31	6.26	6.36	0.49	0.9
	2	5.43	5.38	5.48	0.48	
Endrin	1	6.53	6.48	6.58	0.49	4.9
	2	5.71	5.66	5.76	0.47	
Methoxychlor	1	7.46	7.41	7.51	0.47	1.6
	2	6.68	6.63	6.73	0.47	
gamma-BHC (Lindane)	1	4.31	4.26	4.36	0.48	1.9
	2	3.66	3.61	3.71	0.48	
Heptachlor	1	4.90	4.85	4.95	0.52	8.1
	2	4.01	3.96	4.06	0.48	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB169225BSD

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Lab Sample ID: PB169225BSD

Date(s) Analyzed: 08/15/2025 08/15/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Heptachlor epoxide	1	5.66	5.61	5.71	0.49	3.2
	2	4.80	4.75	4.85	0.48	
gamma-Chlordane	1	5.91	5.86	5.96	0.49	1.8
	2	5.05	5.00	5.10	0.48	
Endosulfan II	1	6.75	6.70	6.80	0.51	6.3
	2	6.00	5.95	6.05	0.48	
4,4'-DDD	1	6.67	6.62	6.72	0.49	1.9
	2	5.85	5.80	5.90	0.48	
4,4'-DDT	1	6.98	6.93	7.03	0.48	2.4
	2	6.11	6.06	6.16	0.47	
Endrin aldehyde	1	6.88	6.83	6.93	0.52	3.5
	2	6.18	6.13	6.23	0.54	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TW-22M-W

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Lab Sample ID: Q2815-11

Date(s) Analyzed: 08/15/2025 08/15/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDT	1	6.98	6.93	7.03	0.027	83.4
	2	6.12	6.07	6.17	0.065	
alpha-Chlordane	1	6.00	5.95	6.05	0.016	2.5
	2	5.10	5.05	5.15	0.016	
Dieldrin	1	6.31	6.26	6.36	0.0084	10.2
	2	5.43	5.38	5.48	0.0093	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TW-705R-S

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Lab Sample ID: Q2815-01

Date(s) Analyzed: 08/18/2025 08/18/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.67	6.62	6.72	1.60	13.3
	2	5.85	5.80	5.90	1.40	
4,4'-DDE	1	6.16	6.11	6.21	0.54	17.1
	2	5.30	5.25	5.35	0.46	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TW-705R-SDL

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Lab Sample ID: Q2815-01DL

Date(s) Analyzed: 08/18/2025 08/18/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.67	6.62	6.72	1.50	6.9
	2	5.85	5.80	5.90	1.40	
4,4'-DDE	1	6.16	6.11	6.21	0.54	19.6
	2	5.30	5.25	5.35	0.45	



QC SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	USACE018-44 DOD	Date Received:	
Client Sample ID:	PB169225BL	SDG No.:	Q2815
Lab Sample ID:	PB169225BL	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096822.D	1	08/12/25 10:20	08/15/25 16:07	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.010	U	0.0039	0.010	0.050	ug/L
319-85-7	beta-BHC	0.010	U	0.0049	0.010	0.050	ug/L
319-86-8	delta-BHC	0.025	U	0.011	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.010	U	0.0037	0.010	0.050	ug/L
76-44-8	Heptachlor	0.010	U	0.0027	0.010	0.050	ug/L
309-00-2	Aldrin	0.010	U	0.0036	0.010	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0096	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.010	U	0.0031	0.010	0.050	ug/L
60-57-1	Dieldrin	0.010	U	0.0036	0.010	0.050	ug/L
72-55-9	4,4-DDE	0.010	U	0.0037	0.010	0.050	ug/L
72-20-8	Endrin	0.010	U	0.0032	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0079	0.025	0.050	ug/L
72-54-8	4,4-DDD	0.025	U	0.0071	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.010	U	0.0037	0.010	0.050	ug/L
50-29-3	4,4-DDT	0.010	U	0.0035	0.010	0.050	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0093	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.011	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.010	U	0.0035	0.010	0.050	ug/L
5103-74-2	gamma-Chlordane	0.010	U	0.0039	0.010	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.17	0.50	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	20.6		30 - 135		103%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.7		44 - 124		99%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	USACE018-44 DOD	Date Received:	
Client Sample ID:	PB169225BL	SDG No.:	Q2815
Lab Sample ID:	PB169225BL	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096822.D	1	08/12/25 10:20	08/15/25 16:07	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096822.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 16:07
Operator : AR\AJ
Sample : PB169225BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB169225BL

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.530	2.824	62200378	94174771	19.555	19.711
28) SA Decachlor...	9.004	7.987	48781039	89293709	20.456m	20.583

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096822.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 16:07
Operator : AR\AJ
Sample : PB169225BL
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PB169225BL

Manual Integrations
APPROVED


Reviewed By : Abdul Mirza 08/18/2025
Supervised By : mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Response_ Signal: PL096822.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 62200378
Conc: 19.56 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BL

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025


Time Response_ Signal: PL096822.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: -0.004 min
Response: 94174771
Conc: 19.71 ng/ml


Response_ Signal: PL096822.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.012 min
Response: 48781039
Conc: 20.46 ng/ml m

Time Response_ Signal: PL096822.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 89293709
Conc: 20.58 ng/ml

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	07/28/25
Project:	USACE018-44 DOD	Date Received:	07/28/25
Client Sample ID:	PIBLK-PL096593.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096593.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096593.D	1		07/28/25	pl072825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.010	U	0.0039	0.010	0.050	ug/L
319-85-7	beta-BHC	0.010	U	0.0049	0.010	0.050	ug/L
319-86-8	delta-BHC	0.025	U	0.011	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.010	U	0.0037	0.010	0.050	ug/L
76-44-8	Heptachlor	0.010	U	0.0027	0.010	0.050	ug/L
309-00-2	Aldrin	0.010	U	0.0036	0.010	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0096	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.010	U	0.0031	0.010	0.050	ug/L
60-57-1	Dieldrin	0.010	U	0.0036	0.010	0.050	ug/L
72-55-9	4,4-DDE	0.010	U	0.0037	0.010	0.050	ug/L
72-20-8	Endrin	0.010	U	0.0032	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0079	0.025	0.050	ug/L
72-54-8	4,4-DDD	0.025	U	0.0071	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.010	U	0.0037	0.010	0.050	ug/L
50-29-3	4,4-DDT	0.010	U	0.0035	0.010	0.050	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0093	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.011	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.010	U	0.0035	0.010	0.050	ug/L
5103-74-2	gamma-Chlordane	0.010	U	0.0039	0.010	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.17	0.50	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	19.2		30 - 135		96%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.6		44 - 124		93%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	07/28/25
Project:	USACE018-44 DOD	Date Received:	07/28/25
Client Sample ID:	PIBLK-PL096593.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096593.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Decanted:	
		Final Vol:	10000
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096593.D	1		07/28/25	pl072825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096593.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 16:11
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 08:56:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 08:58:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.828	59301554	87523738	18.644	18.319
2)	SA Decachlor...	9.016	7.993	45808507	82684633	19.210	19.059

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096593.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 16:11
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 08:56:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 08:58:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 59301554
Conc: 18.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 87523738
Conc: 18.32 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.016 min
Delta R.T.: 0.000 min
Response: 45808507
Conc: 19.21 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.993 min
Delta R.T.: 0.000 min
Response: 82684633
Conc: 19.06 ng/ml

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/15/25
Project:	USACE018-44 DOD	Date Received:	08/15/25
Client Sample ID:	PIBLK-PL096817.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096817.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096817.D	1		08/15/25	pl081625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.010	U	0.0039	0.010	0.050	ug/L
319-85-7	beta-BHC	0.010	U	0.0049	0.010	0.050	ug/L
319-86-8	delta-BHC	0.025	U	0.011	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.010	U	0.0037	0.010	0.050	ug/L
76-44-8	Heptachlor	0.010	U	0.0027	0.010	0.050	ug/L
309-00-2	Aldrin	0.010	U	0.0036	0.010	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0096	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.010	U	0.0031	0.010	0.050	ug/L
60-57-1	Dieldrin	0.010	U	0.0036	0.010	0.050	ug/L
72-55-9	4,4-DDE	0.010	U	0.0037	0.010	0.050	ug/L
72-20-8	Endrin	0.010	U	0.0032	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0079	0.025	0.050	ug/L
72-54-8	4,4-DDD	0.025	U	0.0071	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.010	U	0.0037	0.010	0.050	ug/L
50-29-3	4,4-DDT	0.010	U	0.0035	0.010	0.050	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0093	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.011	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.010	U	0.0035	0.010	0.050	ug/L
5103-74-2	gamma-Chlordane	0.010	U	0.0039	0.010	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.17	0.50	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	22.6		30 - 135		113%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.1		44 - 124		105%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/15/25
Project:	USACE018-44 DOD	Date Received:	08/15/25
Client Sample ID:	PIBLK-PL096817.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096817.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096817.D	1		08/15/25	pl081625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096817.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 09:11
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:54:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.529	2.824	67085722	95299446	21.091	19.946
28) SA Decachlor...	9.005	7.986	53916338	83831500	22.610m	19.323

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096817.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 09:11
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
I.BLK

Manual Integrations
APPROVED


Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:54:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Response_ Signal: PL096817.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 3.529 min
Delta R.T.: -0.006 min
Response: 67085722
Conc: 21.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025


Time Response_ Signal: PL096817.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: -0.004 min
Response: 95299446
Conc: 19.95 ng/ml


Time Response_ Signal: PL096817.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 9.005 min
Delta R.T.: -0.011 min
Response: 53916338
Conc: 22.61 ng/ml m

Time Response_ Signal: PL096817.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 83831500
Conc: 19.32 ng/ml

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/15/25
Project:	USACE018-44 DOD	Date Received:	08/15/25
Client Sample ID:	PIBLK-PL096827.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096827.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096827.D	1		08/15/25	pl081625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.010	U	0.0039	0.010	0.050	ug/L
319-85-7	beta-BHC	0.010	U	0.0049	0.010	0.050	ug/L
319-86-8	delta-BHC	0.025	U	0.011	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.010	U	0.0037	0.010	0.050	ug/L
76-44-8	Heptachlor	0.010	U	0.0027	0.010	0.050	ug/L
309-00-2	Aldrin	0.010	U	0.0036	0.010	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0096	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.010	U	0.0031	0.010	0.050	ug/L
60-57-1	Dieldrin	0.010	U	0.0036	0.010	0.050	ug/L
72-55-9	4,4-DDE	0.010	U	0.0037	0.010	0.050	ug/L
72-20-8	Endrin	0.010	U	0.0032	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0079	0.025	0.050	ug/L
72-54-8	4,4-DDD	0.025	U	0.0071	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.010	U	0.0037	0.010	0.050	ug/L
50-29-3	4,4-DDT	0.010	U	0.0035	0.010	0.050	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0093	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.011	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.010	U	0.0035	0.010	0.050	ug/L
5103-74-2	gamma-Chlordane	0.010	U	0.0039	0.010	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.17	0.50	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	22.1		30 - 135		110%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.6		44 - 124		103%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/15/25
Project:	USACE018-44 DOD	Date Received:	08/15/25
Client Sample ID:	PIBLK-PL096827.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096827.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096827.D	1		08/15/25	pl081625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096827.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 17:51
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.530	2.824	64368055	98181459	20.237	20.550
2) SA Decachlor...	9.004	7.987	52185238	95702144	21.884m	22.060

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096827.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 17:51
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
I.BLK

Manual Integrations
APPROVED


Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Response_ Signal: PL096827.D\ECD1A.ch



#1 Tetrachloro-m-xylene

R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 64368055
Conc: 20.24 ng/ml

Instrument :

ECD_L

ClientSampleId :

I.BLK


Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025


Time Response_ Signal: PL096827.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: -0.004 min
Response: 98181459
Conc: 20.55 ng/ml


Time Response_ Signal: PL096827.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.012 min
Response: 52185238
Conc: 21.88 ng/ml m

Time Response_ Signal: PL096827.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 95702144
Conc: 22.06 ng/ml

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/15/25
Project:	USACE018-44 DOD	Date Received:	08/15/25
Client Sample ID:	PIBLK-PL096835.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096835.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096835.D	1		08/15/25	pl081625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.010	U	0.0039	0.010	0.050	ug/L
319-85-7	beta-BHC	0.010	U	0.0049	0.010	0.050	ug/L
319-86-8	delta-BHC	0.025	U	0.011	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.010	U	0.0037	0.010	0.050	ug/L
76-44-8	Heptachlor	0.010	U	0.0027	0.010	0.050	ug/L
309-00-2	Aldrin	0.010	U	0.0036	0.010	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0096	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.010	U	0.0031	0.010	0.050	ug/L
60-57-1	Dieldrin	0.010	U	0.0036	0.010	0.050	ug/L
72-55-9	4,4-DDE	0.010	U	0.0037	0.010	0.050	ug/L
72-20-8	Endrin	0.010	U	0.0032	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0079	0.025	0.050	ug/L
72-54-8	4,4-DDD	0.025	U	0.0071	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.010	U	0.0037	0.010	0.050	ug/L
50-29-3	4,4-DDT	0.010	U	0.0035	0.010	0.050	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0093	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.011	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.010	U	0.0035	0.010	0.050	ug/L
5103-74-2	gamma-Chlordane	0.010	U	0.0039	0.010	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.17	0.50	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	21.5		30 - 135		107%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.5		44 - 124		102%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/15/25
Project:	USACE018-44 DOD	Date Received:	08/15/25
Client Sample ID:	PIBLK-PL096835.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096835.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096835.D	1		08/15/25	pl081625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096835.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 19:54
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:56:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.531	2.826	64836130	97845915	20.384	20.479
2) SA Decachlor...	9.002	7.986	51218863	90215642	21.478m	20.795

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096835.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 19:54
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
I.BLK

Manual Integrations
APPROVED


Reviewed By : Abdul Mirza 08/18/2025
Supervised By : mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:56:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Response_ Signal: PL096835.D\ECD1A.ch



#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 64836130
Conc: 20.38 ng/ml

Instrument :

ECD_L

ClientSampleId :

I.BLK


Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025


Time Response_ Signal: PL096835.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 97845915
Conc: 20.48 ng/ml


Response_ Signal: PL096835.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 9.002 min
Delta R.T.: -0.014 min
Response: 51218863
Conc: 21.48 ng/ml m

Time Response_ Signal: PL096835.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 90215642
Conc: 20.80 ng/ml

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/15/25
Project:	USACE018-44 DOD	Date Received:	08/15/25
Client Sample ID:	PIBLK-PL096840.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096840.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096840.D	1		08/15/25	pl081625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.010	U	0.0039	0.010	0.050	ug/L
319-85-7	beta-BHC	0.010	U	0.0049	0.010	0.050	ug/L
319-86-8	delta-BHC	0.025	U	0.011	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.010	U	0.0037	0.010	0.050	ug/L
76-44-8	Heptachlor	0.010	U	0.0027	0.010	0.050	ug/L
309-00-2	Aldrin	0.010	U	0.0036	0.010	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0096	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.010	U	0.0031	0.010	0.050	ug/L
60-57-1	Dieldrin	0.010	U	0.0036	0.010	0.050	ug/L
72-55-9	4,4-DDE	0.010	U	0.0037	0.010	0.050	ug/L
72-20-8	Endrin	0.010	U	0.0032	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0079	0.025	0.050	ug/L
72-54-8	4,4-DDD	0.025	U	0.0071	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.010	U	0.0037	0.010	0.050	ug/L
50-29-3	4,4-DDT	0.010	U	0.0035	0.010	0.050	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0093	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.011	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.010	U	0.0035	0.010	0.050	ug/L
5103-74-2	gamma-Chlordane	0.010	U	0.0039	0.010	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.17	0.50	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	22.8		30 - 135		114%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.7		44 - 124		103%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/15/25
Project:	USACE018-44 DOD	Date Received:	08/15/25
Client Sample ID:	PIBLK-PL096840.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096840.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096840.D	1		08/15/25	pl081625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096840.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 21:56
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:57:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.531	2.825	65743372	98589616	20.669	20.635
28) SA Decachlor...	9.002	7.986	54443636	95339152	22.831m	21.976

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096840.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 21:56
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 08/18/2025
Supervised By : mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:57:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Response_ Signal: PL096840.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 65743372
Conc: 20.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK

Manual Integrations
APPROVED


Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Time 3.20 3.30 3.40 3.50 3.60 3.70 3.80 3.90

Response_ Signal: PL096840.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 98589616
Conc: 20.64 ng/ml




Time 2.65 2.70 2.75 2.80 2.85 2.90 2.95 3.00

Response_ Signal: PL096840.D\ECD1A.ch

#28 Decachlorobiphenyl

R.T.: 9.002 min
Delta R.T.: -0.014 min
Response: 54443636
Conc: 22.83 ng/ml m




Time 8.80 8.90 9.00 9.10 9.20

Response_ Signal: PL096840.D\ECD2B.ch

#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 95339152
Conc: 21.98 ng/ml



Time 7.80 7.90 8.00 8.10 8.20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/18/25
Project:	USACE018-44 DOD	Date Received:	08/18/25
Client Sample ID:	PIBLK-PL096849.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096849.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096849.D	1		08/18/25	PL081825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.010	U	0.0039	0.010	0.050	ug/L
319-85-7	beta-BHC	0.010	U	0.0049	0.010	0.050	ug/L
319-86-8	delta-BHC	0.025	U	0.011	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.010	U	0.0037	0.010	0.050	ug/L
76-44-8	Heptachlor	0.010	U	0.0027	0.010	0.050	ug/L
309-00-2	Aldrin	0.010	U	0.0036	0.010	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0096	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.010	U	0.0031	0.010	0.050	ug/L
60-57-1	Dieldrin	0.010	U	0.0036	0.010	0.050	ug/L
72-55-9	4,4-DDE	0.010	U	0.0037	0.010	0.050	ug/L
72-20-8	Endrin	0.010	U	0.0032	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0079	0.025	0.050	ug/L
72-54-8	4,4-DDD	0.025	U	0.0071	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.010	U	0.0037	0.010	0.050	ug/L
50-29-3	4,4-DDT	0.010	U	0.0035	0.010	0.050	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0093	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.011	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.010	U	0.0035	0.010	0.050	ug/L
5103-74-2	gamma-Chlordane	0.010	U	0.0039	0.010	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.17	0.50	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	22.3		30 - 135		111%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.3		44 - 124		107%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/18/25
Project:	USACE018-44 DOD	Date Received:	08/18/25
Client Sample ID:	PIBLK-PL096849.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096849.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096849.D	1		08/18/25	PL081825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096849.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 12:39
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 16:13:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.530	2.824	67219078	101.9E6	21.133	21.329
28) SA Decachlor...	9.007	7.987	53146388	91268181	22.287	21.038

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096849.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 12:39
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 16:13:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm




Response_ Signal: PL096849.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 67219078
Conc: 21.13 ng/ml


Instrument :
ECD_L
ClientSampleId :
I.BLK



Time Response_ Signal: PL096849.D\ECD2B.ch

#1 Tetrachloro-m-xylene


R.T.: 2.824 min
Delta R.T.: -0.004 min
Response: 101903504
Conc: 21.33 ng/ml



Time Response_ Signal: PL096849.D\ECD1A.ch

#28 Decachlorobiphenyl


R.T.: 9.007 min
Delta R.T.: -0.009 min
Response: 53146388
Conc: 22.29 ng/ml



Time Response_ Signal: PL096849.D\ECD2B.ch

#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 91268181
Conc: 21.04 ng/ml



Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/18/25
Project:	USACE018-44 DOD	Date Received:	08/18/25
Client Sample ID:	PIBLK-PL096857.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096857.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096857.D	1		08/18/25	pl081825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.010	U	0.0039	0.010	0.050	ug/L
319-85-7	beta-BHC	0.010	U	0.0049	0.010	0.050	ug/L
319-86-8	delta-BHC	0.025	U	0.011	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.010	U	0.0037	0.010	0.050	ug/L
76-44-8	Heptachlor	0.010	U	0.0027	0.010	0.050	ug/L
309-00-2	Aldrin	0.010	U	0.0036	0.010	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.025	U	0.0096	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.010	U	0.0031	0.010	0.050	ug/L
60-57-1	Dieldrin	0.010	U	0.0036	0.010	0.050	ug/L
72-55-9	4,4-DDE	0.010	U	0.0037	0.010	0.050	ug/L
72-20-8	Endrin	0.010	U	0.0032	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.025	U	0.0079	0.025	0.050	ug/L
72-54-8	4,4-DDD	0.025	U	0.0071	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.010	U	0.0037	0.010	0.050	ug/L
50-29-3	4,4-DDT	0.010	U	0.0035	0.010	0.050	ug/L
72-43-5	Methoxychlor	0.025	U	0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.025	U	0.0093	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.025	U	0.011	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.010	U	0.0035	0.010	0.050	ug/L
5103-74-2	gamma-Chlordane	0.010	U	0.0039	0.010	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.17	0.50	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	22.5		30 - 135		112%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.1		44 - 124		105%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	08/18/25
Project:	USACE018-44 DOD	Date Received:	08/18/25
Client Sample ID:	PIBLK-PL096857.D	SDG No.:	Q2815
Lab Sample ID:	I.BLK-PL096857.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096857.D	1		08/18/25	p1081825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096857.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 16:57
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	3.531	2.825	66964840	100.5E6	21.053	21.041
28) SA Decachlor...	9.002	7.984	53610964	95610379	22.482m	22.039m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096857.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 16:57
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 66964840
Conc: 21.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/19/2025
Supervised By :mohammad ahmed 08/20/2025




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 100526992
Conc: 21.04 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.002 min
Delta R.T.: -0.013 min
Response: 53610964
Conc: 22.48 ng/ml m



#28 Decachlorobiphenyl

R.T.: 7.984 min
Delta R.T.: -0.008 min
Response: 95610379
Conc: 22.04 ng/ml m

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	USACE018-44 DOD	Date Received:	
Client Sample ID:	PB169225BS	SDG No.:	Q2815
Lab Sample ID:	PB169225BS	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096823.D	1	08/12/25 10:20	08/15/25 16:21	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.50		0.0039	0.010	0.050	ug/L
319-85-7	beta-BHC	0.50		0.0049	0.010	0.050	ug/L
319-86-8	delta-BHC	0.50		0.011	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.50		0.0037	0.010	0.050	ug/L
76-44-8	Heptachlor	0.53		0.0027	0.010	0.050	ug/L
309-00-2	Aldrin	0.50		0.0036	0.010	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.51		0.0096	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.49		0.0031	0.010	0.050	ug/L
60-57-1	Dieldrin	0.49		0.0036	0.010	0.050	ug/L
72-55-9	4,4-DDE	0.49		0.0037	0.010	0.050	ug/L
72-20-8	Endrin	0.51		0.0032	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.51		0.0079	0.025	0.050	ug/L
72-54-8	4,4-DDD	0.50		0.0071	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.48		0.0037	0.010	0.050	ug/L
50-29-3	4,4-DDT	0.50		0.0035	0.010	0.050	ug/L
72-43-5	Methoxychlor	0.49		0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.52		0.0093	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.55		0.011	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.50		0.0035	0.010	0.050	ug/L
5103-74-2	gamma-Chlordane	0.50		0.0039	0.010	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.17	0.50	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	20.9		30 - 135		105%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.7		44 - 124		99%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	USACE018-44 DOD	Date Received:	
Client Sample ID:	PB169225BS	SDG No.:	Q2815
Lab Sample ID:	PB169225BS	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096823.D	1	08/12/25 10:20	08/15/25 16:21	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096823.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 16:21
 Operator : AR\AJ
 Sample : PB169225BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB169225BS

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 08/20/2025
 Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:55:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.825	60775217	94199127	19.107	19.716
28)	SA Decachlor...	9.003	7.987	48814775	90894687	20.470m	20.952
Target Compounds							
2)	A alpha-BHC	3.976	3.330	231.9E6	354.3E6	50.129m	50.066
3)	MA gamma-BHC...	4.303	3.662	219.1E6	330.6E6	49.539m	50.059
4)	MA Heptachlor	4.896	4.011	219.5E6	326.3E6	52.820	48.974
5)	MB Aldrin	5.235	4.293	212.1E6	307.8E6	49.311	49.594
6)	B beta-BHC	4.490	3.958	90346488	140.1E6	50.054m	49.652
7)	B delta-BHC	4.737	4.191	200.6E6	323.8E6	50.286	50.023
8)	B Heptachlo...	5.655	4.795	194.9E6	279.8E6	50.536	48.985
9)	A Endosulfan I	6.036	5.166	176.2E6	263.2E6	48.990	47.574
10)	B gamma-Chl...	5.908	5.047	191.4E6	291.6E6	50.305	49.595
11)	B alpha-Chl...	5.989	5.111	192.7E6	287.4E6	49.982	48.722
12)	B 4,4'-DDE	6.159	5.299	157.6E6	272.8E6	49.062	49.477
13)	MA Dieldrin	6.307	5.430	183.1E6	290.5E6	49.329m	49.185
14)	MA Endrin	6.534	5.705	153.6E6	261.5E6	50.761m	48.378
15)	B Endosulfa...	6.748	5.996	162.5E6	250.1E6	50.742	48.697
16)	A 4,4'-DDD	6.667	5.852	127.2E6	232.6E6	50.315	49.441
17)	MA 4,4'-DDT	6.982	6.104	142.3E6	245.6E6	49.616	48.551
18)	B Endrin al...	6.876	6.174	114.1E6	199.7E6	53.165	54.945
19)	B Endosulfa...	7.109	6.398	138.2E6	246.6E6	48.092m	48.484
20)	A Methoxychlor	7.454	6.676	71535659	132.8E6	48.718	48.452
21)	B Endrin ke...	7.589	6.902	147.3E6	288.7E6	48.986	51.834
22)	Mirex	8.067	7.092	119.6E6	213.1E6	48.231m	48.857

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096823.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 16:21
Operator : AR\AJ
Sample : PB169225BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PB169225BS

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 08/20/2025

Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 60775217
Conc: 19.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 94199127
Conc: 19.72 ng/ml




#2 alpha-BHC

R.T.: 3.976 min
Delta R.T.: -0.007 min
Response: 231916800
Conc: 50.13 ng/ml m



#2 alpha-BHC

R.T.: 3.330 min
Delta R.T.: -0.003 min
Response: 354317224
Conc: 50.07 ng/ml



#3 gamma-BHC (Lindane)

R.T.: 4.303 min
Delta R.T.: -0.007 min
Response: 219130977
Conc: 49.54 ng/ml

Instrument :

ECD_L

ClientSampleId :


PB169225BS

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 08/20/2025

Supervised By :mohammad ahmed 08/21/2025




#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 330559345
Conc: 50.06 ng/ml




#4 Heptachlor

R.T.: 4.896 min
Delta R.T.: -0.007 min
Response: 219501777
Conc: 52.82 ng/ml



#4 Heptachlor

R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 326259099
Conc: 48.97 ng/ml




#5 Aldrin

R.T.: 5.235 min
Delta R.T.: -0.008 min
Response: 212087085
Conc: 49.31 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#5 Aldrin

R.T.: 4.293 min
Delta R.T.: -0.004 min
Response: 307833389
Conc: 49.59 ng/ml




#6 beta-BHC

R.T.: 4.490 min
Delta R.T.: -0.007 min
Response: 90346488
Conc: 50.05 ng/ml m



#6 beta-BHC

R.T.: 3.958 min
Delta R.T.: -0.003 min
Response: 140064434
Conc: 49.65 ng/ml




#7 delta-BHC

R.T.: 4.737 min
Delta R.T.: -0.005 min
Response: 200634572
Conc: 50.29 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#7 delta-BHC

R.T.: 4.191 min
Delta R.T.: -0.004 min
Response: 323827369
Conc: 50.02 ng/ml




#8 Heptachlor epoxide

R.T.: 5.655 min
Delta R.T.: -0.007 min
Response: 194893536
Conc: 50.54 ng/ml



#8 Heptachlor epoxide

R.T.: 4.795 min
Delta R.T.: -0.004 min
Response: 279768553
Conc: 48.98 ng/ml




#9 Endosulfan I

R.T.: 6.036 min
Delta R.T.: -0.007 min
Response: 176186217
Conc: 48.99 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#9 Endosulfan I

R.T.: 5.166 min
Delta R.T.: -0.003 min
Response: 263153744
Conc: 47.57 ng/ml




#10 gamma-Chlordane

R.T.: 5.908 min
Delta R.T.: -0.007 min
Response: 191421188
Conc: 50.30 ng/ml



#10 gamma-Chlordane

R.T.: 5.047 min
Delta R.T.: -0.004 min
Response: 291623122
Conc: 49.59 ng/ml




#11 alpha-Chlordane

R.T.: 5.989 min
Delta R.T.: -0.008 min
Response: 192662406
Conc: 49.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#11 alpha-Chlordane

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 287404609
Conc: 48.72 ng/ml




#12 4,4'-DDE

R.T.: 6.159 min
Delta R.T.: -0.007 min
Response: 157550711
Conc: 49.06 ng/ml



#12 4,4'-DDE

R.T.: 5.299 min
Delta R.T.: -0.005 min
Response: 272836065
Conc: 49.48 ng/ml




#13 Dieldrin

R.T.: 6.307 min
Delta R.T.: -0.009 min
Response: 183068866
Conc: 49.33 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#13 Dieldrin

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 290510751
Conc: 49.18 ng/ml




#14 Endrin

R.T.: 6.534 min
Delta R.T.: -0.009 min
Response: 153643102
Conc: 50.76 ng/ml m



#14 Endrin

R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 261530591
Conc: 48.38 ng/ml




#15 Endosulfan II

R.T.: 6.748 min
Delta R.T.: -0.007 min
Response: 162507755
Conc: 50.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025




#15 Endosulfan II

R.T.: 5.996 min
Delta R.T.: -0.005 min
Response: 250120283
Conc: 48.70 ng/ml



#16 4,4'-DDD

R.T.: 6.667 min
Delta R.T.: -0.008 min
Response: 127213853
Conc: 50.31 ng/ml




#16 4,4'-DDD

R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 232633507
Conc: 49.44 ng/ml

Response_ Signal: PL096823.D\ECD1A.ch

#17 4,4'-DDT



R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 142282054
Conc: 49.62 ng/ml


Instrument :
ECD_L
ClientSampleId :
PB169225BS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025

Response_ Signal: PL096823.D\ECD2B.ch


#17 4,4'-DDT



R.T.: 6.104 min
Delta R.T.: -0.005 min
Response: 245583576
Conc: 48.55 ng/ml

Response_ Signal: PL096823.D\ECD1A.ch


#18 Endrin aldehyde



R.T.: 6.876 min
Delta R.T.: -0.008 min
Response: 114087942
Conc: 53.17 ng/ml


Response_ Signal: PL096823.D\ECD2B.ch

#18 Endrin aldehyde



R.T.: 6.174 min
Delta R.T.: -0.005 min
Response: 199729538
Conc: 54.95 ng/ml

Response_ Signal: PL096823.D\ECD1A.ch



#19 Endosulfan Sulfate


R.T.: 7.109 min
Delta R.T.: -0.009 min
Response: 138200880
Conc: 48.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025


Time Response_ Signal: PL096823.D\ECD2B.ch



#19 Endosulfan Sulfate

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 246600399
Conc: 48.48 ng/ml


Response_ Signal: PL096823.D\ECD1A.ch



#20 Methoxychlor

R.T.: 7.454 min
Delta R.T.: -0.008 min
Response: 71535659
Conc: 48.72 ng/ml


Time Response_ Signal: PL096823.D\ECD2B.ch



#20 Methoxychlor

R.T.: 6.676 min
Delta R.T.: -0.005 min
Response: 132781179
Conc: 48.45 ng/ml

Response_ Signal: PL096823.D\ECD1A.ch



#21 Endrin ketone


R.T.: 7.589 min
Delta R.T.: -0.008 min
Response: 147288355
Conc: 48.99 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/20/2025
Supervised By :mohammad ahmed 08/21/2025


Time Response_ Signal: PL096823.D\ECD2B.ch



#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 288700248
Conc: 51.83 ng/ml


Time Response_ Signal: PL096823.D\ECD1A.ch



#22 Mirex

R.T.: 8.067 min
Delta R.T.: -0.009 min
Response: 119581307
Conc: 48.23 ng/ml m


Time Response_ Signal: PL096823.D\ECD2B.ch



#22 Mirex

R.T.: 7.092 min
Delta R.T.: -0.005 min
Response: 213057480
Conc: 48.86 ng/ml

Signal: PL096823.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 9.003 min
Delta R.T.: -0.013 min
Response: 48814775
Conc: 20.47 ng/ml

Instrument :

ECD_L

ClientSampleId :

PB169225BS


Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/20/2025

Supervised By :mohammad ahmed 08/21/2025

Signal: PL096823.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.006 min
Response: 90894687
Conc: 20.95 ng/ml

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	USACE018-44 DOD	Date Received:	
Client Sample ID:	PB169225BSD	SDG No.:	Q2815
Lab Sample ID:	PB169225BSD	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096824.D	1	08/12/25 10:20	08/15/25 16:34	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
319-84-6	alpha-BHC	0.48		0.0039	0.010	0.050	ug/L
319-85-7	beta-BHC	0.49		0.0049	0.010	0.050	ug/L
319-86-8	delta-BHC	0.49		0.011	0.025	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.48		0.0037	0.010	0.050	ug/L
76-44-8	Heptachlor	0.52		0.0027	0.010	0.050	ug/L
309-00-2	Aldrin	0.48		0.0036	0.010	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.49		0.0096	0.025	0.050	ug/L
959-98-8	Endosulfan I	0.48		0.0031	0.010	0.050	ug/L
60-57-1	Dieldrin	0.49		0.0036	0.010	0.050	ug/L
72-55-9	4,4-DDE	0.50		0.0037	0.010	0.050	ug/L
72-20-8	Endrin	0.49		0.0032	0.010	0.050	ug/L
33213-65-9	Endosulfan II	0.51		0.0079	0.025	0.050	ug/L
72-54-8	4,4-DDD	0.49		0.0071	0.025	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.48		0.0037	0.010	0.050	ug/L
50-29-3	4,4-DDT	0.48		0.0035	0.010	0.050	ug/L
72-43-5	Methoxychlor	0.47		0.011	0.025	0.050	ug/L
53494-70-5	Endrin ketone	0.49		0.0093	0.025	0.050	ug/L
7421-93-4	Endrin aldehyde	0.54		0.011	0.025	0.050	ug/L
5103-71-9	alpha-Chlordane	0.49		0.0035	0.010	0.050	ug/L
5103-74-2	gamma-Chlordane	0.49		0.0039	0.010	0.050	ug/L
8001-35-2	Toxaphene	0.50	U	0.17	0.50	1.00	ug/L
SURROGATES							
2051-24-3	Decachlorobiphenyl	20.8		30 - 135		104%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.3		44 - 124		96%	SPK: 20

Report of Analysis

Client:	First Environment, Inc.	Date Collected:	
Project:	USACE018-44 DOD	Date Received:	
Client Sample ID:	PB169225BSD	SDG No.:	Q2815
Lab Sample ID:	PB169225BSD	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
Prep Method :	3510C	Test:	Pesticide-TCL
		PH :	
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096824.D	1	08/12/25 10:20	08/15/25 16:34	PB169225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096824.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 16:34
 Operator : AR\AJ
 Sample : PB169225BSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB169225BSD

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 08/18/2025
 Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:55:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.826	59543401	91996171	18.720	19.255
28)	SA Decachlor...	9.004	7.987	48180370	90356362	20.204m	20.828
Target Compounds							
2)	A alpha-BHC	3.978	3.331	223.2E6	342.8E6	48.236	48.442
3)	MA gamma-BHC...	4.306	3.663	210.3E6	320.0E6	47.536	48.462
4)	MA Heptachlor	4.897	4.011	214.5E6	317.1E6	51.625	47.601
5)	MB Aldrin	5.237	4.294	207.2E6	299.2E6	48.180	48.202
6)	B beta-BHC	4.491	3.959	87600808	135.8E6	48.533m	48.135
7)	B delta-BHC	4.738	4.192	195.2E6	313.4E6	48.931	48.408
8)	B Heptachlo...	5.656	4.796	190.6E6	273.5E6	49.435	47.880
9)	A Endosulfan I	6.037	5.167	173.6E6	255.4E6	48.277	46.174
10)	B gamma-Chl...	5.910	5.048	186.9E6	283.8E6	49.128	48.272
11)	B alpha-Chl...	5.990	5.112	188.2E6	280.2E6	48.827	47.494
12)	B 4,4'-DDE	6.160	5.300	159.5E6	265.9E6	49.682	48.218
13)	MA Dieldrin	6.310	5.431	180.2E6	284.1E6	48.545	48.096
14)	MA Endrin	6.534	5.705	148.9E6	253.1E6	49.189m	46.827
15)	B Endosulfa...	6.749	5.997	162.8E6	245.2E6	50.827	47.742
16)	A 4,4'-DDD	6.669	5.853	124.7E6	227.7E6	49.301	48.390
17)	MA 4,4'-DDT	6.983	6.105	138.2E6	238.1E6	48.196	47.076
18)	B Endrin al...	6.878	6.175	112.1E6	196.6E6	52.223	54.094
19)	B Endosulfa...	7.110	6.398	136.6E6	241.4E6	47.520m	47.452
20)	A Methoxychlor	7.455	6.677	68285598	129.5E6	46.505	47.241
21)	B Endrin ke...	7.590	6.903	143.9E6	273.2E6	47.846	49.057
22)	Mirex	8.067	7.092	118.2E6	210.0E6	47.689m	48.160

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096824.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 16:34
Operator : AR\AJ
Sample : PB169225BSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PB169225BSD

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 59543401
Conc: 18.72 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 91996171
Conc: 19.26 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 223157039
Conc: 48.24 ng/ml



#2 alpha-BHC

R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 342828244
Conc: 48.44 ng/ml



#3 gamma-BHC (Lindane)

R.T.: 4.306 min
Delta R.T.: -0.004 min
Response: 210273174
Conc: 47.54 ng/ml

Instrument :

ECD_L

ClientSampleId :


PB169225BSD

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 08/18/2025

Supervised By :mohammad ahmed 08/21/2025




#3 gamma-BHC (Lindane)

R.T.: 3.663 min
Delta R.T.: -0.002 min
Response: 320012087
Conc: 48.46 ng/ml




#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.005 min
Response: 214539353
Conc: 51.63 ng/ml



#4 Heptachlor

R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 317116974
Conc: 47.60 ng/ml




#5 Aldrin

R.T.: 5.237 min
Delta R.T.: -0.006 min
Response: 207221296
Conc: 48.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#5 Aldrin

R.T.: 4.294 min
Delta R.T.: -0.003 min
Response: 299191263
Conc: 48.20 ng/ml




#6 beta-BHC

R.T.: 4.491 min
Delta R.T.: -0.006 min
Response: 87600808
Conc: 48.53 ng/ml m



#6 beta-BHC

R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 135784500
Conc: 48.14 ng/ml




#7 delta-BHC

R.T.: 4.738 min
Delta R.T.: -0.004 min
Response: 195226463
Conc: 48.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#7 delta-BHC

R.T.: 4.192 min
Delta R.T.: -0.003 min
Response: 313372209
Conc: 48.41 ng/ml



#8 Heptachlor epoxide

R.T.: 5.656 min
Delta R.T.: -0.006 min
Response: 190649519
Conc: 49.44 ng/ml



#8 Heptachlor epoxide

R.T.: 4.796 min
Delta R.T.: -0.003 min
Response: 273458365
Conc: 47.88 ng/ml

Response_ Signal: PL096824.D\ECD1A.ch

#9 Endosulfan I

R.T.: 6.037 min
Delta R.T.: -0.006 min
Response: 173622932
Conc: 48.28 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Time

Response_ Signal: PL096824.D\ECD2B.ch

#9 Endosulfan I

R.T.: 5.167 min
Delta R.T.: -0.002 min
Response: 255405216
Conc: 46.17 ng/ml

Time

Response_ Signal: PL096824.D\ECD1A.ch

#10 gamma-Chlordane

R.T.: 5.910 min
Delta R.T.: -0.006 min
Response: 186943063
Conc: 49.13 ng/ml


Time

Response_ Signal: PL096824.D\ECD2B.ch

#10 gamma-Chlordane

R.T.: 5.048 min
Delta R.T.: -0.003 min
Response: 283843264
Conc: 48.27 ng/ml

Time




#11 alpha-Chlordane

R.T.: 5.990 min
Delta R.T.: -0.006 min
Response: 188211879
Conc: 48.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#11 alpha-Chlordane

R.T.: 5.112 min
Delta R.T.: -0.003 min
Response: 280162453
Conc: 47.49 ng/ml




#12 4,4'-DDE

R.T.: 6.160 min
Delta R.T.: -0.006 min
Response: 159539795
Conc: 49.68 ng/ml



#12 4,4'-DDE

R.T.: 5.300 min
Delta R.T.: -0.004 min
Response: 265892329
Conc: 48.22 ng/ml




#13 Dieldrin

R.T.: 6.310 min
Delta R.T.: -0.006 min
Response: 180156209
Conc: 48.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#13 Dieldrin

R.T.: 5.431 min
Delta R.T.: -0.003 min
Response: 284082332
Conc: 48.10 ng/ml




#14 Endrin

R.T.: 6.534 min
Delta R.T.: -0.008 min
Response: 148884541
Conc: 49.19 ng/ml m



#14 Endrin

R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 253142704
Conc: 46.83 ng/ml




#15 Endosulfan II

R.T.: 6.749 min
Delta R.T.: -0.006 min
Response: 162781600
Conc: 50.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025




#15 Endosulfan II

R.T.: 5.997 min
Delta R.T.: -0.004 min
Response: 245214434
Conc: 47.74 ng/ml



#16 4,4'-DDD

R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 124651862
Conc: 49.30 ng/ml



#16 4,4'-DDD

R.T.: 5.853 min
Delta R.T.: -0.003 min
Response: 227687881
Conc: 48.39 ng/ml

Response_ Signal: PL096824.D\ECD1A.ch

#17 4,4'-DDT

R.T.: 6.983 min
Delta R.T.: -0.006 min
Response: 138211624
Conc: 48.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Time 6.60 6.70 6.80 6.90 7.00 7.10 7.20 7.30

Response_ Signal: PL096824.D\ECD2B.ch

#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 238126866
Conc: 47.08 ng/ml

Time 5.90 6.00 6.10 6.20 6.30

Response_ Signal: PL096824.D\ECD1A.ch

#18 Endrin aldehyde

R.T.: 6.878 min
Delta R.T.: -0.006 min
Response: 112065931
Conc: 52.22 ng/ml

Time 6.70 6.80 6.90 7.00 7.10

Response_ Signal: PL096824.D\ECD2B.ch


#18 Endrin aldehyde

R.T.: 6.175 min
Delta R.T.: -0.004 min
Response: 196634214
Conc: 54.09 ng/ml

Time 5.90 6.00 6.10 6.20 6.30 6.40

Response_ Signal: PL096824.D\ECD1A.ch

#19 Endosulfan Sulfate



R.T.: 7.110 min
Delta R.T.: -0.008 min
Response: 136556540
Conc: 47.52 ng/ml


Instrument :
ECD_L
ClientSampleId :
PB169225BSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025

Response_ Signal: PL096824.D\ECD2B.ch


#19 Endosulfan Sulfate



R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 241352693
Conc: 47.45 ng/ml

Response_ Signal: PL096824.D\ECD1A.ch


#20 Methoxychlor



R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 68285598
Conc: 46.50 ng/ml


Response_ Signal: PL096824.D\ECD2B.ch

#20 Methoxychlor



R.T.: 6.677 min
Delta R.T.: -0.004 min
Response: 129462927
Conc: 47.24 ng/ml

Response_ Signal: PL096824.D\ECD1A.ch



#21 Endrin ketone


R.T.: 7.590 min
Delta R.T.: -0.007 min
Response: 143858655
Conc: 47.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/18/2025
Supervised By :mohammad ahmed 08/21/2025


Time Response_ Signal: PL096824.D\ECD2B.ch



#21 Endrin ketone

R.T.: 6.903 min
Delta R.T.: -0.004 min
Response: 273236792
Conc: 49.06 ng/ml


Time Response_ Signal: PL096824.D\ECD1A.ch



#22 Mirex

R.T.: 8.067 min
Delta R.T.: -0.009 min
Response: 118237062
Conc: 47.69 ng/ml m

Time Response_ Signal: PL096824.D\ECD2B.ch




#22 Mirex

R.T.: 7.092 min
Delta R.T.: -0.005 min
Response: 210019786
Conc: 48.16 ng/ml

Response_

Signal: PL096824.D\ECD1A.ch

#28 Decachlorobiphenyl



R.T.: 9.004 min
Delta R.T.: -0.012 min
Response: 48180370
Conc: 20.20 ng/ml

Instrument :

ECD_L

ClientSampleId :

PB169225BSD

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 08/18/2025


Supervised By :mohammad ahmed 08/21/2025

Time

Response_

Signal: PL096824.D\ECD2B.ch

#28 Decachlorobiphenyl



R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 90356362
Conc: 20.83 ng/ml

Time

Manual Integration Report

Sequence:	pl072825	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL096594.D	4,4"-DDD	Abdul	7/29/2025 9:49:39 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
PEM	PL096594.D	4,4"-DDD #2	Abdul	7/29/2025 9:49:39 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
PEM	PL096594.D	Endrin aldehyde	Abdul	7/29/2025 9:49:39 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
PEM	PL096594.D	Endrin ketone	Abdul	7/29/2025 9:49:39 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
PEM	PL096594.D	Endrin ketone #2	Abdul	7/29/2025 9:49:39 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
RESCHK	PL096595.D	Endosulfan I #2	Abdul	7/29/2025 9:49:43 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
PSTDICC025	PL096599.D	Endrin aldehyde #2	Abdul	7/29/2025 9:49:47 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
PSTDICC025	PL096599.D	Endrin ketone #2	Abdul	7/29/2025 9:49:47 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
PSTDICC005	PL096600.D	Aldrin	Abdul	7/29/2025 9:49:51 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
PSTDICC005	PL096600.D	Endosulfan II	Abdul	7/29/2025 9:49:51 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
PSTDICC005	PL096600.D	Endrin aldehyde #2	Abdul	7/29/2025 9:49:51 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
PSTDICC005	PL096600.D	Heptachlor	Abdul	7/29/2025 9:49:51 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software
PSTDICC005	PL096600.D	Heptachlor epoxide	Abdul	7/29/2025 9:49:51 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software

Manual Integration Report

Sequence:

pl072825

Instrument

ECD_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDICV050	PL096611.D	delta-BHC	Abdul	7/29/2025 9:49:55 AM	mohammad	7/30/2025 2:11:50	Peak Integrated by Software

Manual Integration Report

Sequence:	pl081625	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
I.BLK	PL096817.D	Decachlorobiphenyl	Abdul	8/18/2025 8:11:13 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096818.D	4,4"-DDD	Abdul	8/20/2025 8:50:45 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096818.D	4,4"-DDD #2	Abdul	8/20/2025 8:50:45 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096818.D	Endrin	Abdul	8/20/2025 8:50:45 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096818.D	Endrin aldehyde	Abdul	8/20/2025 8:50:45 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096818.D	Endrin aldehyde #2	Abdul	8/20/2025 8:50:45 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096818.D	Endrin ketone	Abdul	8/20/2025 8:50:45 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096818.D	Endrin ketone #2	Abdul	8/20/2025 8:50:45 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096818.D	Methoxychlor #2	Abdul	8/20/2025 8:50:45 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096819.D	4,4"-DDD	Abdul	8/18/2025 8:11:20 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096819.D	4,4"-DDT #2	Abdul	8/18/2025 8:11:20 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096819.D	alpha-BHC	Abdul	8/18/2025 8:11:20 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096819.D	beta-BHC	Abdul	8/18/2025 8:11:20 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software

Manual Integration Report

Sequence:	pl081625	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PL096819.D	Endosulfan II	Abdul	8/18/2025 8:11:20 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096819.D	Endrin	Abdul	8/18/2025 8:11:20 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096819.D	Endrin aldehyde	Abdul	8/18/2025 8:11:20 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096819.D	gamma-BHC (Lindane)	Abdul	8/18/2025 8:11:20 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PCHLORCCC500	PL096820.D	Chlordane-1	Abdul	8/18/2025 8:11:22 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PTOXCCC500	PL096821.D	Decachlorobiphenyl #2	Abdul	8/18/2025 8:11:26 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BL	PL096822.D	Decachlorobiphenyl	Abdul	8/18/2025 8:11:31 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BS	PL096823.D	alpha-BHC	Abdul	8/20/2025 8:50:41 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BS	PL096823.D	beta-BHC	Abdul	8/20/2025 8:50:41 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BS	PL096823.D	Decachlorobiphenyl	Abdul	8/20/2025 8:50:41 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BS	PL096823.D	Dieldrin	Abdul	8/20/2025 8:50:41 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BS	PL096823.D	Endosulfan Sulfate	Abdul	8/20/2025 8:50:41 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BS	PL096823.D	Endrin	Abdul	8/20/2025 8:50:41 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software

Manual Integration Report

Sequence:	pl081625	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PB169225BS	PL096823.D	gamma-BHC (Lindane)	Abdul	8/20/2025 8:50:41 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BS	PL096823.D	Mirex	Abdul	8/20/2025 8:50:41 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BSD	PL096824.D	beta-BHC	Abdul	8/18/2025 8:11:42 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BSD	PL096824.D	Decachlorobiphenyl	Abdul	8/18/2025 8:11:42 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BSD	PL096824.D	Endosulfan Sulfate	Abdul	8/18/2025 8:11:42 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BSD	PL096824.D	Endrin	Abdul	8/18/2025 8:11:42 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PB169225BSD	PL096824.D	Mirex	Abdul	8/18/2025 8:11:42 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
I.BLK	PL096827.D	Decachlorobiphenyl	Abdul	8/18/2025 8:11:56 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096828.D	beta-BHC	Abdul	8/18/2025 8:12:01 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096828.D	Decachlorobiphenyl	Abdul	8/18/2025 8:12:01 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096828.D	Dieldrin #2	Abdul	8/18/2025 8:12:01 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096828.D	Endosulfan II	Abdul	8/18/2025 8:12:01 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096828.D	Endrin	Abdul	8/18/2025 8:12:01 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software

Manual Integration Report

Sequence:	pl081625	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PL096828.D	gamma-BHC (Lindane)	Abdul	8/18/2025 8:12:01 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096828.D	Heptachlor	Abdul	8/18/2025 8:12:01 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096828.D	Mirex	Abdul	8/18/2025 8:12:01 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PCHLORCCC500	PL096829.D	Chlordane-1	Abdul	8/18/2025 8:12:04 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PCHLORCCC500	PL096829.D	Decachlorobiphenyl	Abdul	8/18/2025 8:12:04 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PTOXCCC500	PL096830.D	Decachlorobiphenyl #2	Abdul	8/18/2025 8:12:07 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
I.BLK	PL096835.D	Decachlorobiphenyl	Abdul	8/18/2025 8:12:26 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096836.D	4,4"-DDD	Abdul	8/20/2025 8:50:38 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096836.D	beta-BHC	Abdul	8/20/2025 8:50:38 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096836.D	Endrin	Abdul	8/20/2025 8:50:38 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096836.D	Endrin aldehyde	Abdul	8/20/2025 8:50:38 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096836.D	Endrin aldehyde #2	Abdul	8/20/2025 8:50:38 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PEM	PL096836.D	Endrin ketone #2	Abdul	8/20/2025 8:50:38 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software

Manual Integration Report

Sequence:	pl081625	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL096836.D	Methoxychlor #2	Abdul	8/20/2025 8:50:38 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096837.D	4,4"-DDD	Abdul	8/18/2025 8:12:34 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096837.D	Decachlorobiphenyl	Abdul	8/18/2025 8:12:34 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096837.D	Dieldrin #2	Abdul	8/18/2025 8:12:34 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096837.D	Endosulfan II	Abdul	8/18/2025 8:12:34 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096837.D	Endrin	Abdul	8/18/2025 8:12:34 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096837.D	Methoxychlor #2	Abdul	8/18/2025 8:12:34 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096837.D	Mirex	Abdul	8/18/2025 8:12:34 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
Q2815-11	PL096839.D	alpha-Chlordane	Abdul	8/18/2025 8:13:08 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
Q2815-11	PL096839.D	alpha-Chlordane #2	Abdul	8/18/2025 8:13:08 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
Q2815-11	PL096839.D	Decachlorobiphenyl	Abdul	8/18/2025 8:13:08 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
Q2815-11	PL096839.D	Decachlorobiphenyl #2	Abdul	8/18/2025 8:13:08 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
Q2815-11	PL096839.D	Dieldrin	Abdul	8/18/2025 8:13:08 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software

Manual Integration Report

Sequence:	pl081625	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q2815-11	PL096839.D	Dieldrin #2	Abdul	8/18/2025 8:13:08 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
Q2815-11	PL096839.D	Tetrachloro-m-xylene	Abdul	8/18/2025 8:13:08 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
I.BLK	PL096840.D	Decachlorobiphenyl	Abdul	8/18/2025 8:13:00 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096841.D	4,4"-DDD	Abdul	8/20/2025 8:50:35 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096841.D	beta-BHC	Abdul	8/20/2025 8:50:35 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096841.D	Decachlorobiphenyl	Abdul	8/20/2025 8:50:35 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096841.D	Dieldrin	Abdul	8/20/2025 8:50:35 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096841.D	Dieldrin #2	Abdul	8/20/2025 8:50:35 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096841.D	Endosulfan II	Abdul	8/20/2025 8:50:35 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096841.D	Endosulfan Sulfate	Abdul	8/20/2025 8:50:35 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096841.D	Endrin	Abdul	8/20/2025 8:50:35 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096841.D	Methoxychlor #2	Abdul	8/20/2025 8:50:35 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software
PSTDCCC050	PL096841.D	Mirex	Abdul	8/20/2025 8:50:35 AM	mohammad	8/19/2025 1:27:28	Peak Integrated by Software

Manual Integration Report

Sequence:

pl081625

Instrument

ECD_I

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:	PL081825	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
I.BLK	PL096843.D	Decachlorobiphenyl	Abdul	8/19/2025 8:37:33 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096844.D	alpha-BHC	Abdul	8/19/2025 8:37:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096844.D	beta-BHC	Abdul	8/19/2025 8:37:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096844.D	Endrin	Abdul	8/19/2025 8:37:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096844.D	Endrin aldehyde	Abdul	8/19/2025 8:37:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096844.D	gamma-BHC (Lindane)	Abdul	8/19/2025 8:37:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096844.D	Methoxychlor #2	Abdul	8/19/2025 8:37:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096845.D	alpha-BHC	Abdul	8/19/2025 8:37:46 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096845.D	beta-BHC	Abdul	8/19/2025 8:37:46 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096845.D	Decachlorobiphenyl	Abdul	8/19/2025 8:37:46 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096845.D	gamma-BHC (Lindane)	Abdul	8/19/2025 8:37:46 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096845.D	Heptachlor	Abdul	8/19/2025 8:37:46 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096845.D	Mirex	Abdul	8/19/2025 8:37:46 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software

Manual Integration Report

Sequence:	PL081825	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PL096845.D	Mirex #2	Abdul	8/19/2025 8:37:46 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096845.D	Tetrachloro-m-xylene	Abdul	8/19/2025 8:37:46 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096850.D	alpha-BHC	Abdul	8/19/2025 8:38:05 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096850.D	beta-BHC	Abdul	8/19/2025 8:38:05 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096850.D	Decachlorobiphenyl	Abdul	8/19/2025 8:38:05 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096850.D	Endosulfan Sulfate	Abdul	8/19/2025 8:38:05 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096850.D	Endrin	Abdul	8/19/2025 8:38:05 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096850.D	gamma-BHC (Lindane)	Abdul	8/19/2025 8:38:05 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096850.D	Mirex	Abdul	8/19/2025 8:38:05 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096850.D	Mirex #2	Abdul	8/19/2025 8:38:05 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PCHLORCCC500	PL096851.D	Chlordane-1	Abdul	8/19/2025 8:38:10 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PCHLORCCC500	PL096851.D	Decachlorobiphenyl	Abdul	8/19/2025 8:38:10 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PCHLORCCC500	PL096851.D	Decachlorobiphenyl #2	Abdul	8/19/2025 8:38:10 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software

Manual Integration Report

Sequence:	PL081825	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PTOXCCC500	PL096852.D	Decachlorobiphenyl #2	Abdul	8/19/2025 8:38:14 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PTOXCCC500	PL096852.D	Toxaphene-1	Abdul	8/19/2025 8:38:14 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PTOXCCC500	PL096852.D	Toxaphene-1 #2	Abdul	8/19/2025 8:38:14 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
Q2815-01	PL096855.D	Decachlorobiphenyl	Abdul	8/19/2025 8:38:25 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
Q2815-01	PL096855.D	Tetrachloro-m-xylene	Abdul	8/19/2025 8:38:25 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
Q2815-01DL	PL096856.D	Tetrachloro-m-xylene	Abdul	8/19/2025 8:38:29 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
I.BLK	PL096857.D	Decachlorobiphenyl	Abdul	8/19/2025 8:38:34 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
I.BLK	PL096857.D	Decachlorobiphenyl #2	Abdul	8/19/2025 8:38:34 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096858.D	beta-BHC	Abdul	8/19/2025 8:38:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096858.D	Decachlorobiphenyl	Abdul	8/19/2025 8:38:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096858.D	Decachlorobiphenyl #2	Abdul	8/19/2025 8:38:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096858.D	Dieldrin	Abdul	8/19/2025 8:38:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096858.D	Endosulfan II	Abdul	8/19/2025 8:38:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software

Manual Integration Report

Sequence:	PL081825	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PL096858.D	Endosulfan Sulfate	Abdul	8/19/2025 8:38:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096858.D	Endrin	Abdul	8/19/2025 8:38:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096858.D	Methoxychlor #2	Abdul	8/19/2025 8:38:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096858.D	Mirex	Abdul	8/19/2025 8:38:38 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PCHLORCCC500	PL096859.D	Chlordane-1	Abdul	8/19/2025 8:38:42 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PCHLORCCC500	PL096859.D	Decachlorobiphenyl	Abdul	8/19/2025 8:38:42 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
I.BLK	PL096866.D	Decachlorobiphenyl	Abdul	8/19/2025 8:39:11 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096867.D	4,4"-DDE	Abdul	8/19/2025 8:39:14 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096867.D	4,4"-DDE #2	Abdul	8/19/2025 8:39:14 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096867.D	beta-BHC	Abdul	8/19/2025 8:39:14 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096867.D	Endrin	Abdul	8/19/2025 8:39:14 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096867.D	Endrin aldehyde	Abdul	8/19/2025 8:39:14 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096867.D	Endrin ketone #2	Abdul	8/19/2025 8:39:14 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software

Manual Integration Report

Sequence:	PL081825	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL096867.D	gamma-BHC (Lindane)	Abdul	8/19/2025 8:39:14 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PEM	PL096867.D	Methoxychlor #2	Abdul	8/19/2025 8:39:14 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096868.D	beta-BHC	Abdul	8/19/2025 8:39:18 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096868.D	Decachlorobiphenyl	Abdul	8/19/2025 8:39:18 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096868.D	Dieldrin	Abdul	8/19/2025 8:39:18 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096868.D	Endrin	Abdul	8/19/2025 8:39:18 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096868.D	gamma-BHC (Lindane)	Abdul	8/19/2025 8:39:18 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096868.D	Methoxychlor #2	Abdul	8/19/2025 8:39:18 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096868.D	Mirex	Abdul	8/19/2025 8:39:18 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
I.BLK	PL096878.D	Decachlorobiphenyl	Abdul	8/19/2025 8:36:02 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
I.BLK	PL096878.D	Decachlorobiphenyl #2	Abdul	8/19/2025 8:36:02 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096879.D	4,4"-DDD	Abdul	8/19/2025 8:36:06 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096879.D	alpha-BHC	Abdul	8/19/2025 8:36:06 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software

Manual Integration Report

Sequence:	PL081825	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PL096879.D	beta-BHC	Abdul	8/19/2025 8:36:06 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096879.D	Decachlorobiphenyl	Abdul	8/19/2025 8:36:06 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096879.D	Dieldrin	Abdul	8/19/2025 8:36:06 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096879.D	Endosulfan Sulfate	Abdul	8/19/2025 8:36:06 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096879.D	Endrin	Abdul	8/19/2025 8:36:06 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096879.D	gamma-BHC (Lindane)	Abdul	8/19/2025 8:36:06 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software
PSTDCCC050	PL096879.D	Mirex	Abdul	8/19/2025 8:36:06 AM	mohammad	8/20/2025 3:05:16	Peak Integrated by Software

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL072825

Review By	Abdul	Review On	7/29/2025 9:50:24 AM
Supervise By	mohammad	Supervise On	7/30/2025 2:11:50 AM
SubDirectory	PL072825	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL096592.D	28 Jul 2025 15:57	AR\AJ	Ok
2	I.BLK	PL096593.D	28 Jul 2025 16:11	AR\AJ	Ok
3	PEM	PL096594.D	28 Jul 2025 16:25	AR\AJ	Ok,M
4	RESCHK	PL096595.D	28 Jul 2025 16:38	AR\AJ	Ok,M
5	PSTDICC100	PL096596.D	28 Jul 2025 16:52	AR\AJ	Ok
6	PSTDICC075	PL096597.D	28 Jul 2025 17:06	AR\AJ	Ok
7	PSTDICC050	PL096598.D	28 Jul 2025 17:19	AR\AJ	Ok
8	PSTDICC025	PL096599.D	28 Jul 2025 17:33	AR\AJ	Ok,M
9	PSTDICC005	PL096600.D	28 Jul 2025 17:47	AR\AJ	Ok,M
10	PCHLORICC1000	PL096601.D	28 Jul 2025 18:00	AR\AJ	Ok
11	PCHLORICC750	PL096602.D	28 Jul 2025 18:14	AR\AJ	Ok
12	PCHLORICC500	PL096603.D	28 Jul 2025 18:27	AR\AJ	Ok
13	PCHLORICC250	PL096604.D	28 Jul 2025 18:41	AR\AJ	Ok
14	PCHLORICC050	PL096605.D	28 Jul 2025 18:55	AR\AJ	Ok
15	PTOXICC1000	PL096606.D	28 Jul 2025 19:08	AR\AJ	Ok
16	PTOXICC750	PL096607.D	28 Jul 2025 19:22	AR\AJ	Ok
17	PTOXICC500	PL096608.D	28 Jul 2025 19:36	AR\AJ	Ok
18	PTOXICC250	PL096609.D	28 Jul 2025 19:49	AR\AJ	Ok
19	PTOXICC100	PL096610.D	28 Jul 2025 20:03	AR\AJ	Ok
20	PSTDICV050	PL096611.D	28 Jul 2025 20:17	AR\AJ	Ok,M
21	PCHLORICV500	PL096612.D	28 Jul 2025 20:30	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL072825

Review By	Abdul	Review On	7/29/2025 9:50:24 AM
Supervise By	mohammad	Supervise On	7/30/2025 2:11:50 AM
SubDirectory	PL072825	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PL096613.D	28 Jul 2025 20:44	AR\AJ	Ok
----	------------	------------	-------------------	-------	----

M : Manual Integration

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081625

Review By	Abdul	Review On	8/18/2025 8:13:57 AM
Supervise By	mohammad	Supervise On	8/19/2025 1:27:28 AM
SubDirectory	PL081625	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL096816.D	15 Aug 2025 08:58	AR\AJ	Ok
2	I.BLK	PL096817.D	15 Aug 2025 09:11	AR\AJ	Ok,M
3	PEM	PL096818.D	15 Aug 2025 10:13	AR\AJ	Ok,M
4	PSTDCCC050	PL096819.D	15 Aug 2025 10:58	AR\AJ	Ok,M
5	PCHLORCCC500	PL096820.D	15 Aug 2025 12:28	AR\AJ	Ok,M
6	PTOXCCC500	PL096821.D	15 Aug 2025 15:53	AR\AJ	Ok,M
7	PB169225BL	PL096822.D	15 Aug 2025 16:07	AR\AJ	Ok,M
8	PB169225BS	PL096823.D	15 Aug 2025 16:21	AR\AJ	Ok,M
9	PB169225BSD	PL096824.D	15 Aug 2025 16:34	AR\AJ	Ok,M
10	PB169225BS	PL096825.D	15 Aug 2025 16:48	AR\AJ	Ok,M
11	PB169225BS	PL096826.D	15 Aug 2025 17:36	AR\AJ	Ok,M
12	I.BLK	PL096827.D	15 Aug 2025 17:51	AR\AJ	Ok,M
13	PSTDCCC050	PL096828.D	15 Aug 2025 18:05	AR\AJ	Ok,M
14	PCHLORCCC500	PL096829.D	15 Aug 2025 18:32	AR\AJ	Ok,M
15	PTOXCCC500	PL096830.D	15 Aug 2025 18:45	AR\AJ	Ok,M
16	Q2814-01	PL096831.D	15 Aug 2025 18:59	AR\AJ	Ok,M
17	Q2814-03	PL096832.D	15 Aug 2025 19:13	AR\AJ	Ok,M
18	Q2814-16	PL096833.D	15 Aug 2025 19:26	AR\AJ	Ok,M
19	Q2814-17	PL096834.D	15 Aug 2025 19:40	AR\AJ	Ok,M
20	I.BLK	PL096835.D	15 Aug 2025 19:54	AR\AJ	Ok,M
21	PEM	PL096836.D	15 Aug 2025 20:07	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081625

Review By	Abdul	Review On	8/18/2025 8:13:57 AM
Supervise By	mohammad	Supervise On	8/19/2025 1:27:28 AM
SubDirectory	PL081625	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PSTDCCC050	PL096837.D	15 Aug 2025 20:48	AR\AJ	Ok,M
23	Q2815-01	PL096838.D	15 Aug 2025 21:02	AR\AJ	Not Ok
24	Q2815-11	PL096839.D	15 Aug 2025 21:15	AR\AJ	Ok,M
25	I.BLK	PL096840.D	15 Aug 2025 21:56	AR\AJ	Ok,M
26	PSTDCCC050	PL096841.D	15 Aug 2025 22:10	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081825

Review By	Abdul	Review On	8/19/2025 8:40:55 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:05:16 AM
SubDirectory	PL081825	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL096842.D	18 Aug 2025 09:58	AR\AJ	Ok
2	I.BLK	PL096843.D	18 Aug 2025 10:12	AR\AJ	Ok,M
3	PEM	PL096844.D	18 Aug 2025 10:26	AR\AJ	Ok,M
4	PSTDCCC050	PL096845.D	18 Aug 2025 11:25	AR\AJ	Ok,M
5	PB169283BL	PL096846.D	18 Aug 2025 11:56	AR\AJ	Ok
6	PB169283BS	PL096847.D	18 Aug 2025 12:10	AR\AJ	Not Ok
7	Q2883-01	PL096848.D	18 Aug 2025 12:24	AR\AJ	Ok,M
8	I.BLK	PL096849.D	18 Aug 2025 12:39	AR\AJ	Ok
9	PSTDCCC050	PL096850.D	18 Aug 2025 12:53	AR\AJ	Ok,M
10	PCHLORCCC500	PL096851.D	18 Aug 2025 14:10	AR\AJ	Ok,M
11	PTOXCCC500	PL096852.D	18 Aug 2025 14:48	AR\AJ	Not Ok
12	Q2864-03	PL096853.D	18 Aug 2025 15:58	AR\AJ	Ok
13	PB169264BL	PL096854.D	18 Aug 2025 16:12	AR\AJ	Ok,M
14	Q2815-01	PL096855.D	18 Aug 2025 16:30	AR\AJ	Dilution
15	Q2815-01DL	PL096856.D	18 Aug 2025 16:44	AR\AJ	Ok,M
16	I.BLK	PL096857.D	18 Aug 2025 16:57	AR\AJ	Ok,M
17	PSTDCCC050	PL096858.D	18 Aug 2025 17:11	AR\AJ	Ok,M
18	PCHLORCCC500	PL096859.D	18 Aug 2025 17:25	AR\AJ	Ok,M
19	Q2879-01	PL096860.D	18 Aug 2025 17:38	AR\AJ	Ok,M
20	Q2879-03	PL096861.D	18 Aug 2025 17:52	AR\AJ	Ok,M
21	Q2879-05	PL096862.D	18 Aug 2025 18:06	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081825

Review By	Abdul	Review On	8/19/2025 8:40:55 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:05:16 AM
SubDirectory	PL081825	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2879-07	PL096863.D	18 Aug 2025 18:19	AR\AJ	Ok,M
23	Q2879-09	PL096864.D	18 Aug 2025 18:33	AR\AJ	Ok,M
24	Q2879-11	PL096865.D	18 Aug 2025 18:46	AR\AJ	Ok,M
25	I.BLK	PL096866.D	18 Aug 2025 19:00	AR\AJ	Ok,M
26	PEM	PL096867.D	18 Aug 2025 19:14	AR\AJ	Ok,M
27	PSTDCCC050	PL096868.D	18 Aug 2025 19:41	AR\AJ	Ok,M
28	Q2879-13	PL096869.D	18 Aug 2025 20:22	AR\AJ	Not Ok
29	Q2879-15	PL096870.D	18 Aug 2025 20:35	AR\AJ	Not Ok
30	Q2879-17	PL096871.D	18 Aug 2025 20:49	AR\AJ	Not Ok
31	Q2884-01	PL096872.D	18 Aug 2025 21:02	AR\AJ	Not Ok
32	Q2884-03	PL096873.D	18 Aug 2025 21:16	AR\AJ	Not Ok
33	Q2888-01	PL096874.D	18 Aug 2025 21:30	AR\AJ	Not Ok
34	Q2888-01MS	PL096875.D	18 Aug 2025 21:43	AR\AJ	Not Ok
35	Q2888-01MSD	PL096876.D	18 Aug 2025 21:57	AR\AJ	Not Ok
36	Q2889-01	PL096877.D	18 Aug 2025 22:10	AR\AJ	Not Ok
37	I.BLK	PL096878.D	18 Aug 2025 22:38	AR\AJ	Ok,M
38	PSTDCCC050	PL096879.D	18 Aug 2025 22:51	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL072825

Review By	Abdul	Review On	7/29/2025 9:50:24 AM
Supervise By	mohammad	Supervise On	7/30/2025 2:11:50 AM
SubDirectory	PL072825	HP Acquire Method	HP Processing Method
			p1072825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL096592.D	28 Jul 2025 15:57		AR\AJ	Ok
2	I.BLK	I.BLK	PL096593.D	28 Jul 2025 16:11		AR\AJ	Ok
3	PEM	PEM	PL096594.D	28 Jul 2025 16:25		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL096595.D	28 Jul 2025 16:38		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PL096596.D	28 Jul 2025 16:52		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL096597.D	28 Jul 2025 17:06		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL096598.D	28 Jul 2025 17:19	Comp#18 kept on L.R. for 2nd column	AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL096599.D	28 Jul 2025 17:33		AR\AJ	Ok,M
9	PSTDICC005	PSTDICC005	PL096600.D	28 Jul 2025 17:47		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL096601.D	28 Jul 2025 18:00		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL096602.D	28 Jul 2025 18:14		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL096603.D	28 Jul 2025 18:27		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL096604.D	28 Jul 2025 18:41		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL096605.D	28 Jul 2025 18:55		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL096606.D	28 Jul 2025 19:08		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL096607.D	28 Jul 2025 19:22		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL096608.D	28 Jul 2025 19:36		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL072825

Review By	Abdul	Review On	7/29/2025 9:50:24 AM
Supervise By	mohammad	Supervise On	7/30/2025 2:11:50 AM
SubDirectory	PL072825	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	PTOXICC250	PTOXICC250	PL096609.D	28 Jul 2025 19:49		AR\AJ	Ok
19	PTOXICC100	PTOXICC100	PL096610.D	28 Jul 2025 20:03		AR\AJ	Ok
20	PSTDICV050	ICVPL072825	PL096611.D	28 Jul 2025 20:17		AR\AJ	Ok,M
21	PCHLORICV500	ICVPL072825CHLOR	PL096612.D	28 Jul 2025 20:30		AR\AJ	Ok
22	PTOXICV500	ICVPL072825TOX	PL096613.D	28 Jul 2025 20:44		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081625

Review By	Abdul	Review On	8/18/2025 8:13:57 AM
Supervise By	mohammad	Supervise On	8/19/2025 1:27:28 AM
SubDirectory	PL081625	HP Acquire Method	HP Processing Method p1072825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL096816.D	15 Aug 2025 08:58		AR\AJ	Ok
2	I.BLK	I.BLK	PL096817.D	15 Aug 2025 09:11		AR\AJ	Ok,M
3	PEM	PEM	PL096818.D	15 Aug 2025 10:13		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL096819.D	15 Aug 2025 10:58		AR\AJ	Ok,M
5	PCHLORCCC500	PCHLORCCC500	PL096820.D	15 Aug 2025 12:28		AR\AJ	Ok,M
6	PTOXCCC500	PTOXCCC500	PL096821.D	15 Aug 2025 15:53		AR\AJ	Ok,M
7	PB169225BL	PB169225BL	PL096822.D	15 Aug 2025 16:07		AR\AJ	Ok,M
8	PB169225BS	PB169225BS	PL096823.D	15 Aug 2025 16:21		AR\AJ	Ok,M
9	PB169225BSD	PB169225BSD	PL096824.D	15 Aug 2025 16:34		AR\AJ	Ok,M
10	PB169225BS	PB169225BS	PL096825.D	15 Aug 2025 16:48	CHLOR BS	AR\AJ	Ok,M
11	PB169225BS	PB169225BS	PL096826.D	15 Aug 2025 17:36	TOX BS	AR\AJ	Ok,M
12	I.BLK	I.BLK	PL096827.D	15 Aug 2025 17:51		AR\AJ	Ok,M
13	PSTDCCC050	PSTDCCC050	PL096828.D	15 Aug 2025 18:05		AR\AJ	Ok,M
14	PCHLORCCC500	PCHLORCCC500	PL096829.D	15 Aug 2025 18:32		AR\AJ	Ok,M
15	PTOXCCC500	PTOXCCC500	PL096830.D	15 Aug 2025 18:45		AR\AJ	Ok,M
16	Q2814-01	TW-84SB-E	PL096831.D	15 Aug 2025 18:59		AR\AJ	Ok,M
17	Q2814-03	TW-17M-W	PL096832.D	15 Aug 2025 19:13		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081625

Review By	Abdul	Review On	8/18/2025 8:13:57 AM
Supervise By	mohammad	Supervise On	8/19/2025 1:27:28 AM
SubDirectory	PL081625	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q2814-16	FB	PL096833.D	15 Aug 2025 19:26		AR\AJ	Ok,M
19	Q2814-17	TW-518R-S	PL096834.D	15 Aug 2025 19:40		AR\AJ	Ok,M
20	I.BLK	I.BLK	PL096835.D	15 Aug 2025 19:54		AR\AJ	Ok,M
21	PEM	PEM	PL096836.D	15 Aug 2025 20:07		AR\AJ	Ok,M
22	PSTDCCC050	PSTDCCC050	PL096837.D	15 Aug 2025 20:48		AR\AJ	Ok,M
23	Q2815-01	TW-705R-S	PL096838.D	15 Aug 2025 21:02	All surrogate fail , need dilution	AR\AJ	Not Ok
24	Q2815-11	TW-22M-W	PL096839.D	15 Aug 2025 21:15		AR\AJ	Ok,M
25	I.BLK	I.BLK	PL096840.D	15 Aug 2025 21:56		AR\AJ	Ok,M
26	PSTDCCC050	PSTDCCC050	PL096841.D	15 Aug 2025 22:10		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081825

Review By	Abdul	Review On	8/19/2025 8:40:55 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:05:16 AM
SubDirectory	PL081825	HP Acquire Method	HP Processing Method p1072825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL096842.D	18 Aug 2025 09:58		AR\AJ	Ok
2	I.BLK	I.BLK	PL096843.D	18 Aug 2025 10:12		AR\AJ	Ok,M
3	PEM	PEM	PL096844.D	18 Aug 2025 10:26		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL096845.D	18 Aug 2025 11:25		AR\AJ	Ok,M
5	PB169283BL	PB169283BL	PL096846.D	18 Aug 2025 11:56		AR\AJ	Ok
6	PB169283BS	PB169283BS	PL096847.D	18 Aug 2025 12:10	Recovery Fail delta-BHC-I	AR\AJ	Not Ok
7	Q2883-01	RBR2520553	PL096848.D	18 Aug 2025 12:24		AR\AJ	Ok,M
8	I.BLK	I.BLK	PL096849.D	18 Aug 2025 12:39		AR\AJ	Ok
9	PSTDCCC050	PSTDCCC050	PL096850.D	18 Aug 2025 12:53		AR\AJ	Ok,M
10	PCHLORCCC500	PCHLORCCC500	PL096851.D	18 Aug 2025 14:10		AR\AJ	Ok,M
11	PTOXCCC500	PTOXCCC500	PL096852.D	18 Aug 2025 14:48	CCC FAIL	AR\AJ	Not Ok
12	Q2864-03	LAW-25-113-121	PL096853.D	18 Aug 2025 15:58		AR\AJ	Ok
13	PB169264BL	PB169264BL	PL096854.D	18 Aug 2025 16:12		AR\AJ	Ok,M
14	Q2815-01	TW-705R-S	PL096855.D	18 Aug 2025 16:30	need dilution	AR\AJ	Dilution
15	Q2815-01DL	TW-705R-SDL	PL096856.D	18 Aug 2025 16:44		AR\AJ	Ok,M
16	I.BLK	I.BLK	PL096857.D	18 Aug 2025 16:57		AR\AJ	Ok,M
17	PSTDCCC050	PSTDCCC050	PL096858.D	18 Aug 2025 17:11		AR\AJ	Ok,M
18	PCHLORCCC500	PCHLORCCC500	PL096859.D	18 Aug 2025 17:25		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081825

Review By	Abdul	Review On	8/19/2025 8:40:55 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:05:16 AM
SubDirectory	PL081825	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2879-01	OU4-TS-GRILLO-TSCF	PL096860.D	18 Aug 2025 17:38		AR\AJ	Ok,M
20	Q2879-03	OU4-TS-GRILLO-TSCF	PL096861.D	18 Aug 2025 17:52		AR\AJ	Ok,M
21	Q2879-05	OU4-TS-GRILLO-TSCF	PL096862.D	18 Aug 2025 18:06		AR\AJ	Ok,M
22	Q2879-07	OU4-TS-GRILLO-TSCF	PL096863.D	18 Aug 2025 18:19	DCB low in 2nd column	AR\AJ	Ok,M
23	Q2879-09	OU4-TS-GRILLO-TSCF	PL096864.D	18 Aug 2025 18:33	DCB low in 2nd column	AR\AJ	Ok,M
24	Q2879-11	OU4-TS-GRILLO-TSCF	PL096865.D	18 Aug 2025 18:46		AR\AJ	Ok,M
25	I.BLK	I.BLK	PL096866.D	18 Aug 2025 19:00		AR\AJ	Ok,M
26	PEM	PEM	PL096867.D	18 Aug 2025 19:14		AR\AJ	Ok,M
27	PSTDCCC050	PSTDCCC050	PL096868.D	18 Aug 2025 19:41		AR\AJ	Ok,M
28	Q2879-13	OU4-TS-GRILLO-TSCF	PL096869.D	18 Aug 2025 20:22	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok
29	Q2879-15	OU4-TS-GRILLO-TSCF	PL096870.D	18 Aug 2025 20:35	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok
30	Q2879-17	OU4-TS-GRILLO-TSCF	PL096871.D	18 Aug 2025 20:49	Hit for Comp#17,END CCC FAIL FOR COM#17,DCB Low in both column	AR\AJ	Not Ok
31	Q2884-01	VNJ-232	PL096872.D	18 Aug 2025 21:02	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok
32	Q2884-03	302	PL096873.D	18 Aug 2025 21:16	END CCC FAIL FOR COM#17	AR\AJ	Not Ok
33	Q2888-01	VNJ-210	PL096874.D	18 Aug 2025 21:30	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok
34	Q2888-01MS	VNJ-210MS	PL096875.D	18 Aug 2025 21:43	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok
35	Q2888-01MSD	VNJ-210MSD	PL096876.D	18 Aug 2025 21:57	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081825

Review By	Abdul	Review On	8/19/2025 8:40:55 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:05:16 AM
SubDirectory	PL081825	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

36	Q2889-01	PL-02-081525	PL096877.D	18 Aug 2025 22:10	END CCC FAIL FOR COM#17	AR\AJ	Not Ok
37	I.BLK	I.BLK	PL096878.D	18 Aug 2025 22:38		AR\AJ	Ok,M
38	PSTDCCC050	PSTDCCC050	PL096879.D	18 Aug 2025 22:51	Comp#17,20 low in 2nd column	AR\AJ	Ok,M

M : Manual Integration

SOP ID: M3510C,3580A-Extraction Pesticide-17

Clean Up SOP #: Florisil

Matrix : Water

Weigh By: N/A

Balance check: N/A

Balance ID: N/A

pH Strip Lot#: E3880

Extraction By: RS

Filter By: RJ

pH Meter ID: N/A

Hood ID: 4,6,7

Extraction Start Date : 08/12/2025

Extraction Start Time : 10:20

Extraction End Date : 08/12/2025

Extraction End Time : 14:45

Concentration By: RS

Supervisor By : RUPESH

Extraction Method: ☒ Separatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP24627
Surrogate	1.0ML	200 PPB	PP24663
Spike Sol 2	2.0ML	1000 PPB	PP24622
Spike Sol 3	2.0ML	1000 PPB	PP24621
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3954
Baked Na2SO4	N/A	EP2632
Hexane	N/A	E3962
Florisil	N/A	E3927
9:1 Hexane:Acetone Mixture	N/A	EP2596
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40 BTS723, Q2814 samples Extract out of Holding Time.

KD Bath ID: WATER BATH-1,2

KD Bath Temperature: 60 °C

Envap ID: NEVAP-02


Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
8/12/25	RS (Ext Lab)	RS (Ext Lab)
14:50	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction Pesticide-17

Concentration Date: 08/12/2025

Sample ID	Client Sample ID	Test	g /mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB169225BL	PBLK225	Pesticide-TCL	1000	6	RUPESH	ritesh	10			SEP-1
PB169225BS	PLCS225	Pesticide-TCL	1000	6	RUPESH	ritesh	10			2
PB169225BS D	PLCSD225	Pesticide-TCL	1000	6	RUPESH	ritesh	10			3
Q2814-01	TW-84SB-E	Pesticide-TCL	1000	6	RUPESH	ritesh	10	G		4
Q2814-03	TW-17M-W	Pesticide-TCL	990	6	RUPESH	ritesh	10	G		5
Q2814-16	FB	Pesticide-TCL	500	6	RUPESH	ritesh	5	E		6
Q2814-17	TW-518R-S	Pesticide-TCL	990	6	RUPESH	ritesh	10	G		7
Q2815-01	TW-705R-S	Pesticide-TCL	1000	6	RUPESH	ritesh	10	E		8
Q2815-11	TW-22M-W	Pesticide-TCL	990	6	RUPESH	ritesh	10	E		9
	CHLORDANE	1	1000	6	1	1	10			10
	TOXAPHENE	1	1000	6	1	1	10			11



RS
8/12

* Extracts relinquished on the same date as received.

169228
10-20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q2814P WorkList ID : 191236 Department : Extraction Date : 08-12-2025 10:14:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2814-01	TW-84SB-E	Water	Pesticide-TCL	Cool 4 deg C	FIRS02	D41	08/05/2025	8081B
Q2814-03	TW-17M-W	Water	Pesticide-TCL	Cool 4 deg C	FIRS02	D41	08/05/2025	8081B
Q2814-16	FB	Water	Pesticide-TCL	Cool 4 deg C	FIRS02	D41	08/04/2025	8081B
Q2814-17	TW-518R-S	Water	Pesticide-TCL	Cool 4 deg C	FIRS02	D41	08/05/2025	8081B
Q2815-01	TW-705R-S	Water	Pesticide-TCL	Cool 4 deg C	FIRS02	D41	08/06/2025	8081B
Q2815-11	TW-22M-W	Water	Pesticide-TCL	Cool 4 deg C	FIRS02	D41	08/08/2025	8081B

Date/Time 8/12/25 10:15
Raw Sample Received by: RS (Ext-lab)
Raw Sample Relinquished by: CP SM

Date/Time 8/12/25 10:50
Raw Sample Received by: CP SM
Raw Sample Relinquished by: RS (Ext-lab)

Prep Standard - Chemical Standard Summary

Order ID : Q2815
Test : Pesticide-TCL
Prepbatch ID : PB169225,
Sequence ID/Qc Batch ID: pl081625,pl081825,

Standard ID :
EP2596,EP2632,PP24257,PP24259,PP24329,PP24433,PP24621,PP24622,PP24627,PP24651,PP24663,PP24738,PP24739,PP24740,PP24741,PP24742,PP24744,PP24745,PP24746,PP24747,PP24748,PP24749,PP24750,PP24751,PP24752,PP24753,PP24754,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764,

Chemical ID :
E3875,E3877,E3914,E3915,E3927,E3937,E3941,E3944,E3954,E3956,E3962,P12604,P12610,P12611,P13037,P13038,P13041,P13195,P13196,P13246,P13356,P13774,P13786,P13788,P13861,P13862,P9053,W3177,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1215	FLOROSIL CLEAN UP-WASHING SOLN	EP2596	03/27/2025	09/19/2025	Rajesh Parikh	None	None	Evelyn Huang
								03/27/2025

FROM 100.00000ml of E3915 + 900.00000ml of E3914 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2632	08/11/2025	01/28/2026	RUPESEKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel
								08/11/2025

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1472	20 PPM Pest Stock Solution 2nd Source	PP24257	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 1.00000ml of P13037 + 9.00000ml of E3877 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3663	20 PPM MIREX Stock STD (Secondary source)	PP24259	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.20000ml of P13195 + 9.80000ml of E3877 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
518	Pest/PCB I.BLK 20 PPB	PP24433	03/31/2025	08/22/2025	Abdul Mirza	None	None	Yogesh Patel
								04/02/2025

FROM 99.90000ml of E3914 + 0.10000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3878	1000 PPB TOXAPHENE SPIKE (RESTEK)	PP24621	06/04/2025	09/10/2025	Abdul Mirza	None	None	Yogesh Patel
								06/11/2025

FROM 0.10000ml of P13861 + 99.90000ml of E3937 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1501	1000 ppb CHLORDANE SPIKE (RESTEK)	PP24622	06/04/2025	09/09/2025	Abdul Mirza	None	None	Yogesh Patel
								06/11/2025

FROM 0.10000ml of P12611 + 99.90000ml of E3937 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
79	500 PPB Pesticide Spike Solution	PP24627	06/10/2025	08/12/2025	Abdul Mirza	None	None	Yogesh Patel
								06/11/2025

FROM 95.00000ml of E3937 + 2.50000ml of PP24257 + 2.50000ml of PP24259 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4027	Pesticide resolution Check Mixture 8081	PP24651	06/16/2025	12/11/2025	Abdul Mirza	None	None	Yogesh Patel
								07/22/2025

FROM 1.00000ml of P13246 + 99.00000ml of E3941 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24663	06/24/2025	12/24/2025	Abdul Mirza	None	None	Yogesh Patel
								07/21/2025

FROM 1.00000ml of P13786 + 999.00000ml of E3944 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3629	20 PPM PEST stock Solution 1st source(RESTEK)	PP24738	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13038 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1472	20 PPM Pest Stock Solution 2nd Source	PP24739	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13041 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1273	20 PPM Mirex Stock (Primary Source)	PP24740	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P9053 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3663	20 PPM MIREX Stock STD (Secondary source)	PP24741	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 1.00000ml of P13196 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24742	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 1.00000ml of P13788 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3630	100/100 PPB PEST Working std.1st Source(RESTEK)	PP24744	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 98.50000ml of E3956 + 0.50000ml of PP24738 + 0.50000ml of PP24740 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
80	100/100 PPB Pesticide Working Solution 2nd Source	PP24745	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 98.50000ml of E3956 + 0.50000ml of PP24739 + 0.50000ml of PP24741 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
386	1000/100 PPB Chlordane STD (Restek)	PP24746	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P12604 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3746	1000/100 ppb Chlordane STD-RESTEK 2ND SOURCE	PP24747	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P12610 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
383	1000/100 PPB Toxaphene STD (Restek)	PP24748	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P13774 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3669	1000/100 PPB TOXAPHENE STD 2nd source (RESTEK)	PP24749	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P13862 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3631	75 PPB ICAL PEST STD(RESTEK)	PP24750	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.25000ml of E3956 + 0.75000ml of PP24744 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3632	50 PPB ICAL PEST STD(RESTEK)	PP24751	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24744 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3633	25 PPB ICAL PEST STD(RESTEK)	PP24752	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.75000ml of E3956 + 0.25000ml of PP24744 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3634	5 PPB ICAL PEST STD(RESTEK)	PP24753	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.90000ml of E3956 + 0.10000ml of PP24751 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3988	50 PPB PEST ICV STD(RESTEK)	PP24754	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24745 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
528	CHLOR 750 PPB STD	PP24755	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.25000ml of E3956 + 0.75000ml of PP24746 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
529	CHLOR 500 PPB STD	PP24756	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24746 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
530	CHLOR 250 PPB STD	PP24757	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.75000ml of E3956 + 0.25000ml of PP24746 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3408	CHLOR 50 PPB STD	PP24758	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.90000ml of E3956 + 0.10000ml of PP24756 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
532	CHLOR 500 PPB ICV STD	PP24759	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24747 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
533	TOX 750 PPB STD	PP24760	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.25000ml of E3956 + 0.75000ml of PP24748 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
534	TOX 500 PPB STD	PP24761	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24748 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
535	TOX 250 PPB STD	PP24762	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.75000ml of E3956 + 0.25000ml of PP24748 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2217	TOX 100 PPB STD	PP24763	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.90000ml of E3956 + 0.10000ml of PP24748 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3670	TOX 500 PPB ICV std (RESTEK)	PP24764	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
FROM 0.50000ml of PP24749 = Final Quantity: 1.000 ml								

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	01/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	09/19/2025	03/19/2025 / RUPESH	03/13/2025 / RUPESH	E3914

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	09/26/2025	03/26/2025 / Rajesh	03/19/2025 / RUPESH	E3915

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
phenomenex	FS0006 / Cleanert SPE Silica, 1000 mg/6ml	Z0830QB1	04/18/2026	05/30/2025 / RUPESH	03/13/2025 / RUPESH	E3927

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	11/22/2025	05/22/2025 / RUPESH	05/14/2025 / RUPESH	E3937

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	12/11/2025	06/11/2025 / Rajesh	06/04/2025 / Rajesh	E3941

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	06/20/2025 / RUPESH	05/14/2025 / RUPESH	E3944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25B1862001	03/19/2026	07/14/2025 / RUPESH	06/11/2025 / RUPESH	E3954

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	07/16/2025 / RUPESH	07/16/2025 / RUPESH	E3956

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	08/05/2025 / RUPESH	07/30/2025 / RUPESH	E3962

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	01/21/2026	07/21/2025 / Abdul	07/03/2023 / Abdul	P12604

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	01/21/2026	07/21/2025 / Abdul	07/03/2023 / Abdul	P12610

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0193299	09/09/2025	03/10/2025 / Abdul	07/03/2023 / Abdul	P12611

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0200423	09/10/2025	03/10/2025 / Abdul	12/26/2023 / Abdul	P13037

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0200423	01/21/2026	07/21/2025 / Abdul	12/26/2023 / Abdul	P13038

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0199099	07/21/2026	07/21/2025 / Abdul	12/26/2023 / Abdul	P13041

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	042022	09/10/2025	03/10/2025 / Abdul	01/17/2024 / Abdul	P13195

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	042022	01/21/2026	07/21/2025 / Abdul	01/17/2024 / Abdul	P13196

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	19161 / 8081 pesticide resolution check mixture	013124	12/17/2025	06/17/2025 / Abdul	02/09/2024 / Abdul	P13246

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0203038	01/21/2026	07/21/2025 / Abdul	05/03/2024 / Ankita	P13774

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	12/24/2025	06/24/2025 / Abdul	11/19/2024 / Ankita	P13786

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	01/21/2026	07/21/2025 / Abdul	11/19/2024 / Ankita	P13788

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0210240	09/10/2025	03/10/2025 / Abdul	12/09/2024 / Abdul	P13861

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0210240	01/21/2026	07/21/2025 / Abdul	12/09/2024 / Abdul	P13862

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	112018	01/21/2026	07/21/2025 / Abdul	11/01/2019 / Stephen	P9053

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd. by RP on 2/12/25

 E3877

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd by RS on 3/14/25



E3914

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
 If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Recd. by RS on 3/19/25

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E3915

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Cleanert Florisil

1g/6ml 30/pkg

LOT#: Z0830QB1

MFG#: G01256

CAT# FS0006



固相萃取产品

Made in China

Agela Technologies

E3927



Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E3937

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Harout Sahagian

Recd. by RI on 6/11/25

E3941

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
 If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date: 2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3944

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 25B1862001
Manufactured Date: 2024-12-18
Expiration Date: 2026-03-19
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	99.9 %
Color (APHA)	≤ 10	5
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.3 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	<0.1
Chloride (Cl)	$\leq 10 \text{ ppm}$	<5 ppm
Water (by KF, coulometric)	$\leq 0.02 \%$	<0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

RS
7/14/25

E3954

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/16/25

E3956

Jamie Croak
Director Quality Operations, Bioscience Production

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/30/25

E3962

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

RESTEK

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0193299
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : April 30, 2029 **Storage:** 10°C or colder
Ship: Ambient

P12616
↓
P12615
Five
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	—%	1,010.0 µg/mL	+/- 56.0475

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Bryan Snyder
Bryan Snyder - Operations Tech I

Date Mixed: 06-Jan-2023

Balance Serial # B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

CRm
P 12611
↓
P 12615 } *5*
CRout
7/3/2023



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 Lot No.: A0199099

Description : Organochlorine Pesticide Mix AB #1
Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL Pkg Amt: > 1 mL

Expiration Date : June 30, 2027 Storage: 10°C or colder

Ship: Ambient

P130397
↓
P130431
12-26-2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.0 µg/mL	+/- 8.9732
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	200.1 µg/mL	+/- 8.9762
3	beta-BHC	319-85-7	BCCC6425	99%	200.3 µg/mL	+/- 8.9844
4	delta-BHC	319-86-8	14450800	98%	200.0 µg/mL	+/- 8.9740
5	Heptachlor	76-44-8	813251	99%	200.1 µg/mL	+/- 8.9754
6	Aldrin	309-00-2	14389400	98%	200.0 µg/mL	+/- 8.9718
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.1 µg/mL	+/- 8.9754
8	trans-Chlordane	5103-74-2	32943	98%	199.9 µg/mL	+/- 8.9696
9	cis-Chlordane	5103-71-9	31766	98%	200.1 µg/mL	+/- 8.9762
10	Endosulfan I	959-98-8	BCCF4060	99%	200.1 µg/mL	+/- 8.9754
11	4,4'-DDE	72-55-9	GHYQG	99%	200.1 µg/mL	+/- 8.9777
12	Dieldrin	60-57-1	11129900	98%	200.0 µg/mL	+/- 8.9718
13	Endrin	72-20-8	14123200	98%	199.9 µg/mL	+/- 8.9696
14	4,4'-DDD	72-54-8	HAN02	99%	200.1 µg/mL	+/- 8.9777
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	200.0 µg/mL	+/- 8.9718

17	Endrin aldehyde	7421-93-4	30720	98%	200.1 µg/mL	+/- 8.9784
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.0 µg/mL	+/- 8.9732
19	Methoxychlor	72-43-5	13668200	99%	200.1 µg/mL	+/- 8.9777
20	Endrin ketone	53494-70-5	1-ABS-16-7	98%	200.0 µg/mL	+/- 8.9740

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P 13039
 ↓
 P 13043
 1
 5
 12/26/23

Quality Confirmation Test

Column:
 30m x .25mm x .2um
 Rtx-CLP II (cat.# 11323)

Carrier Gas:
 helium-constant pressure 20 psi.

Temp. Program:
 150°C to 300°C
 @ 4°C/min. (hold 5 min.)


Inj. Temp:
 200°C

Det. Temp:
 300°C

Det. Type:
 ECD

Split Vent:
 Split ratio 50:1

Inj. Vol
 1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Josh McCloskey
 Josh McCloskey - Operations Technician I

Date Mixed: 19-Jun-2023 **Balance Serial #** 1128360905

Jennifer Pollino
 Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 23-Jun-2023

Manufactured under Restek's ISO 9001:2015
 Registered Quality System
 Certificate #FM 80397



CERTIFIED WEIGHT REPORT

Part Number: **79136**
Lot Number: **042022**
Description: **Mirex**

Solvent(s):
Acetone
Lot#: **81025**

		042022
Formulated By:	Prashant Chauhan	DATE
		042022
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: **042027**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **1000**
NIST Test ID#: **6UTB**




5E-05 Balance Uncertainty
0.006 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): **50.0**

SDS Information
(Solvent Safety Info. On Attached pg.)
Expanded Uncertainty (+/-) (µg/mL)
CAS# OSHA PEL (TWA) LD50

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Mirex	437	9492400	1000	99.4	0.5	0.05034	0.05040	1001.1	10.3	2385-85-5	N/A	ori-rat 306mg/kg

Method GC7MSD-1.M: Column: SPB-608 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 150°C (4min.), Temp 2 = 290°C (13.5 min.), Rate = 8°C/min., Injector B = 200°C, Detector B = 290°C. Split Ratio = 100:1, Scan Rate = 2. Analysis performed by Candice Warren.



P13195
↓
P13199
↓
(5)
↓
W. H. H. H.
01/17/2024

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5 % of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

01/17/2024
HMF
D 13/199
D 13/199
5



CERTIFIED WEIGHT REPORT

Part Number: **79136**
Lot Number: **042022**
Description: **Mirex**Solvent(s):
Acetone Lot#
81025

		042022
Formulated By:	Prashant Chauhan	DATE
		042022
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 042027

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration (µg/mL): 1000

NIST Test ID#: 6UTB




5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 50.0 0.006 Flask Uncertainty

Expanded
SDS Information
(Solvent Safety Info. On Attached pg.)

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Mirex	437	9492400	1000	99.4	0.5	0.05034	0.05040	1001.1	10.3	2385-85-5	N/A	ori-rat 306mg/kg

Method GC7MSD-1.M: Column: SPB-608 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 150°C (4min.), Temp 2 = 290°C (13.5 min.), Rate = 8°C/min., Injector B = 200°C, Detector B = 290°C. Split Ratio = 100:1, Scan Rate = 2. Analysis performed by Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5 % of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

P13195
↓
P13199
↓
(5)
↓
W. H. H. H.
01/17/2024

01/17/2024
HMF
D 13/199
D 13/199
5



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0197993
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Ship: Ambient

P 12603
↓
P 12605
[3]
Kauf
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----%	1,005.0 µg/mL	+/- 55.7700

Solvent: Hexane
CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Morgan Craighead - Mix Technician


Date Mixed: 11-May-2023

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 12603 } (3)
↓
P 12605

7/3/2023



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *chromatographic plus*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0197993
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Ship: Ambient

Handwritten: P 12606, P 12610, 5 Five, 7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. { hold 10 min. }

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 11-May-2023

Balance Serial # 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 **Lot No.:** A0200423

Description : Organochlorine Pesticide Mix AB #1
Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 10°C or colder

Ship: Ambient

P 13034
↓
P 13038
✓
12.26.2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.5 µg/mL	+/- 8.9956
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	199.9 µg/mL	+/- 8.9696
3	beta-BHC	319-85-7	BCCC6425	99%	200.0 µg/mL	+/- 8.9732
4	delta-BHC	319-86-8	14450800	98%	199.9 µg/mL	+/- 8.9696
5	Heptachlor	76-44-8	813251	99%	202.0 µg/mL	+/- 9.0629
6	Aldrin	309-00-2	14389400	98%	200.9 µg/mL	+/- 9.0136
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.0 µg/mL	+/- 8.9732
8	trans-Chlordane	5103-74-2	34616	99%	200.5 µg/mL	+/- 8.9956
9	cis-Chlordane	5103-71-9	31766	98%	201.4 µg/mL	+/- 9.0356
10	Endosulfan I	959-98-8	BCCF4060	99%	200.0 µg/mL	+/- 8.9732
11	4,4'-DDE	72-55-9	GHYQG	99%	201.5 µg/mL	+/- 9.0405
12	Dieldrin	60-57-1	14515000	98%	199.9 µg/mL	+/- 8.9696
13	Endrin	72-20-8	14485300	98%	200.4 µg/mL	+/- 8.9916
14	4,4'-DDD	72-54-8	HAN02	99%	200.5 µg/mL	+/- 8.9956
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	201.9 µg/mL	+/- 9.0575

17	Endrin aldehyde	7421-93-4	30720	98%	201.4 µg/mL	+/- 9.0356
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.5 µg/mL	+/- 8.9956
19	Methoxychlor	72-43-5	14563200	98%	200.9 µg/mL	+/- 9.0136
20	Endrin ketone	53494-70-5	14537700	98%	199.9 µg/mL	+/- 8.9696

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13034
P13038
1
5
12/26/2023

Quality Confirmation Test

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 50:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 31-Jul-2023

Balance Serial # B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 03-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 **Lot No.:** A0200423

Description : Organochlorine Pesticide Mix AB #1
Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 10°C or colder

Ship: Ambient

P 13034
↓
P 13038
12.26.2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.5 µg/mL	+/- 8.9956
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	199.9 µg/mL	+/- 8.9696
3	beta-BHC	319-85-7	BCCC6425	99%	200.0 µg/mL	+/- 8.9732
4	delta-BHC	319-86-8	14450800	98%	199.9 µg/mL	+/- 8.9696
5	Heptachlor	76-44-8	813251	99%	202.0 µg/mL	+/- 9.0629
6	Aldrin	309-00-2	14389400	98%	200.9 µg/mL	+/- 9.0136
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.0 µg/mL	+/- 8.9732
8	trans-Chlordane	5103-74-2	34616	99%	200.5 µg/mL	+/- 8.9956
9	cis-Chlordane	5103-71-9	31766	98%	201.4 µg/mL	+/- 9.0356
10	Endosulfan I	959-98-8	BCCF4060	99%	200.0 µg/mL	+/- 8.9732
11	4,4'-DDE	72-55-9	GHYQG	99%	201.5 µg/mL	+/- 9.0405
12	Dieldrin	60-57-1	14515000	98%	199.9 µg/mL	+/- 8.9696
13	Endrin	72-20-8	14485300	98%	200.4 µg/mL	+/- 8.9916
14	4,4'-DDD	72-54-8	HAN02	99%	200.5 µg/mL	+/- 8.9956
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	201.9 µg/mL	+/- 9.0575

17	Endrin aldehyde	7421-93-4	30720	98%	201.4 µg/mL	+/- 9.0356
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.5 µg/mL	+/- 8.9956
19	Methoxychlor	72-43-5	14563200	98%	200.9 µg/mL	+/- 9.0136
20	Endrin ketone	53494-70-5	14537700	98%	199.9 µg/mL	+/- 8.9696

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13034
P13038
1
5
12/26/2023

Quality Confirmation Test

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 50:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 31-Jul-2023

Balance Serial # B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 03-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED WEIGHT REPORT

Part Number: **19161**
 Lot Number: **013124**
 Description: **CLP Pesticides & PCB's Resolution Check Standard**
 9 components
 Expiration Date: **013129**
 Recommended Storage: **Refrigerate (4 °C)**
 Nominal Concentration (µg/mL): **Varied**
 NIST Test ID#: **6UTB**

Solvent(s):
 Hexane 273615 (50%)
 Toluene 28508 (50%)

Volume(s) shown below were combined and diluted to (mL): 100.0 5E-05 Balance Uncertainty 0.021 Flask Uncertainty

		013124
Formulated By:	Lawrence Barry	DATE
		013124
Reviewed By:	Pedro L. Rentas	DATE

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc.(ug/mL)	Final Conc.(ug/mL)	Expanded Uncertainty (+/-) µg/mL	SDS Information (Solvent Safety Info. On Attached pg.)		
									CAS#	OSHA PEL (TWA)	LD50
1. trans-Chlordane	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	5103-74-2	0.5mg/m3 (skin)	ori-rat 500mg/kg
2. Endosulfan I	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	959-98-8	0.1mg/m3 (skin)	ori-rat 18mg/kg
3. 4,4'-DDE	19361	013124	0.010	1.00	0.004	201.6	2.0	0.03	72-55-9	N/A	ori-rat 880mg/kg
4. Dieldrin	19361	013124	0.010	1.00	0.004	202.8	2.0	0.03	60-57-1	0.25mg/m3 (skin)	ori-rat 38300ug/kg
5. Endosulfan sulfate	19361	013124	0.010	1.00	0.004	204.2	2.0	0.03	1031-07-8	N/A	ori-rat 18mg/kg
6. Endrin ketone	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	53494-70-5	N/A	N/A
7. 4,4'-Methoxychlor	19361	013124	0.010	1.00	0.004	1000.7	10.0	0.09	72-43-5	10mg/m3	ori-rat 6000mg/kg
8. 2,4,5,6-Tetrachloro-m-xylene	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	877-09-8	N/A	N/A
9. Decachlorobiphenyl (209)	19361	013124	0.010	1.00	0.004	202.0	2.0	0.03	2051-24-3	N/A	N/A

• The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 • Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
 • Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 • All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 • Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

P13243 } (5)
 ↓
 P13247 }
 ↓
 02/9/2024



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



110 Benner Circle
Belleville, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0203038
Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : January 31, 2028 **Storage:** 10°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	—%	1,009.0 µg/mL	+/- 55.9920

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13773
↓
P13779
AI
11/19/24
05-06-2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I


Date Mixed: 10-Oct-2023

Balance Serial # 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397


Handwritten notes and signatures are present at the bottom of the page. On the left, 'P13388' is written with a downward arrow pointing to 'P13389'. To the right, 'P13773' is written with a downward arrow pointing to 'P13779'. Below these, 'AJ' is written above the date '11/19/24'. At the bottom, there is a signature and the date '05-06-2024'.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
↓
P19789 AJ
11/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

heliu-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A.O.E.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024

Balance Serial # B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
↓
P19789 AJ
11/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A.O.E.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024

Balance Serial # B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0210240

Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	----%	1,009.3 µg/mL	+/- 56.0105

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13861
P13862

12/9/2024

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Amanda Miller - Operations Tech III - ARM QC

Date Mixed: 11-Apr-2024 **Balance Serial #** B442140311


Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 26-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13861 } ②
P13862 }
↑

12/9/2024



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0210240

Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	----%	1,009.3 µg/mL	+/- 56.0105

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13861
P13862

12/9/2024

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Amanda Miller - Operations Tech III - ARM QC

Date Mixed: 11-Apr-2024 **Balance Serial #** B442140311



Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 26-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13861 } ②
P13862 }
↑

12/9/2024



CERTIFIED WEIGHT REPORT

Part Number: **72072**
Lot Number: **112018**
Description: **n-Tetracosane-d50**

Solvent(s):
Methylene chloride
Lot#: **102669**

Expiration Date: **112028**
Recommended Storage: **Ambient (20 °C)**
Nominal Concentration (µg/mL): **1000**
NIST Test ID#: **2684186**

Weight(s) shown below were combined and diluted to (mL):




200.0
0.058 Balance Uncertainty
0.058 Flask Uncertainty

Received by
SG on 11/1/19
P9044 - P9053


<i>Prashant Chauhan</i>		112018
Formulated By:	Prashant Chauhan	DATE
<i>Pedro Rentas</i>		112018
Reviewed By:	Pedro Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. n-Tetracosane-d50	2072	PR-17753/09216TC1	1000	98	0.2	0.20411	0.20415	1000.2	4.2	16416-32-3	N/A	N/A

Method GC8MSD-3.M: Column:SPB-5 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 50°C (1min.), Temp 2 = 300°C (9min.), Rate = 10°C/min., Injector B= 250°C, Detector B = 275°C, Split Ratio = 100:1, Scan Rate = 2. Analysis performed by: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Run 40, "P72072 L112018 [1000µg/mL in MeCl2]"

Run Length: 35.00 min, 20999 points at 10 points/second.
Created: Thu, Nov 22, 2018 at 7:23:18 AM.
Sampled: Sequence "112018-GC4M1", Method "GC4-M1".
Analyzed using Method "GC4-M1".

Comments

GC4-M1 Analysis by Melissa Stonier


Column ID SPB5 L#60062-01A : 30 meter x 0.53mm x 1.5µm Film Thickness

Flow rates: Total Flow = 300 mL/min, Helium (carrier) = 6.5 mL, Helium (make-up) = 25 mL, Hydrogen (detector) = 30 mL,
Air (detector) = 360 mL

Oven Temp 1 = 50°C (1 min), Rate = 10°C/min, Oven Temp 2 = 300°C (9 min), Total Run Time = 35 Minutes.

Injector Temp = 200°C, FID Temp = 300°C, FID Signal = eDag Channel 1.

Gas Chromatograph = HP 5890, Auto Sampler = HP 7673, Standard Injection = 0.5 µL, Range = 3



n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

W3147
W3147
02/03/2023
JP

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

J. Croak
Jamie Croak
Director Quality Operations, Bioscience
479 of 752 on



SHIPPING DOCUMENTS

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: _____
ADDRESS: _____
CITY: _____ STATE: _____ ZIP: _____
ATTENTION: _____
PHONE: _____ FAX: _____

CLIENT PROJECT INFORMATION

PROJECT NAME: _____
PROJECT NO.: _____ LOCATION: _____
PROJECT MANAGER: _____
e-mail: _____
PHONE: _____ FAX: _____

CLIENT BILLING INFORMATION

BILL TO: _____ PO#: _____
ADDRESS: _____
CITY: _____ STATE: _____ ZIP: _____
ATTENTION: _____ PHONE: _____

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

1 2 3 4 5 6 7 8 9
✓
509C
Metals

PRESERVATIVES

COMMENTS

← Specify Preservatives
A-HCl D-NaOH
B-HNO3 E-ICE
C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	TW-705R-S	GW		4	8/6/20	11:21	3	2	1								
2.	TW-705R-S																
3.	TW-10PC-W	GW		4	8/6/20	01:30	3	2	1								
4.	TW-10P-E	GW		4	8/6/20	01:40	3	2	1								
5.	TW-10P-S	GW		4	8/6/20	01:50	3	2	1								
6.	TW-10P-W	GW		4	8/6/20	2:00	3	2	1								
7.	TW-10P-N	GW		4	8/6/20	2:10	3	2	1								
8.	TW-88H-E	GW		4	8/7/20	8:35	3	2	1								
9.	TW-88H-N	GW		4	8/7/20	8:45	3	2	1								
10.	TW-88H-W	GW		4	8/7/20	11:10	3	2	1								

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. GMS	DATE/TIME: 6:10	RECEIVED BY: 1. K. J. 8.8.25	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☒ COOLER TEMP 3.5°C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	Comments:
RELINQUISHED BY SAMPLER: 3.	DATE/TIME:	RECEIVED BY: 3.	Comments:

Page ____ of ____

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete

☐ YES ☐ NO

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: _____
ADDRESS: _____
CITY: _____ STATE: _____ ZIP: _____
ATTENTION: _____
PHONE: _____ FAX: _____

CLIENT PROJECT INFORMATION

PROJECT NAME: _____
PROJECT NO.: _____ LOCATION: _____
PROJECT MANAGER: _____
e-mail: _____
PHONE: _____ FAX: _____

CLIENT BILLING INFORMATION

BILL TO: _____ PO#: _____
ADDRESS: _____
CITY: _____ STATE: _____ ZIP: _____
ATTENTION: _____ PHONE: _____

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

1	2	3	4	5	6	7	8	9

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER	
1.	TW-88H-S	GW		4	8/23	11:10	3	2	1									
2.	TW-22M-W	GW		4	8/25	11:10	3	2	1	3								
3.	TW-22M-S	GW		4	8/25	11:23	3	2	1									
4.	TW-22M-E	GW		4	8/25	11:33	3	2	1									
5.	TW-22M-N	GW		4	8/25	11:43	3	2	1									
6.																		
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. GME	DATE/TIME: 6:01	RECEIVED BY: 1. [Signature] 8.8.25	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☒ COOLER TEMP 3.5°C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	Comments:
RELINQUISHED BY SAMPLER: 3.	DATE/TIME:	RECEIVED BY: 3.	Page ____ of

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO

From: Yazmeen Gomez
Sent: Monday, August 11, 2025 2:55 PM
To: 'Daniel Ligon'
Cc: 'Alfred Smith'; Jordan Hedvat
Subject: RE: Bottleware Order - USACE018-44
Attachments: q2815.pdf; q2814.pdf

Good afternoon,

A few things -

- Below mentions Filtered and Unfiltered for 8 GW samples, however, we only received three unpreserved metals bottles - TW-84SB-E, TW-17M-W, and TW-11M-W.
- 8 samples not listed on the COC were received – TW-17M-E, TW-17M-S, TW-84SB-S, TW-84SB-W, TW-11M-W, TW-11M-E, TW-11M-S, TW-11M-N.
- PCB and PESTICIDE are not mentioned the GW COC's however, I went off the below request and do have PEST and PCB logged for some of the samples.
- Once I sign off the login summaries will be sent to you – please confirm everything looks good, or if anything needs to be updated.

Best Regards,



Yazmeen Gomez
Sr. Project Manager
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3147
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com   

From: Yazmeen Gomez <Yazmeen.Gomez@alliancetg.com>
Sent: Friday, August 1, 2025 11:03 AM
To: Jordan Hedvat <Jordan.Hedvat@AllianceTG.com>; Daniel Ligon <DLigon@firstenvironment.com>
Cc: Alfred Smith <asmith@firstenvironment.com>
Subject: RE: Bottleware Order - USACE018-44

Hi Daniel,

Bottle order delivery for today confirmed.

Best Regards,



Yazmeen Gomez
Sr. Project Manager
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3147
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com   

From: Jordan Hedvat <Jordan.Hedvat@alliancetg.com>
Sent: Friday, August 1, 2025 8:14 AM
To: Daniel Ligon <DLigon@firstenvironment.com>
Cc: Alfred Smith <asmith@firstenvironment.com>; Yazmeen Gomez <Yazmeen.Gomez@alliancetg.com>
Subject: Re: Bottleware Order - USACE018-44

Hi Daniel,

We will process the bottle order for delivery. Do you need TeraCore or Encores for VOC soil? If not we will provide jars for those tests as well. Please reach out if we can help with anything else.

Regards,

Jordan

Jordan Hedvat
Account Executive, Environmental Laboratories
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3144
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

From: Daniel Ligon <DLigon@firstenvironment.com>
Sent: Thursday, July 31, 2025 4:53 PM
To: Jordan Hedvat <Jordan.Hedvat@alliancetg.com>
Cc: Alfred Smith <asmith@firstenvironment.com>
Subject: Bottleware Order - USACE018-44

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Hi Jordan – I have a bottleware order for some sampling we have next week in Long Island, NY. We'd like to have this delivered tomorrow if possible – sorry for the late email. First thing Monday morning would work as well for delivery to our office in Butler. See the following for # of samples we need with the parameters:

- 48x soil VOCs via EPA Method 8260
- 48x soil SVOCs via EPA Method 8270

- 8x soil Pesticides/PCBs via EPA Methods 8081/8082
- 8x soil TAL Metals (filtered and unfiltered)
- 48x groundwater VOCs via EPA Method 8260
- 48x groundwater SVOCs via EPA Method 8270
- 8x groundwater Pesticides/PCBs via EPA Methods 8081/8082
- 8x groundwater TAL Metals (filtered and unfiltered)
- 3 soil, 3 groundwater field duplicates
- 6 field blanks
- Trip blanks in every cooler

Let me know if this can be accommodated. Thanks!

Daniel Ligon
Environmental Specialist



First Environment, Inc.
10 Park Place, Bldg 1A, Suite 504
Butler, NJ 07405
Ph: 973-334-0003 ext. 224
Cell: 973-873-8515
DLigon@firstenvironment.com
www.firstenvironment.com

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2815 **FIRS02**

Order Date : 8/11/2025 10:33:09 AM

Project Mgr :

Client Name : First Environment, Inc.

Project Name : USACE018-44 DOD

Report Type : Level 4

Client Contact : Al Smith

Receive DateTime : 8/8/2025 6:01:00 PM

EDD Type : EQUIS

Invoice Name : First Environment, Inc.

Purchase Order :

Hard Copy Date :

Invoice Contact : Al Smith

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2815-01	TW-705R-S	Water	08/06/2025	11:21	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-02	TW-10PC-W	Water	08/06/2025	01:30	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-03	TW-10P-E	Water	08/06/2025	01:40	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-04	TW-10P-S	Water	08/06/2025	01:50	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-05	TW-10P-W	Water	08/06/2025	02:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-06	TW-10P-N	Water	08/06/2025	02:10	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-07	TW-88H-E	Water	08/07/2025	08:35	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-08	TW-88H-N	Water	08/07/2025	08:45					

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2815 FIRS02

Order Date : 8/11/2025 10:33:09 AM

Project Mgr :

Client Name : First Environment, Inc.

Project Name : USACE018-44 DOD

Report Type : Level 4

Client Contact : Al Smith

Receive DateTime : 8/8/2025 6:01:00 PM

EDD Type : EQUIS

Invoice Name : First Environment, Inc.

Purchase Order :

Hard Copy Date :

Invoice Contact : Al Smith

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2815-09	TW-88H-W	Water	08/07/2025	11:10	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-10	TW-88H-S	Water	08/07/2025	11:10	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-11	TW-22M-W	Water	08/08/2025	11:10	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-12	TW-22M-S	Water	08/08/2025	11:23	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-13	TW-22M-E	Water	08/08/2025	11:33	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-14	TW-22M-N	Water	08/08/2025	11:43	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-15	TW-17M-E	Water	08/08/2025	00:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2815	FIRS02	Order Date : 8/11/2025 10:33:09 AM	Project Mgr :
Client Name : First Environment, Inc.		Project Name : USACE018-44 DOD	Report Type : Level 4
Client Contact : Al Smith		Receive DateTime : 8/8/2025 6:01:00 PM	EDD Type : EQUIS
Invoice Name : First Environment, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Al Smith			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2815-16	TW-17M-S	Water	08/08/2025	00:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-17	TW-84SB-S	Water	08/08/2025	00:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-18	TW-84SB-W	Water	08/08/2025	00:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-19	DUP	Water	08/08/2025	00:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-20	TW-11M-W	Water	08/08/2025	00:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-21	TW-11M-E	Water	08/08/2025	00:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-22	TW-11M-S	Water	08/08/2025	00:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-23	TW-11M-N	Water	08/08/2025	00:00					

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2815	FIRS02	Order Date : 8/11/2025 10:33:09 AM	Project Mgr :
Client Name : First Environment, Inc.		Project Name : USACE018-44 DOD	Report Type : Level 4
Client Contact : Al Smith		Receive DateTime : 8/8/2025 6:01:00 PM	EDD Type : EQUIS
Invoice Name : First Environment, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Al Smith			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2815-24	TB	Water	08/08/2025	00:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
Q2815-26	FB	Water	08/08/2025	00:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
					VOC-TCLVOA-10		8260-Low	10 Bus. Days	

Relinquished By :

Date / Time :

Received By :

Date / Time :

Storage Area : VOA Refridgerator Room

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096594.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 16:25
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 08:57:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 08:58:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.828	66676938	99261210	20.962	20.776
28)	SA Decachlor...	9.017	7.992	50725219	90556842	21.271	20.874
Target Compounds							
2)	A alpha-BHC	3.983	3.333	48281852	72680702	10.436	10.270
3)	MA gamma-BHC...	4.311	3.666	46672698	68921884	10.551	10.437
6)	B beta-BHC	4.498	3.962	18987747	30564619	10.520	10.835
14)	MA Endrin	6.543	5.710	163.5E6	289.9E6	54.016	53.632
17)	MA 4,4'-DDT	6.990	6.109	314.0E6	562.5E6	109.497	111.208
18)	B Endrin al...	0.000	6.180	0	12998853	N.D.	0.219 #
20)	A Methoxychlor	7.463	6.682	392.3E6	694.1E6	267.202	253.290



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096594.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 16:25
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 08:57:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 08:58:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 66676938
Conc: 20.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 99261210
Conc: 20.78 ng/ml




#2 alpha-BHC

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 48281852
Conc: 10.44 ng/ml



#2 alpha-BHC


R.T.: 3.333 min
Delta R.T.: 0.000 min
Response: 72680702
Conc: 10.27 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 46672698
Conc: 10.55 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#3 gamma-BHC (Lindane)

R.T.: 3.666 min
Delta R.T.: 0.000 min
Response: 68921884
Conc: 10.44 ng/ml




#6 beta-BHC

R.T.: 4.498 min
Delta R.T.: 0.000 min
Response: 18987747
Conc: 10.52 ng/ml



#6 beta-BHC


R.T.: 3.962 min
Delta R.T.: 0.000 min
Response: 30564619
Conc: 10.84 ng/ml



#14 Endrin


R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 163494961
Conc: 54.02 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#14 Endrin

R.T.: 5.710 min
Delta R.T.: 0.000 min
Response: 289935231
Conc: 53.63 ng/ml




#17 4,4'-DDT

R.T.: 6.990 min
Delta R.T.: 0.000 min
Response: 314003596
Conc: 109.50 ng/ml



#17 4,4'-DDT


R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 562525716
Conc: 111.21 ng/ml



#18 Endrin aldehyde


R.T.: 0.000 min
Exp R.T.: 6.884 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PEM




#18 Endrin aldehyde

R.T.: 6.180 min
Delta R.T.: 0.001 min
Response: 12998853
Conc: 0.22 ng/ml




#20 Methoxychlor

R.T.: 7.463 min
Delta R.T.: 0.000 min
Response: 392347777
Conc: 267.20 ng/ml



#20 Methoxychlor


R.T.: 6.682 min
Delta R.T.: 0.000 min
Response: 694139453
Conc: 253.29 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.017 min
Delta R.T.: 0.000 min
Response: 50725219
Conc: 21.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM



#28 Decachlorobiphenyl

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 90556842
Conc: 20.87 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 16:38
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 RESCHK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 08:57:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 08:58:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.828	53837096	81066871	16.926	16.968
28)	SA Decachlor...	9.016	7.993	41967810	75251539	17.599	17.346
Target Compounds							
9)	A Endosulfan I	6.045	5.170	28022040	48036256	7.792	8.684
10)	B gamma-Chl...	5.916	5.052	31638592	48929496	8.314	8.321
12)	B 4,4'-DDE	6.166	5.304	54660848	96001382	17.022	17.409
13)	MA Dieldrin	6.316	5.435	63095073	99918926	17.001	16.917
19)	B Endosulfa...	7.118	6.403	50325612	88476396	17.513	17.395
20)	A Methoxychlor	7.463	6.682	128.9E6	238.4E6	87.784	87.000
21)	B Endrin ke...	7.598	6.907	51693529	95186865	17.193	17.090
25)	Chlordane-3	5.916	5.052	31638592	48929496	47.368	70.372 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


```
Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\  
Data File : PL096595.D  
Signal(s) : Signal #1: ECD1A.ch  Signal #2: ECD2B.ch  
Acq On    : 28 Jul 2025  16:38  
Operator  : AR\AJ  
Sample    : RESCHK  
Misc      :  
ALS Vial  : 4    Sample Multiplier: 1
```

Instrument :
ECD_L
ClientSampleId :
RESCHK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 08:57:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 08:58:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 53837096
Conc: 16.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 81066871
Conc: 16.97 ng/ml




#9 Endosulfan I

R.T.: 6.045 min
Delta R.T.: 0.001 min
Response: 28022040
Conc: 7.79 ng/ml



#9 Endosulfan I


R.T.: 5.170 min
Delta R.T.: 0.000 min
Response: 48036256
Conc: 8.68 ng/ml



#10 gamma-Chlordane


R.T.: 5.916 min
Delta R.T.: 0.000 min
Response: 31638592
Conc: 8.31 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#10 gamma-Chlordane

R.T.: 5.052 min
Delta R.T.: 0.000 min
Response: 48929496
Conc: 8.32 ng/ml




#12 4,4'-DDE

R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 54660848
Conc: 17.02 ng/ml



#12 4,4'-DDE


R.T.: 5.304 min
Delta R.T.: 0.000 min
Response: 96001382
Conc: 17.41 ng/ml



#13 Dieldrin


R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 63095073
Conc: 17.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#13 Dieldrin

R.T.: 5.435 min
Delta R.T.: 0.000 min
Response: 99918926
Conc: 16.92 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.118 min
Delta R.T.: 0.000 min
Response: 50325612
Conc: 17.51 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.403 min
Delta R.T.: 0.001 min
Response: 88476396
Conc: 17.40 ng/ml



#20 Methoxychlor


R.T.: 7.463 min
Delta R.T.: 0.000 min
Response: 128897891
Conc: 87.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#20 Methoxychlor

R.T.: 6.682 min
Delta R.T.: 0.000 min
Response: 238424041
Conc: 87.00 ng/ml




#21 Endrin ketone

R.T.: 7.598 min
Delta R.T.: 0.000 min
Response: 51693529
Conc: 17.19 ng/ml



#21 Endrin ketone


R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 95186865
Conc: 17.09 ng/ml



#25 Chlordane-3


R.T.: 5.916 min
Delta R.T.: 0.000 min
Response: 31638592
Conc: 47.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#25 Chlordane-3

R.T.: 5.052 min
Delta R.T.: 0.000 min
Response: 48929496
Conc: 70.37 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.016 min
Delta R.T.: 0.000 min
Response: 41967810
Conc: 17.60 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.993 min
Delta R.T.: 0.000 min
Response: 75251539
Conc: 17.35 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096599.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 17:33
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 07:45:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 07:36:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.828	80667083	120.4E6	25.718	25.447
28)	SA Decachlor...	9.016	7.992	61253778	110.7E6	26.443	26.222
Target Compounds							
2)	A alpha-BHC	3.983	3.333	113.7E6	176.1E6	24.483	24.678
3)	MA gamma-BHC...	4.311	3.665	109.5E6	164.7E6	24.930	24.831
4)	MA Heptachlor	4.903	4.014	103.9E6	167.4E6	25.150	25.138
5)	MB Aldrin	5.242	4.297	106.1E6	154.6E6	24.852	24.810
6)	B beta-BHC	4.497	3.961	46424655	71835591	26.066	25.815
7)	B delta-BHC	4.743	4.195	98875722	161.7E6	24.853	24.903
8)	B Heptachlo...	5.662	4.799	97796537	143.3E6	25.420	25.140
9)	A Endosulfan I	6.044	5.170	89274310	138.5E6	25.243	25.835
10)	B gamma-Chl...	5.916	5.051	95166219	147.9E6	25.039	25.072
11)	B alpha-Chl...	5.996	5.115	96944619	147.4E6	25.499	25.414
12)	B 4,4'-DDE	6.166	5.303	77835138	135.9E6	24.731	24.536
13)	MA Dieldrin	6.316	5.434	91962286	148.1E6	24.768	25.106
14)	MA Endrin	6.542	5.709	76652120	133.6E6	25.664	24.852
15)	B Endosulfa...	6.755	6.001	79391270	129.4E6	26.092	25.256
16)	A 4,4'-DDD	6.675	5.856	62524397	117.6E6	24.773	25.050
17)	MA 4,4'-DDT	6.989	6.109	70933756	126.1E6	24.589	24.645
18)	B Endrin al...	6.884	6.179	54110594	101.7E6	25.464	27.119
19)	B Endosulfa...	7.117	6.403	71756969	128.4E6	25.163	25.362
20)	A Methoxychlor	7.462	6.681	37688830	69721101	25.500	25.524
21)	B Endrin ke...	7.597	6.907	75730573	138.5E6	25.211	25.044
22)	Mirex	8.076	7.097	63693911	111.6E6	26.118	26.231



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096599.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 17:33
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 07:45:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 07:36:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 80667083
Conc: 25.72 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 120387362
Conc: 25.45 ng/ml




#2 alpha-BHC

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 113727400
Conc: 24.48 ng/ml



#2 alpha-BHC


R.T.: 3.333 min
Delta R.T.: 0.000 min
Response: 176134463
Conc: 24.68 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 109455262
Conc: 24.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#3 gamma-BHC (Lindane)

R.T.: 3.665 min
Delta R.T.: 0.000 min
Response: 164738265
Conc: 24.83 ng/ml




#4 Heptachlor

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 103855678
Conc: 25.15 ng/ml



#4 Heptachlor


R.T.: 4.014 min
Delta R.T.: 0.000 min
Response: 167381908
Conc: 25.14 ng/ml



#5 Aldrin


R.T.: 5.242 min
Delta R.T.: 0.000 min
Response: 106077325
Conc: 24.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#5 Aldrin

R.T.: 4.297 min
Delta R.T.: 0.000 min
Response: 154569484
Conc: 24.81 ng/ml




#6 beta-BHC

R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 46424655
Conc: 26.07 ng/ml



#6 beta-BHC


R.T.: 3.961 min
Delta R.T.: 0.000 min
Response: 71835591
Conc: 25.81 ng/ml



#7 delta-BHC


R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 98875722
Conc: 24.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#7 delta-BHC

R.T.: 4.195 min
Delta R.T.: 0.000 min
Response: 161747114
Conc: 24.90 ng/ml




#8 Heptachlor epoxide

R.T.: 5.662 min
Delta R.T.: 0.000 min
Response: 97796537
Conc: 25.42 ng/ml



#8 Heptachlor epoxide


R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 143274077
Conc: 25.14 ng/ml



#9 Endosulfan I


R.T.: 6.044 min
Delta R.T.: 0.000 min
Response: 89274310
Conc: 25.24 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#9 Endosulfan I

R.T.: 5.170 min
Delta R.T.: 0.000 min
Response: 138547710
Conc: 25.84 ng/ml




#10 gamma-Chlordane

R.T.: 5.916 min
Delta R.T.: 0.000 min
Response: 95166219
Conc: 25.04 ng/ml



#10 gamma-Chlordane


R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 147883477
Conc: 25.07 ng/ml



#11 alpha-Chlordane


R.T.: 5.996 min
Delta R.T.: 0.000 min
Response: 96944619
Conc: 25.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#11 alpha-Chlordane

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 147407858
Conc: 25.41 ng/ml




#12 4,4'-DDE

R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 77835138
Conc: 24.73 ng/ml



#12 4,4'-DDE


R.T.: 5.303 min
Delta R.T.: 0.000 min
Response: 135946511
Conc: 24.54 ng/ml



#13 Dieldrin


R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 91962286
Conc: 24.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#13 Dieldrin

R.T.: 5.434 min
Delta R.T.: 0.000 min
Response: 148077198
Conc: 25.11 ng/ml




#14 Endrin

R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 76652120
Conc: 25.66 ng/ml



#14 Endrin


R.T.: 5.709 min
Delta R.T.: 0.000 min
Response: 133610110
Conc: 24.85 ng/ml



#15 Endosulfan II


R.T.: 6.755 min
Delta R.T.: 0.000 min
Response: 79391270
Conc: 26.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#15 Endosulfan II

R.T.: 6.001 min
Delta R.T.: 0.000 min
Response: 129380147
Conc: 25.26 ng/ml




#16 4,4'-DDD

R.T.: 6.675 min
Delta R.T.: 0.000 min
Response: 62524397
Conc: 24.77 ng/ml



#16 4,4'-DDD


R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 117568590
Conc: 25.05 ng/ml



#17 4,4'-DDT


R.T.: 6.989 min
Delta R.T.: 0.000 min
Response: 70933756
Conc: 24.59 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#17 4,4'-DDT

R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 126146252
Conc: 24.65 ng/ml




#18 Endrin aldehyde

R.T.: 6.884 min
Delta R.T.: 0.000 min
Response: 54110594
Conc: 25.46 ng/ml



#18 Endrin aldehyde


R.T.: 6.179 min
Delta R.T.: 0.000 min
Response: 101733323
Conc: 27.12 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.117 min
Delta R.T.: 0.000 min
Response: 71756969
Conc: 25.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#19 Endosulfan Sulfate

R.T.: 6.403 min
Delta R.T.: 0.000 min
Response: 128398736
Conc: 25.36 ng/ml




#20 Methoxychlor

R.T.: 7.462 min
Delta R.T.: 0.000 min
Response: 37688830
Conc: 25.50 ng/ml



#20 Methoxychlor


R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 69721101
Conc: 25.52 ng/ml



#21 Endrin ketone


R.T.: 7.597 min
Delta R.T.: 0.000 min
Response: 75730573
Conc: 25.21 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025




#21 Endrin ketone

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 138475607
Conc: 25.04 ng/ml




#22 Mirex

R.T.: 8.076 min
Delta R.T.: 0.000 min
Response: 63693911
Conc: 26.12 ng/ml



#22 Mirex


R.T.: 7.097 min
Delta R.T.: 0.000 min
Response: 111590978
Conc: 26.23 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.016 min
Delta R.T.: 0.000 min
Response: 61253778
Conc: 26.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC025



#28 Decachlorobiphenyl

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 110748988
Conc: 26.22 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096600.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 17:47
 Operator : AR\AJ
 Sample : PSTDICC005
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC005

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 07:45:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 07:36:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	3.534	2.827	16289913	24636469	5.193	5.207
28) SA Decachlor...	9.016	7.992	13081800	24155779	5.647	5.719
Target Compounds						
2) A alpha-BHC	3.981	3.333	21891026	33671129	4.713	4.718
3) MA gamma-BHC...	4.310	3.664	21804234	32260997	4.966	4.863
4) MA Heptachlor	4.902	4.013	22657647	33277461	5.487	4.998
5) MB Aldrin	5.242	4.296	23335409	30546909	5.467	4.903
6) B beta-BHC	4.497	3.961	9319336	14886777	5.233	5.350
7) B delta-BHC	4.742	4.194	19209773	31362311	4.828	4.829
8) B Heptachlo...	5.662	4.798	21916244	28926139	5.697	5.076
9) A Endosulfan I	6.043	5.169	18404240	30432392	5.204	5.675
10) B gamma-Chl...	5.915	5.050	18527805	28851412	4.875	4.891
11) B alpha-Chl...	5.996	5.115	19648624	31301796	5.168	5.397
12) B 4,4'-DDE	6.165	5.303	16169148	26895387	5.138	4.854
13) MA Dieldrin	6.315	5.434	17886564	29267822	4.817	4.962
14) MA Endrin	6.541	5.709	14803966	27745055	4.957	5.161
15) B Endosulfa...	6.754	6.000	16262556	26018008	5.345	5.079
16) A 4,4'-DDD	6.674	5.856	12297723	23708739	4.873	5.052
17) MA 4,4'-DDT	6.989	6.108	13672011	24000169	4.739	4.689
18) B Endrin al...	6.884	6.178	11040304	29990732	5.195	7.995 #
19) B Endosulfa...	7.117	6.402	14782385	26237528	5.184	5.182
20) A Methoxychlor	7.462	6.681	7310793	14153462	4.946	5.181
21) B Endrin ke...	7.597	6.907	15006830	27568439	4.996	4.986
22) Mirex	8.076	7.097	13350221	24190284	5.474	5.686



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096600.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 17:47
Operator : AR\AJ
Sample : PSTDICC005
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC005

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 07:45:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 07:36:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.534 min
Delta R.T.: 0.000 min
Response: 16289913
Conc: 5.19 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#1 Tetrachloro-m-xylene

R.T.: 2.827 min
Delta R.T.: 0.000 min
Response: 24636469
Conc: 5.21 ng/ml




#2 alpha-BHC

R.T.: 3.981 min
Delta R.T.: 0.000 min
Response: 21891026
Conc: 4.71 ng/ml



#2 alpha-BHC


R.T.: 3.333 min
Delta R.T.: 0.000 min
Response: 33671129
Conc: 4.72 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.310 min
Delta R.T.: 0.000 min
Response: 21804234
Conc: 4.97 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#3 gamma-BHC (Lindane)

R.T.: 3.664 min
Delta R.T.: 0.000 min
Response: 32260997
Conc: 4.86 ng/ml




#4 Heptachlor

R.T.: 4.902 min
Delta R.T.: 0.000 min
Response: 22657647
Conc: 5.49 ng/ml



#4 Heptachlor


R.T.: 4.013 min
Delta R.T.: 0.000 min
Response: 33277461
Conc: 5.00 ng/ml



#5 Aldrin


R.T.: 5.242 min
Delta R.T.: -0.001 min
Response: 23335409
Conc: 5.47 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#5 Aldrin

R.T.: 4.296 min
Delta R.T.: 0.000 min
Response: 30546909
Conc: 4.90 ng/ml




#6 beta-BHC

R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 9319336
Conc: 5.23 ng/ml



#6 beta-BHC


R.T.: 3.961 min
Delta R.T.: 0.000 min
Response: 14886777
Conc: 5.35 ng/ml



#7 delta-BHC


R.T.: 4.742 min
Delta R.T.: 0.000 min
Response: 19209773
Conc: 4.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#7 delta-BHC

R.T.: 4.194 min
Delta R.T.: 0.000 min
Response: 31362311
Conc: 4.83 ng/ml




#8 Heptachlor epoxide

R.T.: 5.662 min
Delta R.T.: 0.000 min
Response: 21916244
Conc: 5.70 ng/ml



#8 Heptachlor epoxide


R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 28926139
Conc: 5.08 ng/ml



#9 Endosulfan I


R.T.: 6.043 min
Delta R.T.: 0.000 min
Response: 18404240
Conc: 5.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#9 Endosulfan I

R.T.: 5.169 min
Delta R.T.: 0.000 min
Response: 30432392
Conc: 5.67 ng/ml




#10 gamma-Chlordane

R.T.: 5.915 min
Delta R.T.: 0.000 min
Response: 18527805
Conc: 4.87 ng/ml



#10 gamma-Chlordane


R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 28851412
Conc: 4.89 ng/ml



#11 alpha-Chlordane


R.T.: 5.996 min
Delta R.T.: 0.000 min
Response: 19648624
Conc: 5.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#11 alpha-Chlordane

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 31301796
Conc: 5.40 ng/ml




#12 4,4'-DDE

R.T.: 6.165 min
Delta R.T.: 0.000 min
Response: 16169148
Conc: 5.14 ng/ml



#12 4,4'-DDE


R.T.: 5.303 min
Delta R.T.: 0.000 min
Response: 26895387
Conc: 4.85 ng/ml



#13 Dieldrin


R.T.: 6.315 min
Delta R.T.: 0.000 min
Response: 17886564
Conc: 4.82 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#13 Dieldrin

R.T.: 5.434 min
Delta R.T.: 0.000 min
Response: 29267822
Conc: 4.96 ng/ml




#14 Endrin

R.T.: 6.541 min
Delta R.T.: -0.001 min
Response: 14803966
Conc: 4.96 ng/ml



#14 Endrin


R.T.: 5.709 min
Delta R.T.: 0.000 min
Response: 27745055
Conc: 5.16 ng/ml



#15 Endosulfan II


R.T.: 6.754 min
Delta R.T.: -0.001 min
Response: 16262556
Conc: 5.34 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#15 Endosulfan II

R.T.: 6.000 min
Delta R.T.: 0.000 min
Response: 26018008
Conc: 5.08 ng/ml




#16 4,4'-DDD

R.T.: 6.674 min
Delta R.T.: 0.000 min
Response: 12297723
Conc: 4.87 ng/ml



#16 4,4'-DDD


R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 23708739
Conc: 5.05 ng/ml



#17 4,4'-DDT


R.T.: 6.989 min
Delta R.T.: 0.000 min
Response: 13672011
Conc: 4.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#17 4,4'-DDT

R.T.: 6.108 min
Delta R.T.: 0.000 min
Response: 24000169
Conc: 4.69 ng/ml




#18 Endrin aldehyde

R.T.: 6.884 min
Delta R.T.: 0.000 min
Response: 11040304
Conc: 5.20 ng/ml



#18 Endrin aldehyde


R.T.: 6.178 min
Delta R.T.: 0.000 min
Response: 29990732
Conc: 7.99 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.117 min
Delta R.T.: 0.000 min
Response: 14782385
Conc: 5.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#19 Endosulfan Sulfate

R.T.: 6.402 min
Delta R.T.: 0.000 min
Response: 26237528
Conc: 5.18 ng/ml




#20 Methoxychlor

R.T.: 7.462 min
Delta R.T.: 0.000 min
Response: 7310793
Conc: 4.95 ng/ml



#20 Methoxychlor


R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 14153462
Conc: 5.18 ng/ml



#21 Endrin ketone


R.T.: 7.597 min
Delta R.T.: 0.000 min
Response: 15006830
Conc: 5.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005




#21 Endrin ketone

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 27568439
Conc: 4.99 ng/ml




#22 Mirex

R.T.: 8.076 min
Delta R.T.: 0.000 min
Response: 13350221
Conc: 5.47 ng/ml



#22 Mirex


R.T.: 7.097 min
Delta R.T.: 0.000 min
Response: 24190284
Conc: 5.69 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.016 min
Delta R.T.: 0.000 min
Response: 13081800
Conc: 5.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDICC005



#28 Decachlorobiphenyl

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 24155779
Conc: 5.72 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
 Data File : PL096611.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2025 20:17
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL072825

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 08:47:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 29 08:20:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.828	161.0E6	247.3E6	50.630	51.756
28)	SA Decachlor...	9.015	7.992	117.1E6	221.9E6	49.100	51.159
Target Compounds							
2)	A alpha-BHC	3.983	3.334	237.2E6	373.0E6	51.282	52.705
3)	MA gamma-BHC...	4.311	3.666	223.6E6	346.0E6	50.545	52.400
4)	MA Heptachlor	4.903	4.014	211.3E6	346.4E6	50.843	51.992
5)	MB Aldrin	5.243	4.297	215.6E6	324.1E6	50.122	52.220
6)	B beta-BHC	4.497	3.961	91859349	145.4E6	50.893	51.555
7)	B delta-BHC	0.000	4.195	0	338.7E6	N.D.	52.325 #
8)	B Heptachlo...	5.662	4.799	194.2E6	295.6E6	50.346	51.754
9)	A Endosulfan I	6.044	5.170	178.4E6	277.2E6	49.613	50.116
10)	B gamma-Chl...	5.916	5.051	193.4E6	305.7E6	50.837	51.997
11)	B alpha-Chl...	5.996	5.115	191.7E6	300.1E6	49.735	50.869
12)	B 4,4'-DDE	6.166	5.304	158.1E6	286.3E6	49.223	51.924
13)	MA Dieldrin	6.316	5.435	186.8E6	306.3E6	50.341	51.858
14)	MA Endrin	6.543	5.709	147.3E6	272.8E6	48.674	50.462
15)	B Endosulfa...	6.755	6.001	158.8E6	265.2E6	49.582	51.632
16)	A 4,4'-DDD	6.675	5.856	126.9E6	242.0E6	50.209	51.423
17)	MA 4,4'-DDT	6.989	6.109	144.5E6	261.7E6	50.400	51.730
18)	B Endrin al...	6.883	6.179	108.5E6	197.4E6	50.570	53.386
19)	B Endosulfa...	7.117	6.402	143.9E6	262.8E6	50.068	51.670
20)	A Methoxychlor	7.462	6.681	74330342	140.5E6	50.622	51.276
21)	B Endrin ke...	7.597	6.907	153.3E6	289.7E6	50.977	52.006
22)	Mirex	8.076	7.097	123.5E6	221.2E6	49.793	50.730
24)	Chlordane-2	5.243f	0.000	215.6E6	0	1231.798	N.D. #
25)	Chlordane-3	5.916	5.051	193.4E6	305.7E6	289.617	439.735 #
26)	Chlordane-4	5.996	5.115	191.7E6	300.1E6	231.618	482.173 #
27)	Chlordane-5	0.000	6.001	0	265.2E6	N.D.	1039.524 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072825\
Data File : PL096611.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2025 20:17
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL072825

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 08:47:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Tue Jul 29 08:20:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 161043183
Conc: 50.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#1 Tetrachloro-m-xylene

R.T.: 2.828 min
Delta R.T.: 0.000 min
Response: 247278851
Conc: 51.76 ng/ml




#2 alpha-BHC

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 237249143
Conc: 51.28 ng/ml



#2 alpha-BHC


R.T.: 3.334 min
Delta R.T.: 0.000 min
Response: 372995453
Conc: 52.71 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 223584100
Conc: 50.55 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#3 gamma-BHC (Lindane)

R.T.: 3.666 min
Delta R.T.: 0.000 min
Response: 346014885
Conc: 52.40 ng/ml




#4 Heptachlor

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 211285693
Conc: 50.84 ng/ml



#4 Heptachlor


R.T.: 4.014 min
Delta R.T.: 0.000 min
Response: 346369326
Conc: 51.99 ng/ml



#5 Aldrin


R.T.: 5.243 min
Delta R.T.: 0.000 min
Response: 215573068
Conc: 50.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#5 Aldrin

R.T.: 4.297 min
Delta R.T.: 0.000 min
Response: 324132499
Conc: 52.22 ng/ml




#6 beta-BHC

R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 91859349
Conc: 50.89 ng/ml



#6 beta-BHC


R.T.: 3.961 min
Delta R.T.: 0.000 min
Response: 145431130
Conc: 51.55 ng/ml



#7 delta-BHC


R.T.: 0.000 min
Exp R.T.: 4.743 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#7 delta-BHC

R.T.: 4.195 min
Delta R.T.: 0.000 min
Response: 338731728
Conc: 52.32 ng/ml




#8 Heptachlor epoxide

R.T.: 5.662 min
Delta R.T.: 0.000 min
Response: 194162959
Conc: 50.35 ng/ml



#8 Heptachlor epoxide


R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 295584103
Conc: 51.75 ng/ml



#9 Endosulfan I


R.T.: 6.044 min
Delta R.T.: 0.000 min
Response: 178427332
Conc: 49.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#9 Endosulfan I

R.T.: 5.170 min
Delta R.T.: 0.000 min
Response: 277213283
Conc: 50.12 ng/ml




#10 gamma-Chlordane

R.T.: 5.916 min
Delta R.T.: 0.000 min
Response: 193445906
Conc: 50.84 ng/ml



#10 gamma-Chlordane


R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 305748780
Conc: 52.00 ng/ml



#11 alpha-Chlordane


R.T.: 5.996 min
Delta R.T.: 0.000 min
Response: 191709507
Conc: 49.73 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#11 alpha-Chlordane

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 300071767
Conc: 50.87 ng/ml




#12 4,4'-DDE

R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 158068181
Conc: 49.22 ng/ml



#12 4,4'-DDE


R.T.: 5.304 min
Delta R.T.: 0.000 min
Response: 286332408
Conc: 51.92 ng/ml



#13 Dieldrin


R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 186821697
Conc: 50.34 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#13 Dieldrin

R.T.: 5.435 min
Delta R.T.: 0.000 min
Response: 306303510
Conc: 51.86 ng/ml




#14 Endrin

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 147327440
Conc: 48.67 ng/ml



#14 Endrin


R.T.: 5.709 min
Delta R.T.: 0.000 min
Response: 272795312
Conc: 50.46 ng/ml



#15 Endosulfan II


R.T.: 6.755 min
Delta R.T.: 0.000 min
Response: 158792104
Conc: 49.58 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#15 Endosulfan II

R.T.: 6.001 min
Delta R.T.: 0.000 min
Response: 265195055
Conc: 51.63 ng/ml




#16 4,4'-DDD

R.T.: 6.675 min
Delta R.T.: 0.000 min
Response: 126947170
Conc: 50.21 ng/ml



#16 4,4'-DDD


R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 241958902
Conc: 51.42 ng/ml



#17 4,4'-DDT


R.T.: 6.989 min
Delta R.T.: 0.000 min
Response: 144531756
Conc: 50.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#17 4,4'-DDT

R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 261665047
Conc: 51.73 ng/ml




#18 Endrin aldehyde

R.T.: 6.883 min
Delta R.T.: 0.000 min
Response: 108518783
Conc: 50.57 ng/ml



#18 Endrin aldehyde


R.T.: 6.179 min
Delta R.T.: 0.000 min
Response: 197425966
Conc: 53.39 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.117 min
Delta R.T.: 0.000 min
Response: 143881166
Conc: 50.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#19 Endosulfan Sulfate

R.T.: 6.402 min
Delta R.T.: 0.000 min
Response: 262804486
Conc: 51.67 ng/ml




#20 Methoxychlor

R.T.: 7.462 min
Delta R.T.: 0.000 min
Response: 74330342
Conc: 50.62 ng/ml



#20 Methoxychlor


R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 140523051
Conc: 51.28 ng/ml



#21 Endrin ketone


R.T.: 7.597 min
Delta R.T.: 0.000 min
Response: 153274359
Conc: 50.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#21 Endrin ketone

R.T.: 6.907 min
Delta R.T.: 0.000 min
Response: 289661773
Conc: 52.01 ng/ml




#22 Mirex

R.T.: 8.076 min
Delta R.T.: 0.000 min
Response: 123454153
Conc: 49.79 ng/ml



#22 Mirex


R.T.: 7.097 min
Delta R.T.: 0.000 min
Response: 221225472
Conc: 50.73 ng/ml



#24 Chlordane-2


R.T.: 5.243 min
Delta R.T.: 0.029 min
Response: 215573068
Conc: 1231.80 ng/m

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 4.418 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.916 min
Delta R.T.: 0.000 min
Response: 193445906
Conc: 289.62 ng/ml



#25 Chlordane-3


R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 305748780
Conc: 439.73 ng/ml



#26 Chlordane-4


R.T.: 5.996 min
Delta R.T.: -0.005 min
Response: 191709507
Conc: 231.62 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825




#26 Chlordane-4

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 300071767
Conc: 482.17 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.838 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 6.001 min
Delta R.T.: -0.007 min
Response: 265195055
Conc: 1039.52 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.015 min
Delta R.T.: -0.001 min
Response: 117087022
Conc: 49.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL072825



#28 Decachlorobiphenyl

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 221944303
Conc: 51.16 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096817.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 09:11
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:54:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.529	2.824	67085722	95299446	21.091	19.946
28)	SA Decachlor...	9.006	7.986	52210698	83831500	21.894	19.323

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096817.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 09:11
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:54:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.529 min
Delta R.T.: -0.006 min
Response: 67085722
Conc: 21.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: -0.004 min
Response: 95299446
Conc: 19.95 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.006 min
Delta R.T.: -0.010 min
Response: 52210698
Conc: 21.89 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 83831500
Conc: 19.32 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096818.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 10:13
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:54:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.826	66146774	96842073	20.796	20.269
28)	SA Decachlor...	9.004	7.987	49966203	85675813	20.953	19.749
Target Compounds							
2)	A alpha-BHC	3.978	3.330	47707827	68862700	10.312	9.730
3)	MA gamma-BHC...	4.306	3.662	46816462	64625617	10.584	9.787
6)	B beta-BHC	4.493	3.959	18826992	30137711	10.431	10.684
8)	B Heptachlo...	5.649	0.000	18535020	0	4.806	N.D. #
13)	MA Dieldrin	0.000	5.413f	0	12133923	N.D.	2.054 #
14)	MA Endrin	6.535	5.706	164.0E6	268.8E6	54.180	49.725
15)	B Endosulfa...	6.733f	0.000	13879018	0	4.334	N.D. #
16)	A 4,4'-DDD	6.673	0.000	8424285	0	3.332	N.D. #
17)	MA 4,4'-DDT	6.983	6.105	307.7E6	523.0E6	107.285	103.402
18)	B Endrin al...	0.000	6.172	0	22480880	N.D.	6.184 #
20)	A Methoxychlor	7.455	6.677	388.4E6	604.7E6	264.496	220.663
21)	B Endrin ke...	0.000	6.897	0	14350987	N.D.	2.577 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096818.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 10:13
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:54:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 66146774
Conc: 20.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 96842073
Conc: 20.27 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 47707827
Conc: 10.31 ng/ml



#2 alpha-BHC


R.T.: 3.330 min
Delta R.T.: -0.003 min
Response: 68862700
Conc: 9.73 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.306 min
Delta R.T.: -0.005 min
Response: 46816462
Conc: 10.58 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 64625617
Conc: 9.79 ng/ml




#6 beta-BHC

R.T.: 4.493 min
Delta R.T.: -0.004 min
Response: 18826992
Conc: 10.43 ng/ml



#6 beta-BHC


R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 30137711
Conc: 10.68 ng/ml



#8 Heptachlor epoxide


R.T.: 5.649 min
Delta R.T.: -0.014 min
Response: 18535020
Conc: 4.81 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 4.799 min
Response: 0
Conc: N.D.




#13 Dieldrin

R.T.: 0.000 min
Exp R.T. : 6.316 min
Response: 0
Conc: N.D.



#13 Dieldrin


R.T.: 5.413 min
Delta R.T.: -0.021 min
Response: 12133923
Conc: 2.05 ng/ml



#14 Endrin


R.T.: 6.535 min
Delta R.T.: -0.007 min
Response: 163991939
Conc: 54.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#14 Endrin

R.T.: 5.706 min
Delta R.T.: -0.003 min
Response: 268809359
Conc: 49.72 ng/ml




#15 Endosulfan II

R.T.: 6.733 min
Delta R.T.: -0.022 min
Response: 13879018
Conc: 4.33 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T.: 6.001 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 6.673 min
Delta R.T.: -0.002 min
Response: 8424285
Conc: 3.33 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T.: 5.856 min
Response: 0
Conc: N.D.




#17 4,4'-DDT

R.T.: 6.983 min
Delta R.T.: -0.006 min
Response: 307659279
Conc: 107.28 ng/ml



#17 4,4'-DDT


R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 523036483
Conc: 103.40 ng/ml



#18 Endrin aldehyde


R.T.: 0.000 min
Exp R.T.: 6.884 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PEM




#18 Endrin aldehyde

R.T.: 6.172 min
Delta R.T.: -0.007 min
Response: 22480880
Conc: 6.18 ng/ml




#20 Methoxychlor

R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 388373800
Conc: 264.50 ng/ml



#20 Methoxychlor


R.T.: 6.677 min
Delta R.T.: -0.004 min
Response: 604725336
Conc: 220.66 ng/ml



#21 Endrin ketone


R.T.: 0.000 min
Exp R.T.: 7.597 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PEM




#21 Endrin ketone

R.T.: 6.897 min
Delta R.T.: -0.010 min
Response: 14350987
Conc: 2.58 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.011 min
Response: 49966203
Conc: 20.95 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 85675813
Conc: 19.75 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096819.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 10:58
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:54:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.822	173.6E6	258.7E6	54.574	54.143
28)	SA Decachlor...	9.009	7.987	127.8E6	209.1E6	53.584	48.192
Target Compounds							
2)	A alpha-BHC	3.977	3.328	258.8E6	377.1E6	55.941	53.290
3)	MA gamma-BHC...	4.305	3.660	230.8E6	349.6E6	52.174	52.938
4)	MA Heptachlor	4.897	4.009	248.7E6	345.4E6	59.841	51.850
5)	MB Aldrin	5.237	4.291	238.5E6	317.1E6	55.458	51.080
6)	B beta-BHC	4.492	3.956	92716647	148.2E6	51.367	52.522
7)	B delta-BHC	4.738	4.189	224.7E6	337.8E6	56.322	52.180
8)	B Heptachlo...	5.657	4.794	206.2E6	291.8E6	53.473	51.083
9)	A Endosulfan I	6.039	5.165	198.1E6	277.0E6	55.090	50.075
10)	B gamma-Chl...	5.911	5.046	214.3E6	310.4E6	56.305	52.789
11)	B alpha-Chl...	5.992	5.110	213.7E6	303.9E6	55.429	51.520
12)	B 4,4'-DDE	6.162	5.299	183.4E6	286.5E6	57.098	51.952
13)	MA Dieldrin	6.311	5.430	201.5E6	302.9E6	54.299	51.283
14)	MA Endrin	6.537	5.704	164.0E6	275.4E6	54.168	50.951
15)	B Endosulfa...	6.750	5.996	181.7E6	291.5E6	56.739	56.746
16)	A 4,4'-DDD	6.670	5.852	153.5E6	240.2E6	60.694	51.048
17)	MA 4,4'-DDT	6.985	6.105	158.5E6	279.8E6	55.256	55.319
18)	B Endrin al...	6.879	6.174	122.6E6	188.4E6	57.114	51.827
19)	B Endosulfa...	7.113	6.398	151.1E6	252.7E6	52.569	49.688
20)	A Methoxychlor	7.457	6.677	80618417	138.3E6	54.904	50.460
21)	B Endrin ke...	7.592	6.903	159.7E6	280.3E6	53.123	50.322
22)	Mirex	8.071	7.092	128.9E6	216.7E6	51.989	49.693
24)	Chlordane-2	5.237f	4.439f	238.5E6	18561319	1362.958	76.513 #
25)	Chlordane-3	5.911	5.046	214.3E6	310.4E6	320.770	446.433 #
26)	Chlordane-4	5.992	5.110	213.7E6	303.9E6	258.136	488.341 #
27)	Chlordane-5	0.000	5.996	0	291.5E6	N.D.	1142.484 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


```
Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096819.D
Signal(s) : Signal #1: ECD1A.ch  Signal #2: ECD2B.ch
Acq On    : 15 Aug 2025  10:58
Operator  : AR\AJ
Sample    : PSTDCCC050
Misc      :
ALS Vial  : 4    Sample Multiplier: 1
```

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:54:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 173588834
Conc: 54.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.822 min
Delta R.T.: -0.006 min
Response: 258683720
Conc: 54.14 ng/ml




#2 alpha-BHC

R.T.: 3.977 min
Delta R.T.: -0.005 min
Response: 258805129
Conc: 55.94 ng/ml



#2 alpha-BHC


R.T.: 3.328 min
Delta R.T.: -0.005 min
Response: 377136964
Conc: 53.29 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.305 min
Delta R.T.: -0.005 min
Response: 230787677
Conc: 52.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.660 min
Delta R.T.: -0.005 min
Response: 349570640
Conc: 52.94 ng/ml




#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.005 min
Response: 248682326
Conc: 59.84 ng/ml



#4 Heptachlor


R.T.: 4.009 min
Delta R.T.: -0.005 min
Response: 345419725
Conc: 51.85 ng/ml



#5 Aldrin


R.T.: 5.237 min
Delta R.T.: -0.006 min
Response: 238527024
Conc: 55.46 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.291 min
Delta R.T.: -0.006 min
Response: 317057424
Conc: 51.08 ng/ml




#6 beta-BHC

R.T.: 4.492 min
Delta R.T.: -0.005 min
Response: 92716647
Conc: 51.37 ng/ml



#6 beta-BHC


R.T.: 3.956 min
Delta R.T.: -0.005 min
Response: 148157872
Conc: 52.52 ng/ml



#7 delta-BHC


R.T.: 4.738 min
Delta R.T.: -0.004 min
Response: 224715371
Conc: 56.32 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.189 min
Delta R.T.: -0.006 min
Response: 337794712
Conc: 52.18 ng/ml




#8 Heptachlor epoxide

R.T.: 5.657 min
Delta R.T.: -0.005 min
Response: 206221416
Conc: 53.47 ng/ml



#8 Heptachlor epoxide


R.T.: 4.794 min
Delta R.T.: -0.005 min
Response: 291750031
Conc: 51.08 ng/ml



#9 Endosulfan I


R.T.: 6.039 min
Delta R.T.: -0.004 min
Response: 198123782
Conc: 55.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.165 min
Delta R.T.: -0.004 min
Response: 276983051
Conc: 50.07 ng/ml




#10 gamma-Chlordane

R.T.: 5.911 min
Delta R.T.: -0.005 min
Response: 214254477
Conc: 56.31 ng/ml



#10 gamma-Chlordane


R.T.: 5.046 min
Delta R.T.: -0.005 min
Response: 310405952
Conc: 52.79 ng/ml



#11 alpha-Chlordane


R.T.: 5.992 min
Delta R.T.: -0.005 min
Response: 213658778
Conc: 55.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.110 min
Delta R.T.: -0.005 min
Response: 303910082
Conc: 51.52 ng/ml




#12 4,4'-DDE

R.T.: 6.162 min
Delta R.T.: -0.005 min
Response: 183355642
Conc: 57.10 ng/ml



#12 4,4'-DDE


R.T.: 5.299 min
Delta R.T.: -0.005 min
Response: 286484870
Conc: 51.95 ng/ml



#13 Dieldrin


R.T.: 6.311 min
Delta R.T.: -0.005 min
Response: 201510129
Conc: 54.30 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 302906139
Conc: 51.28 ng/ml




#14 Endrin

R.T.: 6.537 min
Delta R.T.: -0.005 min
Response: 163954583
Conc: 54.17 ng/ml



#14 Endrin


R.T.: 5.704 min
Delta R.T.: -0.005 min
Response: 275437287
Conc: 50.95 ng/ml



#15 Endosulfan II


R.T.: 6.750 min
Delta R.T.: -0.005 min
Response: 181715837
Conc: 56.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 5.996 min
Delta R.T.: -0.005 min
Response: 291461407
Conc: 56.75 ng/ml




#16 4,4'-DDD

R.T.: 6.670 min
Delta R.T.: -0.005 min
Response: 153457756
Conc: 60.69 ng/ml



#16 4,4'-DDD


R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 240194386
Conc: 51.05 ng/ml



#17 4,4'-DDT


R.T.: 6.985 min
Delta R.T.: -0.004 min
Response: 158456518
Conc: 55.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 279819657
Conc: 55.32 ng/ml




#18 Endrin aldehyde

R.T.: 6.879 min
Delta R.T.: -0.004 min
Response: 122561784
Conc: 57.11 ng/ml



#18 Endrin aldehyde


R.T.: 6.174 min
Delta R.T.: -0.005 min
Response: 188393647
Conc: 51.83 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.113 min
Delta R.T.: -0.005 min
Response: 151067060
Conc: 52.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 252725524
Conc: 49.69 ng/ml




#20 Methoxychlor

R.T.: 7.457 min
Delta R.T.: -0.005 min
Response: 80618417
Conc: 54.90 ng/ml



#20 Methoxychlor


R.T.: 6.677 min
Delta R.T.: -0.004 min
Response: 138285984
Conc: 50.46 ng/ml



#21 Endrin ketone


R.T.: 7.592 min
Delta R.T.: -0.005 min
Response: 159725679
Conc: 53.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.903 min
Delta R.T.: -0.004 min
Response: 280283510
Conc: 50.32 ng/ml




#22 Mirex

R.T.: 8.071 min
Delta R.T.: -0.005 min
Response: 128899173
Conc: 51.99 ng/ml



#22 Mirex


R.T.: 7.092 min
Delta R.T.: -0.005 min
Response: 216703290
Conc: 49.69 ng/ml



#24 Chlordane-2


R.T.: 5.237 min
Delta R.T.: 0.023 min
Response: 238527024
Conc: 1362.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 4.439 min
Delta R.T.: 0.021 min
Response: 18561319
Conc: 76.51 ng/ml




#25 Chlordane-3

R.T.: 5.911 min
Delta R.T.: -0.006 min
Response: 214254477
Conc: 320.77 ng/ml



#25 Chlordane-3


R.T.: 5.046 min
Delta R.T.: -0.006 min
Response: 310405952
Conc: 446.43 ng/ml



#26 Chlordane-4


R.T.: 5.992 min
Delta R.T.: -0.009 min
Response: 213658778
Conc: 258.14 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 5.110 min
Delta R.T.: -0.005 min
Response: 303910082
Conc: 488.34 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.838 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.996 min
Delta R.T.: -0.012 min
Response: 291461407
Conc: 1142.48 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.009 min
Delta R.T.: -0.006 min
Response: 127780741
Conc: 53.58 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 209072671
Conc: 48.19 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096822.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 16:07
 Operator : AR\AJ
 Sample : PB169225BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB169225BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:55:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.824	62200378	94174771	19.555	19.711
28)	SA Decachlor...	9.005	7.987	48085872	89293709	20.165	20.583
Target Compounds							
4)	MA Heptachlor	4.925f	0.000	5440554	0	1.309	N.D. #
8)	B Heptachlo...	5.647	0.000	4482918	0	1.162	N.D. #
14)	MA Endrin	0.000	5.714	0	9802454	N.D.	1.813 #
15)	B Endosulfa...	6.736f	0.000	10273648	0	3.208	N.D. #
18)	B Endrin al...	0.000	6.167	0	24550070	N.D.	6.754 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096822.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 16:07
Operator : AR\AJ
Sample : PB169225BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB169225BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 62200378
Conc: 19.56 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BL




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: -0.004 min
Response: 94174771
Conc: 19.71 ng/ml




#4 Heptachlor

R.T.: 4.925 min
Delta R.T.: 0.022 min
Response: 5440554
Conc: 1.31 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T.: 4.014 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.647 min
Delta R.T.: -0.015 min
Response: 4482918
Conc: 1.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BL




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 4.799 min
Response: 0
Conc: N.D.




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.542 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.714 min
Delta R.T.: 0.005 min
Response: 9802454
Conc: 1.81 ng/ml



#15 Endosulfan II


R.T.: 6.736 min
Delta R.T.: -0.019 min
Response: 10273648
Conc: 3.21 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BL




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.001 min
Response: 0
Conc: N.D.




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.884 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.167 min
Delta R.T.: -0.012 min
Response: 24550070
Conc: 6.75 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.005 min
Delta R.T.: -0.011 min
Response: 48085872
Conc: 20.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BL



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 89293709
Conc: 20.58 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096823.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 16:21
 Operator : AR\AJ
 Sample : PB169225BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB169225BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:55:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.825	60775217	94199127	19.107	19.716
28)	SA Decachlor...	9.004	7.987	47285056	90894687	19.829	20.952
Target Compounds							
2)	A alpha-BHC	3.977	3.330	227.6E6	354.3E6	49.191	50.066
3)	MA gamma-BHC...	4.305	3.662	213.4E6	330.6E6	48.244	50.059
4)	MA Heptachlor	4.896	4.011	219.5E6	326.3E6	52.820	48.974
5)	MB Aldrin	5.235	4.293	212.1E6	307.8E6	49.311	49.594
6)	B beta-BHC	4.491	3.958	85696559	140.1E6	47.478	49.652
7)	B delta-BHC	4.737	4.191	200.6E6	323.8E6	50.286	50.023
8)	B Heptachlo...	5.655	4.795	194.9E6	279.8E6	50.536	48.985
9)	A Endosulfan I	6.036	5.166	176.2E6	263.2E6	48.990	47.574
10)	B gamma-Chl...	5.908	5.047	191.4E6	291.6E6	50.305	49.595
11)	B alpha-Chl...	5.989	5.111	192.7E6	287.4E6	49.982	48.722
12)	B 4,4'-DDE	6.159	5.299	157.6E6	272.8E6	49.062	49.477
13)	MA Dieldrin	6.309	5.430	183.7E6	290.5E6	49.491	49.185
14)	MA Endrin	6.535	5.705	145.6E6	261.5E6	48.103	48.378
15)	B Endosulfa...	6.748	5.996	162.5E6	250.1E6	50.742	48.697
16)	A 4,4'-DDD	6.667	5.852	127.2E6	232.6E6	50.315	49.441
17)	MA 4,4'-DDT	6.982	6.104	142.3E6	245.6E6	49.616	48.551
18)	B Endrin al...	6.876	6.174	114.1E6	199.7E6	53.165	54.945
19)	B Endosulfa...	7.110	6.398	137.5E6	246.6E6	47.845	48.484
20)	A Methoxychlor	7.454	6.676	71535659	132.8E6	48.718	48.452
21)	B Endrin ke...	7.589	6.902	147.3E6	288.7E6	48.986	51.834
22)	Mirex	8.068	7.092	114.7E6	213.1E6	46.258	48.857
24)	Chlordane-2	5.235f	0.000	212.1E6	0	1211.879	N.D. #
25)	Chlordane-3	5.908	5.047	191.4E6	291.6E6	286.586	419.419 #
26)	Chlordane-4	5.989	5.111	192.7E6	287.4E6	232.769	461.819 #
27)	Chlordane-5	0.000	5.996	0	250.1E6	N.D.	980.433 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096823.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 16:21
Operator : AR\AJ
Sample : PB169225BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB169225BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 60775217
Conc: 19.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 94199127
Conc: 19.72 ng/ml




#2 alpha-BHC

R.T.: 3.977 min
Delta R.T.: -0.005 min
Response: 227574161
Conc: 49.19 ng/ml



#2 alpha-BHC


R.T.: 3.330 min
Delta R.T.: -0.003 min
Response: 354317224
Conc: 50.07 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.305 min
Delta R.T.: -0.006 min
Response: 213403666
Conc: 48.24 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 330559345
Conc: 50.06 ng/ml




#4 Heptachlor

R.T.: 4.896 min
Delta R.T.: -0.007 min
Response: 219501777
Conc: 52.82 ng/ml



#4 Heptachlor


R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 326259099
Conc: 48.97 ng/ml



#5 Aldrin


R.T.: 5.235 min
Delta R.T.: -0.008 min
Response: 212087085
Conc: 49.31 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#5 Aldrin

R.T.: 4.293 min
Delta R.T.: -0.004 min
Response: 307833389
Conc: 49.59 ng/ml




#6 beta-BHC

R.T.: 4.491 min
Delta R.T.: -0.006 min
Response: 85696559
Conc: 47.48 ng/ml



#6 beta-BHC


R.T.: 3.958 min
Delta R.T.: -0.003 min
Response: 140064434
Conc: 49.65 ng/ml



#7 delta-BHC


R.T.: 4.737 min
Delta R.T.: -0.005 min
Response: 200634572
Conc: 50.29 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#7 delta-BHC

R.T.: 4.191 min
Delta R.T.: -0.004 min
Response: 323827369
Conc: 50.02 ng/ml




#8 Heptachlor epoxide

R.T.: 5.655 min
Delta R.T.: -0.007 min
Response: 194893536
Conc: 50.54 ng/ml



#8 Heptachlor epoxide


R.T.: 4.795 min
Delta R.T.: -0.004 min
Response: 279768553
Conc: 48.98 ng/ml



#9 Endosulfan I


R.T.: 6.036 min
Delta R.T.: -0.007 min
Response: 176186217
Conc: 48.99 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#9 Endosulfan I

R.T.: 5.166 min
Delta R.T.: -0.003 min
Response: 263153744
Conc: 47.57 ng/ml




#10 gamma-Chlordane

R.T.: 5.908 min
Delta R.T.: -0.007 min
Response: 191421188
Conc: 50.30 ng/ml



#10 gamma-Chlordane


R.T.: 5.047 min
Delta R.T.: -0.004 min
Response: 291623122
Conc: 49.59 ng/ml



#11 alpha-Chlordane


R.T.: 5.989 min
Delta R.T.: -0.008 min
Response: 192662406
Conc: 49.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#11 alpha-Chlordane

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 287404609
Conc: 48.72 ng/ml




#12 4,4'-DDE

R.T.: 6.159 min
Delta R.T.: -0.007 min
Response: 157550711
Conc: 49.06 ng/ml



#12 4,4'-DDE


R.T.: 5.299 min
Delta R.T.: -0.005 min
Response: 272836065
Conc: 49.48 ng/ml



#13 Dieldrin


R.T.: 6.309 min
Delta R.T.: -0.007 min
Response: 183668334
Conc: 49.49 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#13 Dieldrin

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 290510751
Conc: 49.18 ng/ml




#14 Endrin

R.T.: 6.535 min
Delta R.T.: -0.007 min
Response: 145598600
Conc: 48.10 ng/ml



#14 Endrin


R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 261530591
Conc: 48.38 ng/ml



#15 Endosulfan II


R.T.: 6.748 min
Delta R.T.: -0.007 min
Response: 162507755
Conc: 50.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#15 Endosulfan II

R.T.: 5.996 min
Delta R.T.: -0.005 min
Response: 250120283
Conc: 48.70 ng/ml




#16 4,4'-DDD

R.T.: 6.667 min
Delta R.T.: -0.008 min
Response: 127213853
Conc: 50.31 ng/ml



#16 4,4'-DDD


R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 232633507
Conc: 49.44 ng/ml



#17 4,4'-DDT


R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 142282054
Conc: 49.62 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#17 4,4'-DDT

R.T.: 6.104 min
Delta R.T.: -0.005 min
Response: 245583576
Conc: 48.55 ng/ml




#18 Endrin aldehyde

R.T.: 6.876 min
Delta R.T.: -0.008 min
Response: 114087942
Conc: 53.17 ng/ml



#18 Endrin aldehyde


R.T.: 6.174 min
Delta R.T.: -0.005 min
Response: 199729538
Conc: 54.95 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.110 min
Delta R.T.: -0.008 min
Response: 137490676
Conc: 47.84 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#19 Endosulfan Sulfate

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 246600399
Conc: 48.48 ng/ml




#20 Methoxychlor

R.T.: 7.454 min
Delta R.T.: -0.008 min
Response: 71535659
Conc: 48.72 ng/ml



#20 Methoxychlor


R.T.: 6.676 min
Delta R.T.: -0.005 min
Response: 132781179
Conc: 48.45 ng/ml



#21 Endrin ketone


R.T.: 7.589 min
Delta R.T.: -0.008 min
Response: 147288355
Conc: 48.99 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 288700248
Conc: 51.83 ng/ml




#22 Mirex

R.T.: 8.068 min
Delta R.T.: -0.008 min
Response: 114690781
Conc: 46.26 ng/ml



#22 Mirex


R.T.: 7.092 min
Delta R.T.: -0.005 min
Response: 213057480
Conc: 48.86 ng/ml



#24 Chlordane-2


R.T.: 5.235 min
Delta R.T.: 0.022 min
Response: 212087085
Conc: 1211.88 ng/m

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 4.418 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.908 min
Delta R.T.: -0.008 min
Response: 191421188
Conc: 286.59 ng/ml



#25 Chlordane-3


R.T.: 5.047 min
Delta R.T.: -0.005 min
Response: 291623122
Conc: 419.42 ng/ml



#26 Chlordane-4


R.T.: 5.989 min
Delta R.T.: -0.012 min
Response: 192662406
Conc: 232.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS




#26 Chlordane-4

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 287404609
Conc: 461.82 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.838 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.996 min
Delta R.T.: -0.012 min
Response: 250120283
Conc: 980.43 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.012 min
Response: 47285056
Conc: 19.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BS



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.006 min
Response: 90894687
Conc: 20.95 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096824.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 16:34
 Operator : AR\AJ
 Sample : PB169225BSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB169225BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:55:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.826	59543401	91996171	18.720	19.255
28)	SA Decachlor...	9.005	7.987	46912605	90356362	19.673	20.828
Target Compounds							
2)	A alpha-BHC	3.978	3.331	223.2E6	342.8E6	48.236	48.442
3)	MA gamma-BHC...	4.306	3.663	210.3E6	320.0E6	47.536	48.462
4)	MA Heptachlor	4.897	4.011	214.5E6	317.1E6	51.625	47.601
5)	MB Aldrin	5.237	4.294	207.2E6	299.2E6	48.180	48.202
6)	B beta-BHC	4.493	3.959	84541325	135.8E6	46.838	48.135
7)	B delta-BHC	4.738	4.192	195.2E6	313.4E6	48.931	48.408
8)	B Heptachlo...	5.656	4.796	190.6E6	273.5E6	49.435	47.880
9)	A Endosulfan I	6.037	5.167	173.6E6	255.4E6	48.277	46.174
10)	B gamma-Chl...	5.910	5.048	186.9E6	283.8E6	49.128	48.272
11)	B alpha-Chl...	5.990	5.112	188.2E6	280.2E6	48.827	47.494
12)	B 4,4'-DDE	6.160	5.300	159.5E6	265.9E6	49.682	48.218
13)	MA Dieldrin	6.310	5.431	180.2E6	284.1E6	48.545	48.096
14)	MA Endrin	6.536	5.705	142.2E6	253.1E6	46.989	46.827
15)	B Endosulfa...	6.749	5.997	162.8E6	245.2E6	50.827	47.742
16)	A 4,4'-DDD	6.669	5.853	124.7E6	227.7E6	49.301	48.390
17)	MA 4,4'-DDT	6.983	6.105	138.2E6	238.1E6	48.196	47.076
18)	B Endrin al...	6.878	6.175	112.1E6	196.6E6	52.223	54.094
19)	B Endosulfa...	7.111	6.398	135.4E6	241.4E6	47.106	47.452
20)	A Methoxychlor	7.455	6.677	68285598	129.5E6	46.505	47.241
21)	B Endrin ke...	7.590	6.903	143.9E6	273.2E6	47.846	49.057
22)	Mirex	8.068	7.092	113.2E6	210.0E6	45.656	48.160
24)	Chlordane-2	5.237f	0.000	207.2E6	0	1184.075	N.D. #
25)	Chlordane-3	5.910	5.048	186.9E6	283.8E6	279.881	408.230 #
26)	Chlordane-4	5.990	5.112	188.2E6	280.2E6	227.392	450.182 #
27)	Chlordane-5	0.000	5.997	0	245.2E6	N.D.	961.203 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096824.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 16:34
Operator : AR\AJ
Sample : PB169225BSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB169225BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 59543401
Conc: 18.72 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 91996171
Conc: 19.26 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 223157039
Conc: 48.24 ng/ml



#2 alpha-BHC


R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 342828244
Conc: 48.44 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.306 min
Delta R.T.: -0.004 min
Response: 210273174
Conc: 47.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#3 gamma-BHC (Lindane)

R.T.: 3.663 min
Delta R.T.: -0.002 min
Response: 320012087
Conc: 48.46 ng/ml




#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.005 min
Response: 214539353
Conc: 51.63 ng/ml



#4 Heptachlor


R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 317116974
Conc: 47.60 ng/ml



#5 Aldrin


R.T.: 5.237 min
Delta R.T.: -0.006 min
Response: 207221296
Conc: 48.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#5 Aldrin

R.T.: 4.294 min
Delta R.T.: -0.003 min
Response: 299191263
Conc: 48.20 ng/ml




#6 beta-BHC

R.T.: 4.493 min
Delta R.T.: -0.004 min
Response: 84541325
Conc: 46.84 ng/ml



#6 beta-BHC


R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 135784500
Conc: 48.14 ng/ml



#7 delta-BHC


R.T.: 4.738 min
Delta R.T.: -0.004 min
Response: 195226463
Conc: 48.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#7 delta-BHC

R.T.: 4.192 min
Delta R.T.: -0.003 min
Response: 313372209
Conc: 48.41 ng/ml




#8 Heptachlor epoxide

R.T.: 5.656 min
Delta R.T.: -0.006 min
Response: 190649519
Conc: 49.44 ng/ml



#8 Heptachlor epoxide


R.T.: 4.796 min
Delta R.T.: -0.003 min
Response: 273458365
Conc: 47.88 ng/ml



#9 Endosulfan I


R.T.: 6.037 min
Delta R.T.: -0.006 min
Response: 173622932
Conc: 48.28 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#9 Endosulfan I

R.T.: 5.167 min
Delta R.T.: -0.002 min
Response: 255405216
Conc: 46.17 ng/ml




#10 gamma-Chlordane

R.T.: 5.910 min
Delta R.T.: -0.006 min
Response: 186943063
Conc: 49.13 ng/ml



#10 gamma-Chlordane


R.T.: 5.048 min
Delta R.T.: -0.003 min
Response: 283843264
Conc: 48.27 ng/ml



#11 alpha-Chlordane


R.T.: 5.990 min
Delta R.T.: -0.006 min
Response: 188211879
Conc: 48.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#11 alpha-Chlordane

R.T.: 5.112 min
Delta R.T.: -0.003 min
Response: 280162453
Conc: 47.49 ng/ml




#12 4,4'-DDE

R.T.: 6.160 min
Delta R.T.: -0.006 min
Response: 159539795
Conc: 49.68 ng/ml



#12 4,4'-DDE


R.T.: 5.300 min
Delta R.T.: -0.004 min
Response: 265892329
Conc: 48.22 ng/ml



#13 Dieldrin


R.T.: 6.310 min
Delta R.T.: -0.006 min
Response: 180156209
Conc: 48.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#13 Dieldrin

R.T.: 5.431 min
Delta R.T.: -0.003 min
Response: 284082332
Conc: 48.10 ng/ml




#14 Endrin

R.T.: 6.536 min
Delta R.T.: -0.007 min
Response: 142225524
Conc: 46.99 ng/ml



#14 Endrin


R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 253142704
Conc: 46.83 ng/ml



#15 Endosulfan II


R.T.: 6.749 min
Delta R.T.: -0.006 min
Response: 162781600
Conc: 50.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#15 Endosulfan II

R.T.: 5.997 min
Delta R.T.: -0.004 min
Response: 245214434
Conc: 47.74 ng/ml




#16 4,4'-DDD

R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 124651862
Conc: 49.30 ng/ml



#16 4,4'-DDD


R.T.: 5.853 min
Delta R.T.: -0.003 min
Response: 227687881
Conc: 48.39 ng/ml



#17 4,4'-DDT


R.T.: 6.983 min
Delta R.T.: -0.006 min
Response: 138211624
Conc: 48.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 238126866
Conc: 47.08 ng/ml




#18 Endrin aldehyde

R.T.: 6.878 min
Delta R.T.: -0.006 min
Response: 112065931
Conc: 52.22 ng/ml



#18 Endrin aldehyde


R.T.: 6.175 min
Delta R.T.: -0.004 min
Response: 196634214
Conc: 54.09 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.111 min
Delta R.T.: -0.006 min
Response: 135367424
Conc: 47.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#19 Endosulfan Sulfate

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 241352693
Conc: 47.45 ng/ml




#20 Methoxychlor

R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 68285598
Conc: 46.50 ng/ml



#20 Methoxychlor


R.T.: 6.677 min
Delta R.T.: -0.004 min
Response: 129462927
Conc: 47.24 ng/ml



#21 Endrin ketone


R.T.: 7.590 min
Delta R.T.: -0.007 min
Response: 143858655
Conc: 47.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#21 Endrin ketone

R.T.: 6.903 min
Delta R.T.: -0.004 min
Response: 273236792
Conc: 49.06 ng/ml




#22 Mirex

R.T.: 8.068 min
Delta R.T.: -0.007 min
Response: 113196979
Conc: 45.66 ng/ml



#22 Mirex


R.T.: 7.092 min
Delta R.T.: -0.005 min
Response: 210019786
Conc: 48.16 ng/ml



#24 Chlordane-2


R.T.: 5.237 min
Delta R.T.: 0.023 min
Response: 207221296
Conc: 1184.08 ng/m

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 4.418 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.910 min
Delta R.T.: -0.006 min
Response: 186943063
Conc: 279.88 ng/ml



#25 Chlordane-3


R.T.: 5.048 min
Delta R.T.: -0.004 min
Response: 283843264
Conc: 408.23 ng/ml



#26 Chlordane-4


R.T.: 5.990 min
Delta R.T.: -0.011 min
Response: 188211879
Conc: 227.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD




#26 Chlordane-4

R.T.: 5.112 min
Delta R.T.: -0.003 min
Response: 280162453
Conc: 450.18 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.838 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.997 min
Delta R.T.: -0.012 min
Response: 245214434
Conc: 961.20 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.005 min
Delta R.T.: -0.010 min
Response: 46912605
Conc: 19.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169225BSD



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 90356362
Conc: 20.83 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096827.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 17:51
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:55:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.530	2.824	64368055	98181459	20.237	20.550
28) SA Decachlor...	9.006	7.987	52183738	95702144	21.883	22.060
Target Compounds						
14) MA Endrin	0.000	5.714	0	6592782	N.D.	1.220 #
15) B Endosulfa...	6.734f	0.000	4706082	0	1.469	N.D. #
18) B Endrin al...	0.000	6.163f	0	20519104	N.D.	5.645 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096827.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 17:51
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 64368055
Conc: 20.24 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: -0.004 min
Response: 98181459
Conc: 20.55 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.542 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.714 min
Delta R.T.: 0.005 min
Response: 6592782
Conc: 1.22 ng/ml



#15 Endosulfan II


R.T.: 6.734 min
Delta R.T.: -0.021 min
Response: 4706082
Conc: 1.47 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.001 min
Response: 0
Conc: N.D.




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.884 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.163 min
Delta R.T.: -0.016 min
Response: 20519104
Conc: 5.64 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.006 min
Delta R.T.: -0.010 min
Response: 52183738
Conc: 21.88 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 95702144
Conc: 22.06 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096828.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 18:05
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:55:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.826	180.8E6	262.6E6	56.837	54.955
28)	SA Decachlor...	9.005	7.986	124.9E6	230.0E6	52.360	53.015
Target Compounds							
2)	A alpha-BHC	3.978	3.331	265.6E6	391.8E6	57.410	55.367
3)	MA gamma-BHC...	4.305	3.663	249.0E6	364.2E6	56.284	55.153
4)	MA Heptachlor	4.897	4.011	251.2E6	358.3E6	60.447	53.779
5)	MB Aldrin	5.237	4.294	244.0E6	337.3E6	56.736	54.343
6)	B beta-BHC	4.492	3.959	100.3E6	155.5E6	55.575	55.137
7)	B delta-BHC	4.738	4.192	233.2E6	357.2E6	58.436	55.177
8)	B Heptachlo...	5.656	4.796	224.7E6	306.0E6	58.256	53.572
9)	A Endosulfan I	6.038	5.166	198.5E6	281.6E6	55.197	50.904
10)	B gamma-Chl...	5.910	5.047	217.9E6	319.1E6	57.272	54.270
11)	B alpha-Chl...	5.990	5.112	217.6E6	313.0E6	56.444	53.056
12)	B 4,4'-DDE	6.160	5.300	179.1E6	295.5E6	55.767	53.587
13)	MA Dieldrin	6.310	5.431	208.1E6	315.8E6	56.087	53.465
14)	MA Endrin	6.536	5.705	165.8E6	285.1E6	54.784	52.735
15)	B Endosulfa...	6.749	5.997	192.5E6	273.0E6	60.113	53.148
16)	A 4,4'-DDD	6.669	5.853	148.8E6	252.4E6	58.869	53.649
17)	MA 4,4'-DDT	6.983	6.105	158.8E6	268.7E6	55.393	53.128
18)	B Endrin al...	6.878	6.175	125.7E6	207.6E6	58.576	57.110
19)	B Endosulfa...	7.111	6.398	154.1E6	268.2E6	53.613	52.730
20)	A Methoxychlor	7.456	6.677	78302387	143.6E6	53.327	52.384
21)	B Endrin ke...	7.590	6.902	162.8E6	314.7E6	54.131	56.503
22)	Mirex	8.069	7.092	127.1E6	232.2E6	51.279	53.257
24)	Chlordane-2	5.237f	0.000	244.0E6	0	1394.359	N.D. #
25)	Chlordane-3	5.910	5.047	217.9E6	319.1E6	326.278	458.954 #
26)	Chlordane-4	5.990	5.112	217.6E6	313.0E6	262.866	502.899 #
27)	Chlordane-5	0.000	5.997	0	273.0E6	N.D.	1070.039 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096828.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 18:05
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:55:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 180784769
Conc: 56.84 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 262561642
Conc: 54.95 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 265600436
Conc: 57.41 ng/ml



#2 alpha-BHC


R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 391837029
Conc: 55.37 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.305 min
Delta R.T.: -0.005 min
Response: 248967627
Conc: 56.28 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.663 min
Delta R.T.: -0.002 min
Response: 364194496
Conc: 55.15 ng/ml




#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.006 min
Response: 251200000
Conc: 60.45 ng/ml



#4 Heptachlor


R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 358273285
Conc: 53.78 ng/ml



#5 Aldrin


R.T.: 5.237 min
Delta R.T.: -0.006 min
Response: 244022336
Conc: 56.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.294 min
Delta R.T.: -0.003 min
Response: 337308377
Conc: 54.34 ng/ml




#6 beta-BHC

R.T.: 4.492 min
Delta R.T.: -0.005 min
Response: 100310566
Conc: 55.57 ng/ml



#6 beta-BHC


R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 155535841
Conc: 55.14 ng/ml



#7 delta-BHC


R.T.: 4.738 min
Delta R.T.: -0.004 min
Response: 233150376
Conc: 58.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.192 min
Delta R.T.: -0.003 min
Response: 357192657
Conc: 55.18 ng/ml




#8 Heptachlor epoxide

R.T.: 5.656 min
Delta R.T.: -0.006 min
Response: 224665911
Conc: 58.26 ng/ml



#8 Heptachlor epoxide


R.T.: 4.796 min
Delta R.T.: -0.003 min
Response: 305967969
Conc: 53.57 ng/ml



#9 Endosulfan I


R.T.: 6.038 min
Delta R.T.: -0.006 min
Response: 198508279
Conc: 55.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.166 min
Delta R.T.: -0.003 min
Response: 281573435
Conc: 50.90 ng/ml




#10 gamma-Chlordane

R.T.: 5.910 min
Delta R.T.: -0.006 min
Response: 217933444
Conc: 57.27 ng/ml



#10 gamma-Chlordane


R.T.: 5.047 min
Delta R.T.: -0.004 min
Response: 319111731
Conc: 54.27 ng/ml



#11 alpha-Chlordane


R.T.: 5.990 min
Delta R.T.: -0.006 min
Response: 217573409
Conc: 56.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.112 min
Delta R.T.: -0.003 min
Response: 312970390
Conc: 53.06 ng/ml




#12 4,4'-DDE

R.T.: 6.160 min
Delta R.T.: -0.006 min
Response: 179083022
Conc: 55.77 ng/ml



#12 4,4'-DDE


R.T.: 5.300 min
Delta R.T.: -0.004 min
Response: 295499974
Conc: 53.59 ng/ml



#13 Dieldrin


R.T.: 6.310 min
Delta R.T.: -0.006 min
Response: 208146865
Conc: 56.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.431 min
Delta R.T.: -0.003 min
Response: 315794823
Conc: 53.47 ng/ml




#14 Endrin

R.T.: 6.536 min
Delta R.T.: -0.006 min
Response: 165819585
Conc: 54.78 ng/ml



#14 Endrin


R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 285083802
Conc: 52.74 ng/ml



#15 Endosulfan II


R.T.: 6.749 min
Delta R.T.: -0.006 min
Response: 192519100
Conc: 60.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 5.997 min
Delta R.T.: -0.004 min
Response: 272979738
Conc: 53.15 ng/ml




#16 4,4'-DDD

R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 148842821
Conc: 58.87 ng/ml



#16 4,4'-DDD


R.T.: 5.853 min
Delta R.T.: -0.003 min
Response: 252433717
Conc: 53.65 ng/ml



#17 4,4'-DDT


R.T.: 6.983 min
Delta R.T.: -0.006 min
Response: 158848972
Conc: 55.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 268739257
Conc: 53.13 ng/ml




#18 Endrin aldehyde

R.T.: 6.878 min
Delta R.T.: -0.006 min
Response: 125699655
Conc: 58.58 ng/ml



#18 Endrin aldehyde


R.T.: 6.175 min
Delta R.T.: -0.004 min
Response: 207597724
Conc: 57.11 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.111 min
Delta R.T.: -0.007 min
Response: 154065949
Conc: 53.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 268193619
Conc: 52.73 ng/ml




#20 Methoxychlor

R.T.: 7.456 min
Delta R.T.: -0.007 min
Response: 78302387
Conc: 53.33 ng/ml



#20 Methoxychlor


R.T.: 6.677 min
Delta R.T.: -0.004 min
Response: 143559352
Conc: 52.38 ng/ml



#21 Endrin ketone


R.T.: 7.590 min
Delta R.T.: -0.007 min
Response: 162756760
Conc: 54.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 314706868
Conc: 56.50 ng/ml




#22 Mirex

R.T.: 8.069 min
Delta R.T.: -0.007 min
Response: 127139267
Conc: 51.28 ng/ml



#22 Mirex


R.T.: 7.092 min
Delta R.T.: -0.005 min
Response: 232248246
Conc: 53.26 ng/ml



#24 Chlordane-2


R.T.: 5.237 min
Delta R.T.: 0.023 min
Response: 244022336
Conc: 1394.36 ng/m

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 4.418 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.910 min
Delta R.T.: -0.007 min
Response: 217933444
Conc: 326.28 ng/ml



#25 Chlordane-3


R.T.: 5.047 min
Delta R.T.: -0.005 min
Response: 319111731
Conc: 458.95 ng/ml



#26 Chlordane-4


R.T.: 5.990 min
Delta R.T.: -0.011 min
Response: 217573409
Conc: 262.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 5.112 min
Delta R.T.: -0.003 min
Response: 312970390
Conc: 502.90 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.838 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.997 min
Delta R.T.: -0.012 min
Response: 272979738
Conc: 1070.04 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.005 min
Delta R.T.: -0.011 min
Response: 124860925
Conc: 52.36 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 229997908
Conc: 53.02 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096835.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 19:54
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:56:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.531	2.826	64836130	97845915	20.384	20.479
28) SA Decachlor...	9.003	7.986	49757231	90215642	20.866	20.795
Target Compounds						
14) MA Endrin	0.000	5.714	0	13009158	N.D.	2.406 #
15) B Endosulfa...	6.737f	0.000	8422327	0	2.630	N.D. #
18) B Endrin al...	0.000	6.169	0	12212942	N.D.	3.360 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096835.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 19:54
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:56:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 64836130
Conc: 20.38 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 97845915
Conc: 20.48 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.542 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.714 min
Delta R.T.: 0.005 min
Response: 13009158
Conc: 2.41 ng/ml



#15 Endosulfan II


R.T.: 6.737 min
Delta R.T.: -0.017 min
Response: 8422327
Conc: 2.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.001 min
Response: 0
Conc: N.D.




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.884 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.169 min
Delta R.T.: -0.010 min
Response: 12212942
Conc: 3.36 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.003 min
Delta R.T.: -0.012 min
Response: 49757231
Conc: 20.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 90215642
Conc: 20.80 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096836.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 20:07
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:56:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.530	2.825	69382655	100.9E6	21.813	21.115
28) SA Decachlor...	9.004	7.986	51106982	92284478	21.432	21.272
Target Compounds						
2) A alpha-BHC	3.977	3.331	48480683	71648677	10.479	10.124
3) MA gamma-BHC...	4.305	3.662	47716628	67995426	10.787	10.297
6) B beta-BHC	4.492	3.959	19679251	31540061	10.903	11.181
8) B Heptachlo...	5.647	0.000	11719005	0	3.039	N.D. #
14) MA Endrin	6.535	5.705	163.7E6	278.7E6	54.090	51.558
15) B Endosulfa...	6.736f	0.000	23772141	0	7.423	N.D. #
16) A 4,4'-DDD	6.667	5.852	17825677	11739415	7.050	2.495 #
17) MA 4,4'-DDT	6.982	6.105	297.6E6	495.8E6	103.770	98.008
18) B Endrin al...	0.000	6.174	0	16854287	N.D.	4.637 #
20) A Methoxychlor	7.455	6.677	371.2E6	604.6E6	252.799	220.632
21) B Endrin ke...	7.590	6.901	5861879	22729518	1.950	4.081 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096836.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 20:07
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:56:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 69382655
Conc: 21.81 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 100884757
Conc: 21.12 ng/ml




#2 alpha-BHC

R.T.: 3.977 min
Delta R.T.: -0.005 min
Response: 48480683
Conc: 10.48 ng/ml



#2 alpha-BHC


R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 71648677
Conc: 10.12 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.305 min
Delta R.T.: -0.005 min
Response: 47716628
Conc: 10.79 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 67995426
Conc: 10.30 ng/ml




#6 beta-BHC

R.T.: 4.492 min
Delta R.T.: -0.005 min
Response: 19679251
Conc: 10.90 ng/ml



#6 beta-BHC


R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 31540061
Conc: 11.18 ng/ml



#8 Heptachlor epoxide


R.T.: 5.647 min
Delta R.T.: -0.015 min
Response: 11719005
Conc: 3.04 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 4.799 min
Response: 0
Conc: N.D.




#14 Endrin

R.T.: 6.535 min
Delta R.T.: -0.007 min
Response: 163718919
Conc: 54.09 ng/ml



#14 Endrin


R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 278722026
Conc: 51.56 ng/ml



#15 Endosulfan II


R.T.: 6.736 min
Delta R.T.: -0.019 min
Response: 23772141
Conc: 7.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.001 min
Response: 0
Conc: N.D.




#16 4,4'-DDD

R.T.: 6.667 min
Delta R.T.: -0.008 min
Response: 17825677
Conc: 7.05 ng/ml



#16 4,4'-DDD


R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 11739415
Conc: 2.49 ng/ml



#17 4,4'-DDT


R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 297579199
Conc: 103.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 495753032
Conc: 98.01 ng/ml




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.884 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.174 min
Delta R.T.: -0.005 min
Response: 16854287
Conc: 4.64 ng/ml



#20 Methoxychlor


R.T.: 7.455 min
Delta R.T.: -0.008 min
Response: 371198217
Conc: 252.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#20 Methoxychlor

R.T.: 6.677 min
Delta R.T.: -0.004 min
Response: 604640883
Conc: 220.63 ng/ml




#21 Endrin ketone

R.T.: 7.590 min
Delta R.T.: -0.007 min
Response: 5861879
Conc: 1.95 ng/ml



#21 Endrin ketone


R.T.: 6.901 min
Delta R.T.: -0.006 min
Response: 22729518
Conc: 4.08 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.012 min
Response: 51106982
Conc: 21.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 92284478
Conc: 21.27 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096837.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 20:21
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050


Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:56:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.825	183.2E6	261.2E6	57.592	54.666
28)	SA Decachlor...	9.005	7.986	156.0E6	256.3E6	65.408	59.073
Target Compounds							
2)	A alpha-BHC	3.978	3.331	265.3E6	387.0E6	57.342	54.683
3)	MA gamma-BHC...	4.305	3.662	248.0E6	357.8E6	56.070	54.183
4)	MA Heptachlor	4.897	4.011	249.2E6	346.7E6	59.978	52.039
5)	MB Aldrin	5.236	4.293	243.6E6	330.9E6	56.646	53.315
6)	B beta-BHC	4.492	3.959	96682513	150.1E6	53.565	53.201
7)	B delta-BHC	4.738	4.192	234.0E6	348.3E6	58.652	53.810
8)	B Heptachlo...	5.655	4.795	221.2E6	293.1E6	57.355	51.316
9)	A Endosulfan I	6.037	5.166	198.3E6	280.3E6	55.138	50.681
10)	B gamma-Chl...	5.909	5.047	215.5E6	304.9E6	56.630	51.859
11)	B alpha-Chl...	5.989	5.111	222.0E6	299.8E6	57.592	50.816
12)	B 4,4'-DDE	6.159	5.300	191.0E6	291.0E6	59.486	52.773
13)	MA Dieldrin	6.309	5.431	202.1E6	310.7E6	54.458	52.604
14)	MA Endrin	6.536	5.705	169.2E6	266.8E6	55.912	49.357
15)	B Endosulfa...	6.748	5.997	189.2E6	259.7E6	59.089	50.562
16)	A 4,4'-DDD	6.668	5.852	146.9E6	245.3E6	58.084	52.135
17)	MA 4,4'-DDT	6.982	6.105	154.6E6	248.7E6	53.905	49.162
18)	B Endrin al...	6.878	6.175	121.7E6	198.5E6	56.703	54.621
19)	B Endosulfa...	7.110	6.398	149.2E6	257.9E6	51.921	50.706
20)	A Methoxychlor	7.455	6.676	73357486	130.4E6	49.959	47.600
21)	B Endrin ke...	7.590	6.902	162.3E6	293.1E6	53.985	52.632
22)	Mirex	8.068	7.092	127.9E6	218.8E6	51.581	50.172
24)	Chlordane-2	5.236f	0.000	243.6E6	0	1392.132	N.D. #
25)	Chlordane-3	5.909	5.047	215.5E6	304.9E6	322.623	438.568 #
26)	Chlordane-4	5.989	5.111	222.0E6	299.8E6	268.210	481.674 #
27)	Chlordane-5	0.000	5.997	0	259.7E6	N.D.	1017.978 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.



#1 Tetrachloro-m-xylene


R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 183186274
Conc: 57.59 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 261180108
Conc: 54.67 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.005 min
Response: 265285541
Conc: 57.34 ng/ml



#2 alpha-BHC


R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 386992270
Conc: 54.68 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.305 min
Delta R.T.: -0.005 min
Response: 248023524
Conc: 56.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 357789794
Conc: 54.18 ng/ml




#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.006 min
Response: 249249228
Conc: 59.98 ng/ml



#4 Heptachlor


R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 346682651
Conc: 52.04 ng/ml



#5 Aldrin


R.T.: 5.236 min
Delta R.T.: -0.006 min
Response: 243632550
Conc: 56.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.293 min
Delta R.T.: -0.003 min
Response: 330929439
Conc: 53.31 ng/ml




#6 beta-BHC

R.T.: 4.492 min
Delta R.T.: -0.005 min
Response: 96682513
Conc: 53.56 ng/ml



#6 beta-BHC


R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 150074487
Conc: 53.20 ng/ml



#7 delta-BHC


R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 234013055
Conc: 58.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.192 min
Delta R.T.: -0.003 min
Response: 348342823
Conc: 53.81 ng/ml




#8 Heptachlor epoxide

R.T.: 5.655 min
Delta R.T.: -0.007 min
Response: 221190087
Conc: 57.35 ng/ml



#8 Heptachlor epoxide


R.T.: 4.795 min
Delta R.T.: -0.004 min
Response: 293085163
Conc: 51.32 ng/ml



#9 Endosulfan I


R.T.: 6.037 min
Delta R.T.: -0.006 min
Response: 198295830
Conc: 55.14 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.166 min
Delta R.T.: -0.003 min
Response: 280339392
Conc: 50.68 ng/ml




#10 gamma-Chlordane

R.T.: 5.909 min
Delta R.T.: -0.007 min
Response: 215492172
Conc: 56.63 ng/ml



#10 gamma-Chlordane


R.T.: 5.047 min
Delta R.T.: -0.004 min
Response: 304937077
Conc: 51.86 ng/ml



#11 alpha-Chlordane


R.T.: 5.989 min
Delta R.T.: -0.007 min
Response: 221997369
Conc: 57.59 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 299761391
Conc: 50.82 ng/ml




#12 4,4'-DDE

R.T.: 6.159 min
Delta R.T.: -0.007 min
Response: 191025321
Conc: 59.49 ng/ml



#12 4,4'-DDE


R.T.: 5.300 min
Delta R.T.: -0.004 min
Response: 291012828
Conc: 52.77 ng/ml



#13 Dieldrin


R.T.: 6.309 min
Delta R.T.: -0.007 min
Response: 202100038
Conc: 54.46 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.431 min
Delta R.T.: -0.003 min
Response: 310705546
Conc: 52.60 ng/ml




#14 Endrin

R.T.: 6.536 min
Delta R.T.: -0.006 min
Response: 169234012
Conc: 55.91 ng/ml



#14 Endrin


R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 266823917
Conc: 49.36 ng/ml



#15 Endosulfan II


R.T.: 6.748 min
Delta R.T.: -0.007 min
Response: 189241749
Conc: 59.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 5.997 min
Delta R.T.: -0.004 min
Response: 259698451
Conc: 50.56 ng/ml




#16 4,4'-DDD

R.T.: 6.668 min
Delta R.T.: -0.007 min
Response: 146856670
Conc: 58.08 ng/ml



#16 4,4'-DDD


R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 245312145
Conc: 52.14 ng/ml



#17 4,4'-DDT


R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 154583352
Conc: 53.91 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 248673819
Conc: 49.16 ng/ml




#18 Endrin aldehyde

R.T.: 6.878 min
Delta R.T.: -0.006 min
Response: 121679916
Conc: 56.70 ng/ml



#18 Endrin aldehyde


R.T.: 6.175 min
Delta R.T.: -0.004 min
Response: 198549833
Conc: 54.62 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.110 min
Delta R.T.: -0.007 min
Response: 149205288
Conc: 51.92 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 257902949
Conc: 50.71 ng/ml




#20 Methoxychlor

R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 73357486
Conc: 49.96 ng/ml



#20 Methoxychlor


R.T.: 6.676 min
Delta R.T.: -0.005 min
Response: 130447321
Conc: 47.60 ng/ml



#21 Endrin ketone


R.T.: 7.590 min
Delta R.T.: -0.007 min
Response: 162317226
Conc: 53.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 293146987
Conc: 52.63 ng/ml




#22 Mirex

R.T.: 8.068 min
Delta R.T.: -0.007 min
Response: 127887459
Conc: 51.58 ng/ml



#22 Mirex


R.T.: 7.092 min
Delta R.T.: -0.005 min
Response: 218791981
Conc: 50.17 ng/ml



#24 Chlordane-2


R.T.: 5.236 min
Delta R.T.: 0.023 min
Response: 243632550
Conc: 1392.13 ng/m

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 4.418 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.909 min
Delta R.T.: -0.007 min
Response: 215492172
Conc: 322.62 ng/ml



#25 Chlordane-3


R.T.: 5.047 min
Delta R.T.: -0.005 min
Response: 304937077
Conc: 438.57 ng/ml



#26 Chlordane-4


R.T.: 5.989 min
Delta R.T.: -0.012 min
Response: 221997369
Conc: 268.21 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 5.111 min
Delta R.T.: -0.003 min
Response: 299761391
Conc: 481.67 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.838 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.997 min
Delta R.T.: -0.012 min
Response: 259698451
Conc: 1017.98 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.005 min
Delta R.T.: -0.011 min
Response: 155975968
Conc: 65.41 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 256275819
Conc: 59.07 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096839.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 21:15
 Operator : AR\AJ
 Sample : Q2815-11
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TW-22M-W

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:57:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.826	90354924	85851950	28.406	17.969 #
28)	SA Decachlor...	9.004	7.987	29279186	53534074	12.278	12.340
Target Compounds							
4)	MA Heptachlor	4.921f	0.000	32926209	0	7.923	N.D. #
9)	A Endosulfan I	0.000	5.178	0	81096586	N.D.	14.661 #
11)	B alpha-Chl...	5.996	5.103	11436760	7016053	2.967	1.189 #
12)	B 4,4'-DDE	6.137f	5.305	58713135	23424148	18.284	4.248 #
14)	MA Endrin	0.000	5.714	0	5202198	N.D.	0.962 #
17)	MA 4,4'-DDT	6.982	6.120	7635862	32723031	2.663	6.469 #
18)	B Endrin al...	0.000	6.166	0	13730005	N.D.	3.777 #
23)	Chlordane-1	0.000	3.826	0	22493702	N.D.	107.298 #
26)	Chlordane-4	5.996	5.103	11436760	7016053	13.818	11.274



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096839.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 21:15
Operator : AR\AJ
Sample : Q2815-11
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TW-22M-W

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:57:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 90354924
Conc: 28.41 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-22M-W




#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 85851950
Conc: 17.97 ng/ml




#4 Heptachlor

R.T.: 4.921 min
Delta R.T.: 0.018 min
Response: 32926209
Conc: 7.92 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T.: 4.014 min
Response: 0
Conc: N.D.



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 6.044 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TW-22M-W




#9 Endosulfan I

R.T.: 5.178 min
Delta R.T.: 0.009 min
Response: 81096586
Conc: 14.66 ng/ml




#11 alpha-Chlordane

R.T.: 5.996 min
Delta R.T.: 0.000 min
Response: 11436760
Conc: 2.97 ng/ml



#11 alpha-Chlordane


R.T.: 5.103 min
Delta R.T.: -0.012 min
Response: 7016053
Conc: 1.19 ng/ml



#12 4,4'-DDE


R.T.: 6.137 min
Delta R.T.: -0.029 min
Response: 58713135
Conc: 18.28 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-22M-W




#12 4,4'-DDE

R.T.: 5.305 min
Delta R.T.: 0.001 min
Response: 23424148
Conc: 4.25 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.542 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.714 min
Delta R.T.: 0.005 min
Response: 5202198
Conc: 0.96 ng/ml



#17 4,4'-DDT


R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 7635862
Conc: 2.66 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-22M-W




#17 4,4'-DDT

R.T.: 6.120 min
Delta R.T.: 0.011 min
Response: 32723031
Conc: 6.47 ng/ml




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.884 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.166 min
Delta R.T.: -0.013 min
Response: 13730005
Conc: 3.78 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 4.690 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TW-22M-W




#23 Chlordane-1

R.T.: 3.826 min
Delta R.T.: -0.012 min
Response: 22493702
Conc: 107.30 ng/ml




#26 Chlordane-4

R.T.: 5.996 min
Delta R.T.: -0.005 min
Response: 11436760
Conc: 13.82 ng/ml



#26 Chlordane-4


R.T.: 5.103 min
Delta R.T.: -0.012 min
Response: 7016053
Conc: 11.27 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.011 min
Response: 29279186
Conc: 12.28 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-22M-W



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 53534074
Conc: 12.34 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096840.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 21:56
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:57:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.531	2.825	65743372	98589616	20.669	20.635
28) SA Decachlor...	9.003	7.986	52727049	95339152	22.111	21.976
Target Compounds						
14) MA Endrin	0.000	5.714	0	16670766	N.D.	3.084 #
15) B Endosulfa...	6.736f	0.000	9162922	0	2.861	N.D. #
18) B Endrin al...	0.000	6.170	0	11141610	N.D.	3.065 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096840.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 21:56
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:57:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 65743372
Conc: 20.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 98589616
Conc: 20.64 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.542 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.714 min
Delta R.T.: 0.005 min
Response: 16670766
Conc: 3.08 ng/ml



#15 Endosulfan II


R.T.: 6.736 min
Delta R.T.: -0.019 min
Response: 9162922
Conc: 2.86 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.001 min
Response: 0
Conc: N.D.




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.884 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.170 min
Delta R.T.: -0.009 min
Response: 11141610
Conc: 3.07 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.003 min
Delta R.T.: -0.012 min
Response: 52727049
Conc: 22.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 95339152
Conc: 21.98 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096841.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 22:10
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:57:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.825	180.6E6	264.5E6	56.769	55.355
28)	SA Decachlor...	9.004	7.986	128.5E6	232.4E6	53.888	53.559
Target Compounds							
2)	A alpha-BHC	3.978	3.331	267.2E6	392.6E6	57.750	55.472
3)	MA gamma-BHC...	4.306	3.663	250.1E6	364.4E6	56.533	55.182
4)	MA Heptachlor	4.897	4.011	245.6E6	348.9E6	59.102	52.366
5)	MB Aldrin	5.236	4.294	242.7E6	338.4E6	56.432	54.523
6)	B beta-BHC	4.492	3.959	103.1E6	155.4E6	57.118	55.083
7)	B delta-BHC	4.738	4.191	232.8E6	356.1E6	58.353	55.011
8)	B Heptachlo...	5.656	4.795	220.0E6	304.9E6	57.034	53.391
9)	A Endosulfan I	6.037	5.166	197.9E6	279.5E6	55.031	50.530
10)	B gamma-Chl...	5.909	5.047	217.0E6	318.9E6	57.015	54.231
11)	B alpha-Chl...	5.990	5.111	216.1E6	312.8E6	56.054	53.024
12)	B 4,4'-DDE	6.160	5.300	181.3E6	294.5E6	56.444	53.398
13)	MA Dieldrin	6.309	5.430	203.4E6	312.0E6	54.815	52.824
14)	MA Endrin	6.536	5.705	158.2E6	280.2E6	52.251	51.832
15)	B Endosulfa...	6.748	5.997	193.0E6	270.0E6	60.262	52.570
16)	A 4,4'-DDD	6.668	5.852	154.9E6	257.4E6	61.268	54.710
17)	MA 4,4'-DDT	6.982	6.104	147.8E6	243.3E6	51.533	48.098
18)	B Endrin al...	6.878	6.175	123.9E6	201.7E6	57.736	55.476
19)	B Endosulfa...	7.111	6.398	153.4E6	264.4E6	53.374	51.978
20)	A Methoxychlor	7.455	6.676	71143490	128.0E6	48.451	46.700
21)	B Endrin ke...	7.590	6.902	165.3E6	307.3E6	54.974	55.181
22)	Mirex	8.068	7.091	127.4E6	230.2E6	51.398	52.777
24)	Chlordane-2	5.236f	0.000	242.7E6	0	1386.874	N.D. #
25)	Chlordane-3	5.909	5.047	217.0E6	318.9E6	324.816	458.624 #
26)	Chlordane-4	5.990	5.111	216.1E6	312.8E6	261.049	502.603 #
27)	Chlordane-5	0.000	5.997	0	270.0E6	N.D.	1058.401 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096841.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 22:10
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:57:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 180569649
Conc: 56.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 264474564
Conc: 55.36 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 267171388
Conc: 57.75 ng/ml



#2 alpha-BHC


R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 392574947
Conc: 55.47 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.306 min
Delta R.T.: -0.005 min
Response: 250071176
Conc: 56.53 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.663 min
Delta R.T.: -0.002 min
Response: 364386933
Conc: 55.18 ng/ml




#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.005 min
Response: 245610481
Conc: 59.10 ng/ml



#4 Heptachlor


R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 348861586
Conc: 52.37 ng/ml



#5 Aldrin


R.T.: 5.236 min
Delta R.T.: -0.007 min
Response: 242712510
Conc: 56.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.294 min
Delta R.T.: -0.003 min
Response: 338425766
Conc: 54.52 ng/ml




#6 beta-BHC

R.T.: 4.492 min
Delta R.T.: -0.005 min
Response: 103096953
Conc: 57.12 ng/ml



#6 beta-BHC


R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 155383025
Conc: 55.08 ng/ml



#7 delta-BHC


R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 232818505
Conc: 58.35 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.191 min
Delta R.T.: -0.004 min
Response: 356121645
Conc: 55.01 ng/ml




#8 Heptachlor epoxide

R.T.: 5.656 min
Delta R.T.: -0.007 min
Response: 219954792
Conc: 57.03 ng/ml



#8 Heptachlor epoxide


R.T.: 4.795 min
Delta R.T.: -0.004 min
Response: 304932521
Conc: 53.39 ng/ml



#9 Endosulfan I


R.T.: 6.037 min
Delta R.T.: -0.006 min
Response: 197913152
Conc: 55.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.166 min
Delta R.T.: -0.003 min
Response: 279504144
Conc: 50.53 ng/ml




#10 gamma-Chlordane

R.T.: 5.909 min
Delta R.T.: -0.007 min
Response: 216956458
Conc: 57.02 ng/ml



#10 gamma-Chlordane


R.T.: 5.047 min
Delta R.T.: -0.004 min
Response: 318882402
Conc: 54.23 ng/ml



#11 alpha-Chlordane


R.T.: 5.990 min
Delta R.T.: -0.006 min
Response: 216069565
Conc: 56.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 312786250
Conc: 53.02 ng/ml




#12 4,4'-DDE

R.T.: 6.160 min
Delta R.T.: -0.006 min
Response: 181255368
Conc: 56.44 ng/ml



#12 4,4'-DDE


R.T.: 5.300 min
Delta R.T.: -0.004 min
Response: 294458512
Conc: 53.40 ng/ml



#13 Dieldrin


R.T.: 6.309 min
Delta R.T.: -0.007 min
Response: 203426141
Conc: 54.81 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 312007966
Conc: 52.82 ng/ml




#14 Endrin

R.T.: 6.536 min
Delta R.T.: -0.007 min
Response: 158153855
Conc: 52.25 ng/ml



#14 Endrin


R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 280200055
Conc: 51.83 ng/ml



#15 Endosulfan II


R.T.: 6.748 min
Delta R.T.: -0.007 min
Response: 192997115
Conc: 60.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 5.997 min
Delta R.T.: -0.004 min
Response: 270010863
Conc: 52.57 ng/ml




#16 4,4'-DDD

R.T.: 6.668 min
Delta R.T.: -0.007 min
Response: 154906864
Conc: 61.27 ng/ml



#16 4,4'-DDD


R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 257426326
Conc: 54.71 ng/ml



#17 4,4'-DDT


R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 147780944
Conc: 51.53 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.104 min
Delta R.T.: -0.005 min
Response: 243296006
Conc: 48.10 ng/ml




#18 Endrin aldehyde

R.T.: 6.878 min
Delta R.T.: -0.006 min
Response: 123895984
Conc: 57.74 ng/ml



#18 Endrin aldehyde


R.T.: 6.175 min
Delta R.T.: -0.004 min
Response: 201657062
Conc: 55.48 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.111 min
Delta R.T.: -0.007 min
Response: 153379636
Conc: 53.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 264373350
Conc: 51.98 ng/ml




#20 Methoxychlor

R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 71143490
Conc: 48.45 ng/ml



#20 Methoxychlor


R.T.: 6.676 min
Delta R.T.: -0.005 min
Response: 127981208
Conc: 46.70 ng/ml



#21 Endrin ketone


R.T.: 7.590 min
Delta R.T.: -0.007 min
Response: 165290236
Conc: 54.97 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 307346587
Conc: 55.18 ng/ml




#22 Mirex

R.T.: 8.068 min
Delta R.T.: -0.008 min
Response: 127434169
Conc: 51.40 ng/ml



#22 Mirex


R.T.: 7.091 min
Delta R.T.: -0.006 min
Response: 230155234
Conc: 52.78 ng/ml



#24 Chlordane-2


R.T.: 5.236 min
Delta R.T.: 0.023 min
Response: 242712510
Conc: 1386.87 ng/m

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 4.418 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.909 min
Delta R.T.: -0.007 min
Response: 216956458
Conc: 324.82 ng/ml



#25 Chlordane-3


R.T.: 5.047 min
Delta R.T.: -0.005 min
Response: 318882402
Conc: 458.62 ng/ml



#26 Chlordane-4


R.T.: 5.990 min
Delta R.T.: -0.011 min
Response: 216069565
Conc: 261.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 312786250
Conc: 502.60 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.838 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.997 min
Delta R.T.: -0.012 min
Response: 270010863
Conc: 1058.40 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.012 min
Response: 128504088
Conc: 53.89 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 232358139
Conc: 53.56 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
 Data File : PL096844.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2025 10:26
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 12:38:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.531	2.825	73626080	104.4E6	23.147	21.848
28) SA Decachlor...	9.006	7.987	54106762	94409704	22.690	21.762
Target Compounds						
2) A alpha-BHC	3.978	3.330	52309665	74616765	11.307	10.544
3) MA gamma-BHC...	4.305	3.662	50236408	70745882	11.357	10.714
6) B beta-BHC	4.493	3.958	21236931	32707689	11.766	11.595
14) MA Endrin	6.536	5.705	164.6E6	271.5E6	54.367	50.230
16) A 4,4'-DDD	6.669	5.852	9335872	17541435	3.692	3.728
17) MA 4,4'-DDT	6.983	6.105	308.3E6	511.2E6	107.516	101.064
18) B Endrin al...	0.000	6.175	0	14799158	N.D.	4.071 #
20) A Methoxychlor	7.455	6.677	374.7E6	603.2E6	255.163	220.112
21) B Endrin ke...	7.590	6.902	8908592	21316876	2.963	3.827 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096844.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 10:26
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 12:38:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 73626080
Conc: 23.15 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 104384511
Conc: 21.85 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.005 min
Response: 52309665
Conc: 11.31 ng/ml



#2 alpha-BHC


R.T.: 3.330 min
Delta R.T.: -0.003 min
Response: 74616765
Conc: 10.54 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.305 min
Delta R.T.: -0.005 min
Response: 50236408
Conc: 11.36 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 70745882
Conc: 10.71 ng/ml




#6 beta-BHC

R.T.: 4.493 min
Delta R.T.: -0.004 min
Response: 21236931
Conc: 11.77 ng/ml



#6 beta-BHC


R.T.: 3.958 min
Delta R.T.: -0.003 min
Response: 32707689
Conc: 11.59 ng/ml



#14 Endrin


R.T.: 6.536 min
Delta R.T.: -0.006 min
Response: 164558169
Conc: 54.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#14 Endrin

R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 271542500
Conc: 50.23 ng/ml




#16 4,4'-DDD

R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 9335872
Conc: 3.69 ng/ml



#16 4,4'-DDD


R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 17541435
Conc: 3.73 ng/ml



#17 4,4'-DDT


R.T.: 6.983 min
Delta R.T.: -0.006 min
Response: 308321781
Conc: 107.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 511212306
Conc: 101.06 ng/ml




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.884 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.175 min
Delta R.T.: -0.004 min
Response: 14799158
Conc: 4.07 ng/ml



#20 Methoxychlor


R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 374669518
Conc: 255.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#20 Methoxychlor

R.T.: 6.677 min
Delta R.T.: -0.004 min
Response: 603217501
Conc: 220.11 ng/ml




#21 Endrin ketone

R.T.: 7.590 min
Delta R.T.: -0.006 min
Response: 8908592
Conc: 2.96 ng/ml



#21 Endrin ketone


R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 21316876
Conc: 3.83 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.006 min
Delta R.T.: -0.009 min
Response: 54106762
Conc: 22.69 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 94409704
Conc: 21.76 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
 Data File : PL096850.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2025 12:53
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 16:14:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.825	180.5E6	258.1E6	56.743	54.013
28)	SA Decachlor...	9.005	7.987	125.2E6	215.8E6	52.509	49.743
Target Compounds							
2)	A alpha-BHC	3.978	3.331	264.2E6	382.2E6	57.098	54.010
3)	MA gamma-BHC...	4.305	3.663	246.4E6	354.3E6	55.695	53.650
4)	MA Heptachlor	4.897	4.011	243.4E6	342.8E6	58.563	51.453
5)	MB Aldrin	5.236	4.293	239.5E6	326.2E6	55.696	52.556
6)	B beta-BHC	4.492	3.959	100.4E6	151.9E6	55.639	53.854
7)	B delta-BHC	4.738	4.191	230.8E6	345.4E6	57.846	53.355
8)	B Heptachlo...	5.656	4.795	216.8E6	295.1E6	56.227	51.661
9)	A Endosulfan I	6.037	5.166	194.7E6	272.8E6	54.131	49.314
10)	B gamma-Chl...	5.909	5.047	214.7E6	307.6E6	56.428	52.309
11)	B alpha-Chl...	5.990	5.111	213.9E6	301.7E6	55.491	51.139
12)	B 4,4'-DDE	6.160	5.300	177.0E6	284.8E6	55.124	51.643
13)	MA Dieldrin	6.309	5.431	202.8E6	303.2E6	54.636	51.330
14)	MA Endrin	6.536	5.705	153.1E6	259.3E6	50.576	47.971
15)	B Endosulfa...	6.748	5.997	179.8E6	260.0E6	56.140	50.628
16)	A 4,4'-DDD	6.669	5.852	144.7E6	245.1E6	57.213	52.092
17)	MA 4,4'-DDT	6.983	6.105	146.7E6	240.9E6	51.141	47.630
18)	B Endrin al...	6.877	6.175	119.2E6	190.7E6	55.544	52.471
19)	B Endosulfa...	7.110	6.398	150.1E6	251.8E6	52.224	49.499
20)	A Methoxychlor	7.455	6.676	72047040	127.7E6	49.067	46.583
21)	B Endrin ke...	7.590	6.902	162.4E6	297.4E6	54.012	53.401
22)	Mirex	8.069	7.093	125.4E6	218.4E6	50.561	50.088
24)	Chlordane-2	5.236f	0.000	239.5E6	0	1368.793	N.D. #
25)	Chlordane-3	5.909	5.047	214.7E6	307.6E6	321.472	442.377 #
26)	Chlordane-4	5.990	5.111	213.9E6	301.7E6	258.427	484.731 #
27)	Chlordane-5	0.000	5.997	0	260.0E6	N.D.	1019.308 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096850.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 12:53
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 16:14:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 180486809
Conc: 56.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 258062587
Conc: 54.01 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 264153310
Conc: 57.10 ng/ml



#2 alpha-BHC


R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 382230166
Conc: 54.01 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.305 min
Delta R.T.: -0.005 min
Response: 246361977
Conc: 55.69 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.663 min
Delta R.T.: -0.002 min
Response: 354267631
Conc: 53.65 ng/ml




#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.006 min
Response: 243370604
Conc: 58.56 ng/ml



#4 Heptachlor


R.T.: 4.011 min
Delta R.T.: -0.003 min
Response: 342778742
Conc: 51.45 ng/ml



#5 Aldrin


R.T.: 5.236 min
Delta R.T.: -0.006 min
Response: 239548150
Conc: 55.70 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.293 min
Delta R.T.: -0.004 min
Response: 326218108
Conc: 52.56 ng/ml




#6 beta-BHC

R.T.: 4.492 min
Delta R.T.: -0.005 min
Response: 100427450
Conc: 55.64 ng/ml



#6 beta-BHC


R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 151916219
Conc: 53.85 ng/ml



#7 delta-BHC


R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 230798856
Conc: 57.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.191 min
Delta R.T.: -0.004 min
Response: 345402486
Conc: 53.36 ng/ml




#8 Heptachlor epoxide

R.T.: 5.656 min
Delta R.T.: -0.006 min
Response: 216841548
Conc: 56.23 ng/ml



#8 Heptachlor epoxide


R.T.: 4.795 min
Delta R.T.: -0.004 min
Response: 295053706
Conc: 51.66 ng/ml



#9 Endosulfan I


R.T.: 6.037 min
Delta R.T.: -0.006 min
Response: 194673996
Conc: 54.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.166 min
Delta R.T.: -0.003 min
Response: 272775922
Conc: 49.31 ng/ml




#10 gamma-Chlordane

R.T.: 5.909 min
Delta R.T.: -0.006 min
Response: 214722800
Conc: 56.43 ng/ml



#10 gamma-Chlordane


R.T.: 5.047 min
Delta R.T.: -0.004 min
Response: 307585622
Conc: 52.31 ng/ml



#11 alpha-Chlordane


R.T.: 5.990 min
Delta R.T.: -0.007 min
Response: 213900002
Conc: 55.49 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 301663826
Conc: 51.14 ng/ml




#12 4,4'-DDE

R.T.: 6.160 min
Delta R.T.: -0.006 min
Response: 177016085
Conc: 55.12 ng/ml



#12 4,4'-DDE


R.T.: 5.300 min
Delta R.T.: -0.004 min
Response: 284781762
Conc: 51.64 ng/ml



#13 Dieldrin


R.T.: 6.309 min
Delta R.T.: -0.006 min
Response: 202760836
Conc: 54.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.431 min
Delta R.T.: -0.003 min
Response: 303181110
Conc: 51.33 ng/ml




#14 Endrin

R.T.: 6.536 min
Delta R.T.: -0.007 min
Response: 153081763
Conc: 50.58 ng/ml



#14 Endrin


R.T.: 5.705 min
Delta R.T.: -0.004 min
Response: 259330218
Conc: 47.97 ng/ml



#15 Endosulfan II


R.T.: 6.748 min
Delta R.T.: -0.007 min
Response: 179795338
Conc: 56.14 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 5.997 min
Delta R.T.: -0.004 min
Response: 260037643
Conc: 50.63 ng/ml




#16 4,4'-DDD

R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 144655392
Conc: 57.21 ng/ml



#16 4,4'-DDD


R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 245109783
Conc: 52.09 ng/ml



#17 4,4'-DDT


R.T.: 6.983 min
Delta R.T.: -0.006 min
Response: 146657638
Conc: 51.14 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.105 min
Delta R.T.: -0.004 min
Response: 240928658
Conc: 47.63 ng/ml




#18 Endrin aldehyde

R.T.: 6.877 min
Delta R.T.: -0.006 min
Response: 119191848
Conc: 55.54 ng/ml



#18 Endrin aldehyde


R.T.: 6.175 min
Delta R.T.: -0.004 min
Response: 190732922
Conc: 52.47 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.110 min
Delta R.T.: -0.007 min
Response: 150077065
Conc: 52.22 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 251764095
Conc: 49.50 ng/ml




#20 Methoxychlor

R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 72047040
Conc: 49.07 ng/ml



#20 Methoxychlor


R.T.: 6.676 min
Delta R.T.: -0.005 min
Response: 127661189
Conc: 46.58 ng/ml



#21 Endrin ketone


R.T.: 7.590 min
Delta R.T.: -0.007 min
Response: 162398327
Conc: 54.01 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 297432233
Conc: 53.40 ng/ml




#22 Mirex

R.T.: 8.069 min
Delta R.T.: -0.007 min
Response: 125358145
Conc: 50.56 ng/ml



#22 Mirex


R.T.: 7.093 min
Delta R.T.: -0.004 min
Response: 218428879
Conc: 50.09 ng/ml



#24 Chlordane-2


R.T.: 5.236 min
Delta R.T.: 0.023 min
Response: 239548150
Conc: 1368.79 ng/m

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 4.418 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.909 min
Delta R.T.: -0.007 min
Response: 214722800
Conc: 321.47 ng/ml



#25 Chlordane-3


R.T.: 5.047 min
Delta R.T.: -0.005 min
Response: 307585622
Conc: 442.38 ng/ml



#26 Chlordane-4


R.T.: 5.990 min
Delta R.T.: -0.011 min
Response: 213900002
Conc: 258.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 301663826
Conc: 484.73 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.838 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.997 min
Delta R.T.: -0.012 min
Response: 260037643
Conc: 1019.31 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.005 min
Delta R.T.: -0.011 min
Response: 125215431
Conc: 52.51 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 215802739
Conc: 49.74 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
 Data File : PL096855.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2025 16:30
 Operator : AR\AJ
 Sample : Q2815-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TW-705R-S

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 19 06:00:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.823	73980277	68001961	23.259	14.233 #
28)	SA Decachlor...	9.007	7.987	24573818	40225957	10.305	9.272
Target Compounds							
4)	MA Heptachlor	4.924f	0.000	31607524	0	7.606	N.D. #
10)	B gamma-Chl...	0.000	5.050	0	45455635	N.D.	7.730 #
12)	B 4,4'-DDE	6.161	5.299	173.6E6	251.1E6	54.062	45.534
13)	MA Dieldrin	0.000	5.461f	0	225.6E6	N.D.	38.197 #
14)	MA Endrin	0.000	5.713	0	6216594	N.D.	1.150 #
16)	A 4,4'-DDD	6.669	5.852	405.9E6	635.7E6	160.537	135.109
17)	MA 4,4'-DDT	7.003	6.088f	7040332	10916286	2.455	2.158
18)	B Endrin al...	0.000	6.177	0	10043705	N.D.	2.763 #
19)	B Endosulfa...	7.124	0.000	7455655	0	2.594	N.D. #
25)	Chlordane-3	0.000	5.050	0	45455635	N.D.	65.375 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096855.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 16:30
Operator : AR\AJ
Sample : Q2815-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TW-705R-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 73980277
Conc: 23.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-705R-S




#1 Tetrachloro-m-xylene

R.T.: 2.823 min
Delta R.T.: -0.005 min
Response: 68001961
Conc: 14.23 ng/ml




#4 Heptachlor

R.T.: 4.924 min
Delta R.T.: 0.022 min
Response: 31607524
Conc: 7.61 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T.: 4.014 min
Response: 0
Conc: N.D.



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.916 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TW-705R-S




#10 gamma-Chlordane

R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 45455635
Conc: 7.73 ng/ml




#12 4,4'-DDE

R.T.: 6.161 min
Delta R.T.: -0.006 min
Response: 173606744
Conc: 54.06 ng/ml



#12 4,4'-DDE


R.T.: 5.299 min
Delta R.T.: -0.005 min
Response: 251094235
Conc: 45.53 ng/ml



#13 Dieldrin


R.T.: 0.000 min
Exp R.T.: 6.316 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TW-705R-S




#13 Dieldrin

R.T.: 5.461 min
Delta R.T.: 0.027 min
Response: 225612345
Conc: 38.20 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T.: 6.542 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.713 min
Delta R.T.: 0.004 min
Response: 6216594
Conc: 1.15 ng/ml



#16 4,4'-DDD


R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 405896514
Conc: 160.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-705R-S




#16 4,4'-DDD

R.T.: 5.852 min
Delta R.T.: -0.004 min
Response: 635727139
Conc: 135.11 ng/ml




#17 4,4'-DDT

R.T.: 7.003 min
Delta R.T.: 0.014 min
Response: 7040332
Conc: 2.46 ng/ml



#17 4,4'-DDT


R.T.: 6.088 min
Delta R.T.: -0.021 min
Response: 10916286
Conc: 2.16 ng/ml



#18 Endrin aldehyde


R.T.: 0.000 min
Exp R.T.: 6.884 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TW-705R-S




#18 Endrin aldehyde

R.T.: 6.177 min
Delta R.T.: -0.002 min
Response: 10043705
Conc: 2.76 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.124 min
Delta R.T.: 0.006 min
Response: 7455655
Conc: 2.59 ng/ml



#19 Endosulfan Sulfate


R.T.: 0.000 min
Exp R.T.: 6.402 min
Response: 0
Conc: N.D.



#25 Chlordane-3


R.T.: 0.000 min
Exp R.T.: 5.916 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TW-705R-S




#25 Chlordane-3

R.T.: 5.050 min
Delta R.T.: -0.002 min
Response: 45455635
Conc: 65.38 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.007 min
Delta R.T.: -0.009 min
Response: 24573818
Conc: 10.30 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 40225957
Conc: 9.27 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
 Data File : PL096856.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2025 16:44
 Operator : AR\AJ
 Sample : Q2815-01DL 2X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TW-705R-SDL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 19 06:00:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.825	41074504	35190031	12.913	7.365 #
28)	SA Decachlor...	9.005	7.986	12736470	21675838	5.341	4.996
Target Compounds							
4)	MA Heptachlor	4.895	0.000	20229589	0	4.868	N.D. #
10)	B gamma-Chl...	0.000	5.050	0	24918013	N.D.	4.238 #
12)	B 4,4'-DDE	6.159	5.299	87221532	123.0E6	27.161	22.307
13)	MA Dieldrin	0.000	5.461f	0	115.9E6	N.D.	19.630 #
14)	MA Endrin	0.000	5.713	0	4297567	N.D.	0.795 #
16)	A 4,4'-DDD	6.668	5.851	190.3E6	320.5E6	75.265	68.109
18)	B Endrin al...	0.000	6.176	0	9061843	N.D.	2.493 #
25)	Chlordane-3	0.000	5.050	0	24918013	N.D.	35.838 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096856.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 16:44
Operator : AR\AJ
Sample : Q2815-01DL 2X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TW-705R-SDL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.530 min
Delta R.T.: -0.005 min
Response: 41074504
Conc: 12.91 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-705R-SDL




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 35190031
Conc: 7.37 ng/ml




#4 Heptachlor

R.T.: 4.895 min
Delta R.T.: -0.007 min
Response: 20229589
Conc: 4.87 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T.: 4.014 min
Response: 0
Conc: N.D.



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.916 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TW-705R-SDL




#10 gamma-Chlordane

R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 24918013
Conc: 4.24 ng/ml




#12 4,4'-DDE

R.T.: 6.159 min
Delta R.T.: -0.007 min
Response: 87221532
Conc: 27.16 ng/ml



#12 4,4'-DDE


R.T.: 5.299 min
Delta R.T.: -0.005 min
Response: 123007575
Conc: 22.31 ng/ml



#13 Dieldrin


R.T.: 0.000 min
Exp R.T.: 6.316 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TW-705R-SDL




#13 Dieldrin

R.T.: 5.461 min
Delta R.T.: 0.027 min
Response: 115946893
Conc: 19.63 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T.: 6.542 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.713 min
Delta R.T.: 0.004 min
Response: 4297567
Conc: 0.79 ng/ml



#16 4,4'-DDD


R.T.: 6.668 min
Delta R.T.: -0.007 min
Response: 190297206
Conc: 75.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-705R-SDL




#16 4,4'-DDD

R.T.: 5.851 min
Delta R.T.: -0.005 min
Response: 320473038
Conc: 68.11 ng/ml




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.884 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.176 min
Delta R.T.: -0.003 min
Response: 9061843
Conc: 2.49 ng/ml



#25 Chlordane-3


R.T.: 0.000 min
Exp R.T.: 5.916 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TW-705R-SDL




#25 Chlordane-3

R.T.: 5.050 min
Delta R.T.: -0.002 min
Response: 24918013
Conc: 35.84 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.005 min
Delta R.T.: -0.011 min
Response: 12736470
Conc: 5.34 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 21675838
Conc: 5.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096857.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 16:57
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.825	66964840	100.5E6	21.053	21.041
28)	SA Decachlor...	9.004	7.986	51749831	94895525	21.701	21.874

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096857.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 16:57
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 66964840
Conc: 21.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 100526992
Conc: 21.04 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.012 min
Response: 51749831
Conc: 21.70 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 94895525
Conc: 21.87 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
 Data File : PL096858.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2025 17:11
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 19 06:00:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.825	178.3E6	255.7E6	56.053	53.513
28)	SA Decachlor...	9.004	7.986	123.9E6	222.0E6	51.944	51.165
Target Compounds							
2)	A alpha-BHC	3.978	3.330	264.4E6	379.1E6	57.144	53.569
3)	MA gamma-BHC...	4.306	3.662	247.9E6	351.4E6	56.045	53.208
4)	MA Heptachlor	4.897	4.010	239.3E6	336.9E6	57.586	50.577
5)	MB Aldrin	5.236	4.293	238.3E6	326.2E6	55.402	52.559
6)	B beta-BHC	4.492	3.958	96617749	151.1E6	53.529	53.581
7)	B delta-BHC	4.738	4.191	230.8E6	344.7E6	57.852	53.254
8)	B Heptachlo...	5.655	4.794	223.7E6	294.9E6	58.001	51.630
9)	A Endosulfan I	6.037	5.165	194.5E6	275.2E6	54.088	49.757
10)	B gamma-Chl...	5.909	5.046	214.6E6	309.2E6	56.404	52.576
11)	B alpha-Chl...	5.989	5.110	212.3E6	303.4E6	55.083	51.430
12)	B 4,4'-DDE	6.159	5.299	177.6E6	286.7E6	55.293	51.999
13)	MA Dieldrin	6.309	5.430	199.3E6	304.1E6	53.708	51.481
14)	MA Endrin	6.534	5.704	150.4E6	265.6E6	49.705	49.140
15)	B Endosulfa...	6.748	5.996	177.5E6	263.1E6	55.422	51.215
16)	A 4,4'-DDD	6.667	5.851	144.8E6	250.5E6	57.289	53.228
17)	MA 4,4'-DDT	6.982	6.104	140.9E6	234.9E6	49.125	46.430
18)	B Endrin al...	6.876	6.174	117.8E6	192.7E6	54.905	52.999
19)	B Endosulfa...	7.110	6.397	150.0E6	255.9E6	52.188	50.320
20)	A Methoxychlor	7.454	6.676	69792447	123.5E6	47.531	45.054
21)	B Endrin ke...	7.589	6.901	163.4E6	294.6E6	54.341	52.884
22)	Mirex	8.068	7.091	125.3E6	222.9E6	50.536	51.123
24)	Chlordane-2	5.236f	0.000	238.3E6	0	1361.562	N.D. #
25)	Chlordane-3	5.909	5.046	214.6E6	309.2E6	321.332	444.628 #
26)	Chlordane-4	5.989	5.110	212.3E6	303.4E6	256.527	487.493 #
27)	Chlordane-5	0.000	5.996	0	263.1E6	N.D.	1031.131 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081825\
Data File : PL096858.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 17:11
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:00:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 178292660
Conc: 56.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 255673747
Conc: 53.51 ng/ml




#2 alpha-BHC

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 264370535
Conc: 57.14 ng/ml



#2 alpha-BHC


R.T.: 3.330 min
Delta R.T.: -0.003 min
Response: 379109531
Conc: 53.57 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.306 min
Delta R.T.: -0.005 min
Response: 247910221
Conc: 56.04 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 351354786
Conc: 53.21 ng/ml




#4 Heptachlor

R.T.: 4.897 min
Delta R.T.: -0.006 min
Response: 239310105
Conc: 57.59 ng/ml



#4 Heptachlor


R.T.: 4.010 min
Delta R.T.: -0.004 min
Response: 336938392
Conc: 50.58 ng/ml



#5 Aldrin


R.T.: 5.236 min
Delta R.T.: -0.007 min
Response: 238282661
Conc: 55.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.293 min
Delta R.T.: -0.004 min
Response: 326237655
Conc: 52.56 ng/ml




#6 beta-BHC

R.T.: 4.492 min
Delta R.T.: -0.005 min
Response: 96617749
Conc: 53.53 ng/ml



#6 beta-BHC


R.T.: 3.958 min
Delta R.T.: -0.003 min
Response: 151147750
Conc: 53.58 ng/ml



#7 delta-BHC


R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 230820906
Conc: 57.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.191 min
Delta R.T.: -0.004 min
Response: 344747683
Conc: 53.25 ng/ml




#8 Heptachlor epoxide

R.T.: 5.655 min
Delta R.T.: -0.007 min
Response: 223683999
Conc: 58.00 ng/ml



#8 Heptachlor epoxide


R.T.: 4.794 min
Delta R.T.: -0.005 min
Response: 294878831
Conc: 51.63 ng/ml



#9 Endosulfan I


R.T.: 6.037 min
Delta R.T.: -0.007 min
Response: 194520543
Conc: 54.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.165 min
Delta R.T.: -0.004 min
Response: 275227850
Conc: 49.76 ng/ml




#10 gamma-Chlordane

R.T.: 5.909 min
Delta R.T.: -0.007 min
Response: 214629910
Conc: 56.40 ng/ml



#10 gamma-Chlordane


R.T.: 5.046 min
Delta R.T.: -0.005 min
Response: 309151010
Conc: 52.58 ng/ml



#11 alpha-Chlordane


R.T.: 5.989 min
Delta R.T.: -0.008 min
Response: 212327187
Conc: 55.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.110 min
Delta R.T.: -0.005 min
Response: 303382700
Conc: 51.43 ng/ml




#12 4,4'-DDE

R.T.: 6.159 min
Delta R.T.: -0.007 min
Response: 177559416
Conc: 55.29 ng/ml



#12 4,4'-DDE


R.T.: 5.299 min
Delta R.T.: -0.005 min
Response: 286745809
Conc: 52.00 ng/ml



#13 Dieldrin


R.T.: 6.309 min
Delta R.T.: -0.007 min
Response: 199318111
Conc: 53.71 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 304075067
Conc: 51.48 ng/ml




#14 Endrin

R.T.: 6.534 min
Delta R.T.: -0.008 min
Response: 150445307
Conc: 49.70 ng/ml



#14 Endrin


R.T.: 5.704 min
Delta R.T.: -0.005 min
Response: 265648440
Conc: 49.14 ng/ml



#15 Endosulfan II


R.T.: 6.748 min
Delta R.T.: -0.007 min
Response: 177495778
Conc: 55.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 5.996 min
Delta R.T.: -0.005 min
Response: 263053855
Conc: 51.22 ng/ml




#16 4,4'-DDD

R.T.: 6.667 min
Delta R.T.: -0.008 min
Response: 144848017
Conc: 57.29 ng/ml



#16 4,4'-DDD


R.T.: 5.851 min
Delta R.T.: -0.005 min
Response: 250453109
Conc: 53.23 ng/ml



#17 4,4'-DDT


R.T.: 6.982 min
Delta R.T.: -0.007 min
Response: 140873930
Conc: 49.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.104 min
Delta R.T.: -0.005 min
Response: 234856582
Conc: 46.43 ng/ml




#18 Endrin aldehyde

R.T.: 6.876 min
Delta R.T.: -0.007 min
Response: 117821788
Conc: 54.91 ng/ml



#18 Endrin aldehyde


R.T.: 6.174 min
Delta R.T.: -0.005 min
Response: 192653257
Conc: 53.00 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.110 min
Delta R.T.: -0.008 min
Response: 149972589
Conc: 52.19 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.397 min
Delta R.T.: -0.005 min
Response: 255940668
Conc: 50.32 ng/ml




#20 Methoxychlor

R.T.: 7.454 min
Delta R.T.: -0.008 min
Response: 69792447
Conc: 47.53 ng/ml



#20 Methoxychlor


R.T.: 6.676 min
Delta R.T.: -0.005 min
Response: 123471413
Conc: 45.05 ng/ml



#21 Endrin ketone


R.T.: 7.589 min
Delta R.T.: -0.008 min
Response: 163387029
Conc: 54.34 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.901 min
Delta R.T.: -0.006 min
Response: 294551207
Conc: 52.88 ng/ml




#22 Mirex

R.T.: 8.068 min
Delta R.T.: -0.008 min
Response: 125296545
Conc: 50.54 ng/ml



#22 Mirex


R.T.: 7.091 min
Delta R.T.: -0.006 min
Response: 222941586
Conc: 51.12 ng/ml



#24 Chlordane-2


R.T.: 5.236 min
Delta R.T.: 0.022 min
Response: 238282661
Conc: 1361.56 ng/m

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 4.418 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.909 min
Delta R.T.: -0.007 min
Response: 214629910
Conc: 321.33 ng/ml



#25 Chlordane-3


R.T.: 5.046 min
Delta R.T.: -0.006 min
Response: 309151010
Conc: 444.63 ng/ml



#26 Chlordane-4


R.T.: 5.989 min
Delta R.T.: -0.012 min
Response: 212327187
Conc: 256.53 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 5.110 min
Delta R.T.: -0.005 min
Response: 303382700
Conc: 487.49 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.838 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.996 min
Delta R.T.: -0.013 min
Response: 263053855
Conc: 1031.13 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.012 min
Response: 123868220
Conc: 51.94 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.007 min
Response: 221969720
Conc: 51.16 ng/ml