


LB136788
INSTRUMENT ID:CV1

7470A

μ Abs.:

41723

Linear



A= 0.0000e+000

B= 2.3988e-004

C= 3.3618e-002

Rho= 0.9999451

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.033	0.033	-3	0.000	-3	0				
0.2	0.200	0.218	0.018	768	0.0 %	768					
2.5	2.500	2.485	-0.015	10218	0.0 %	10218	0				
5.0	5.000	4.926	-0.074	20393	0.0 %	20393					
7.5	7.500	7.497	-0.003	31112	0.0 %	31112					
10.0	10.000	10.042	0.042	41723	0.0 %	41723					

LB136788 INSTRUMENT ID:CV1

Method: 7470A Operator: Admin Date of Analysis: 12 Aug 2025 13:54:25

Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50.0	-	-3	PPB	Std	0.0000	7470A	12 Aug 2025 14:07:24
S	0.2 - 1	50.2	-	768	PPB	Std	0.2000	7470A	12 Aug 2025 14:09:41
S	2.5 - 1	52.5	-	10218	PPB	Std	2.5000	7470A	12 Aug 2025 14:11:59
S	5.0 - 1	55.0	-	20393	PPB	Std	5.0000	7470A	12 Aug 2025 14:14:15
S	7.5 - 1	57.5	-	31112	PPB	Std	7.5000	7470A	12 Aug 2025 14:16:32
S	10.0 - 1	510.0	-	41723	PPB	Std	10.0000	7470A	12 Aug 2025 14:18:50
U	ICV10 - 1	ICV10	4.1399	17118	PPB	SMPL	-7470A		12 Aug 2025 14:21:34
U	ICB10 - 1	ICB10	-0.0369	-294	PPB	SMPL	-7470A		12 Aug 2025 14:23:49
U	CCV29 - 1	CCV29	4.9296	20410	PPB	SMPL	-7470A		12 Aug 2025 14:26:06
U	CCB29 - 1	CCB29	-0.0273	-254	PPB	SMPL	-7470A		12 Aug 2025 14:28:24
U	CRA - 1	CRA	0.2118	743	PPB	SMPL	-7470A		12 Aug 2025 14:30:43
U	HighStd - 1	HighStd	10.1276	42079	PPB	SMPL	-7470A		12 Aug 2025 14:33:01
U	ChkStd - 1	ChkStd	7.3618	30549	PPB	SMPL	-7470A		12 Aug 2025 14:35:20
U	PB169202BL - 1	PBW	-0.0513	-354	PPB	SMPL	-7470A		12 Aug 2025 14:37:39
U	PB169202BS - 1	LCSW	4.1394	17116	PPB	SMPL	-7470A		12 Aug 2025 14:42:43
U	Q2759-04 - 1	LIQUID-DRUM	21.1310	87949	PPB	SMPL	-7470A		12 Aug 2025 14:44:59
U	Q2776-01 - 1	72825	-0.0371	-295	PPB	SMPL	-7470A		12 Aug 2025 14:47:27
U	Q2814-01 - 1	TW-84SB-E	0.0497	67	PPB	SMPL	-7470A		12 Aug 2025 14:50:03
U	Q2814-03 - 1	TW-17M-W	0.1411	448	PPB	SMPL	-7470A		12 Aug 2025 14:52:25
U	Q2814-16 - 1	FB	0.0020	-132	PPB	SMPL	-7470A		12 Aug 2025 14:54:43
U	CCV30 - 1	CCV30	4.9843	20638	PPB	SMPL	-7470A		12 Aug 2025 14:56:58
U	CCB30 - 1	CCB30	-0.0326	-276	PPB	SMPL	-7470A		12 Aug 2025 14:59:14
U	Q2814-17 - 1	TW-518R-S	0.2783	1020	PPB	SMPL	-7470A		12 Aug 2025 15:01:32
U	Q2815-01 - 1	TW-705R-S	0.3579	1352	PPB	SMPL	-7470A		12 Aug 2025 15:03:47
U	Q2815-11 - 1	TW-22M-W	0.0674	141	PPB	SMPL	-7470A		12 Aug 2025 15:06:05
U	Q2815-20 - 1	TW-11M-W	0.1461	469	PPB	SMPL	-7470A		12 Aug 2025 15:08:23
U	Q2815-26 - 1	FB	0.0103	-97	PPB	SMPL	-7470A		12 Aug 2025 15:10:39
U	Q2821-05 - 1	GW-COMP	0.0351	6	PPB	SMPL	-7470A		12 Aug 2025 15:12:55
U	Q2821-08 - 1	DM-COMP	-0.0110	-186	PPB	SMPL	-7470A		12 Aug 2025 15:15:16
U	Q2821-08DUP - 1	DM-COMPDUP	0.0387	21	PPB	SMPL	-7470A		12 Aug 2025 15:17:34
U	Q2821-08MS - 1	DM-COMPMS	3.7016	15291	PPB	SMPL	-7470A		12 Aug 2025 15:19:51
U	Q2821-08MSD - 1	DM-COMPMSD	3.8177	15775	PPB	SMPL	-7470A		12 Aug 2025 15:22:06
U	CCV31 - 1	CCV31	4.9118	20336	PPB	SMPL	-7470A		12 Aug 2025 15:24:24
U	CCB31 - 1	CCB31	-0.0954	-538	PPB	SMPL	-7470A		12 Aug 2025 15:26:41
U	Q2821-08LX5 - 1		0.0806	196	PPB	SMPL	-7470A		12 Aug 2025 15:29:00
U	Q2821-08A - 1		4.1375	17108	PPB	SMPL	-7470A		12 Aug 2025 15:31:15
U	Q2759-04DLX5 - 1		2.8880	11899	PPB	SMPL	-7470A		12 Aug 2025 15:33:31
U	CCV32 - 1	CCV32	4.7818	19794	PPB	SMPL	-7470A		12 Aug 2025 15:35:50
U	CCB32 - 1	CCB32	-0.0132	-195	PPB	SMPL	-7470A		12 Aug 2025 15:38:08