


7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.007	-0.007	-920	0.000	-920	0				0
0.2	0.200	0.184	-0.016	-234	0.0 %	-234					0
2.5	2.500	2.507	0.007	8083	0.0 %	8083	0				0
5.0	5.000	4.982	-0.018	16942	0.0 %	16942					0
7.5	7.500	7.613	0.113	26361	0.0 %	26361					0
10.0	10.000	9.921	-0.079	34627	0.0 %	34627					0

LB136836 INSTRUMENT ID:CV1

Method: 7470A Operator: Admin Date of Analysis: 15 Aug 2025 10:34:38

Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50.0	-	-920	PPB	Std	0.0000	7470A	15 Aug 2025 11:25:54
S	0.2 - 1	50.2	-	-234	PPB	Std	0.2000	7470A	15 Aug 2025 11:28:11
S	2.5 - 1	52.5	-	8083	PPB	Std	2.5000	7470A	15 Aug 2025 11:30:28
S	5.0 - 1	55.0	-	16942	PPB	Std	5.0000	7470A	15 Aug 2025 11:32:44
S	7.5 - 1	57.5	-	26361	PPB	Std	7.5000	7470A	15 Aug 2025 11:35:03
S	10.0 - 1	510.0	-	34627	PPB	Std	10.0000	7470A	15 Aug 2025 11:37:20
U	ICV18 - 1	ICV18	4.0785	13708	PPB	SMPL	-7470A		15 Aug 2025 11:40:06
U	ICB18 - 1	ICB18	0.0028	-884	PPB	SMPL	-7470A		15 Aug 2025 11:45:30
U	CCV54 - 1	CCV54	5.2390	17863	PPB	SMPL	-7470A		15 Aug 2025 12:11:25
U	CCB54 - 1	CCB54	-0.0595	-1107	PPB	SMPL	-7470A		15 Aug 2025 12:13:45
U	CRA - 1	CRA	0.1877	-222	PPB	SMPL	-7470A		15 Aug 2025 12:16:02
U	HighStd - 1	HighStd	9.9914	34878	PPB	SMPL	-7470A		15 Aug 2025 12:18:20
U	ChkStd - 1	ChkStd	7.6262	26410	PPB	SMPL	-7470A		15 Aug 2025 12:20:35
U	PB169249BL - 1	PBW	-0.1458	-1416	PPB	SMPL	-7470A		15 Aug 2025 12:22:53
U	PB169249BS - 1	LCSW	4.1949	14125	PPB	SMPL	-7470A		15 Aug 2025 12:25:11
U	Q2814-22 - 1	TW-84SB-E	-0.0447	-1054	PPB	SMPL	-7470A		15 Aug 2025 12:27:28
U	Q2814-23 - 1	TW-17M-W	0.0132	-847	PPB	SMPL	-7470A		15 Aug 2025 12:29:47
U	Q2815-25 - 1	TW-11M-W	0.1841	-235	PPB	SMPL	-7470A		15 Aug 2025 12:32:04
U	Q2837-01 - 1	TW-WTS-13	0.0059	-873	PPB	SMPL	-7470A		15 Aug 2025 12:34:21
U	Q2840-03 - 1	0804-E	0.6824	1549	PPB	SMPL	-7470A		15 Aug 2025 12:36:38
U	CCV55 - 1	CCV55	4.8250	16381	PPB	SMPL	-7470A		15 Aug 2025 12:38:53
U	CCB55 - 1	CCB55	-0.0111	-934	PPB	SMPL	-7470A		15 Aug 2025 12:41:15
U	Q2842-03 - 1	MW-17B-55.5-081225	0.0226	-813	PPB	SMPL	-7470A		15 Aug 2025 12:43:33
U	Q2842-03DUP - 1	MW-17B-55.5-081225DUP	0.0145	-842	PPB	SMPL	-7470A		15 Aug 2025 12:45:48
U	Q2842-04 - 1	Q2842-03MS	3.7508	12535	PPB	SMPL	-7470A		15 Aug 2025 12:48:03
U	Q2842-05 - 1	Q2842-03MSD	3.8882	13027	PPB	SMPL	-7470A		15 Aug 2025 12:50:19
U	Q2849-01 - 1	BUR-25-0071	0.0821	-600	PPB	SMPL	-7470A		15 Aug 2025 12:52:36
U	Q2856-01 - 1	#452	0.5153	951	PPB	SMPL	-7470A		15 Aug 2025 12:54:54
U	Q2856-03 - 1	#456	-0.0533	-1085	PPB	SMPL	-7470A		15 Aug 2025 12:57:13
U	Q2857-01 - 1	4023	-0.0290	-998	PPB	SMPL	-7470A		15 Aug 2025 12:59:34
U	Q2857-02 - 1	3919	0.0037	-881	PPB	SMPL	-7470A		15 Aug 2025 13:01:53
U	Q2842-03LX5 - 1		0.0129	-848	PPB	SMPL	-7470A		15 Aug 2025 13:04:09
U	CCV56 - 1	CCV56	5.0932	17341	PPB	SMPL	-7470A		15 Aug 2025 13:09:07
U	CCB56 - 1	CCB56	-0.0477	-1065	PPB	SMPL	-7470A		15 Aug 2025 13:11:23
U	Q2842-03A - 1		3.9251	13159	PPB	SMPL	-7470A		15 Aug 2025 13:13:41
U	CCV57 - 1	CCV57	5.1775	17643	PPB	SMPL	-7470A		15 Aug 2025 13:15:59
U	CCB57 - 1	CCB57	-0.0651	-1127	PPB	SMPL	-7470A		15 Aug 2025 13:18:16