

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** First Environment, Inc.

**SDG No.:** Q2815

**Contract:** FIRS02

**Lab Code:** ACE

**Instrument ID:**

**Date:**

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |            |            |           |           |
|-----------|-------------------------|------------------------------------------|------------|------------|-----------|-----------|
|           |                         | Al                                       | Ca         | Fe         | Mg        | Ag        |
| Aluminum  | 396.100                 | 0.0000000                                | -0.0002060 | 0.0000000  | 0.0000000 | 0.0000000 |
| Antimony  | 206.833                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Arsenic   | 193.759                 | 0.0000000                                | 0.0000000  | -0.0000440 | 0.0000000 | 0.0000000 |
| Barium    | 493.409                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Beryllium | 234.861                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Cadmium   | 226.502                 | 0.0000000                                | 0.0000000  | 0.0000930  | 0.0000000 | 0.0000000 |
| Calcium   | 373.690                 | 0.0000000                                | 0.0000000  | -0.0075970 | 0.0000000 | 0.0000000 |
| Chromium  | 267.716                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Cobalt    | 228.616                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Copper    | 224.700                 | 0.0000000                                | 0.0000000  | 0.0007850  | 0.0000000 | 0.0000000 |
| Iron      | 240.488                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Lead      | 220.353                 | -0.0000920                               | 0.0000000  | 0.0000380  | 0.0000000 | 0.0000000 |
| Magnesium | 279.079                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Manganese | 257.610                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Nickel    | 231.604                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Potassium | 766.490                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Selenium  | 196.090                 | 0.0000000                                | 0.0000000  | -0.0001440 | 0.0000000 | 0.0000000 |
| Silver    | 328.068                 | 0.0000000                                | 0.0000000  | -0.0001490 | 0.0000000 | 0.0000000 |
| Sodium    | 589.592                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Thallium  | 190.856                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Vanadium  | 292.402                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Zinc      | 213.800                 | 0.0000000                                | 0.0000000  | 0.0001050  | 0.0000000 | 0.0000000 |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** First Environment, Inc.

**SDG No.:** Q2815

**Contract:** FIRS02

**Lab Code:** ACE

**Instrument ID:**

**Date:**

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |           |           |           |            |
|-----------|-------------------------|------------------------------------------|-----------|-----------|-----------|------------|
|           |                         | As                                       | Ba        | Be        | Cd        | Co         |
| Aluminum  | 396.100                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Antimony  | 206.833                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Arsenic   | 193.759                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Barium    | 493.409                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Beryllium | 234.861                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Cadmium   | 226.502                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0002870  |
| Calcium   | 373.690                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Chromium  | 267.716                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Cobalt    | 228.616                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Copper    | 224.700                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0009530  |
| Iron      | 240.488                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | -0.0039600 |
| Lead      | 220.353                 | 0.0000000                                | 0.0003170 | 0.0000000 | 0.0000000 | 0.0000000  |
| Magnesium | 279.079                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Manganese | 257.610                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Nickel    | 231.604                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Potassium | 766.490                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Selenium  | 196.090                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | -0.0003570 |
| Silver    | 328.068                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Sodium    | 589.592                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Thallium  | 190.856                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0054900  |
| Vanadium  | 292.402                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Zinc      | 213.800                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** First Environment, Inc.

**SDG No.:** Q2815

**Contract:** FIRS02

**Lab Code:** ACE

**Instrument ID:**

**Date:**

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |           |           |            |            |
|-----------|-------------------------|------------------------------------------|-----------|-----------|------------|------------|
|           |                         | Cr                                       | Cu        | K         | Mn         | Mo         |
| Aluminum  | 396.100                 | 0.0000000                                | 0.0000000 | 0.0000590 | 0.0000000  | 0.0396900  |
| Antimony  | 206.833                 | 0.0122000                                | 0.0000000 | 0.0000000 | 0.0000000  | 0.0000000  |
| Arsenic   | 193.759                 | -0.0029000                               | 0.0000000 | 0.0000000 | 0.0000000  | 0.0004900  |
| Barium    | 493.409                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000  | 0.0000000  |
| Beryllium | 234.861                 | 0.0000000                                | 0.0000000 | 0.0000000 | -0.0000710 | -0.0003400 |
| Cadmium   | 226.502                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000  | 0.0000000  |
| Calcium   | 373.690                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000  | 0.0000000  |
| Chromium  | 267.716                 | 0.0000000                                | 0.0000000 | 0.0000070 | 0.0002200  | 0.0000000  |
| Cobalt    | 228.616                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000  | -0.0007860 |
| Copper    | 224.700                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0006510  | 0.0020500  |
| Iron      | 240.488                 | 0.0000000                                | 0.0000000 | 0.0000730 | 0.0000000  | -0.0015250 |
| Lead      | 220.353                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0001400  | -0.0008600 |
| Magnesium | 279.079                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000  | 0.0000000  |
| Manganese | 257.610                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000  | 0.0000000  |
| Nickel    | 231.604                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000  | 0.0000000  |
| Potassium | 766.490                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000  | 0.0000000  |
| Selenium  | 196.090                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0007460  | 0.0000000  |
| Silver    | 328.068                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000  | -0.0000120 |
| Sodium    | 589.592                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000  | 0.0000000  |
| Thallium  | 190.856                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0017400  | -0.0100400 |
| Vanadium  | 292.402                 | -0.0025100                               | 0.0000000 | 0.0000000 | 0.0000000  | -0.0072000 |
| Zinc      | 213.800                 | 0.0000000                                | 0.0009010 | 0.0000000 | 0.0000000  | 0.0000000  |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** First Environment, Inc.

**SDG No.:** Q2815

**Contract:** FIRS02

**Lab Code:** ACE

**Instrument ID:**

**Date:**

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |            |           |           |           |
|-----------|-------------------------|------------------------------------------|------------|-----------|-----------|-----------|
|           |                         | Na                                       | Ni         | Pb        | Sb        | Se        |
| Aluminum  | 396.100                 | 0.0000000                                | 0.0000000  | 0.0012800 | 0.0000000 | 0.0000000 |
| Antimony  | 206.833                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Arsenic   | 193.759                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Barium    | 493.409                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Beryllium | 234.861                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Cadmium   | 226.502                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Calcium   | 373.690                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Chromium  | 267.716                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Cobalt    | 228.616                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Copper    | 224.700                 | 0.0000000                                | -0.0047000 | 0.0036100 | 0.0000000 | 0.0000000 |
| Iron      | 240.488                 | 0.0000000                                | -0.0017000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Lead      | 220.353                 | 0.0000000                                | 0.0006580  | 0.0000000 | 0.0000000 | 0.0001290 |
| Magnesium | 279.079                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Manganese | 257.610                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Nickel    | 231.604                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Potassium | 766.490                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Selenium  | 196.090                 | 0.0000000                                | 0.0000000  | 0.0003330 | 0.0000000 | 0.0000000 |
| Silver    | 328.068                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Sodium    | 589.592                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Thallium  | 190.856                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Vanadium  | 292.402                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Zinc      | 213.800                 | 0.0000000                                | 0.0067600  | 0.0000000 | 0.0000000 | 0.0000000 |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** First Environment, Inc.

**SDG No.:** Q2815

**Contract:** FIRS02

**Lab Code:** ACE

**Instrument ID:**

**Date:**

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |            |           |            |           |
|-----------|-------------------------|------------------------------------------|------------|-----------|------------|-----------|
|           |                         | Sn                                       | Ti         | Tl        | V          | Zn        |
| Aluminum  | 396.100                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Antimony  | 206.833                 | -0.0035600                               | -0.0007970 | 0.0000000 | -0.0018900 | 0.0000000 |
| Arsenic   | 193.759                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Barium    | 493.409                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Beryllium | 234.861                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Cadmium   | 226.502                 | 0.0000000                                | 0.0000630  | 0.0001280 | 0.0000000  | 0.0000000 |
| Calcium   | 373.690                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Chromium  | 267.716                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0001110  | 0.0000000 |
| Cobalt    | 228.616                 | 0.0000000                                | 0.0018800  | 0.0000000 | 0.0000000  | 0.0000000 |
| Copper    | 224.700                 | 0.0000000                                | 0.0003840  | 0.0000000 | 0.0000000  | 0.0000000 |
| Iron      | 240.488                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Lead      | 220.353                 | 0.0000000                                | -0.0003610 | 0.0000000 | 0.0000000  | 0.0000000 |
| Magnesium | 279.079                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Manganese | 257.610                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Nickel    | 231.604                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Potassium | 766.490                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Selenium  | 196.090                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Silver    | 328.068                 | 0.0000000                                | -0.0007420 | 0.0000000 | 0.0000000  | 0.0000000 |
| Sodium    | 589.592                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |
| Thallium  | 190.856                 | 0.0000000                                | -0.0039700 | 0.0000000 | -0.0115600 | 0.0000000 |
| Vanadium  | 292.402                 | 0.0000000                                | 0.0005320  | 0.0000000 | 0.0000000  | 0.0000000 |
| Zinc      | 213.800                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000  | 0.0000000 |