

CALIBRATION VERIFICATION SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2815</u>
Continuing Calib Date:	<u>08/13/2025</u>	Initial Calibration Date(s):	<u>08/01/2025</u> <u>08/01/2025</u>
Continuing Calib Time:	<u>09:51</u>	Initial Calibration Time(s):	<u>12:05</u> <u>20:28</u>

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.66	6.66	6.56	6.76	0.00
Aroclor-1254-2	(2)	6.88	6.87	6.77	6.97	-0.01
Aroclor-1254-3	(3)	7.24	7.24	7.14	7.34	0.00
Aroclor-1254-4	(4)	7.52	7.52	7.42	7.62	0.00
Aroclor-1254-5	(5)	7.94	7.93	7.83	8.03	-0.01
Tetrachloro-m-xylene		4.66	4.66	4.56	4.76	0.00
Decachlorobiphenyl		10.44	10.43	10.33	10.53	-0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Continuing Calib Date: 08/13/2025

Initial Calibration Date(s): 08/01/2025

08/01/2025

Continuing Calib Time: 09:51

Initial Calibration Time(s): 12:05

20:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.69	5.69	5.59	5.79	0.01
Aroclor-1254-2 (2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1254-3 (3)	6.24	6.24	6.14	6.34	0.00
Aroclor-1254-4 (4)	6.46	6.46	6.36	6.56	0.00
Aroclor-1254-5 (5)	6.88	6.88	6.78	6.98	0.00
Tetrachloro-m-xylene	3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl	8.82	8.82	8.72	8.92	0.00