

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112924.D **Time Analyzed:** 20:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.750	4.655	4.855	521.250	500.000	4.3
Aroclor-1016-2	4.769	4.674	4.874	530.910	500.000	6.2
Aroclor-1016-3	4.826	4.731	4.931	519.890	500.000	4.0
Aroclor-1016-4	4.945	4.851	5.051	523.180	500.000	4.6
Aroclor-1016-5	5.202	5.108	5.308	582.330	500.000	16.5
Aroclor-1260-1	6.239	6.145	6.345	520.880	500.000	4.2
Aroclor-1260-2	6.429	6.333	6.533	520.880	500.000	4.2
Aroclor-1260-3	6.794	6.700	6.900	541.610	500.000	8.3
Aroclor-1260-4	7.053	6.959	7.159	508.450	500.000	1.7
Aroclor-1260-5	7.297	7.203	7.403	500.470	500.000	0.1
Decachlorobiphenyl	8.686	8.593	8.793	48.470	50.000	-3.1
Tetrachloro-m-xylene	3.666	3.569	3.769	54.370	50.000	8.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112924.D **Time Analyzed:** 20:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.733	4.638	4.838	494.680	500.000	-1.1
Aroclor-1016-2	4.751	4.656	4.856	495.150	500.000	-1.0
Aroclor-1016-3	4.926	4.831	5.031	490.140	500.000	-2.0
Aroclor-1016-4	4.969	4.873	5.073	489.100	500.000	-2.2
Aroclor-1016-5	5.180	5.085	5.285	506.470	500.000	1.3
Aroclor-1260-1	6.208	6.113	6.313	494.280	500.000	-1.1
Aroclor-1260-2	6.397	6.302	6.502	505.530	500.000	1.1
Aroclor-1260-3	6.547	6.453	6.653	515.570	500.000	3.1
Aroclor-1260-4	7.017	6.923	7.123	511.950	500.000	2.4
Aroclor-1260-5	7.259	7.164	7.364	525.140	500.000	5.0
Decachlorobiphenyl	8.633	8.540	8.740	52.140	50.000	4.3
Tetrachloro-m-xylene	3.659	3.562	3.762	52.540	50.000	5.1