

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074373.D **Time Analyzed:** 08:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.808	5.712	5.912	533.120	500.000	6.6
Aroclor-1016-2	5.830	5.733	5.933	526.230	500.000	5.2
Aroclor-1016-3	5.892	5.796	5.996	534.530	500.000	6.9
Aroclor-1016-4	5.989	5.893	6.093	541.630	500.000	8.3
Aroclor-1016-5	6.282	6.186	6.386	523.040	500.000	4.6
Aroclor-1260-1	7.400	7.303	7.503	489.650	500.000	-2.1
Aroclor-1260-2	7.653	7.555	7.755	475.290	500.000	-4.9
Aroclor-1260-3	8.011	7.913	8.113	480.170	500.000	-4.0
Aroclor-1260-4	8.239	8.142	8.342	485.690	500.000	-2.9
Aroclor-1260-5	8.565	8.468	8.668	464.970	500.000	-7.0
Decachlorobiphenyl	10.437	10.338	10.538	47.490	50.000	-5.0
Tetrachloro-m-xylene	4.655	4.560	4.760	56.520	50.000	13.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2815
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074373.D **Time Analyzed:** 08:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.900	4.806	5.006	554.710	500.000	10.9
Aroclor-1016-2	4.957	4.863	5.063	558.540	500.000	11.7
Aroclor-1016-3	5.078	4.983	5.183	559.480	500.000	11.9
Aroclor-1016-4	5.117	5.023	5.223	569.010	500.000	13.8
Aroclor-1016-5	5.332	5.239	5.439	594.260	500.000	18.9
Aroclor-1260-1	6.549	6.455	6.655	544.340	500.000	8.9
Aroclor-1260-2	6.704	6.609	6.809	527.160	500.000	5.4
Aroclor-1260-3	6.913	6.819	7.019	553.670	500.000	10.7
Aroclor-1260-4	7.174	7.079	7.279	532.690	500.000	6.5
Aroclor-1260-5	7.412	7.318	7.518	545.830	500.000	9.2
Decachlorobiphenyl	8.817	8.726	8.926	51.890	50.000	3.8
Tetrachloro-m-xylene	3.799	3.705	3.905	58.630	50.000	17.3