

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL072825

Review By	Abdul	Review On	7/29/2025 9:50:24 AM
Supervise By	mohammad	Supervise On	7/30/2025 2:11:50 AM
SubDirectory	PL072825	HP Acquire Method	HP Processing Method p1072825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL096592.D	28 Jul 2025 15:57		AR\AJ	Ok
2	I.BLK	I.BLK	PL096593.D	28 Jul 2025 16:11		AR\AJ	Ok
3	PEM	PEM	PL096594.D	28 Jul 2025 16:25		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL096595.D	28 Jul 2025 16:38		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PL096596.D	28 Jul 2025 16:52		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL096597.D	28 Jul 2025 17:06		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL096598.D	28 Jul 2025 17:19	Comp#18 kept on L.R. for 2nd column	AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL096599.D	28 Jul 2025 17:33		AR\AJ	Ok,M
9	PSTDICC005	PSTDICC005	PL096600.D	28 Jul 2025 17:47		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL096601.D	28 Jul 2025 18:00		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL096602.D	28 Jul 2025 18:14		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL096603.D	28 Jul 2025 18:27		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL096604.D	28 Jul 2025 18:41		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL096605.D	28 Jul 2025 18:55		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL096606.D	28 Jul 2025 19:08		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL096607.D	28 Jul 2025 19:22		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL096608.D	28 Jul 2025 19:36		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL072825

Review By	Abdul	Review On	7/29/2025 9:50:24 AM
Supervise By	mohammad	Supervise On	7/30/2025 2:11:50 AM
SubDirectory	PL072825	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	PTOXICC250	PTOXICC250	PL096609.D	28 Jul 2025 19:49		AR\AJ	Ok
19	PTOXICC100	PTOXICC100	PL096610.D	28 Jul 2025 20:03		AR\AJ	Ok
20	PSTDICV050	ICVPL072825	PL096611.D	28 Jul 2025 20:17		AR\AJ	Ok,M
21	PCHLORICV500	ICVPL072825CHLOR	PL096612.D	28 Jul 2025 20:30		AR\AJ	Ok
22	PTOXICV500	ICVPL072825TOX	PL096613.D	28 Jul 2025 20:44		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081625

Review By	Abdul	Review On	8/18/2025 8:13:57 AM
Supervise By	mohammad	Supervise On	8/19/2025 1:27:28 AM
SubDirectory	PL081625	HP Acquire Method	HP Processing Method p1072825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL096816.D	15 Aug 2025 08:58		AR\AJ	Ok
2	I.BLK	I.BLK	PL096817.D	15 Aug 2025 09:11		AR\AJ	Ok,M
3	PEM	PEM	PL096818.D	15 Aug 2025 10:13		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL096819.D	15 Aug 2025 10:58		AR\AJ	Ok,M
5	PCHLORCCC500	PCHLORCCC500	PL096820.D	15 Aug 2025 12:28		AR\AJ	Ok,M
6	PTOXCCC500	PTOXCCC500	PL096821.D	15 Aug 2025 15:53		AR\AJ	Ok,M
7	PB169225BL	PB169225BL	PL096822.D	15 Aug 2025 16:07		AR\AJ	Ok,M
8	PB169225BS	PB169225BS	PL096823.D	15 Aug 2025 16:21		AR\AJ	Ok,M
9	PB169225BSD	PB169225BSD	PL096824.D	15 Aug 2025 16:34		AR\AJ	Ok,M
10	PB169225BS	PB169225BS	PL096825.D	15 Aug 2025 16:48	CHLOR BS	AR\AJ	Ok,M
11	PB169225BS	PB169225BS	PL096826.D	15 Aug 2025 17:36	TOX BS	AR\AJ	Ok,M
12	I.BLK	I.BLK	PL096827.D	15 Aug 2025 17:51		AR\AJ	Ok,M
13	PSTDCCC050	PSTDCCC050	PL096828.D	15 Aug 2025 18:05		AR\AJ	Ok,M
14	PCHLORCCC500	PCHLORCCC500	PL096829.D	15 Aug 2025 18:32		AR\AJ	Ok,M
15	PTOXCCC500	PTOXCCC500	PL096830.D	15 Aug 2025 18:45		AR\AJ	Ok,M
16	Q2814-01	TW-84SB-E	PL096831.D	15 Aug 2025 18:59		AR\AJ	Ok,M
17	Q2814-03	TW-17M-W	PL096832.D	15 Aug 2025 19:13		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081625

Review By	Abdul	Review On	8/18/2025 8:13:57 AM
Supervise By	mohammad	Supervise On	8/19/2025 1:27:28 AM
SubDirectory	PL081625	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q2814-16	FB	PL096833.D	15 Aug 2025 19:26		AR\AJ	Ok,M
19	Q2814-17	TW-518R-S	PL096834.D	15 Aug 2025 19:40		AR\AJ	Ok,M
20	I.BLK	I.BLK	PL096835.D	15 Aug 2025 19:54		AR\AJ	Ok,M
21	PEM	PEM	PL096836.D	15 Aug 2025 20:07		AR\AJ	Ok,M
22	PSTDCCC050	PSTDCCC050	PL096837.D	15 Aug 2025 20:48		AR\AJ	Ok,M
23	Q2815-01	TW-705R-S	PL096838.D	15 Aug 2025 21:02	All surrogate fail , need dilution	AR\AJ	Not Ok
24	Q2815-11	TW-22M-W	PL096839.D	15 Aug 2025 21:15		AR\AJ	Ok,M
25	I.BLK	I.BLK	PL096840.D	15 Aug 2025 21:56		AR\AJ	Ok,M
26	PSTDCCC050	PSTDCCC050	PL096841.D	15 Aug 2025 22:10		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081825

Review By	Abdul	Review On	8/19/2025 8:40:55 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:05:16 AM
SubDirectory	PL081825	HP Acquire Method	HP Processing Method p1072825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL096842.D	18 Aug 2025 09:58		AR\AJ	Ok
2	I.BLK	I.BLK	PL096843.D	18 Aug 2025 10:12		AR\AJ	Ok,M
3	PEM	PEM	PL096844.D	18 Aug 2025 10:26		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL096845.D	18 Aug 2025 11:25		AR\AJ	Ok,M
5	PB169283BL	PB169283BL	PL096846.D	18 Aug 2025 11:56		AR\AJ	Ok
6	PB169283BS	PB169283BS	PL096847.D	18 Aug 2025 12:10	Recovery Fail delta-BHC-I	AR\AJ	Not Ok
7	Q2883-01	RBR2520553	PL096848.D	18 Aug 2025 12:24		AR\AJ	Ok,M
8	I.BLK	I.BLK	PL096849.D	18 Aug 2025 12:39		AR\AJ	Ok
9	PSTDCCC050	PSTDCCC050	PL096850.D	18 Aug 2025 12:53		AR\AJ	Ok,M
10	PCHLORCCC500	PCHLORCCC500	PL096851.D	18 Aug 2025 14:10		AR\AJ	Ok,M
11	PTOXCCC500	PTOXCCC500	PL096852.D	18 Aug 2025 14:48	CCC FAIL	AR\AJ	Not Ok
12	Q2864-03	LAW-25-113-121	PL096853.D	18 Aug 2025 15:58		AR\AJ	Ok
13	PB169264BL	PB169264BL	PL096854.D	18 Aug 2025 16:12		AR\AJ	Ok,M
14	Q2815-01	TW-705R-S	PL096855.D	18 Aug 2025 16:30	need dilution	AR\AJ	Dilution
15	Q2815-01DL	TW-705R-SDL	PL096856.D	18 Aug 2025 16:44		AR\AJ	Ok,M
16	I.BLK	I.BLK	PL096857.D	18 Aug 2025 16:57		AR\AJ	Ok,M
17	PSTDCCC050	PSTDCCC050	PL096858.D	18 Aug 2025 17:11		AR\AJ	Ok,M
18	PCHLORCCC500	PCHLORCCC500	PL096859.D	18 Aug 2025 17:25		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081825

Review By	Abdul	Review On	8/19/2025 8:40:55 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:05:16 AM
SubDirectory	PL081825	HP Acquire Method	HP Processing Method pl072825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2879-01	OU4-TS-GRILLO-TSCF	PL096860.D	18 Aug 2025 17:38		AR\AJ	Ok,M
20	Q2879-03	OU4-TS-GRILLO-TSCF	PL096861.D	18 Aug 2025 17:52		AR\AJ	Ok,M
21	Q2879-05	OU4-TS-GRILLO-TSCF	PL096862.D	18 Aug 2025 18:06		AR\AJ	Ok,M
22	Q2879-07	OU4-TS-GRILLO-TSCF	PL096863.D	18 Aug 2025 18:19	DCB low in 2nd column	AR\AJ	Ok,M
23	Q2879-09	OU4-TS-GRILLO-TSCF	PL096864.D	18 Aug 2025 18:33	DCB low in 2nd column	AR\AJ	Ok,M
24	Q2879-11	OU4-TS-GRILLO-TSCF	PL096865.D	18 Aug 2025 18:46		AR\AJ	Ok,M
25	I.BLK	I.BLK	PL096866.D	18 Aug 2025 19:00		AR\AJ	Ok,M
26	PEM	PEM	PL096867.D	18 Aug 2025 19:14		AR\AJ	Ok,M
27	PSTDCCC050	PSTDCCC050	PL096868.D	18 Aug 2025 19:41		AR\AJ	Ok,M
28	Q2879-13	OU4-TS-GRILLO-TSCF	PL096869.D	18 Aug 2025 20:22	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok
29	Q2879-15	OU4-TS-GRILLO-TSCF	PL096870.D	18 Aug 2025 20:35	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok
30	Q2879-17	OU4-TS-GRILLO-TSCF	PL096871.D	18 Aug 2025 20:49	Hit for Comp#17,END CCC FAIL FOR COM#17,DCB Low in both column	AR\AJ	Not Ok
31	Q2884-01	VNJ-232	PL096872.D	18 Aug 2025 21:02	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok
32	Q2884-03	302	PL096873.D	18 Aug 2025 21:16	END CCC FAIL FOR COM#17	AR\AJ	Not Ok
33	Q2888-01	VNJ-210	PL096874.D	18 Aug 2025 21:30	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok
34	Q2888-01MS	VNJ-210MS	PL096875.D	18 Aug 2025 21:43	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok
35	Q2888-01MSD	VNJ-210MSD	PL096876.D	18 Aug 2025 21:57	Hit for Comp#17,END CCC FAIL FOR COM#17	AR\AJ	Not Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL081825

Review By	Abdul	Review On	8/19/2025 8:40:55 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:05:16 AM
SubDirectory	PL081825	HP Acquire Method	HP Processing Method pl072825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

36	Q2889-01	PL-02-081525	PL096877.D	18 Aug 2025 22:10	END CCC FAIL FOR COM#17	AR\AJ	Not Ok
37	I.BLK	I.BLK	PL096878.D	18 Aug 2025 22:38		AR\AJ	Ok,M
38	PSTDCCC050	PSTDCCC050	PL096879.D	18 Aug 2025 22:51	Comp#17,20 low in 2nd column	AR\AJ	Ok,M

M : Manual Integration