

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2815

Analytical Method: 8081B

Client: First Environment, Inc.

Datafile : PL096823.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169225BS (Column 1)	alpha-BHC	0.5	0.50	ug/L	100				54	138	
	beta-BHC	0.5	0.50	ug/L	100				56	136	
	delta-BHC	0.5	0.50	ug/L	101				52	142	
	gamma-BHC (Lindane)	0.5	0.50	ug/L	99				59	134	
	Heptachlor	0.5	0.53	ug/L	106				54	130	
	Aldrin	0.5	0.49	ug/L	99				45	134	
	Heptachlor epoxide	0.5	0.51	ug/L	101				61	133	
	Endosulfan I	0.5	0.49	ug/L	98				62	126	
	Dieldrin	0.5	0.49	ug/L	99				60	136	
	4,4'-DDE	0.5	0.49	ug/L	98				57	135	
	Endrin	0.5	0.51	ug/L	102				60	138	
	Endosulfan II	0.5	0.51	ug/L	101				52	135	
	4,4'-DDD	0.5	0.50	ug/L	101				56	143	
	Endosulfan sulfate	0.5	0.48	ug/L	96				62	133	
	4,4'-DDT	0.5	0.50	ug/L	99				51	143	
	Methoxychlor	0.5	0.49	ug/L	97				54	145	
	Endrin ketone	0.5	0.49	ug/L	98				58	134	
	Endrin aldehyde	0.5	0.53	ug/L	106				51	132	
	alpha-Chlordane	0.5	0.50	ug/L	100				60	129	
	gamma-Chlordane	0.5	0.50	ug/L	101				56	136	
	alpha-BHC	0.5	0.50	ug/L	100				54	138	
PB169225BS (Column 2)	beta-BHC	0.5	0.50	ug/L	99				56	136	
	delta-BHC	0.5	0.50	ug/L	100				52	142	
	gamma-BHC (Lindane)	0.5	0.50	ug/L	100				59	134	
	Heptachlor	0.5	0.49	ug/L	98				54	130	
	Aldrin	0.5	0.50	ug/L	99				45	134	
	Heptachlor epoxide	0.5	0.49	ug/L	98				61	133	
	Endosulfan I	0.5	0.48	ug/L	95				62	126	
	Dieldrin	0.5	0.49	ug/L	98				60	136	
	4,4'-DDE	0.5	0.49	ug/L	99				57	135	
	Endrin	0.5	0.48	ug/L	97				60	138	
	Endosulfan II	0.5	0.49	ug/L	97				52	135	
	4,4'-DDD	0.5	0.49	ug/L	99				56	143	
	Endosulfan sulfate	0.5	0.48	ug/L	97				62	133	
	4,4'-DDT	0.5	0.49	ug/L	97				51	143	
	Methoxychlor	0.5	0.48	ug/L	97				54	145	
	Endrin ketone	0.5	0.52	ug/L	104				58	134	
	Endrin aldehyde	0.5	0.55	ug/L	110				51	132	
	alpha-Chlordane	0.5	0.49	ug/L	97				60	129	
	gamma-Chlordane	0.5	0.50	ug/L	99				56	136	