

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2815</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>08/20/2025 18:27</u>
Lab File ID:	<u>BP025512.D</u>	Init. Calib. Date(s):	<u>08/14/2025 08/14/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>12:33 17:24</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.204	1.232		2.3	
Benzaldehyde	1.082	1.279		18.2	
Phenol-d6	1.544	1.491		-3.4	
Phenol	1.620	1.535		-5.2	20.0
bis(2-Chloroethyl)ether	1.263	1.201		-4.9	
2-Chlorophenol	1.359	1.324		-2.6	
2-Methylphenol	1.125	1.059		-5.9	
2,2-oxybis(1-Chloropropane)	1.692	1.647		-2.7	
Acetophenone	0.502	0.508		1.2	
3+4-Methylphenols	1.560	1.396		-10.5	
n-Nitroso-di-n-propylamine	1.023	0.895	0.050	-12.5	
Nitrobenzene-d5	0.395	0.409		3.5	
Hexachloroethane	0.563	0.578		2.7	
Nitrobenzene	0.353	0.367		4.0	
Isophorone	0.700	0.664		-5.1	
2-Nitrophenol	0.181	0.188		3.9	20.0
2,4-Dimethylphenol	0.312	0.308		-1.3	
bis(2-Chloroethoxy)methane	0.423	0.413		-2.4	
2,4-Dichlorophenol	0.310	0.310		0.0	20.0
Naphthalene	1.040	1.033		-0.7	
4-Chloroaniline	0.459	0.429		-6.5	
Hexachlorobutadiene	0.211	0.225		6.6	20.0
Caprolactam	0.114	0.092		-19.3	
4-Chloro-3-methylphenol	0.355	0.324		-8.7	20.0
2-Methylnaphthalene	0.676	0.633		-6.4	
Hexachlorocyclopentadiene	0.349	0.359	0.050	2.9	
2,4,6-Trichlorophenol	0.390	0.414		6.2	20.0
2-Fluorobiphenyl	1.463	1.526		4.3	
2,4,5-Trichlorophenol	0.427	0.446		4.4	
1,1-Biphenyl	1.453	1.507		3.7	
2-Chloronaphthalene	1.131	1.165		3.0	
2-Nitroaniline	0.329	0.343		4.3	
Dimethylphthalate	1.465	1.399		-4.5	
Acenaphthylene	1.871	1.871		0.0	
2,6-Dinitrotoluene	0.309	0.313		1.3	
3-Nitroaniline	0.347	0.329		-5.2	
Acenaphthene	1.098	1.086		-1.1	20.0
2,4-Dinitrophenol	0.165	0.172	0.050	4.2	
4-Nitrophenol	0.285	0.266	0.050	-6.7	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2815</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>08/20/2025 18:27</u>
Lab File ID:	<u>BP025512.D</u>	Init. Calib. Date(s):	<u>08/14/2025 08/14/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>12:33 17:24</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.736	1.703		-1.9	
2,4-Dinitrotoluene	0.440	0.429		-2.5	
Diethylphthalate	1.488	1.402		-5.8	
4-Chlorophenyl-phenylether	0.681	0.651		-4.4	
Fluorene	1.370	1.306		-4.7	
4-Nitroaniline	0.356	0.332		-6.7	
4,6-Dinitro-2-methylphenol	0.125	0.127		1.6	
n-Nitrosodiphenylamine	0.610	0.609		-0.2	20.0
2,4,6-Tribromophenol	0.269	0.257		-4.5	
4-Bromophenyl-phenylether	0.220	0.224		1.8	
Hexachlorobenzene	0.255	0.256		0.4	
Atrazine	0.228	0.221		-3.1	
Pentachlorophenol	0.161	0.167		3.7	20.0
Phenanthrene	1.099	1.085		-1.3	
Anthracene	1.122	1.084		-3.4	
Carbazole	1.057	1.022		-3.3	
Di-n-butylphthalate	1.300	1.314		1.1	
Fluoranthene	1.311	1.267		-3.4	20.0
Pyrene	1.300	1.220		-6.2	
Terphenyl-d14	1.093	1.021		-6.6	
Butylbenzylphthalate	0.589	0.600		1.9	
3,3-Dichlorobenzidine	0.497	0.503		1.2	
Benzo (a) anthracene	1.319	1.279		-3.0	
Chrysene	1.235	1.213		-1.8	
Bis(2-ethylhexyl)phthalate	0.867	0.895		3.2	
Di-n-octyl phthalate	1.492	1.648		10.5	20.0
Benzo (b) fluoranthene	1.179	1.121		-4.9	
Benzo (k) fluoranthene	1.180	1.119		-5.2	
Benzo (a) pyrene	1.148	1.115		-2.9	20.0
Indeno (1,2,3-cd) pyrene	1.467	1.484		1.2	
Dibenzo (a,h) anthracene	1.199	1.220		1.8	
Benzo (g,h,i) perylene	1.178	1.213		3.0	
1,2,4,5-Tetrachlorobenzene	0.578	0.626		8.3	
1,4-Dioxane	0.472	0.489		3.6	20.0
2,3,4,6-Tetrachlorophenol	0.377	0.372		-1.3	

All other compounds must meet a minimum RRF of 0.010.