

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081425\
 Data File : BF143405.D
 Acq On : 15 Aug 2025 06:51
 Operator : RC/JU
 Sample : Q2815-03
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TW-10P-E

Quant Time: Aug 15 07:48:34 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 12 13:25:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.934	152	142072	20.000	ng	0.00
21) Naphthalene-d8	8.210	136	539316	20.000	ng	0.00
39) Acenaphthene-d10	9.969	164	291303	20.000	ng	0.00
64) Phenanthrene-d10	11.457	188	417383	20.000	ng	0.00
76) Chrysene-d12	14.098	240	249094	20.000	ng	0.00
86) Perylene-d12	15.592	264	314728	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.569	112	780126	94.955	ng	0.00
7) Phenol-d6	6.563	99	830604	78.837	ng	0.00
23) Nitrobenzene-d5	7.492	82	807058	93.189	ng	0.00
42) 2,4,6-Tribromophenol	10.757	330	318911	149.800	ng	0.00
45) 2-Fluorobiphenyl	9.286	172	1424272	82.911	ng	0.00
79) Terphenyl-d14	13.039	244	1004920	72.026	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081425\
Data File : BF143405.D
Acq On : 15 Aug 2025 06:51
Operator : RC/JU
Sample : Q2815-03
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TW-10P-E

Quant Time: Aug 15 07:48:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 12 13:25:39 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.934 min Scan# 806
Delta R.T. -0.000 min
Lab File: BF143405.D
Acq: 15 Aug 2025 06:51




Instrument :
BNA_F
ClientSampleId :
TW-10P-E


Tgt Ion:152 Resp: 142072
Ion Ratio Lower Upper
152 100
150 154.5 129.8 194.8
115 57.0 47.8 71.8



#5
2-Fluorophenol
Concen: 94.955 ng
RT: 5.569 min Scan# 574
Delta R.T. 0.006 min
Lab File: BF143405.D
Acq: 15 Aug 2025 06:51

Tgt Ion:112 Resp: 780126
Ion Ratio Lower Upper
112 100
64 65.5 52.1 78.1
63 33.8 26.6 39.8









#7
Phenol-d6
Concen: 78.837 ng
RT: 6.563 min Scan# 743
Delta R.T. -0.000 min
Lab File: BF143405.D
Acq: 15 Aug 2025 06:51




Instrument :
BNA_F
ClientSampleId :
TW-10P-E

Tgt Ion: 99 Resp: 830604
Ion Ratio Lower Upper
99 100
42 20.4 16.6 25.0
71 33.6 26.2 39.2



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.210 min Scan# 1023
Delta R.T. -0.000 min
Lab File: BF143405.D
Acq: 15 Aug 2025 06:51

Tgt Ion: 136 Resp: 539316
Ion Ratio Lower Upper
136 100
137 10.8 8.6 13.0
54 10.2 8.4 12.6
68 6.8 5.9 8.9





#23

Nitrobenzene-d5

Concen: 93.189 ng

RT: 7.492 min Scan# 901

Delta R.T. -0.000 min

Lab File: BF143405.D

Acq: 15 Aug 2025 06:51

Instrument :

BNA_F

ClientSampleId :

TW-10P-E





Tgt Ion: 82 Resp: 807058

Ion Ratio Lower Upper

82 100

128 44.7 33.5 50.3

54 56.6 46.7 70.1



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.969 min Scan# 1322

Delta R.T. -0.000 min

Lab File: BF143405.D

Acq: 15 Aug 2025 06:51




Tgt Ion: 164 Resp: 291303


Ion Ratio Lower Upper

164 100

162 98.6 78.6 118.0

160 42.7 34.4 51.6





#42

2,4,6-Tribromophenol

Concen: 149.800 ng

RT: 10.757 min Scan# 1456

Delta R.T. -0.000 min

Lab File: BF143405.D

Acq: 15 Aug 2025 06:51

Instrument :

BNA_F

ClientSampleId :

TW-10P-E





Tgt Ion:330 Resp: 318911

Ion Ratio Lower Upper

330 100

332 98.3 78.1 117.1

141 29.3 25.4 38.0



#45

2-Fluorobiphenyl

Concen: 82.911 ng

RT: 9.286 min Scan# 1206

Delta R.T. -0.000 min

Lab File: BF143405.D

Acq: 15 Aug 2025 06:51




Tgt Ion:172 Resp: 1424272


Ion Ratio Lower Upper

172 100

171 35.6 28.4 42.6

170 23.6 19.1 28.7









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.457 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143405.D
Acq: 15 Aug 2025 06:51




Instrument :
BNA_F
ClientSampleId :
TW-10P-E


Tgt Ion:188 Resp: 417383
Ion Ratio Lower Upper
188 100
94 10.1 8.8 13.2
80 10.8 9.7 14.5



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.098 min Scan# 2024
Delta R.T. 0.006 min
Lab File: BF143405.D
Acq: 15 Aug 2025 06:51

Tgt Ion:240 Resp: 249094
Ion Ratio Lower Upper
240 100
120 10.9 9.0 13.6
236 24.8 19.9 29.9









#79
 Terphenyl-d14
 Concen: 72.026 ng
 RT: 13.039 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: BF143405.D
 Acq: 15 Aug 2025 06:51

Instrument :
 BNA_F
 ClientSampleId :
 TW-10P-E

Tgt Ion:244 Resp: 1004920
 Ion Ratio Lower Upper
 244 100
 212 6.5 5.1 7.7
 122 10.3 8.9 13.3



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.592 min Scan# 2278
 Delta R.T. 0.012 min
 Lab File: BF143405.D
 Acq: 15 Aug 2025 06:51

Tgt Ion:264 Resp: 314728
 Ion Ratio Lower Upper
 264 100
 260 23.4 18.7 28.1
 265 21.5 17.2 25.8

