

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2815

Client ID : SSTDCCC040

Date Analyzed: 08/20/2025

Lab File ID: BF143498.D

Time Analyzed: 22:05

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	138596	6.928	522619	8.21	286971	9.97
UPPER LIMIT	277192	7.428	1045240	8.71	573942	10.469
LOWER LIMIT	69298	6.428	261310	7.71	143486	9.469
EPA SAMPLE NO.						
01 TW-88H-N	131555	6.93	502957	8.20	275165	9.96
02 TW-22M-W	127726	6.93	485532	8.20	258859	9.96
03 TW-22M-S	131776	6.93	484366	8.21	235399	9.97

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.