

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2815

Client ID : SSTDCCC040

Date Analyzed: 08/21/2025

Lab File ID: BF143518.D

Time Analyzed: 09:21

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	118001	6.928	448774	8.21	244595	9.96
UPPER LIMIT	236002	7.428	897548	8.71	489190	10.463
LOWER LIMIT	59000.5	6.428	224387	7.71	122298	9.463
EPA SAMPLE NO.						
01 TW-17M-E	123661	6.93	463978	8.20	243046	9.96

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.