

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143497.D
Instrument ID: BNA_F

Contract: FIRS02
SDG NO.: Q2815
DFTPP Injection Date: 08/20/2025
DFTPP Injection Time: 21:36

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (2) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	83.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF143498.D	08/20/2025	22:05
TW-22M-S	Q2815-12	BF143502.D	08/21/2025	00:00
TW-22M-W	Q2815-11	BF143514.D	08/21/2025	05:47
TW-88H-N	Q2815-08	BF143515.D	08/21/2025	06:15
SSTDCCC040EC	SSTDCCC040	BF143516.D	08/21/2025	06:44