

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025486.D
Instrument ID: BNA_P

Contract: FIRS02
SDG NO.: Q2815
DFTPP Injection Date: 08/19/2025
DFTPP Injection Time: 23:05

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	78.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.2 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP025487.D	08/19/2025	23:46
TW-84SB-S	Q2815-17	BP025498.D	08/20/2025	07:16
TW-17M-S	Q2815-16	BP025499.D	08/20/2025	07:57
SSTDCCC040EC	SSTDCCC040	BP025502.D	08/20/2025	10:41