

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: FIRS02
 Lab Code: ACE SDG NO.: Q2815
 Lab File ID: VX047309.D Date Analyzed: 08/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:44
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	202458	5.56	343367	6.76	311895	10.05
UPPER LIMIT	404916	6.056	686734	7.263	623790	10.549
LOWER LIMIT	101229	5.056	171684	6.263	155948	9.549
EPA SAMPLE NO.						
TW-10P-N	208024	5.57	379571	6.78	354515	10.06
TW-88H-E	226623	5.57	419826	6.78	390050	10.06
VX0813WBL01	202155	5.56	360108	6.76	331541	10.06
VX0813WBS01	201153	5.57	351465	6.77	321390	10.06

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.