

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: FIRS02
 Lab Code: ACE SDG NO.: Q2815
 Lab File ID: VX047309.D Date Analyzed: 08/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:44
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	156059	12.018				
UPPER LIMIT	312118	12.518				
LOWER LIMIT	78029.5	11.518				
EPA SAMPLE NO.						
TW-10P-N	179591	12.02				
TW-88H-E	198183	12.02				
VX0813WBL01	172532	12.02				
VX0813WBS01	166290	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.