

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2815

Lab File ID: VX047371.D

Date Analyzed: 08/18/2025

Instrument ID: MSVOA X

Time Analyzed: 10:00

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	190181	12.018				
UPPER LIMIT	380362	12.518				
LOWER LIMIT	95090.5	11.518				
EPA SAMPLE NO.						
TW-705R-S	295876	12.02				
TW-10PC-W	292088	12.02				
TW-10P-E	294795	12.02				
TW-10P-S	297143	12.02				
TW-10P-W	297817	12.02				
TW-10P-NRE	236692	12.02				
TW-88H-N	209517	12.02				
TW-88H-W	285001	12.02				
TW-88H-S	291956	12.02				
TW-22M-E	264053	12.02				
TW-22M-N	227147	12.02				
TW-17M-E	241466	12.02				
TW-17M-S	253687	12.02				
TW-84SB-S	220943	12.02				
VX0818WBL01	309313	12.02				
VX0818WBS01	197596	12.02				
VX0818WBSD01	191427	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: FIRS02
 Lab Code: ACE SDG NO.: Q2815
 Lab File ID: VX047371.D Date Analyzed: 08/18/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	190181	12.018				
UPPER LIMIT	380362	12.518				
LOWER LIMIT	95090.5	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.